

*Similarity Declaration*

**From:** Shenzhen D-Health Medical Technology Co., Ltd.

Unit 501, No. 1 Juyin Technology Industrial Plant, Jihua Street, Longgang District, Shenzhen, China

Shanghai Berry Electronic Tech Co., Ltd. as owner of the technology and manufacturer of the products, For the Patient Monitor tested in this application, the main test models are PM6100, and the covering models are PM6100A and AM6200. Except for the differences in product models, software and appearance (see the table below for specific differences), the structure and internacircuit of the three models are completely the same.

The PM6100 patient monitor can monitor physiological signals including HR (Heart Rate), NIBP (non-invasive blood pressure), SpO2, PR (pulse rate), TEMP (body temperature) and RESP (Respiration rate). The PM6100A patient monitor can monitor physiological signals including NIBP (non-invasive blood pressure), SpO2, PR (pulse rate), and RESP (Respiration rate). The AM6200 series Veterinary monitor can monitor physiological signals including HR (Heart Rate), NIBP (non-invasive blood pressure), SpO2, PR (pulse rate), TEMP (body temperature) and RESP (Respiration rate).

Therefore, PM6100 model was selected in the test to cover all models.

| Model<br>Item         | PM6100   | PM6100A  | AM6200   |
|-----------------------|--|--|--|
| SW Version            | V2.53.00.00  | V2.53.00.00  | V1.55.00.00  |
| Function              | SpO <sub>2</sub> , PR/HR, NIBP, TEMP, RESP measurement   | SpO <sub>2</sub> , NIBP, RESP measurement  | SpO <sub>2</sub> , PR/HR, NIBP, TEMP, RESP measurement   |
| Communication         | Bluetooth  | Bluetooth  | Bluetooth  |
| Appearance            | The same   |  | Different  |
| Using objects         | Human  | Human  | Animals  |
| Attachment components | 1. Patient monitor host<br>2. Heart rate and body temperature integrated lead<br>3. SpO <sub>2</sub> sensor<br>4. Blood pressure cuff<br>5. Charging cable<br>6. Charging base<br>7. Disposable ECG electrodes | 1. Patient monitor host<br>2. SpO <sub>2</sub> sensor<br>3. Blood pressure cuff<br>4. Charging cable<br>5. Charging base | 1. Patient monitor host<br>2. Heart rate and body temperature integrated lead<br>3. SpO <sub>2</sub> sensor<br>4. Blood pressure cuff<br>5. Charging cable<br>6. Charging base<br>7. Disposable ECG electrodes |
| Label                 | The same   |  | Different  |

Regards,



Full Name:Wuqiao

Job title:RA

E-mail:wuqiao@shberry.com

Phone number:+86 18201831713