

SAR Test Exclusion Evaluation

FCC ID: 2BF43-PY010

Product Name: Vibrating Mesh Nebulizer system

Model No: PY-010

The standalone RFExp exclusion shall be refer to FCC § 1.1307 (b)(3)(i)(A) SAR-Based and MPE-based exemption which device determined the All distance from antenna to user/bystander.

Frequency (GHz)	Tune-up max power (mW)	Exclusion Power(mW)
2.402	0.501	1

The EUT's maximum time-averaged power is less than 1mW so it preclude RFExp requirement.

Sincerely Yours,



Mr. Ben Cheng
Manager
AUDIX Technology Corporation