

NOTES, UNLESS OTHERWISE SPECIFIED:

1. ARTWORK SHOWN IS FOR REFERENCE ONLY.
2. DYNAMIC DATA FIELDS ARE REPRESENTED WITH PLACEHOLDER INFORMATION INDICATING FORMAT.
DATE OF MANUFACTURING: YYYY-MM-DD
SERIAL NUMBER: XXXXXX

3. UDI DATA MATRIX MUST BE IN GS1 FORMATTING CONTAINING THE GTIN NUMBER, DATE OF MANUFACTURING, AND SERIAL NUMBER.

GTIN NUMBER: 95200000000019

THIS IS A PLACEHOLDER NUMBER

THE LAST DIGIT IS A CHECK DIGIT


DATE OF MANUFACTURING: YYMMDD

SERIAL NUMBER: XXXXXX

4. UDI DATA MATRIX MUST HAVE HUMAN READABLE NUMBERS ADJACENT TO THE MACHINE READABLE INFORMATION WITH GS1 APPLICATION IDENTIFIERS IN PARENTHESES.

5. QR CODE: "HTTPS://SPM.CARE/A2rL34z"
LAST 7 CHARACTERS ARE CASE SENSITIVE

6. REFER TO DC-00218 AND DC-00220 FOR BARTENDER FILES.



Note: FCC ID is 4-point type size font and is permanently affixed via adhesive

65149-01	C	AW, LABEL, PRODUCT, CHARGER	65149-01 REV C.AI
PART NO.	PART REV.	DESCRIPTION	ARTWORK FILE


REV	CO NO.	DESCRIPTION	REVISED BY
A	CHO-01717	INITIAL RELEASE	B. STEVENSON
B	CHO-02106	REMOVED SYMBOLS: NRTL, BF APPLIED PART. ADDED RX SYMBOL. CORRECTED BLACK COLOR TO CMYK #00000 IN MASTER FILE	B. STEVENSON
C	CHO-02248	UPDATED FCC ID. ADDED PHONE NUMBER, ADDED JAMA REQ. TRACEABILITY FILE.	B. STEVENSON

This document is **CONFIDENTIAL** and contains designs and/or other information that are the property of SetPoint Medical. This document and the designs and/or other information contained therein, may not, in whole or part, be duplicated or disclosed or used for the manufacture of any part(s) disclosed herein, or for any other purpose, without the prior written permission of SetPoint Medical.

UNLESS OTHERWISE SPECIFIED ALL DIMENSIONS ARE IN INCHES. ALL DIMENSIONS APPLY AFTER FINISH.	SetPoint MEDICAL	25101 RYE CANYON LOOP VALENCIA, CA 91355 (661) 750-6140 (661) 257-3828 FAX
TOLERANCES LINEAR: .X ±.1 .XX ±.01 .XXX ±.005 ANGULAR: X° ±.1 X.X° ±.5		SURFACE FINISH PER ANSI-B46-1 125/
DO NOT SCALE SHEET 1 of 2	DRAWN BY B. STEVENSON	DATE 21JUN2023
SIZE B	DRAWING NUMBER DWG-65149	REV C

NOTES:

- 20. DYNAMIC DATA MUST BE WITHIN THE BOUNDARIES OF THE BOXES SHOWN.
- 21. DATA FIELDS MAY NOT OVERLAP WITH OTHER IMAGE OBJECTS.
- 22. DATA MATRIX X-DIMENSION 0.014IN MINIMUM.
- 23. QR CODE X-DIMENSION 0.010IN MINIMUM.



-01 DYNAMIC DATA PLACEMENT