

Label and Label Location Exhibit

Model Number: GUS-1000-A

FCC ID: 2BHMUGUS1000A


Applicant: Bright Uro

Date: August 28, 2025

1. Label Information

The product cannot bear a **permanent nameplate or etched label** showing the required regulatory identifiers. The User manual includes the FCC ID information:

- **FCC ID:** 2BHMUGUS1000A



Abdominal Sensor




Label design for the package

2. Label Location

The label is located on the **pouch and carton** that packages the device. The device is too small to add labeling.

Label Location Photo



The final shipping box at red arrow below has another label


 Bright Uro, Inc.
3 Goddard
Irvine, CA 92618 USA
TEL: +1 (949) 216-0873

Patents:
www.brighturo.com/patents

Refer to Owner's Manual
for Symbols Glossary.



www.brighturo.com/owners_manual



Glean Urodynamics Abdominal Sensor

QTY 5


REF GUS-1000-A

 2026-08-31

LOT 22222222



(01) 0 0860009 80063 (17) 260831 (10) 22222222 (240) 0000129



FCC ID: 2BHMUGUS1000A

PD-08-028 Rev 2

0000129

3. Permanence of Label

The label is:

- **Permanently affixed** to the product's "Pouch" and "Carton" using one of the following:
 - Adhesive label (from Uline S-7413 which uses Permanent emulsion acrylic which is UL969 compliant)

⚠ 5. Notes

- Devices with small form factors or lacking space for full labeling may apply for **alternative labeling options** (e.g., user manual disclosure, packaging label).