

# Shanghai Medconn Medical Technology Co., Ltd.

## MPE ASSESSMENT REPORT

**Report Type:**  
FCC MPE assessment report

**Model:**  
MQ-3000, MQ-3000PT

**REPORT NUMBER:**  
2407B2158SHA-002

**ISSUE DATE:**  
August 20, 2024

**DOCUMENT CONTROL NUMBER:**  
TTRFFCCMPE-01\_V1 © 2018 Intertek





Total Quality. Assured.

## TEST REPORT

Intertek Testing Services (Shanghai FTZ) Co., Ltd.  
Building No.86, 1198 Qinzhou Road (North)  
Caohejing Development Zone  
Shanghai 200233, China

Telephone: 86 21 6127 8200  
[www.intertek.com](http://www.intertek.com)

Report no.: 2407B2158SHA-002

**Applicant:** Shanghai Medconn Medical Technology Co., Ltd.  
No.1018 WeiChang Road, ShanYang, JinShan District, Shanghai, China

**Manufacturer:** Shanghai Medconn Medical Technology Co., Ltd.  
No.1018 WeiChang Road, ShanYang, JinShan District, Shanghai, China

**Factory:** Shanghai Medconn Medical Technology Co., Ltd.  
Rm 1001-1003, 1F, East of Building4 & 2F, East of Building4, No. 1018  
Weichang Road, Shanyang Town, Jinshan District, Shanghai, China

**FCC ID:** 2BHV20907

### SUMMARY:

The equipment complies with the requirements according to the following standard(s) or Specification:

KDB447498 D01 General RF Exposure Guidance v06  
FCC Part2.1091, FCC Part2.1093 FCC Part1.1307(b)

**PREPARED BY:**

Project Engineer  
Alexander Li

**REVIEWED BY:**

Reviewer  
Wakeyou Wang

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to permit copying or distribution of this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.

**TEST REPORT****Revision History**

Report No.	Version	Description	Issued Date
2407B2158SHA-002	Rev. 01	Initial issue of report	August 20, 2024

**TEST REPORT****1 GENERAL INFORMATION****1.1 Description of Equipment Under Test (EUT)**

Product name:	Glycated Hemoglobin Analyzer
Type/Model:	MQ-3000, MQ-3000PT
Description of EUT:	The EUT is a Glycated Hemoglobin Analyzer with NFC function, both models are identical except the appearance color.
Rating:	100-240V~, 50/60Hz, 150VA
EUT type:	<input checked="" type="checkbox"/> Table top <input type="checkbox"/> Floor standing
Software Version:	-
Hardware Version:	-
Serial numbers:	A240621-02-002
Sample received date:	June 28, 2024
Date of test:	June 28, 2024 to August 5, 2024

**1.2 Technical Specification**

Frequency Range:	13.56 MHz ~ 13.56 MHz
Modulation:	ASK
Antenna gain:	PCB antenna

**TEST REPORT****1.3 Description of Test Facility**

Name:	Intertek Testing Services (Shanghai FTZ) Co., Ltd.
Address:	Building 86, No. 1198 Qinzhou Road(North), Shanghai 200233, P.R. China
Telephone:	86 21 61278200
Telefax:	86 21 54262353

The test facility is recognized, certified, or accredited by these organizations:	CNAS Accreditation Lab Registration No. CNAS L21189
	FCC Accredited Lab Designation Number: CN0175
	IC Registration Lab CAB identifier.: CN0014
	VCCI Registration Lab Member No.: 3598 (Registration No.: R-14243, G-10845, C-14723, T-12252)
	A2LA Accreditation Lab Certificate Number: 3309.02

**TEST REPORT****2 MPE Assessment****Test result:** Pass**2.1 MPE Assessment Limit****Mobile device exposure for standalone operations:**

According to §1.1310, the limit for general population/uncontrolled exposures

Frequency range (MHz)	Electric field strength (V/m)	Magnetic field strength (A/m)	Power density (mW/cm <sup>2</sup> )	Averaging time (minutes)
0.3-1.34	614	1.63	*(100)	30
1.34-30	824/f	2.19/f	*(180/f <sup>2</sup> )	30
30-300	27.5	0.073	0.2	30
300-1500	/	/	f/1500	30
1500-100,000	/	/	1.0	30

Note: Limit for 13.56MHz is 60.77 V/m

**Mobile device exposure for simultaneous transmission operations: the sum of the MPE ratios for all simultaneously transmitting antennas incorporated in a host device is ≤ 1.0**

## TEST REPORT

### 2.2 Assessment Results

Power density (S) is calculated according to the formula:

$$S = P / (4\pi R^2)$$

Where S = power density in mW/cm<sup>2</sup>

P = Radiated transmit power in mW

R = distance (cm)

As we can see from the test report 2407B2158SHA-001:

50.1dBuV/m@3m, @20cm=@3m+40log(3/0.2)=97.14dBuV/m=0.072V/m<60.77V/m.

Therefore, the SAR requirement is deemed to be satisfied without test.

**TEST REPORT**

## **Appendix I**

Definition below must be outlined in the User Manual:

To satisfy FCC RF exposure requirements, a separation distance of 20 cm or more should be maintained between the antenna of this device and persons during device operation. To ensure compliance, operations at closer than this distance is not recommended.

\*\*\*\*\*END\*\*\*\*\*