

Fresenius Medical Care AG

Hafenstr. 9, 97424 Schweinfurt, Germany

Date: May 20, 2025

Federal Communications Commission
7435 Oakland Mills Road
Columbia MD 21046

Requesting permissive change

To Whom It May Concern:

The purpose of this letter is to request a Class II Permissive change for
FCC ID: 2BK5J-DGF109FME, Original grant date: January 21, 2025

The major change filed under this application is :

1. The subject approved module is integrated in a new host (Product: [5008X Monitor with CDX without remote control](#) Model: [F50015749](#)).
2. Evaluation has been done to ensure compliance of EMC and RF exposure.
3. The hardware design of this transmitter that may affect compliance is remained unchanged in this permissive change application.
4. The security mechanism according to section II of KDB 594280 D02 is remained unchanged as it reported to the FCC at the initial certification.
5. WLAN transmitter only supports with passive scanning in UNII-2A & UNII-2C bands without radar detection, hence radar detection is exempted from evaluation in this permissive change.

Sincerely,

 2025/05/21

(Signature)

Name: Alexander Schreiber

Title: Senior Manager Design Quality Assurance

On behalf of: Fresenius Medical Care AG

Telephone: +49 (9721) 678-67806

Email: alexander.schreiber@freseniusmedicalcare.com