

Appendix C: System Check Plots

RF SAFETY LABORATORY

System Verification

Verification Conditions

Verification Source	Power [dBm]	Phantom Section	Test Distance [mm]	Signal, UID	Frequency [MHz]	ConvF (x, y, z)
D2450V2 – SN1112	17.0	Flat	10	CW, 0	2450.000	7.84, 7.84,7.84

Test Setup

Test Date	Ambient Temp [°C]	TSL Temp [°C]	Measurement SW	Phantom	TSL Type	Cond. [S/m]	Perm.	Probe, Cal. Date	DAE, Cal. Date
2025-03-12	24.5	22.7	16.4.0.5005	Twin-SAM V8.0 – 2159	Head Simulating Liquid	1.81	37.3	EX3DV4 – SN7872, 2024-03-15	DAE4ip Sn1843, 2024-08-14

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 80.0	30.0 x 30.0 x 31.2
Grid Steps [mm]	10.0 x 10.0	5.0 x 5.0 x 1.5
Sensor Surface [mm]	3.0	1.4
Grading Ratio	N/A	1.5

Measurement Results

	SAR [W/kg] or APD [W/m²]	Deviation [%]
psSAR1g	2.50	-1.42
psSAR10g	1.16	-2.34
Peak SAR (Extrapolated)	5.54	6.08
psAPD (4.0 cm², sq)	N/A	N/A

