

Label position

Label size: 50 x 35 mm


Intraoral Scanner

Model:

REF

CN

LOT



Battery: DC 3.6V/3300mAh

Adapter Input: AC 100-240V, 50/60Hz, 0.5A

Adapter Output: DC6V/3A Service Life: 5 years

 Changzhou Sifary Medical Technology Co., Ltd.
No. 26 Yandanghe Road, Xinbei District,
213000 Changzhou, Jiangsu, China

Label position

Label size: 50 x 35 mm

Holder of scanner

Model: EMBS-7

FCC ID: 2B07DEM5BS

This device complies with part 15 of the FCC Rules.

Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference and
- (2) this device must accept any interference received, including interference that may cause undesired operation.