



July 18, 2025

Subject: Cover Letter for FCC Application

Applicant: Becton, Dickinson and Company

FCC Certification Number: **2BQT2-PYX1155**

Product Description: BD Pyxis™ Pro MedStation™ Main 1155-00 with Omnikey 5122

To whom it may concern,

This letter is to request certification for our device, **2BQT2-PYX1155**, to the following certification service.

- System Approval for original certification.

Please grant FCC certification for BD Pyxis Pro MedStation Main System. This System integrates the Omnikey 5122 OEM PCBA to provide 13.56 Mhz contactless smart card reading capability.

- OEM Omnikey 5122 PCBA is supplied by HID Global
- OEM Omnikey 5122 antenna is integrated as part of its PCB.
- OEM Omnikey 5122 PCBA is an optional card reader for BD Pyxis Pro MedStation System.

Preliminary testing of the Omnikey 5122 module integrated into **BD Pyxis Pro MedStation Main System** has demonstrated that the RF specifications and performance meet FCC Standard under the system level for original certification.

If you have any questions regarding this request for certification, please feel free to contact me.

Sincerely,

A handwritten signature in black ink that reads "Celiflora Palma". The signature is written in a cursive, flowing style.

Celiflora Palma
Staff Electrical Engineer
Becton, Dickinson and Company
858-617-5564