

Model Difference declaration

Centre Testing International Group Co.,Ltd:

There are six models of Dynamic ECG recorder products submitted by our company, including the main inspection model [EPod Pro] and the covering model [EPod 、 EHL-12 、 EHL-7]. The principle, structure and composition, intended use, key components, and electrical structure of the products are the same, and only the functional configuration is different. Therefore, the model with the most complete functions [EPod Pro] was selected as the main inspection model to cover the other models.

The specific differences are as follows:

Model	EHL-7	EHL-12	EPod	EPod Pro
Screen	●	●	●	●
Leadwire type	3/5-leadwire	3/5/6/10 leadwire	3/5 leadwire	3/5/6/10 leadwire
Maximum storage of a single piece of data (recording time)	Single lead/7-lead: 72h	Single/7/8-lead: 72h; 12-lead: 48h	Single lead/7-lead: 72h	Single/7/8-lead: 72h; 12-lead: 48h
Maximum number of data stored	1	1	10	10
Run time	72 hours	72 hours	72 hours	72 hours
Shell color	Black	Black	White	White
Bluetooth	●	●	●	●
HR	●	●	●	●

Shenzhen Creative Industry Co., Ltd.

2023-08-02

Signature: 