

Information about the Applicant	
Grantee (Company Name)	Geissler Corporation
FCC Contact Person	Randolf Geissler
Address	15665 Medina Rd.
City, State, Zip	Plymouth, MN 55447
Job Number	LGPD0048
Model	101
FCC ID	AHTWRX
Agent	
Approval Type	Original
Equipment Class	DXX Part 15, Low Power Communication Device Transmitter
Rule Part	15C (15.249)

Overview

The Whisper Veterinary Stethoscope receiver, FCC ID AHTWRX, utilizes a 2.4 GHz transceiver under FCC rule part 15.249. This device has a USB interface and its function is to receive data from the Stethoscope.

Confidentiality requested for allowed files. Request is properly addressed and referenced.

Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

Signature

Dave Tolman, TCB Committee
2/14/2012