

Liquid confirmation

Liquid Measurement date: 01/27/02 Liquid depth: 15.15cm

By: Sunny Shih

Simulant	Freq [MHz]	Parameters	Liquid Temp [°C]	Target Value	Measured Value	Deviation [%]	Limit [%]
Head	2450	ϵ	20.6	39.2	41.1	4.85	± 5
		σ	20.6	1.8	1.88	4.44	± 5
Body	2450	ϵ	20.9	52.7	52.8	0.19	± 5
		σ	20.9	1.95	2.03	4.10	± 5