

**APPENDIX B - E-Field Probe Calibration Data**

See attached pages.



## Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

### Calibration Certificate

#### Dosimetric E-Field Probe

Type:

ET3DV5

Serial Number:

1333

Place of Calibration:

Zurich

Date of Calibration:

April 23, 2001

Calibration Interval:

12 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

*Nicolae Neri*

Approved by:

*Alvin Katz*

**Schmid & Partner  
Engineering AG**

Zeughausstrasse 43, 8004 Zurich, Switzerland, Telephone +41 1 245 97 00, Fax +41 1 245 97 79

# Probe ET3DV5

## SN:1333

Manufactured:	December 20, 1997
Last calibration:	April 10, 2000
Recalibrated:	April 23, 2001

Calibrated for System DASY3

ET3DV5 SN:1333

**DASY3 - Parameters of Probe: ET3DV5 SN:1333**

Sensitivity in Free Space

NormX	2.37 $\mu\text{V}/(\text{V}/\text{m})^2$
NormY	2.38 $\mu\text{V}/(\text{V}/\text{m})^2$
NormZ	2.33 $\mu\text{V}/(\text{V}/\text{m})^2$

Diode Compression

DCP X	100 mV
DCP Y	100 mV
DCP Z	100 mV

Sensitivity in Tissue Simulating Liquid

Head                    450 MHz                     $\epsilon_r = 43.5 \pm 5\%$                      $\sigma = 0.87 \pm 10\%$  mho/m

ConvF X	6.25 extrapolated	Boundary effect:
ConvF Y	6.25 extrapolated	Alpha            0.19
ConvF Z	6.25 extrapolated	Depth            3.06

Head                    900 MHz                     $\epsilon_r = 42 \pm 5\%$                      $\sigma = 0.97 \pm 10\%$  mho/m

ConvF X	5.83 $\pm 7\%$ (k=2)	Boundary effect:
ConvF Y	5.83 $\pm 7\%$ (k=2)	Alpha            0.38
ConvF Z	5.83 $\pm 7\%$ (k=2)	Depth            2.70

Brain                    1500 MHz                     $\epsilon_r = 41 \pm 5\%$                      $\sigma = 1.32 \pm 10\%$  mho/m

ConvF X	5.27 interpolated	Boundary effect:
ConvF Y	5.27 interpolated	Alpha            0.63
ConvF Z	5.27 interpolated	Depth            2.23

Brain                    1800 MHz                     $\epsilon_r = 41 \pm 5\%$                      $\sigma = 1.69 \pm 10\%$  mho/m


ConvF X	4.99 $\pm 7\%$ (k=2)	Boundary effect:
ConvF Y	4.99 $\pm 7\%$ (k=2)	Alpha            0.75
ConvF Z	4.99 $\pm 7\%$ (k=2)	Depth            1.99

Sensor Offset


Probe Tip to Sensor Center	2.7	mm
Optical Surface Detection	1.6 $\pm$ 0.2	mm

ET3DV5 SN:1333


Receiving Pattern ( $\phi$ ),  $\theta = 0^\circ$



ET3DV5 SN:1333




Isotropy Error ( $\phi$ ),  $\theta = 0^\circ$





ET3DV5 SN:1333

### Frequency Response of E-Field ( TEM-Cell:ifi110, Waveguide R22)




ET3DV5 SN:1333

**Dynamic Range f(SAR<sub>brain</sub>)**  
( TEM-Cell:ifi1110 )



ET3DV5 SN:1333

### Conversion Factor Assessment



<b>Head</b>	<b>900 MHz</b>	$\epsilon_r = 42 \pm 5\%$	$\sigma = 0.97 \pm 10\%$ mho/m
	ConvF X	<b>5.83</b> $\pm 7\%$ (k=2)	Boundary effect:
	ConvF Y	<b>5.83</b> $\pm 7\%$ (k=2)	Alpha <b>0.38</b>
	ConvF Z	<b>5.83</b> $\pm 7\%$ (k=2)	Depth <b>2.70</b>
<b>Brain</b>	<b>1800 MHz</b>	$\epsilon_r = 41 \pm 5\%$	$\sigma = 1.69 \pm 10\%$ mho/m
	ConvF X	<b>4.99</b> $\pm 7\%$ (k=2)	Boundary effect:
	ConvF Y	<b>4.99</b> $\pm 7\%$ (k=2)	Alpha <b>0.75</b>
	ConvF Z	<b>4.99</b> $\pm 7\%$ (k=2)	Depth <b>1.99</b>