



Date January 9, 2017
TUV SUD BABT TCB
Octagon House,
Segensworth Road,
Fareham,
Hampshire,
PO15 5RL

Limited Modular Approval Request

FCC ID: AXJ02532481

The following attestation addresses the requirements to support modular approval:

Modular approval requirement	Yes (provide brief statement)	No *
(a) The radio elements must have the radio frequency circuitry shielded. Physical components and tuning capacitor(s) may be located external to the shield, but must be on the module assembly	Yes, radio elements with the Control Enclosure.	
(b) The module must have buffered modulation/data inputs to ensure that the device will comply with Part 15 requirements with any type of input signal	Yes. Data inputs are buffered.	
(c) The module must contain power supply regulation on the module		No. STERRAD® 100NX host product supplies +5VDC to power the module.
(d) The module must contain a permanently attached antenna, or contain a unique antenna connector, and be marketed and operated only with specific antenna(s), per Sections 15.203, 15.204(b), 15.204(c), 15.212(a), 2.929(b)	Yes, only one antenna configuration will be used	
(e) The module must demonstrate compliance in a stand-alone configuration		No, Tested in Host model: STERRAD® 100NX
(f) The module must be labelled with its permanently affixed FCC ID label, or use an electronic display (See KDB Publication 784748 about labelling requirements)	Yes, label will be on host system.	
(g) The module must comply with all specific rules applicable to the transmitter. The grantee must provide comprehensive instructions to explain compliance requirements	Yes, User Guide will include information.	

ADVANCED STERILIZATION PRODUCTS®

Division of Ethicon, Inc.

a *Johnson & Johnson* company



Modular approval requirement	Yes (provide brief statement)	No *
(h) The module must comply with RF exposure requirements	Yes.	

* Please provide a detailed explanation if the answer is "No."

Yours sincerely,

A handwritten signature in black ink that reads "Michael Gallucci". To the right of the signature is the date "01/09/17".

Name: Michael Gallucci
Title: Director, Product Management

ADVANCED STERILIZATION PRODUCTS®

Division of Ethicon, Inc.

a *Johnson & Johnson* company

33 Technology Drive • Irvine, CA • 92618 • [Phone] 800.595.0200 • [Fax] 949.581.5997