

Technical Compliance Statement

FCC and ISED Test Report

For the following information

Ref. File No.: C1M1911123

Product : GlucoLeader Enhance 2 Blood Glucose Monitoring System
Model Number : P5B
Applicant : HMD BioMedical Inc.
Manufacturer : HMD BioMedical Inc.
Rules and Standards : 47 CFR FCC Part 15 Subpart B and
ICES-003 Issue 6: 2016
(Class B Limit)

We hereby certify that the above product has been tested by us and complied with above FCC and ICES official limits. The test was performed according to the procedures ANSI C63.4-2014. The equipment might be marketed in US or Canada in accordance with the rules of 47 CFR FCC Part 2 and ISED regulations. The test data and results are issued on the test report **EM-F190482**.

Signature



Alex Deng/Deputy Manager
Date: 2019. 12. 05

Test Laboratory:
AUDIX Technology Corporation, EMC Department
NVLAP Lab. Code: 200077-0
FCC OET Designation: TW1004 & TW1090
Web Site: www.audixtech.com

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.