



CGISS EME Test Laboratory

8000 West Sunrise Blvd
Fort Lauderdale, FL. 33322

S.A.R. EME Compliance Test Report
Part 3 of 5

Date of Report: April 15, 2004
Report Revision: Rev. A
Manufacturer: Motorola
Product Description: CN620; Quad band GSM and Tri band WLAN (802.11a, b, and g)
FCC ID: AZ489FT5829
Device Model: H77UBH6JA5AA

Test Period: 2/19/04-3/24/04
EME Tech: Ed Church
Responsible Eng: Deanna Zakharia (Elect. Principle Staff Eng.)
Author: Michael Sailsman (Global EME Regulatory Affairs Liaison)

Note: Based on the information and the testing results provided herein, the undersigned certifies that when used as stated in the operating instructions supplied, said product complies with the national and international reference standards and guidelines listed in section 2.0 of this report.

Deanna Zakharia signature on file for Ken Enger

4/15/04

Ken Enger
Senior Resource Manager, Laboratory Director, CGISS EME Lab

Date Approved

Note: This report shall not be reproduced without written approval from an officially designated representative of the Motorola EME Laboratory.

Appendix C: Dipole System Performance Check Results (Continued)

SYSTEM VALIDATION

Date: 09/06/2003 Frequency (MHz): 2450
Lab Location: CGISS Mixture Type: 2450-IEEE Head
Robot System: CGISS-2 Ambient Temp.(°C): 22.9
Probe Serial #: 1383 Tissue Temp.(°C): 21.0
DAE Serial #: DAE3V1 SN406

Tissue Characteristics Phantom Type/SN: VAL30242005A
Permittivity: 38.0 Distance (mm): 10
Conductivity: 1.85

Reference Source: Dipole (Dipole/Handset)
Reference SN: 704

Power to Dipole: 250 mW
Power Output (radio): _____ mW

Target SAR Value: 52.4 mW/g, 24.0 mW/g (10g avg.)
(Normalized to 1.0 W)

Measured SAR Value: 13.10 mW/g, 5.99 mW/g (10g avg.)
Power Drift: -0.02 dB

Measured SAR Value: 52.64 mW/g, 24.07 mW/g (10g avg.)
(normalized to 1.0 W,
with drift compensation)

Percent Difference From Target (must be within System Uncertainty): + 0.46 % (1g avg)
+ 0.29 % (10g avg)

Test performed by: Dave Hopper

Initial: WOK

09/06/03

SPEAG DIPOLE D2450V2 ; Test date:09/06/03

Run #: Sys VAL R2-030906-01 Phantom #: VAL30242005A/S12
Model#: D2450V2 SN:704
Robot#: CGISS-2 DAE: DAE3V1SN406 (11/11/02)
Tx Freq: 2450MHz Tester: Dave Hopper
Start power: 250mW Simulated tissue temp: 21.0 C

IEEE Targets:

52.4 mW/g for 1g SAR, 24.0 mW/g for 10g SAR.

SAR calculated 1g is 52.64 mW/g percent from target (including drift) is 0.46 %

SAR Calculated 10g is 24.07 mW/g Percent from target (including drift) is 0.29 %


Flat; Probe: ET3DV6 - SN1383 (Cal Date 26 February 2003); ConvF(5.00,5.00,5.00); Crest factor: 1.0; IEEE

HEAD 2450MHz: $\sigma = 1.85$ mho/m $\epsilon_r = 38.0$ $\rho = 1.00$ g/cm³


Cubes (2): Peak: 27.2 mW/g \pm 0.00 dB, SAR (1g): 13.1 mW/g \pm 0.00 dB, SAR (10g): 5.99 mW/g \pm 0.01 dB, (Worst-case extrapolation)

Penetration depth: 6.5 (6.3, 7.1) [mm]

Powerdrift: -0.02 dB



SAR_{Tot} [mW/g]



SYSTEM PERFORMANCE CHECK TARGET SAR

Date: 09/06/2003 Frequency (MHz): 2450
Lab Location: CGISS Mixture Type: 2450-FCC Body
Robot System: CGISS-2 Ambient Temp.(°C): 22.9
Probe Serial #: 1383 Tissue Temp.(°C): 22.0
DAE Serial #: DAE3V1 SN406

Tissue Characteristics Phantom Type/SN: 40302002A/S10
Permittivity: 53.6 Distance (mm): 10
Conductivity: 2.04

Reference Source: Dipole (Dipole)
Reference SN: 704
Power to Dipole: 250 mW

Measured SAR Value: 12.80 mW/g, 5.92 mW/g (10g avg.)
Power Drift: -0.01 dB

New Target/Measured SAR Value: 51.32 mW/g, 23.73 mW/g (10g avg.)
(Normalized to 1.0 W,
with drift compensation)

Test performed by: Dave Hopper

Initial: WDA


09/06/03

SPEAG DIPOLE D2450V2 ; Test date:09/06/03

Run #: Sys Perf R2-030906-04 Phantom #:40302002A/S10
Model#: D2450V2 SN:704
Robot#: CGISS-2 DAE: DAE3V1SN406 (11/11/02)
Tx Freq: 2450MHz Tester: Dave Hopper
Start power: 250mW Simulated tissue temp: 22.0 C

New Targets:
51.32 mW/g for 1g SAR, 23.73 mW/g for 10g SAR.

Flat; Probe: ET3DV6 - SN1383 (Cal Date 26 February 2003); ConvF(4.70,4.70,4.70); Crest factor: 1.0; FCC
BODY 2450MHz: $\sigma = 2.04$ mho/m $\epsilon_r = 53.6$ $\rho = 1.00$ g/cm³
Cubes (2): Peak: 26.0 mW/g ± 0.00 dB, SAR (1g): 12.8 mW/g ± 0.00 dB, SAR (10g): 5.92 mW/g ± 0.00
dB, (Worst-case extrapolation)
Penetration depth: 7.1 (6.8, 8.0) [mm]
Powerdrift: -0.01 dB



Motorola CGISS EME Lab

SYSTEM VALIDATION

| | | | |
|-----------------|-----------------|--------------------|--------------|
| Date: | <u>3/6/2004</u> | Frequency (MHz): | <u>5200</u> |
| Lab Location: | <u>CGISS</u> | Mixture Type: | <u>Speag</u> |
| Robot System: | <u>GGISS-1</u> | Ambient Temp.(°C): | <u>21.2</u> |
| Probe Serial #: | <u>3512</u> | Tissue Temp.(°C): | <u>21</u> |
| DAE Serial #: | <u>363</u> | | |

Tissue Characteristics

| | | | |
|---------------|-------------|------------------|-------------------------------|
| Permittivity: | <u>34.9</u> | Phantom Type/SN: | <u>SAMTP 1209</u> |
| Conductivity: | <u>4.70</u> | Distance (mm): | <u>10 (tissue/dipole cnt)</u> |

| | | |
|-------------------|----------------|----------|
| Reference Source: | <u>D5GHzV2</u> | (Dipole) |
| Reference SN: | <u>1010</u> | |

| | | |
|-----------------------|-------------------|----|
| Power to Dipole: | <u>100</u> | mW |
| Power Output (radio): | <u> </u> | mW |

| | | | | |
|-----------------------|-------------|-------|-------------|-----------------|
| Target SAR Value: | <u>84.4</u> | mW/g, | <u>23.6</u> | mW/g (10g avg.) |
| (normalized to 1.0 W) | | | | |

| | | | | |
|---------------------|-------------|-------|----------|-----------------|
| Measured SAR Value: | <u>8.29</u> | mW/g, | <u>?</u> | mW/g (10g avg.) |
| Power Drift: | <u>0</u> | dB | | |

| | | | | |
|--|--------------|-------|----------------|-----------------|
| Measured SAR Value: | <u>82.90</u> | mW/g, | <u>#VALUE!</u> | mW/g (10g avg.) |
| (normalized to 1.0 W, including drift) | | | | |

| | | |
|---|----------------|-------------|
| Percent Difference From Target (MUST be within System Uncertainty): | <u>1.78</u> | % (1g ave) |
| | <u>#VALUE!</u> | % (10g ave) |

| | | | |
|--------------------|------------------|----------|-----------|
| Test performed by: | <u>E. Church</u> | Initial: | <u>EC</u> |
|--------------------|------------------|----------|-----------|

03/06/04

SPEAG Dipole 5200MHz. Test Date:03/06/04

Run #: Sys Perf-R1-040306-01 Phantom #: SAMTP 1209
Model #: D5GHzV2 SN: 1010
Robot: CGISS-1 Tester: E. Church
TX Freq: 5200 MHz Sim Tissue Temp: 21.0 (Celsius)
Start Power: 100 mW
DAE3: SN363-VI DAE Cal Date: 05/13/2003

- Comments-

SAR target at 1W is 82.9 mW/g (1g avg, including drift)

Measured cube at zero degrees is 8.18 mW/g
Measured cube at ninety degrees is 8.40 mW/g
Average of cubes is 8.29 mW/g


SAR calculated at 1W is 82.90 mW/g (1g avg). Percent from target (including drift) is 0 %

SAM - Expanded;

Probe: EX3DV3 - SN3512; Probe Cal Date: 23/1/04 ConvF(5.02,5.02,5.02); Crest factor: 1.0; IEEE Head
5200MHz: $\sigma = 4.70$ mho/m $\epsilon_r = 34.9$ $\rho = 1.00$ g/cm³

0

Powerdrift: -0.00 dB



SAR_{Tot} [mW/g]

— 1.32E+0

Motorola CGISS EME Lab

SYSTEM PERFORMANCE CHECK TARGET SAR

Date: 3/19/2004 Frequency (MHz): 5200
Lab Location: CGISS Mixture Type: Speag
Robot System: CGISS-1 Ambient Temp.(°C): 22.4
Probe Serial #: 3512 Tissue Temp.(°C): 20.9
DAE Serial #: 363

Tissue Characteristics

Permittivity: 46.7 Phantom Type/SN: Flat 40302002A-S10
Conductivity: 5.51 Distance (mm): 10 (tissue/dipole cnt)

Reference Source: D5GHzV2 (Dipole)
Reference SN: 1010

Power to Dipole: 100 mW

Measured SAR Value: 8.62 mW/g, _____ mW/g (10g avg.)
Power Drift: 0.28 dB

New Target/Measured

SAR Value: 80.82 mW/g, 0.00 mW/g (10g avg.)
(normalized to 1.0 W, including drift)

Test performed by: E. Church Initial: EC

03/19/04

SPEAG Dipole 5200MHz. Test Date:03/19/04

Run #: Sys Perf-R1-040319-14 Phantom #: 40302002A-S10
Model #: D5GHzV2 SN: 1010
Robot: CGISS-1 Tester: E. Church
TX Freq: 5200 MHz Sim Tissue Temp: 20.9 (Celsius)
Start Power: 100 mW
DAE3: SN363-V1 DAE Cal Date: 05/13/2003

- Comments-

SAR target at 1W is 80.82 mW/g (1g avg, including drift)


Measured cube at zero degrees is 8.58 mW/g
Measured cube at ninety degrees is 8.66 mW/g
Average of cubes is 8.62 mW/g

SAR calculated at 1W is 80.82 mW/g (1g avg). Percent from target (including drift) is + .0 %

Flat Phantom:

Probe: EX3DV3 - SN3512; Probe Cal Date: 23/1/04 ConvF(4.25,4.25,4.25); Crest factor: 1.0; FCC Body 5200 MHz: $\sigma = 5.51 \text{ mho/m}$ $\epsilon_r = 46.7$ $\rho = 1.00 \text{ g/cm}^3$
,, ()

Powerdrift: 0.28 dB



SAR_{tot} [mW/g]

— 1.22E+0

Motorola CGISS EME Lab

APPENDIX D

Probe/Dipole Calibration Certificates

Client **Motorola CGISS**

| CALIBRATION CERTIFICATE | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|---|---|---|--|------------|------|---|-----------------------|-----------------------|--------------|---|------------------------|---------------------|------------|-------------------------------|--------|-----------------------|------------|-----------------------------------|--------|------------------------|------------|-------------------------------|--------|---------------------------|------------|-------------------------------------|------------------------|-----------------------------------|-------------|---------------------------|--------|
| Object(s) | ET3DV6 - SN:1384 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Calibration procedure(s) | QA CAL-01.v2 Calibration procedure for dosimetric E-field probes | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Calibration date: | May 15, 2003 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Condition of the calibrated item | In Tolerance (according to the specific calibration document) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>This calibration statement documents traceability of M&TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Model Type</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>RF generator HP 8684C</td> <td>US3642U01700</td> <td>4-Aug-99 (SPEAG, in house check Aug-02)</td> <td>In house check: Aug-05</td> </tr> <tr> <td>Power sensor E4412A</td> <td>MY41495277</td> <td>2-Apr-03 (METAS, No 252-0250)</td> <td>Apr-04</td> </tr> <tr> <td>Power sensor HP 8481A</td> <td>MY41092180</td> <td>18-Sep-02 (Agilent, No. 20020918)</td> <td>Sep-03</td> </tr> <tr> <td>Power meter EPM E4419B</td> <td>GB41293874</td> <td>2-Apr-03 (METAS, No 252-0250)</td> <td>Apr-04</td> </tr> <tr> <td>Network Analyzer HP 8753E</td> <td>US38432426</td> <td>3-May-00 (Agilent, No. 8702K064602)</td> <td>In house check: May 03</td> </tr> <tr> <td>Fluke Process Calibrator Type 702</td> <td>SN: 6295803</td> <td>3-Sep-01 (ELCAL, No.2360)</td> <td>Sep-03</td> </tr> </tbody> </table> | | | | Model Type | ID # | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration | RF generator HP 8684C | US3642U01700 | 4-Aug-99 (SPEAG, in house check Aug-02) | In house check: Aug-05 | Power sensor E4412A | MY41495277 | 2-Apr-03 (METAS, No 252-0250) | Apr-04 | Power sensor HP 8481A | MY41092180 | 18-Sep-02 (Agilent, No. 20020918) | Sep-03 | Power meter EPM E4419B | GB41293874 | 2-Apr-03 (METAS, No 252-0250) | Apr-04 | Network Analyzer HP 8753E | US38432426 | 3-May-00 (Agilent, No. 8702K064602) | In house check: May 03 | Fluke Process Calibrator Type 702 | SN: 6295803 | 3-Sep-01 (ELCAL, No.2360) | Sep-03 |
| Model Type | ID # | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| RF generator HP 8684C | US3642U01700 | 4-Aug-99 (SPEAG, in house check Aug-02) | In house check: Aug-05 | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor E4412A | MY41495277 | 2-Apr-03 (METAS, No 252-0250) | Apr-04 | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor HP 8481A | MY41092180 | 18-Sep-02 (Agilent, No. 20020918) | Sep-03 | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power meter EPM E4419B | GB41293874 | 2-Apr-03 (METAS, No 252-0250) | Apr-04 | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Network Analyzer HP 8753E | US38432426 | 3-May-00 (Agilent, No. 8702K064602) | In house check: May 03 | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Fluke Process Calibrator Type 702 | SN: 6295803 | 3-Sep-01 (ELCAL, No.2360) | Sep-03 | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Calibrated by: | Name Nicola Vetterli | Function Technician | Signature  | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Approved by: | Name Katja Pokornic | Function Laboratory Director | Signature  | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Date issued: May 15, 2003 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

DASY - Parameters of Probe: ET3DV6 SN:1384**Sensitivity in Free Space**

| | |
|-------|--|
| NormX | $1.76 \mu\text{V}/(\text{V}/\text{m})^2$ |
| NormY | $1.72 \mu\text{V}/(\text{V}/\text{m})^2$ |
| NormZ | $1.89 \mu\text{V}/(\text{V}/\text{m})^2$ |

Diode Compression

| | | |
|-------|-----------|----|
| DCP X | 92 | mV |
| DCP Y | 92 | mV |
| DCP Z | 92 | mV |

Sensitivity in Tissue Simulating Liquid

Head 900 MHz $\epsilon_r = 41.5 \pm 5\%$ $\sigma = 0.97 \pm 5\%$ mho/m

Valid for f=800-1000 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

| | | |
|---------|-----------------------|-------------------|
| ConvF X | $6.6 \pm 9.5\%$ (k=2) | Boundary effect: |
| ConvF Y | $6.6 \pm 9.5\%$ (k=2) | Alpha 0.45 |
| ConvF Z | $6.6 \pm 9.5\%$ (k=2) | Depth 2.42 |

Head 1800 MHz $\epsilon_r = 40.0 \pm 5\%$ $\sigma = 1.40 \pm 5\%$ mho/m

Valid for f=1710-1910 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

| | | |
|---------|-----------------------|-------------------|
| ConvF X | $5.4 \pm 9.5\%$ (k=2) | Boundary effect: |
| ConvF Y | $5.4 \pm 9.5\%$ (k=2) | Alpha 0.55 |
| ConvF Z | $5.4 \pm 9.5\%$ (k=2) | Depth 2.56 |

Boundary Effect

Head 900 MHz Typical SAR gradient: 5 % per mm

| | | | |
|-----------------------|------------------------------|------|------|
| Probe Tip to Boundary | | 1 mm | 2 mm |
| SAR _{be} [%] | Without Correction Algorithm | 11.4 | 6.3 |
| SAR _{be} [%] | With Correction Algorithm | 0.4 | 0.7 |


Head 1800 MHz Typical SAR gradient: 10 % per mm

| | | | |
|-----------------------|------------------------------|------|------|
| Probe Tip to Boundary | | 1 mm | 2 mm |
| SAR _{be} [%] | Without Correction Algorithm | 14.7 | 9.5 |
| SAR _{be} [%] | With Correction Algorithm | 0.1 | 0.0 |

Sensor Offset

| | | |
|----------------------------|------------------|----|
| Probe Tip to Sensor Center | 2.7 | mm |
| Optical Surface Detection | 1.5 ± 0.2 | mm |

Conversion Factor Assessment



Head 900 MHz $\epsilon_r = 41.5 \pm 5\%$ $\sigma = 0.97 \pm 5\%$ mho/m

Valid for f=800-1000 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X


| | | | |
|---------|------------------|------------------|-------------|
| ConvF X | 6.6 ± 9.5% (k=2) | Boundary effect: | |
| ConvF Y | 6.6 ± 9.5% (k=2) | Alpha | 0.45 |
| ConvF Z | 6.6 ± 9.5% (k=2) | Depth | 2.42 |

Head 1800 MHz $\epsilon_r = 40.0 \pm 5\%$ $\sigma = 1.40 \pm 5\%$ mho/m

Valid for f=1710-1910 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

| | | | |
|---------|------------------|------------------|-------------|
| ConvF X | 5.4 ± 9.5% (k=2) | Boundary effect: | |
| ConvF Y | 5.4 ± 9.5% (k=2) | Alpha | 0.55 |
| ConvF Z | 5.4 ± 9.5% (k=2) | Depth | 2.56 |

Conversion Factor Assessment



Body 900 MHz $\epsilon_r = 55.0 \pm 5\%$ $\sigma = 1.05 \pm 5\%$ mho/m

Valid for f=800-1000 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C





| | | | |
|---------|------------------|------------------|-------------|
| ConvF X | 6.5 ± 9.5% (k=2) | Boundary effect: | |
| ConvF Y | 6.5 ± 9.5% (k=2) | Alpha | 0.44 |
| ConvF Z | 6.5 ± 9.5% (k=2) | Depth | 2.51 |



Body 1800 MHz $\epsilon_r = 53.3 \pm 5\%$ $\sigma = 1.52 \pm 5\%$ mho/m

Valid for f=1710-1910 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C


| | | | |
|---------|------------------|------------------|-------------|
| ConvF X | 5.0 ± 9.5% (k=2) | Boundary effect: | |
| ConvF Y | 5.0 ± 9.5% (k=2) | Alpha | 0.64 |
| ConvF Z | 5.0 ± 9.5% (k=2) | Depth | 2.49 |

Receiving Pattern (ϕ), $\theta = 0^\circ$






Isotropy Error (ϕ), $\theta = 0^\circ$





Frequency Response of E-Field

(TEM-Cell:ifi110, Waveguide R22)




Dynamic Range f(SAR_{brain})

(Waveguide R22)



Deviation from Isotropy in HSL

Error (θ, ϕ), $f = 900$ MHz



Client **Motorola CGISS**

| CALIBRATION CERTIFICATE | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|--|---|---|--|------------|------|---|-----------------------|------------------------|------------|-------------------------------|--------|---------------------|------------|-------------------------------|--------|----------------------------|----------------|--------------------------------|--------|-----------------------------------|-------------|-------------------------------------|--------|-----------------------|------------|--|------------------------|------------------------|--------|--|------------------------|---------------------------|------------|--|------------------------|
| Object(s) | EX3DV3 - SN:3512 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Calibration procedure(s) | QA CAL-01.v2 Calibration procedure for dosimetric E-field probes | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Calibration date: | January 23, 2004 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Condition of the calibrated item | In Tolerance (according to the specific calibration document) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Model Type</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter EPM E4419B</td> <td>GB41293874</td> <td>2-Apr-03 (METAS, No 252-0250)</td> <td>Apr-04</td> </tr> <tr> <td>Power sensor E4412A</td> <td>MY41495277</td> <td>2-Apr-03 (METAS, No 252-0250)</td> <td>Apr-04</td> </tr> <tr> <td>Reference 20 dB Attenuator</td> <td>SN: 5086 (20b)</td> <td>3-Apr-03 (METAS, No. 251-0340)</td> <td>Apr-04</td> </tr> <tr> <td>Fluke Process Calibrator Type 702</td> <td>SN: 6295803</td> <td>8-Sep-03 (Sintrel SCS No. E-030020)</td> <td>Sep-04</td> </tr> <tr> <td>Power sensor HP 8481A</td> <td>MY41092190</td> <td>18-Sep-02 (SPEAG, in house check Oct-03)</td> <td>In house check: Oct 05</td> </tr> <tr> <td>RF generator R&S SMT06</td> <td>100058</td> <td>23-May-01 (SPEAG, in house check May-03)</td> <td>In house check: May-05</td> </tr> <tr> <td>Network Analyzer HP 8753E</td> <td>US37390585</td> <td>18-Oct-01 (SPEAG, in house check Oct-03)</td> <td>In house check: Oct 05</td> </tr> </tbody> </table> | | | | Model Type | ID # | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration | Power meter EPM E4419B | GB41293874 | 2-Apr-03 (METAS, No 252-0250) | Apr-04 | Power sensor E4412A | MY41495277 | 2-Apr-03 (METAS, No 252-0250) | Apr-04 | Reference 20 dB Attenuator | SN: 5086 (20b) | 3-Apr-03 (METAS, No. 251-0340) | Apr-04 | Fluke Process Calibrator Type 702 | SN: 6295803 | 8-Sep-03 (Sintrel SCS No. E-030020) | Sep-04 | Power sensor HP 8481A | MY41092190 | 18-Sep-02 (SPEAG, in house check Oct-03) | In house check: Oct 05 | RF generator R&S SMT06 | 100058 | 23-May-01 (SPEAG, in house check May-03) | In house check: May-05 | Network Analyzer HP 8753E | US37390585 | 18-Oct-01 (SPEAG, in house check Oct-03) | In house check: Oct 05 |
| Model Type | ID # | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power meter EPM E4419B | GB41293874 | 2-Apr-03 (METAS, No 252-0250) | Apr-04 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor E4412A | MY41495277 | 2-Apr-03 (METAS, No 252-0250) | Apr-04 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Reference 20 dB Attenuator | SN: 5086 (20b) | 3-Apr-03 (METAS, No. 251-0340) | Apr-04 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Fluke Process Calibrator Type 702 | SN: 6295803 | 8-Sep-03 (Sintrel SCS No. E-030020) | Sep-04 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor HP 8481A | MY41092190 | 18-Sep-02 (SPEAG, in house check Oct-03) | In house check: Oct 05 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| RF generator R&S SMT06 | 100058 | 23-May-01 (SPEAG, in house check May-03) | In house check: May-05 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Network Analyzer HP 8753E | US37390585 | 18-Oct-01 (SPEAG, in house check Oct-03) | In house check: Oct 05 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Calibrated by: | Name Nico Vetterli | Function Technician | Signature  | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Approved by: | Name Katja Pokovic | Function Laboratory Director | Signature  | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Date issued: January 26, 2004 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

DASY - Parameters of Probe: EX3DV3 SN:3512

| Sensitivity in Free Space | | Diode Compression ^A | |
|---------------------------|--|--------------------------------|-------|
| NormX | 0.62 $\mu\text{V}/(\text{V}/\text{m})^2$ | DCP X | 98 mV |
| NormY | 0.61 $\mu\text{V}/(\text{V}/\text{m})^2$ | DCP Y | 98 mV |
| NormZ | 0.70 $\mu\text{V}/(\text{V}/\text{m})^2$ | DCP Z | 98 mV |

Sensitivity in Tissue Simulating Liquid (Conversion Factors)

Please see Page 7.

Boundary Effect

Head **900 MHz** **Typical SAR gradient: 5 % per mm**

| Sensor Center to Phantom Surface Distance | | 2.0 mm | 3.0 mm |
|---|------------------------------|--------|--------|
| SAR _{be} [%] | Without Correction Algorithm | 2.5 | 0.6 |
| SAR _{be} [%] | With Correction Algorithm | 0.9 | 0.4 |

Head **1800 MHz** **Typical SAR gradient: 10 % per mm**

| Sensor to Surface Distance | | 2.0 mm | 3.0 mm |
|----------------------------|------------------------------|--------|--------|
| SAR _{be} [%] | Without Correction Algorithm | 4.9 | 3.1 |
| SAR _{be} [%] | With Correction Algorithm | 1.3 | 0.8 |


Sensor Offset

Probe Tip to Sensor Center **1.0** mm




The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A numerical linearization parameter: uncertainty not required

Frequency Response of E-Field (TEM-Cell:ifi110, Waveguide R22)





Receiving Pattern (ϕ), $\theta = 0^\circ$





Axial Isotropy Error $< \pm 0.2$ dB

Dynamic Range $f(\text{SAR}_{\text{head}})$ (Waveguide R22)



Probe Linearity < ± 0.2 dB

Conversion Factor Assessment




| f [MHz] | Validity [MHz] ^a | Tissue | Permittivity | Conductivity | Alpha | Depth | ConvF | Uncertainty |
|---------|-----------------------------|--------|--------------|--------------|-------|-------|-------|---------------|
| 900 | 800-1000 | Head | 41.5 ± 5% | 0.97 ± 5% | 0.40 | 0.80 | 10.02 | ± 9.5% (k=2) |
| 1800 | 1710-1910 | Head | 40.0 ± 5% | 1.40 ± 5% | 0.40 | 1.05 | 8.60 | ± 9.5% (k=2) |
| 5200 | 4940-5460 | Head | 36.0 ± 5% | 4.66 ± 5% | 0.42 | 1.80 | 5.02 | ± 21.8% (k=2) |
| 5800 | 5510-6090 | Head | 35.3 ± 5% | 5.27 ± 5% | 0.42 | 1.80 | 4.54 | ± 23.4% (k=2) |
| 5200 | 4940-5460 | Body | 49.0 ± 5% | 5.30 ± 5% | 0.45 | 1.90 | 4.25 | ± 21.8% (k=2) |
| 5800 | 5510-6090 | Body | 48.2 ± 5% | 6.00 ± 5% | 0.40 | 1.90 | 4.02 | ± 23.4% (k=2) |

^a The stated uncertainty of calibration is assessed according to P1528.

Deviation from Isotropy in HSL

Error (θ, ϕ), $f = 900$ MHz



Spherical Isotropy Error < ± 0.4 dB