



Accredited by the Swiss Accreditation Service (SAS)  
**The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates**

**Accreditation No.: SCS 0108**

Client

**UL**  
 Fremont, USA

Certificate No.

**EX-7810\_May25**

## CALIBRATION CERTIFICATE

Object

**EX3DV4 - SN:7810**

Calibration procedure(s)

**QA CAL-01.v10, QA CAL-12.v10, QA CAL-14.v7, QA CAL-23.v6,  
 QA CAL-25.v8  
 Calibration procedure for dosimetric E-field probes**

Calibration date

**May 08, 2025**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature ( $22 \pm 3$ ) °C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards           | ID         | Calibration Date (Certificate No.)     | Sched. Cal. |
|-----------------------------|------------|----------------------------------------|-------------|
| Power Sensor R&S NRP-33T    | SN: 100967 | 26-Mar-25 (No. 217-04290)              | Mar-26      |
| Type-N mismatch combination | SN: L1119  | 26-Mar-25 (No. 217-04292)              | Mar-26      |
| OCP DAK-12                  | SN: 1016   | 24-Sept-24 (No. OCP-DAK12-1016_Sep24)  | Sep-25      |
| OCP DAK-3.5                 | SN: 1249   | 23-Sept-24 (No. OCP-DAK3.5-1249_Sep24) | Sep-25      |
| Reference Probe EX3DV4      | SN: 7349   | 10-Jan-25 (No. EX3-7349_Jan25)         | Jan-26      |
| DAE4                        | SN: 1301   | 07-Nov-24 (No. DAE4-1301_Nov24)        | Nov-25      |

| Secondary Standards | ID        | Check Date (in house)                            | Sched. Check |
|---------------------|-----------|--------------------------------------------------|--------------|
| ACAP 2020 Setup 1   | SN: L1404 | 30-Sept-24 (No. Report_ACAP2020E-Cave_20240930s) | Sep-25       |

|               |                    |                       |                      |
|---------------|--------------------|-----------------------|----------------------|
|               | <b>Name</b>        | <b>Function</b>       | <b>Signature</b>     |
| Calibrated by | Aidonia Georgiadou | Laboratory Technician |                      |
| Approved by   | Sven Kühn          | Technical Manager     |                      |
|               |                    |                       | Issued: May 08, 2025 |

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

**Glossary**

|                        |                                                                                                                                                |
|------------------------|------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                    | tissue simulating liquid                                                                                                                       |
| NORM <sub>x,y,z</sub>  | sensitivity in free space                                                                                                                      |
| ConvF                  | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                     |
| DCP                    | diode compression point                                                                                                                        |
| CF                     | crest factor (1/duty_cycle) of the RF signal                                                                                                   |
| A, B, C, D             | modulation dependent linearization parameters                                                                                                  |
| Polarization $\varphi$ | $\varphi$ rotation around probe axis                                                                                                           |
| Polarization $\theta$  | $\theta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\theta = 0$ is normal to probe axis |
| Connector Angle        | information used in DASY system to align probe sensor X to the robot coordinate system                                                         |

**Calibration is Performed According to the Following Standards:**

- IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices – Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Methods Applied and Interpretation of Parameters:**

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\theta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the  $E^2$ -field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub> = NORM<sub>x,y,z</sub> \* frequency\_response** (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>; A, B, C, D** are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

## Parameters of Probe: EX3DV4 - SN:7810

### Basic Calibration Parameters

|                                                           | Sensor X | Sensor Y | Sensor Z | Unc (k = 2) |
|-----------------------------------------------------------|----------|----------|----------|-------------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup> | 0.59     | 0.69     | 0.65     | ±10.1%      |
| DCP (mV) <sup>B</sup>                                     | 110.1    | 107.0    | 106.4    | ±4.7%       |

### Calibration Results for Modulation Response

| UID   | Communication System Name   |   | A<br>dB | B<br>dB $\sqrt{\mu\text{V}}$ | C     | D<br>dB | VR<br>mV | Max<br>dev. | Max<br>Unc <sup>E</sup><br>k = 2 |
|-------|-----------------------------|---|---------|------------------------------|-------|---------|----------|-------------|----------------------------------|
| 0     | CW                          | X | 0.00    | 0.00                         | 1.00  | 0.00    | 143.7    | ±1.2%       | ±4.7%                            |
|       |                             | Y | 0.00    | 0.00                         | 1.00  |         | 130.4    |             |                                  |
|       |                             | Z | 0.00    | 0.00                         | 1.00  |         | 149.4    |             |                                  |
| 10352 | Pulse Waveform (200Hz, 10%) | X | 1.63    | 60.95                        | 6.48  | 10.00   | 60.0     | ±2.6%       | ±9.6%                            |
|       |                             | Y | 1.40    | 60.00                        | 5.96  |         | 60.0     |             |                                  |
|       |                             | Z | 1.47    | 60.43                        | 6.35  |         | 60.0     |             |                                  |
| 10353 | Pulse Waveform (200Hz, 20%) | X | 0.84    | 60.00                        | 4.91  | 6.99    | 80.0     | ±2.4%       | ±9.6%                            |
|       |                             | Y | 0.81    | 60.00                        | 4.80  |         | 80.0     |             |                                  |
|       |                             | Z | 46.00   | 80.00                        | 11.00 |         | 80.0     |             |                                  |
| 10354 | Pulse Waveform (200Hz, 40%) | X | 0.45    | 60.00                        | 3.73  | 3.98    | 95.0     | ±2.7%       | ±9.6%                            |
|       |                             | Y | 0.01    | 125.28                       | 0.33  |         | 95.0     |             |                                  |
|       |                             | Z | 0.16    | 140.59                       | 0.19  |         | 95.0     |             |                                  |
| 10355 | Pulse Waveform (200Hz, 60%) | X | 10.11   | 89.44                        | 0.01  | 2.22    | 120.0    | ±1.7%       | ±9.6%                            |
|       |                             | Y | 5.95    | 159.98                       | 1.45  |         | 120.0    |             |                                  |
|       |                             | Z | 6.49    | 160.00                       | 15.35 |         | 120.0    |             |                                  |
| 10387 | QPSK Waveform, 1 MHz        | X | 0.49    | 64.70                        | 13.41 | 1.00    | 150.0    | ±3.1%       | ±9.6%                            |
|       |                             | Y | 0.45    | 62.15                        | 11.58 |         | 150.0    |             |                                  |
|       |                             | Z | 0.58    | 65.39                        | 13.98 |         | 150.0    |             |                                  |
| 10388 | QPSK Waveform, 10 MHz       | X | 1.31    | 67.40                        | 14.22 | 0.00    | 150.0    | ±1.0%       | ±9.6%                            |
|       |                             | Y | 1.21    | 65.18                        | 13.33 |         | 150.0    |             |                                  |
|       |                             | Z | 1.42    | 67.34                        | 14.78 |         | 150.0    |             |                                  |
| 10396 | 64-QAM Waveform, 100 kHz    | X | 1.77    | 65.53                        | 16.12 | 3.01    | 150.0    | ±1.1%       | ±9.6%                            |
|       |                             | Y | 1.60    | 63.78                        | 15.63 |         | 150.0    |             |                                  |
|       |                             | Z | 1.65    | 64.26                        | 16.05 |         | 150.0    |             |                                  |
| 10399 | 64-QAM Waveform, 40 MHz     | X | 2.77    | 67.03                        | 15.36 | 0.00    | 150.0    | ±1.5%       | ±9.6%                            |
|       |                             | Y | 2.71    | 65.99                        | 14.88 |         | 150.0    |             |                                  |
|       |                             | Z | 2.84    | 66.65                        | 15.41 |         | 150.0    |             |                                  |
| 10414 | WLAN CCDF, 64-QAM, 40 MHz   | X | 3.63    | 66.58                        | 15.33 | 0.00    | 150.0    | ±2.5%       | ±9.6%                            |
|       |                             | Y | 3.80    | 66.44                        | 15.38 |         | 150.0    |             |                                  |
|       |                             | Z | 3.92    | 66.81                        | 15.73 |         | 150.0    |             |                                  |

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Pages 5 and 6).

<sup>B</sup> Linearization parameter uncertainty for maximum specified field strength.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## Parameters of Probe: EX3DV4 - SN:7810

### Sensor Model Parameters

|   | C1<br>fF | C2<br>fF | $\alpha$<br>$V^{-1}$ | T1<br>$ms V^{-2}$ | T2<br>$ms V^{-1}$ | T3<br>ms | T4<br>$V^{-2}$ | T5<br>$V^{-1}$ | T6   |
|---|----------|----------|----------------------|-------------------|-------------------|----------|----------------|----------------|------|
| x | 7.3      | 50.94    | 31.44                | 3.84              | 0.00              | 4.90     | 0.64           | 0.00           | 1.00 |
| y | 8.6      | 62.10    | 33.06                | 2.90              | 0.00              | 4.90     | 0.28           | 0.00           | 1.00 |
| z | 8.9      | 64.32    | 33.53                | 2.32              | 0.00              | 4.90     | 0.30           | 0.00           | 1.00 |

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle                               | 63.5°      |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 9 mm       |
| Tip Diameter                                  | 2.5 mm     |
| Probe Tip to Sensor X Calibration Point       | 1 mm       |
| Probe Tip to Sensor Y Calibration Point       | 1 mm       |
| Probe Tip to Sensor Z Calibration Point       | 1 mm       |
| Recommended Measurement Distance from Surface | 1.4 mm     |

**Note:** Measurement distance from surface can be increased to 3–4 mm for an *Area Scan* job.

## Parameters of Probe: EX3DV4 - SN:7810

### Calibration Parameter Determined in HSL

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 750                  | 41.9                               | 0.89                            | 9.62    | 9.50    | 9.76    | 0.36               | 1.27                    | ±11.0%                   |
| 900                  | 41.5                               | 0.97                            | 9.08    | 8.97    | 9.21    | 0.36               | 1.27                    | ±11.0%                   |
| 1450                 | 40.5                               | 1.20                            | 7.93    | 7.83    | 8.05    | 0.36               | 1.27                    | ±11.0%                   |
| 1640                 | 40.2                               | 1.31                            | 7.74    | 7.65    | 7.86    | 0.36               | 1.27                    | ±11.0%                   |
| 1750                 | 40.1                               | 1.37                            | 7.81    | 7.72    | 7.93    | 0.36               | 1.27                    | ±11.0%                   |
| 1900                 | 40.0                               | 1.40                            | 7.72    | 7.63    | 7.84    | 0.36               | 1.27                    | ±11.0%                   |
| 2100                 | 39.8                               | 1.49                            | 7.51    | 7.42    | 7.62    | 0.36               | 1.27                    | ±11.0%                   |
| 2300                 | 39.5                               | 1.67                            | 7.51    | 7.42    | 7.62    | 0.36               | 1.27                    | ±11.0%                   |
| 2450                 | 39.2                               | 1.80                            | 7.31    | 7.22    | 7.42    | 0.36               | 1.27                    | ±11.0%                   |
| 2600                 | 39.0                               | 1.96                            | 7.16    | 7.08    | 7.27    | 0.36               | 1.27                    | ±11.0%                   |
| 3300                 | 38.2                               | 2.71                            | 6.31    | 6.23    | 6.40    | 0.37               | 1.27                    | ±13.1%                   |
| 3500                 | 37.9                               | 2.91                            | 6.38    | 6.30    | 6.47    | 0.37               | 1.27                    | ±13.1%                   |
| 3700                 | 37.7                               | 3.12                            | 6.37    | 6.29    | 6.46    | 0.37               | 1.27                    | ±13.1%                   |
| 3900                 | 37.5                               | 3.32                            | 6.21    | 6.14    | 6.30    | 0.37               | 1.27                    | ±13.1%                   |
| 4100                 | 37.2                               | 3.53                            | 6.08    | 6.00    | 6.17    | 0.37               | 1.27                    | ±13.1%                   |
| 4200                 | 37.1                               | 3.63                            | 6.14    | 6.07    | 6.23    | 0.37               | 1.27                    | ±13.1%                   |
| 4400                 | 36.9                               | 3.84                            | 5.95    | 5.88    | 6.04    | 0.37               | 1.27                    | ±13.1%                   |
| 4600                 | 36.7                               | 4.04                            | 5.84    | 5.77    | 5.93    | 0.37               | 1.27                    | ±13.1%                   |
| 4800                 | 36.4                               | 4.25                            | 6.00    | 5.93    | 6.09    | 0.37               | 1.27                    | ±13.1%                   |
| 4950                 | 36.3                               | 4.40                            | 5.88    | 5.81    | 5.97    | 0.35               | 1.27                    | ±13.1%                   |
| 5250                 | 35.9                               | 4.71                            | 5.71    | 5.64    | 5.79    | 0.32               | 1.27                    | ±13.1%                   |
| 5600                 | 35.5                               | 5.07                            | 5.28    | 5.22    | 5.36    | 0.29               | 1.27                    | ±13.1%                   |
| 5750                 | 35.4                               | 5.22                            | 5.23    | 5.16    | 5.31    | 0.27               | 1.27                    | ±13.1%                   |
| 5850                 | 35.2                               | 5.32                            | 5.36    | 5.29    | 5.44    | 0.26               | 1.27                    | ±13.1%                   |

<sup>C</sup> Frequency validity above 300 MHz of ±100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ±50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ±10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4–9 MHz, and ConvF assessed at 13 MHz is 9–19 MHz. Above 5 GHz frequency validity can be extended to ±110 MHz.

<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than ±5% from the target values (typically better than ±3%) and are valid for TSL with deviations of up to ±10% if SAR correction is applied.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz and below ±2% for frequencies between 3–6 GHz at any distance larger than half the probe tip diameter from the boundary.

<sup>H</sup> The stated uncertainty is the total calibration uncertainty (k = 2) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.

## Parameters of Probe: EX3DV4 - SN:7810

### Calibration Parameter Determined in HSL

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 6500                 | 34.5                               | 6.07                            | 5.17    | 5.11    | 5.25    | 0.20               | 1.27                    | ±18.6%                   |

<sup>C</sup> Frequency validity is -600/+700 MHz at 6.5 GHz, ±700 MHz at or above 7 GHz, and ±150 MHz at 9.85 GHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than ±10% from the target values (typically better than ±6%) and are valid for TSL with deviations of up to ±10%.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz; below ±2% for frequencies between 3–6 GHz; and below ±4% for frequencies between 6–10 GHz at any distance larger than half the probe tip diameter from the boundary.



<sup>H</sup> The stated uncertainty is the total calibration uncertainty (k = 2) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.

## Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide:R22)




Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )


**Receiving Pattern ( $\phi$ ),  $\vartheta = 0^\circ$** 

100 MHz      600 MHz      1800 MHz      2500 MHz

Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

**Dynamic Range  $f(\text{SAR}_{\text{HSL}})$** (TEM cell,  $f_{\text{eval}} = 1900 \text{ MHz}$ )


—●— not compensated      —●— compensated



—●— not compensated      —●— compensated


Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ), f = 900 MHz





Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

Client

UL

Fremont, USA

Certificate No.

EX-3990\_Feb25

## CALIBRATION CERTIFICATE

Object

EX3DV4 - SN:3990

Calibration procedure(s)

QA CAL-01.v10, QA CAL-12.v10, QA CAL-14.v7, QA CAL-23.v6,  
QA CAL-25.v8

Calibration procedure for dosimetric E-field probes

Calibration date

February 07, 2025

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^\circ\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&amp;TE critical for calibration)

| Primary Standards                    | ID         | Calibration Date (Certificate No.)     | Sched. Cal. |
|--------------------------------------|------------|----------------------------------------|-------------|
| Power Sensor R&S NRP-33T             | SN: 100967 | 28-Mar-24 (No. 217-04038)              | Mar-25      |
| Short [S6019i] + Attenuator [S6020i] | SN: L1119  | 26-Mar-24 (No. 217-04048)              | Mar-25      |
| OCP DAK-12                           | SN: 1016   | 24-Sept-24 (No. OCP-DAK12-1016_Sep24)  | Sep-25      |
| OCP DAK-3.5                          | SN: 1249   | 23-Sept-24 (No. OCP-DAK3.5-1249_Sep24) | Sep-25      |
| Reference Probe EX3DV4               | SN: 7349   | 10-Jan-25 (No. EX3-7349_Jan25)         | Jan-26      |
| DAE4                                 | SN: 1301   | 07-Nov-24 (No. DAE4-1301_Nov24)        | Nov-25      |

| Secondary Standards       | ID        | Check Date (in house)                            | Sched. Check |
|---------------------------|-----------|--------------------------------------------------|--------------|
| ACAP 2020 Calibration Box | SN: L1404 | 30-Sept-24 (No. Report_ACAP2020E-Cave_20240930s) | Sep-25       |

|               | Name       | Function              |
|---------------|------------|-----------------------|
| Calibrated by | Paulo Pina | Laboratory Technician |
| Approved by   | Sven Kühn  | Technical Manager     |

Signature

Issued: February 07, 2025

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)  
**The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates**

**Accreditation No.: SCS 0108**

## Glossary

|                          |                                                                                                                                                      |
|--------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                             |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                            |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                           |
| DCP                      | diode compression point                                                                                                                              |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                         |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                        |
| Polarization $\varphi$   | $\varphi$ rotation around probe axis                                                                                                                 |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                               |

## Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices – Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

## Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

## Parameters of Probe: EX3DV4 - SN:3990

### Basic Calibration Parameters

|                                                           | Sensor X | Sensor Y | Sensor Z | Unc ( $k = 2$ ) |
|-----------------------------------------------------------|----------|----------|----------|-----------------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup> | 0.61     | 0.62     | 0.62     | $\pm 10.1\%$    |
| DCP (mV) <sup>B</sup>                                     | 100.7    | 100.5    | 99.6     | $\pm 4.7\%$     |

### Calibration Results for Modulation Response

| UID   | Communication System Name   |   | A<br>dB | B<br>dB $\sqrt{\mu\text{V}}$ | C     | D<br>dB | VR<br>mV | Max<br>dev. | Max<br>Unc <sup>E</sup><br>$k = 2$ |
|-------|-----------------------------|---|---------|------------------------------|-------|---------|----------|-------------|------------------------------------|
| 0     | CW                          | X | 0.00    | 0.00                         | 1.00  | 0.00    | 145.2    | $\pm 0.7\%$ | $\pm 4.7\%$                        |
|       |                             | Y | 0.00    | 0.00                         | 1.00  |         | 141.1    |             |                                    |
|       |                             | Z | 0.00    | 0.00                         | 1.00  |         | 121.1    |             |                                    |
| 10352 | Pulse Waveform (200Hz, 10%) | X | 84.00   | 108.00                       | 25.00 | 10.00   | 60.0     | $\pm 3.2\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 20.00   | 89.32                        | 19.50 |         | 60.0     |             |                                    |
|       |                             | Z | 20.00   | 94.17                        | 22.38 |         | 60.0     |             |                                    |
| 10353 | Pulse Waveform (200Hz, 20%) | X | 20.00   | 94.32                        | 21.38 | 6.99    | 80.0     | $\pm 1.5\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 20.00   | 90.15                        | 19.01 |         | 80.0     |             |                                    |
|       |                             | Z | 20.00   | 98.86                        | 23.69 |         | 80.0     |             |                                    |
| 10354 | Pulse Waveform (200Hz, 40%) | X | 20.00   | 98.72                        | 22.24 | 3.98    | 95.0     | $\pm 1.1\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 20.00   | 91.78                        | 18.66 |         | 95.0     |             |                                    |
|       |                             | Z | 20.00   | 100.96                       | 23.30 |         | 95.0     |             |                                    |
| 10355 | Pulse Waveform (200Hz, 60%) | X | 20.00   | 103.98                       | 23.46 | 2.22    | 120.0    | $\pm 1.1\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 20.00   | 95.35                        | 19.28 |         | 120.0    |             |                                    |
|       |                             | Z | 20.00   | 105.45                       | 24.08 |         | 120.0    |             |                                    |
| 10387 | QPSK Waveform, 1 MHz        | X | 1.65    | 65.31                        | 14.46 | 1.00    | 150.0    | $\pm 1.8\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 1.69    | 65.94                        | 14.79 |         | 150.0    |             |                                    |
|       |                             | Z | 1.60    | 64.85                        | 14.07 |         | 150.0    |             |                                    |
| 10388 | QPSK Waveform, 10 MHz       | X | 2.17    | 67.14                        | 15.16 | 0.00    | 150.0    | $\pm 1.2\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 2.25    | 67.84                        | 15.53 |         | 150.0    |             |                                    |
|       |                             | Z | 2.11    | 66.59                        | 14.80 |         | 150.0    |             |                                    |
| 10396 | 64-QAM Waveform, 100 kHz    | X | 2.77    | 69.26                        | 18.11 | 3.01    | 150.0    | $\pm 0.7\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 2.80    | 69.79                        | 18.43 |         | 150.0    |             |                                    |
|       |                             | Z | 2.66    | 68.61                        | 17.83 |         | 150.0    |             |                                    |
| 10399 | 64-QAM Waveform, 40 MHz     | X | 3.34    | 66.11                        | 15.17 | 0.00    | 150.0    | $\pm 1.0\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 3.40    | 66.50                        | 15.39 |         | 150.0    |             |                                    |
|       |                             | Z | 3.30    | 65.81                        | 14.97 |         | 150.0    |             |                                    |
| 10414 | WLAN CCDF, 64-QAM, 40 MHz   | X | 4.74    | 65.05                        | 15.14 | 0.00    | 150.0    | $\pm 2.2\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 4.78    | 65.27                        | 15.27 |         | 150.0    |             |                                    |
|       |                             | Z | 4.71    | 64.91                        | 15.04 |         | 150.0    |             |                                    |

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 5 and 6).

<sup>B</sup> Linearization parameter uncertainty for maximum specified field strength.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## Parameters of Probe: EX3DV4 - SN:3990

### Sensor Model Parameters

|   | C1<br>fF | C2<br>fF | $\alpha$<br>V <sup>-1</sup> | T1<br>msV <sup>-2</sup> | T2<br>msV <sup>-1</sup> | T3<br>ms | T4<br>V <sup>-2</sup> | T5<br>V <sup>-1</sup> | T6   |
|---|----------|----------|-----------------------------|-------------------------|-------------------------|----------|-----------------------|-----------------------|------|
| x | 46.4     | 342.58   | 34.81                       | 15.50                   | 0.01                    | 5.09     | 0.95                  | 0.25                  | 1.01 |
| y | 45.6     | 337.22   | 34.81                       | 18.01                   | 0.00                    | 5.05     | 0.91                  | 0.24                  | 1.01 |
| z | 44.9     | 333.81   | 35.15                       | 14.24                   | 0.00                    | 5.10     | 0.82                  | 0.25                  | 1.01 |

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle                               | -47.0°     |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 9 mm       |
| Tip Diameter                                  | 2.5 mm     |
| Probe Tip to Sensor X Calibration Point       | 1 mm       |
| Probe Tip to Sensor Y Calibration Point       | 1 mm       |
| Probe Tip to Sensor Z Calibration Point       | 1 mm       |
| Recommended Measurement Distance from Surface | 1.4 mm     |

**Note:** Measurement distance from surface can be increased to 3–4 mm for an *Area Scan* job.

## Parameters of Probe: EX3DV4 - SN:3990

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 750                  | 41.9                               | 0.89                            | 9.29    | 10.11   | 9.10    | 0.37               | 1.27                    | ±11.0%                   |
| 900                  | 41.5                               | 0.97                            | 8.87    | 9.65    | 8.69    | 0.37               | 1.27                    | ±11.0%                   |
| 1640                 | 40.2                               | 1.31                            | 7.86    | 8.54    | 7.70    | 0.36               | 1.27                    | ±11.0%                   |
| 1750                 | 40.1                               | 1.37                            | 7.89    | 8.58    | 7.73    | 0.36               | 1.27                    | ±11.0%                   |
| 1900                 | 40.0                               | 1.40                            | 7.51    | 8.17    | 7.35    | 0.36               | 1.27                    | ±11.0%                   |
| 2100                 | 39.8                               | 1.49                            | 7.39    | 8.04    | 7.24    | 0.36               | 1.27                    | ±11.0%                   |
| 2300                 | 39.5                               | 1.67                            | 7.19    | 7.82    | 7.04    | 0.36               | 1.27                    | ±11.0%                   |
| 2450                 | 39.2                               | 1.80                            | 6.94    | 7.55    | 6.80    | 0.36               | 1.27                    | ±11.0%                   |
| 2600                 | 39.0                               | 1.96                            | 6.95    | 7.56    | 6.81    | 0.36               | 1.27                    | ±11.0%                   |
| 3300                 | 38.2                               | 2.71                            | 6.44    | 7.00    | 6.31    | 0.35               | 1.27                    | ±13.1%                   |
| 3500                 | 37.9                               | 2.91                            | 6.58    | 7.16    | 6.45    | 0.35               | 1.27                    | ±13.1%                   |
| 3700                 | 37.7                               | 3.12                            | 6.46    | 7.03    | 6.33    | 0.35               | 1.27                    | ±13.1%                   |
| 3900                 | 37.5                               | 3.32                            | 6.51    | 7.08    | 6.38    | 0.35               | 1.27                    | ±13.1%                   |
| 4100                 | 37.2                               | 3.53                            | 6.45    | 7.02    | 6.32    | 0.34               | 1.27                    | ±13.1%                   |
| 4200                 | 37.1                               | 3.63                            | 6.34    | 6.89    | 6.21    | 0.34               | 1.27                    | ±13.1%                   |
| 4400                 | 36.9                               | 3.84                            | 6.16    | 6.70    | 6.03    | 0.34               | 1.27                    | ±13.1%                   |
| 4600                 | 36.7                               | 4.04                            | 6.18    | 6.72    | 6.05    | 0.34               | 1.27                    | ±13.1%                   |
| 4800                 | 36.4                               | 4.25                            | 6.30    | 6.85    | 6.17    | 0.34               | 1.27                    | ±13.1%                   |
| 4950                 | 36.3                               | 4.40                            | 6.18    | 6.72    | 6.06    | 0.33               | 1.27                    | ±13.1%                   |
| 5250                 | 35.9                               | 4.71                            | 5.44    | 5.92    | 5.33    | 0.30               | 1.27                    | ±13.1%                   |
| 5600                 | 35.5                               | 5.07                            | 5.11    | 5.56    | 5.00    | 0.27               | 1.27                    | ±13.1%                   |
| 5750                 | 35.4                               | 5.22                            | 4.94    | 5.37    | 4.84    | 0.26               | 1.27                    | ±13.1%                   |
| 5850                 | 35.2                               | 5.32                            | 4.97    | 5.40    | 4.86    | 0.25               | 1.27                    | ±13.1%                   |

<sup>C</sup> Frequency validity above 300 MHz of ±100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ±50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ±10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4–9 MHz, and ConvF assessed at 13 MHz is 9–19 MHz. Above 5 GHz frequency validity can be extended to ±110 MHz.

<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than ±5% from the target values (typically better than ±3%) and are valid for TSL with deviations of up to ±10% if SAR correction is applied.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz and below ±2% for frequencies between 3–6 GHz at any distance larger than half the probe tip diameter from the boundary.

<sup>H</sup> The stated uncertainty is the total calibration uncertainty (k = 2) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.

## Parameters of Probe: EX3DV4 - SN:3990

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 6500                 | 34.5                               | 6.07                            | 5.23    | 5.69    | 5.12    | 0.20               | 1.27                    | ±18.6%                   |

<sup>C</sup> Frequency validity at 6.5 GHz is -600/+700 MHz, and ±700 MHz at or above 7 GHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than ±10% from the target values (typically better than ±6%) and are valid for TSL with deviations of up to ±10%.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz; below ±2% for frequencies between 3–6 GHz; and below ±4% for frequencies between 6–10 GHz at any distance larger than half the probe tip diameter from the boundary.



<sup>H</sup> The stated uncertainty is the total calibration uncertainty (k = 2) of Norm·ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.

## Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide:R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )


**Receiving Pattern ( $\phi$ ),  $\vartheta = 0^\circ$** 

100 MHz      600 MHz      1800 MHz      2500 MHz

Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )


**Dynamic Range  $f(\text{SAR}_{\text{head}})$** (TEM cell,  $f_{\text{eval}} = 1900 \text{ MHz}$ )Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ), f = 900 MHz





Accredited by the Swiss Accreditation Service (SAS)  
**The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates**

Accreditation No.: **SCS 0108**

Client

**UL**  
Fremont, USA

Certificate No.

**EX-7915\_Mar25**

## CALIBRATION CERTIFICATE

Object **EX3DV4 - SN:7915**

Calibration procedure(s) **QA CAL-01.v10, QA CAL-12.v10, QA CAL-14.v7, QA CAL-23.v6,  
QA CAL-25.v8**  
**Calibration procedure for dosimetric E-field probes**

Calibration date **March 21, 2025**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^\circ\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards                    | ID         | Calibration Date (Certificate No.)     | Sched. Cal. |
|--------------------------------------|------------|----------------------------------------|-------------|
| Power Sensor R&S NRP-33T             | SN: 100967 | 28-Mar-24 (No. 217-04038)              | Mar-25      |
| Short [S6019i] + Attenuator [S6020i] | SN: L1119  | 26-Mar-24 (No. 217-04048)              | Mar-25      |
| OCP DAK-12                           | SN: 1016   | 24-Sept-24 (No. OCP-DAK12-1016_Sep24)  | Sep-25      |
| OCP DAK-3.5                          | SN: 1249   | 23-Sept-24 (No. OCP-DAK3.5-1249_Sep24) | Sep-25      |
| Reference Probe EX3DV4               | SN: 7349   | 10-Jan-25 (No. EX3-7349_Jan25)         | Jan-26      |
| DAE4                                 | SN: 1301   | 07-Nov-24 (No. DAE4-1301_Nov24)        | Nov-25      |

| Secondary Standards       | ID        | Check Date (in house)                            | Sched. Check |
|---------------------------|-----------|--------------------------------------------------|--------------|
| ACAP 2020 Calibration Box | SN: L1404 | 30-Sept-24 (No. Report_ACAP2020E-Cave_20240930s) | Sep-25       |

Calibrated by **Aidonia Georgiadou** Laboratory Technician

Approved by **Sven Kühn** Technical Manager

Signature

Issued: March 26, 2025

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108****Glossary**

|                          |                                                                                                                                                      |
|--------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                             |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                            |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                           |
| DCP                      | diode compression point                                                                                                                              |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                         |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                        |
| Polarization $\varphi$   | $\varphi$ rotation around probe axis                                                                                                                 |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                               |

**Calibration is Performed According to the Following Standards:**

- IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices – Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Methods Applied and Interpretation of Parameters:**

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub> = NORM<sub>x,y,z</sub> \* frequency\_response** (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

## Parameters of Probe: EX3DV4 - SN:7915

### Basic Calibration Parameters

|                                       | Sensor X | Sensor Y | Sensor Z | Unc (k = 2) |
|---------------------------------------|----------|----------|----------|-------------|
| Norm ( $\mu V/(V/m)^2$ ) <sup>A</sup> | 0.65     | 0.58     | 0.61     | ±10.1%      |
| DCP (mV) <sup>B</sup>                 | 104.2    | 105.9    | 107.1    | ±4.7%       |

### Calibration Results for Modulation Response

| UID   | Communication System Name   |   | A<br>dB | B<br>dB $\sqrt{\mu V}$ | C     | D<br>dB | VR<br>mV | Max<br>dev. | Max<br>Unc <sup>E</sup><br>k = 2 |
|-------|-----------------------------|---|---------|------------------------|-------|---------|----------|-------------|----------------------------------|
| 0     | CW                          | X | 0.00    | 0.00                   | 1.00  | 0.00    | 122.3    | ±0.9%       | ±4.7%                            |
|       |                             | Y | 0.00    | 0.00                   | 1.00  |         | 146.4    |             |                                  |
|       |                             | Z | 0.00    | 0.00                   | 1.00  |         | 146.7    |             |                                  |
| 10352 | Pulse Waveform (200Hz, 10%) | X | 1.48    | 60.42                  | 6.34  | 10.00   | 60.0     | ±2.5%       | ±9.6%                            |
|       |                             | Y | 1.48    | 60.40                  | 6.22  |         | 60.0     |             |                                  |
|       |                             | Z | 1.65    | 61.14                  | 6.51  |         | 60.0     |             |                                  |
| 10353 | Pulse Waveform (200Hz, 20%) | X | 44.00   | 80.00                  | 11.00 | 6.99    | 80.0     | ±2.3%       | ±9.6%                            |
|       |                             | Y | 0.81    | 60.00                  | 4.86  |         | 80.0     |             |                                  |
|       |                             | Z | 22.00   | 74.00                  | 9.00  |         | 80.0     |             |                                  |
| 10354 | Pulse Waveform (200Hz, 40%) | X | 0.07    | 130.69                 | 0.59  | 3.98    | 95.0     | ±2.4%       | ±9.6%                            |
|       |                             | Y | 0.08    | 127.37                 | 0.08  |         | 95.0     |             |                                  |
|       |                             | Z | 78.00   | 74.00                  | 7.00  |         | 95.0     |             |                                  |
| 10355 | Pulse Waveform (200Hz, 60%) | X | 7.84    | 159.57                 | 15.11 | 2.22    | 120.0    | ±1.6%       | ±9.6%                            |
|       |                             | Y | 0.38    | 60.00                  | 2.51  |         | 120.0    |             |                                  |
|       |                             | Z | 9.80    | 127.52                 | 4.55  |         | 120.0    |             |                                  |
| 10387 | QPSK Waveform, 1 MHz        | X | 0.57    | 62.99                  | 12.01 | 1.00    | 150.0    | ±3.7%       | ±9.6%                            |
|       |                             | Y | 0.55    | 62.72                  | 11.56 |         | 150.0    |             |                                  |
|       |                             | Z | 0.74    | 66.95                  | 14.01 |         | 150.0    |             |                                  |
| 10388 | QPSK Waveform, 10 MHz       | X | 1.33    | 65.14                  | 13.64 | 0.00    | 150.0    | ±1.2%       | ±9.6%                            |
|       |                             | Y | 1.30    | 64.88                  | 13.32 |         | 150.0    |             |                                  |
|       |                             | Z | 1.51    | 67.24                  | 14.83 |         | 150.0    |             |                                  |
| 10396 | 64-QAM Waveform, 100 kHz    | X | 1.59    | 63.31                  | 15.44 | 3.01    | 150.0    | ±1.1%       | ±9.6%                            |
|       |                             | Y | 1.66    | 63.98                  | 15.52 |         | 150.0    |             |                                  |
|       |                             | Z | 1.78    | 65.60                  | 16.57 |         | 150.0    |             |                                  |
| 10399 | 64-QAM Waveform, 40 MHz     | X | 2.82    | 65.87                  | 14.90 | 0.00    | 150.0    | ±1.8%       | ±9.6%                            |
|       |                             | Y | 2.80    | 65.87                  | 14.80 |         | 150.0    |             |                                  |
|       |                             | Z | 2.95    | 66.82                  | 15.43 |         | 150.0    |             |                                  |
| 10414 | WLAN CCDF, 64-QAM, 40 MHz   | X | 3.99    | 66.34                  | 15.48 | 0.00    | 150.0    | ±3.2%       | ±9.6%                            |
|       |                             | Y | 3.81    | 65.64                  | 15.04 |         | 150.0    |             |                                  |
|       |                             | Z | 3.98    | 66.32                  | 15.53 |         | 150.0    |             |                                  |

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Page 5).

<sup>B</sup> Linearization parameter uncertainty for maximum specified field strength.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## Parameters of Probe: EX3DV4 - SN:7915

### Sensor Model Parameters

|   | C1<br>fF | C2<br>fF | $\alpha$<br>$V^{-1}$ | T1<br>$ms V^{-2}$ | T2<br>$ms V^{-1}$ | T3<br>ms | T4<br>$V^{-2}$ | T5<br>$V^{-1}$ | T6   |
|---|----------|----------|----------------------|-------------------|-------------------|----------|----------------|----------------|------|
| x | 10.5     | 76.64    | 33.89                | 1.64              | 0.00              | 4.90     | 0.12           | 0.02           | 1.00 |
| y | 10.3     | 74.38    | 33.46                | 2.90              | 0.00              | 4.90     | 0.41           | 0.00           | 1.00 |
| z | 11.1     | 79.49    | 33.10                | 3.30              | 0.00              | 4.90     | 0.51           | 0.00           | 1.00 |

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle                               | -3.7°      |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 9 mm       |
| Tip Diameter                                  | 2.5 mm     |
| Probe Tip to Sensor X Calibration Point       | 1 mm       |
| Probe Tip to Sensor Y Calibration Point       | 1 mm       |
| Probe Tip to Sensor Z Calibration Point       | 1 mm       |
| Recommended Measurement Distance from Surface | 1.4 mm     |

**Note:** Measurement distance from surface can be increased to 3–4 mm for an *Area Scan* job.

## Parameters of Probe: EX3DV4 - SN:7915

### Calibration Parameter Determined in HSL

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 750                  | 41.9                               | 0.89                            | 9.54    | 9.48    | 9.60    | 0.35               | 1.27                    | ±11.0%                   |
| 900                  | 41.5                               | 0.97                            | 9.21    | 9.16    | 9.27    | 0.35               | 1.27                    | ±11.0%                   |
| 1750                 | 40.1                               | 1.37                            | 8.01    | 7.96    | 8.06    | 0.34               | 1.27                    | ±11.0%                   |
| 1900                 | 40.0                               | 1.40                            | 7.75    | 7.71    | 7.81    | 0.34               | 1.27                    | ±11.0%                   |
| 2300                 | 39.5                               | 1.67                            | 7.35    | 7.31    | 7.40    | 0.33               | 1.27                    | ±11.0%                   |
| 2450                 | 39.2                               | 1.80                            | 7.20    | 7.16    | 7.25    | 0.33               | 1.27                    | ±11.0%                   |
| 2600                 | 39.0                               | 1.96                            | 7.06    | 7.02    | 7.11    | 0.33               | 1.27                    | ±11.0%                   |
| 5250                 | 35.9                               | 4.71                            | 5.87    | 5.83    | 5.91    | 0.29               | 1.27                    | ±13.1%                   |
| 5600                 | 35.5                               | 5.07                            | 5.34    | 5.31    | 5.38    | 0.27               | 1.27                    | ±13.1%                   |
| 5750                 | 35.4                               | 5.22                            | 5.52    | 5.49    | 5.56    | 0.26               | 1.27                    | ±13.1%                   |


<sup>C</sup> Frequency validity above 300 MHz of ±100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ±50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ±10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4–9 MHz, and ConvF assessed at 13 MHz is 9–19 MHz. Above 5 GHz frequency validity can be extended to ±110 MHz.

<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than ±5% from the target values (typically better than ±3%) and are valid for TSL with deviations of up to ±10% if SAR correction is applied.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz and below ±2% for frequencies between 3–6 GHz at any distance larger than half the probe tip diameter from the boundary.


<sup>H</sup> The stated uncertainty is the total calibration uncertainty (k = 2) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.



Frequency Response of E-Field  
(TEM-Cell:ifi110 EXX, Waveguide:R22)




Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

Receiving Pattern ( $\phi$ ),  $\vartheta = 0^\circ$




**Dynamic Range  $f(\text{SAR}_{\text{HSL}})$** (TEM cell,  $f_{\text{eval}} = 1900 \text{ MHz}$ )Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

Conversion Factor Assessment



Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ), f = 900 MHz





Accredited by the Swiss Accreditation Service (SAS)  
**The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates**

Accreditation No.: **SCS 0108**

Client

**UL**  
Fremont, USA

Certificate No.

**EX-7808\_Mar25**

## CALIBRATION CERTIFICATE

Object **EX3DV4 - SN:7808**

Calibration procedure(s) **QA CAL-01.v10, QA CAL-12.v10, QA CAL-14.v7, QA CAL-23.v6,  
QA CAL-25.v8  
Calibration procedure for dosimetric E-field probes**

Calibration date **March 12, 2025**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^\circ\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards                    | ID         | Calibration Date (Certificate No.)    | Sched. Cal. |
|--------------------------------------|------------|---------------------------------------|-------------|
| Power Sensor R&S NRP-33T             | SN: 100967 | 28-Mar-24 (No. 217-04038)             | Mar-25      |
| Short [S6019I] + Attenuator [S6020I] | SN: L1119  | 26-Mar-24 (No. 217-04048)             | Mar-25      |
| OCP DAK-12                           | SN: 1016   | 24-Sep-24 (No. OCP-DAK12-1016_Sep24)  | Sep-25      |
| OCP DAK-3.5                          | SN: 1249   | 23-Sep-24 (No. OCP-DAK3.5-1249_Sep24) | Sep-25      |
| Reference Probe EX3DV4               | SN: 7349   | 10-Jan-25 (No. EX3-7349_Jan25)        | Jan-26      |
| DAE4                                 | SN: 1301   | 07-Nov-24 (No. DAE4-1301_Nov24)       | Nov-25      |

| Secondary Standards       | ID        | Check Date (in house)                           | Sched. Check |
|---------------------------|-----------|-------------------------------------------------|--------------|
| ACAP 2020 Calibration Box | SN: L1404 | 30-Sep-24 (No. Report_ACAP2020E-Cave_20240930s) | Sep-25       |

|               |                  |                       |           |
|---------------|------------------|-----------------------|-----------|
|               | Name             | Function              | Signature |
| Calibrated by | Krešimir Franjić | Laboratory Technician |           |
| Approved by   | Sven Kühn        | Technical Manager     |           |

Issued: March 12, 2025

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)  
**The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates**

**Accreditation No.: SCS 0108**

## Glossary

|                        |                                                                                                                                                |
|------------------------|------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                    | tissue simulating liquid                                                                                                                       |
| NORM <sub>x,y,z</sub>  | sensitivity in free space                                                                                                                      |
| ConvF                  | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                     |
| DCP                    | diode compression point                                                                                                                        |
| CF                     | crest factor (1/duty_cycle) of the RF signal                                                                                                   |
| A, B, C, D             | modulation dependent linearization parameters                                                                                                  |
| Polarization $\varphi$ | $\varphi$ rotation around probe axis                                                                                                           |
| Polarization $\theta$  | $\theta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\theta = 0$ is normal to probe axis |
| Connector Angle        | information used in DASY system to align probe sensor X to the robot coordinate system                                                         |

## Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices – Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

## Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\theta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

## Parameters of Probe: EX3DV4 - SN:7808

### Basic Calibration Parameters

|                                       | Sensor X | Sensor Y | Sensor Z | Unc ( $k = 2$ ) |
|---------------------------------------|----------|----------|----------|-----------------|
| Norm ( $\mu V/(V/m)^2$ ) <sup>A</sup> | 0.61     | 0.70     | 0.70     | $\pm 10.1\%$    |
| DCP (mV) <sup>B</sup>                 | 108.4    | 103.9    | 104.3    | $\pm 4.7\%$     |

### Calibration Results for Modulation Response

| UID   | Communication System Name   |   | A<br>dB | B<br>dB $\sqrt{\mu V}$ | C     | D<br>dB | VR<br>mV | Max<br>dev. | Max<br>Unc <sup>E</sup><br>$k = 2$ |
|-------|-----------------------------|---|---------|------------------------|-------|---------|----------|-------------|------------------------------------|
| 0     | CW                          | X | 0.00    | 0.00                   | 1.00  | 0.00    | 145.5    | $\pm 1.5\%$ | $\pm 4.7\%$                        |
|       |                             | Y | 0.00    | 0.00                   | 1.00  |         | 132.8    |             |                                    |
|       |                             | Z | 0.00    | 0.00                   | 1.00  |         | 128.4    |             |                                    |
| 10352 | Pulse Waveform (200Hz, 10%) | X | 2.00    | 62.00                  | 7.00  | 10.00   | 60.0     | $\pm 2.9\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 1.57    | 60.77                  | 6.45  |         | 60.0     |             |                                    |
|       |                             | Z | 1.61    | 61.02                  | 6.60  |         | 60.0     |             |                                    |
| 10353 | Pulse Waveform (200Hz, 20%) | X | 10.00   | 72.00                  | 9.00  | 6.99    | 80.0     | $\pm 2.1\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 0.78    | 60.00                  | 4.85  |         | 80.0     |             |                                    |
|       |                             | Z | 0.78    | 60.00                  | 4.92  |         | 80.0     |             |                                    |
| 10354 | Pulse Waveform (200Hz, 40%) | X | 0.40    | 60.00                  | 4.01  | 3.98    | 95.0     | $\pm 2.4\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 0.00    | 125.12                 | 0.64  |         | 95.0     |             |                                    |
|       |                             | Z | 0.28    | 152.25                 | 1.36  |         | 95.0     |             |                                    |
| 10355 | Pulse Waveform (200Hz, 60%) | X | 11.26   | 91.99                  | 0.54  | 2.22    | 120.0    | $\pm 1.7\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 6.17    | 159.97                 | 2.70  |         | 120.0    |             |                                    |
|       |                             | Z | 8.18    | 123.15                 | 10.14 |         | 120.0    |             |                                    |
| 10387 | QPSK Waveform, 1 MHz        | X | 0.51    | 64.72                  | 13.70 | 1.00    | 150.0    | $\pm 2.8\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 0.51    | 62.94                  | 12.03 |         | 150.0    |             |                                    |
|       |                             | Z | 0.49    | 63.98                  | 13.14 |         | 150.0    |             |                                    |
| 10388 | QPSK Waveform, 10 MHz       | X | 1.31    | 67.33                  | 14.10 | 0.00    | 150.0    | $\pm 0.8\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 1.29    | 65.45                  | 13.60 |         | 150.0    |             |                                    |
|       |                             | Z | 1.28    | 66.63                  | 13.89 |         | 150.0    |             |                                    |
| 10396 | 64-QAM Waveform, 100 kHz    | X | 1.77    | 65.29                  | 16.19 | 3.01    | 150.0    | $\pm 1.4\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 1.66    | 64.07                  | 15.53 |         | 150.0    |             |                                    |
|       |                             | Z | 1.54    | 63.29                  | 15.55 |         | 150.0    |             |                                    |
| 10399 | 64-QAM Waveform, 40 MHz     | X | 2.79    | 67.10                  | 15.42 | 0.00    | 150.0    | $\pm 1.4\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 2.79    | 66.11                  | 14.99 |         | 150.0    |             |                                    |
|       |                             | Z | 2.76    | 66.66                  | 15.29 |         | 150.0    |             |                                    |
| 10414 | WLAN CCDF, 64-QAM, 40 MHz   | X | 3.64    | 66.72                  | 15.38 | 0.00    | 150.0    | $\pm 2.4\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 3.75    | 65.83                  | 15.15 |         | 150.0    |             |                                    |
|       |                             | Z | 3.62    | 66.25                  | 15.27 |         | 150.0    |             |                                    |

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 5 to 7).

<sup>B</sup> Linearization parameter uncertainty for maximum specified field strength.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## Parameters of Probe: EX3DV4 - SN:7808

### Sensor Model Parameters

|   | C1<br>fF | C2<br>fF | $\alpha$<br>$V^{-1}$ | T1<br>$ms V^{-2}$ | T2<br>$ms V^{-1}$ | T3<br>ms | T4<br>$V^{-2}$ | T5<br>$V^{-1}$ | T6   |
|---|----------|----------|----------------------|-------------------|-------------------|----------|----------------|----------------|------|
| x | 6.8      | 48.19    | 32.24                | 2.78              | 0.00              | 4.90     | 0.51           | 0.00           | 1.00 |
| y | 9.2      | 66.87    | 33.66                | 1.24              | 0.00              | 4.90     | 0.48           | 0.00           | 1.00 |
| z | 7.4      | 53.66    | 33.78                | 1.85              | 0.00              | 4.90     | 0.00           | 0.01           | 1.00 |

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle                               | 4.4°       |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 9 mm       |
| Tip Diameter                                  | 2.5 mm     |
| Probe Tip to Sensor X Calibration Point       | 1 mm       |
| Probe Tip to Sensor Y Calibration Point       | 1 mm       |
| Probe Tip to Sensor Z Calibration Point       | 1 mm       |
| Recommended Measurement Distance from Surface | 1.4 mm     |

**Note:** Measurement distance from surface can be increased to 3–4 mm for an *Area Scan* job.

## Parameters of Probe: EX3DV4 - SN:7808

### Calibration Parameter Determined in HSL

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 6                    | 55.0                               | 0.75                            | 18.51   | 18.83   | 17.73   | 0.00               | 1.25                    | ±13.3%                   |
| 13                   | 55.0                               | 0.75                            | 18.27   | 18.58   | 17.50   | 0.00               | 1.25                    | ±13.3%                   |
| 30                   | 55.0                               | 0.75                            | 16.78   | 17.06   | 16.06   | 0.00               | 1.25                    | ±13.3%                   |
| 64                   | 54.2                               | 0.75                            | 14.41   | 14.65   | 13.79   | 0.00               | 1.25                    | ±13.3%                   |
| 450                  | 43.5                               | 0.87                            | 10.91   | 10.91   | 10.91   | 0.16               | 1.30                    | ±13.3%                   |
| 750                  | 41.9                               | 0.89                            | 9.80    | 9.74    | 9.36    | 0.37               | 1.27                    | ±11.0%                   |
| 900                  | 41.5                               | 0.97                            | 9.43    | 9.37    | 9.01    | 0.37               | 1.27                    | ±11.0%                   |
| 1450                 | 40.5                               | 1.20                            | 8.36    | 8.31    | 7.99    | 0.36               | 1.27                    | ±11.0%                   |
| 1640                 | 40.2                               | 1.31                            | 8.21    | 8.16    | 7.85    | 0.36               | 1.27                    | ±11.0%                   |
| 1750                 | 40.1                               | 1.37                            | 8.07    | 8.02    | 7.71    | 0.36               | 1.27                    | ±11.0%                   |
| 1900                 | 40.0                               | 1.40                            | 7.88    | 7.84    | 7.53    | 0.36               | 1.27                    | ±11.0%                   |
| 2100                 | 39.8                               | 1.49                            | 7.69    | 7.64    | 7.34    | 0.36               | 1.27                    | ±11.0%                   |
| 2300                 | 39.5                               | 1.67                            | 7.48    | 7.44    | 7.15    | 0.36               | 1.27                    | ±11.0%                   |
| 2450                 | 39.2                               | 1.80                            | 7.32    | 7.27    | 6.99    | 0.36               | 1.27                    | ±11.0%                   |
| 2600                 | 39.0                               | 1.96                            | 7.16    | 7.12    | 6.84    | 0.36               | 1.27                    | ±11.0%                   |
| 3300                 | 38.2                               | 2.71                            | 6.61    | 6.57    | 6.32    | 0.35               | 1.27                    | ±13.1%                   |
| 3500                 | 37.9                               | 2.91                            | 6.52    | 6.48    | 6.23    | 0.35               | 1.27                    | ±13.1%                   |
| 3700                 | 37.7                               | 3.12                            | 6.45    | 6.41    | 6.16    | 0.35               | 1.27                    | ±13.1%                   |
| 3900                 | 37.5                               | 3.32                            | 6.40    | 6.36    | 6.11    | 0.35               | 1.27                    | ±13.1%                   |
| 4100                 | 37.2                               | 3.53                            | 6.30    | 6.27    | 6.02    | 0.35               | 1.27                    | ±13.1%                   |
| 4200                 | 37.1                               | 3.63                            | 6.25    | 6.21    | 5.97    | 0.35               | 1.27                    | ±13.1%                   |
| 4400                 | 36.9                               | 3.84                            | 6.16    | 6.12    | 5.88    | 0.35               | 1.27                    | ±13.1%                   |
| 4600                 | 36.7                               | 4.04                            | 6.06    | 6.02    | 5.79    | 0.34               | 1.27                    | ±13.1%                   |
| 4800                 | 36.4                               | 4.25                            | 6.00    | 5.96    | 5.73    | 0.34               | 1.27                    | ±13.1%                   |
| 4950                 | 36.3                               | 4.40                            | 5.88    | 5.85    | 5.62    | 0.33               | 1.27                    | ±13.1%                   |
| 5250                 | 35.9                               | 4.71                            | 5.68    | 5.65    | 5.43    | 0.30               | 1.27                    | ±13.1%                   |
| 5600                 | 35.5                               | 5.07                            | 5.26    | 5.23    | 5.02    | 0.28               | 1.27                    | ±13.1%                   |
| 5750                 | 35.4                               | 5.22                            | 5.36    | 5.33    | 5.12    | 0.26               | 1.27                    | ±13.1%                   |

<sup>C</sup> Frequency validity above 300 MHz of ±100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ±50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ±10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4–9 MHz, and ConvF assessed at 13 MHz is 9–19 MHz. Above 5 GHz frequency validity can be extended to ±110 MHz.

<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than ±5% from the target values (typically better than ±3%) and are valid for TSL with deviations of up to ±10% if SAR correction is applied.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz and below ±2% for frequencies between 3–6 GHz at any distance larger than half the probe tip diameter from the boundary.

<sup>H</sup> The stated uncertainty is the total calibration uncertainty (k = 2) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.

## Parameters of Probe: EX3DV4 - SN:7808

### Calibration Parameter Determined in HSL

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 5850                 | 35.2                               | 5.32                            | 5.43    | 5.40    | 5.19    | 0.25               | 1.27                    | ±13.1%                   |

<sup>C</sup> Frequency validity above 300 MHz of  $\pm 100$  MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to  $\pm 50$  MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is  $\pm 10$ , 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4–9 MHz, and ConvF assessed at 13 MHz is 9–19 MHz. Above 5 GHz frequency validity can be extended to  $\pm 110$  MHz.

<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than  $\pm 5\%$  from the target values (typically better than  $\pm 3\%$ ) and are valid for TSL with deviations of up to  $\pm 10\%$  if SAR correction is applied.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than  $\pm 1\%$  for frequencies below 3 GHz and below  $\pm 2\%$  for frequencies between 3–6 GHz at any distance larger than half the probe tip diameter from the boundary.

<sup>H</sup> The stated uncertainty is the total calibration uncertainty ( $k = 2$ ) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.

## Parameters of Probe: EX3DV4 - SN:7808

### Calibration Parameter Determined in HSL

| <b>f (MHz)<sup>C</sup></b> | <b>Relative Permittivity<sup>F</sup></b> | <b>Conductivity<sup>F</sup> (S/m)</b> | <b>ConvF X</b> | <b>ConvF Y</b> | <b>ConvF Z</b> | <b>Alpha<sup>G</sup></b> | <b>Depth<sup>G</sup> (mm)</b> | <b>Unc<sup>H</sup> (k = 2)</b> |
|----------------------------|------------------------------------------|---------------------------------------|----------------|----------------|----------------|--------------------------|-------------------------------|--------------------------------|
| 6500                       | 34.5                                     | 6.07                                  | 5.37           | 5.34           | 5.13           | 0.20                     | 1.27                          | ±18.6%                         |

<sup>C</sup> Frequency validity at 6.5 GHz is -600/+700 MHz, and ±700 MHz at or above 7 GHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than ±10% from the target values (typically better than ±6%) and are valid for TSL with deviations of up to ±10%.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz; below ±2% for frequencies between 3–6 GHz; and below ±4% for frequencies between 6–10 GHz at any distance larger than half the probe tip diameter from the boundary.

<sup>H</sup> The stated uncertainty is the total calibration uncertainty (k = 2) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.



## Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide:R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

Receiving Pattern ( $\phi$ ),  $\vartheta = 0^\circ$




Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

**Dynamic Range f(SAR<sub>HSL</sub>)****(TEM cell, f<sub>eval</sub> = 1900MHz)**


Uncertainty of Linearity Assessment: ±0.6% (k=2)

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ), f = 900 MHz





Accredited by the Swiss Accreditation Service (SAS)  
**The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates**

**Accreditation No.: SCS 0108**

**Client**

**UL**  
**Fremont, USA**

**Certificate No.**

**EX-3902\_Mar25**

## CALIBRATION CERTIFICATE

**Object** EX3DV4 - SN:3902

**Calibration procedure(s)** QA CAL-01.v10, QA CAL-12.v10, QA CAL-14.v7, QA CAL-23.v6,  
QA CAL-25.v8  
Calibration procedure for dosimetric E-field probes

**Calibration date** March 10, 2025

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^\circ\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards                    | ID         | Calibration Date (Certificate No.)    | Sched. Cal. |
|--------------------------------------|------------|---------------------------------------|-------------|
| Power Sensor R&S NRP-33T             | SN: 100967 | 28-Mar-24 (No. 217-04038)             | Mar-25      |
| Short [S6019i] + Attenuator [S6020i] | SN: L1119  | 26-Mar-24 (No. 217-04048)             | Mar-25      |
| OCP DAK-12                           | SN: 1016   | 24-Sep-24 (No. OCP-DAK12-1016_Sep24)  | Sep-25      |
| OCP DAK-3.5                          | SN: 1249   | 23-Sep-24 (No. OCP-DAK3.5-1249_Sep24) | Sep-25      |
| Reference Probe EX3DV4               | SN: 7349   | 10-Jan-25 (No. EX3-7349_Jan25)        | Jan-26      |
| DAE4                                 | SN: 1301   | 07-Nov-24 (No. DAE4-1301_Nov24)       | Nov-25      |

| Secondary Standards       | ID        | Check Date (in house)                           | Sched. Check |
|---------------------------|-----------|-------------------------------------------------|--------------|
| ACAP 2020 Calibration Box | SN: L1404 | 30-Sep-24 (No. Report_ACAP2020E-Cave_20240930s) | Sep-25       |

|               | Name           | Function              | Signature |
|---------------|----------------|-----------------------|-----------|
| Calibrated by | Joanna Lleshaj | Laboratory Technician |           |
| Approved by   | Sven Kühn      | Technical Manager     |           |

Issued: March 10, 2025

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

## Calibration Laboratory of

Schmid & Partner  
Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

## Glossary

|                          |                                                                                                                                                      |
|--------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                             |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                            |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                           |
| DCP                      | diode compression point                                                                                                                              |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                         |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                        |
| Polarization $\varphi$   | $\varphi$ rotation around probe axis                                                                                                                 |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                               |

## Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices – Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

## Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub> = NORM<sub>x,y,z</sub> \* frequency\_response** (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

## Parameters of Probe: EX3DV4 - SN:3902

### Basic Calibration Parameters

|                                                           | Sensor X | Sensor Y | Sensor Z | Unc ( $k = 2$ ) |
|-----------------------------------------------------------|----------|----------|----------|-----------------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup> | 0.43     | 0.45     | 0.47     | $\pm 10.1\%$    |
| DCP (mV) <sup>B</sup>                                     | 104.5    | 100.6    | 101.5    | $\pm 4.7\%$     |

### Calibration Results for Modulation Response

| UID   | Communication System Name   |   | A<br>dB | B<br>$\text{dB}\sqrt{\mu\text{V}}$ | C     | D<br>dB | VR<br>mV | Max<br>dev. | Max<br>Unc <sup>E</sup><br>$k = 2$ |
|-------|-----------------------------|---|---------|------------------------------------|-------|---------|----------|-------------|------------------------------------|
| 0     | CW                          | X | 0.00    | 0.00                               | 1.00  | 0.00    | 144.7    | $\pm 0.8\%$ | $\pm 4.7\%$                        |
|       |                             | Y | 0.00    | 0.00                               | 1.00  |         | 121.0    |             |                                    |
|       |                             | Z | 0.00    | 0.00                               | 1.00  |         | 146.1    |             |                                    |
| 10352 | Pulse Waveform (200Hz, 10%) | X | 20.00   | 93.04                              | 22.24 | 10.00   | 60.0     | $\pm 3.1\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 64.00   | 106.00                             | 25.00 |         | 60.0     |             |                                    |
|       |                             | Z | 20.00   | 94.17                              | 22.82 |         | 60.0     |             |                                    |
| 10353 | Pulse Waveform (200Hz, 20%) | X | 20.00   | 95.16                              | 22.31 | 6.99    | 80.0     | $\pm 1.3\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 20.00   | 93.47                              | 20.91 |         | 80.0     |             |                                    |
|       |                             | Z | 20.00   | 96.80                              | 23.25 |         | 80.0     |             |                                    |
| 10354 | Pulse Waveform (200Hz, 40%) | X | 20.00   | 101.30                             | 24.02 | 3.98    | 95.0     | $\pm 1.4\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 20.00   | 96.93                              | 21.11 |         | 95.0     |             |                                    |
|       |                             | Z | 20.00   | 98.53                              | 22.78 |         | 95.0     |             |                                    |
| 10355 | Pulse Waveform (200Hz, 60%) | X | 20.00   | 110.48                             | 27.02 | 2.22    | 120.0    | $\pm 1.4\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 20.00   | 101.53                             | 21.96 |         | 120.0    |             |                                    |
|       |                             | Z | 20.00   | 105.10                             | 24.65 |         | 120.0    |             |                                    |
| 10387 | QPSK Waveform, 1 MHz        | X | 1.72    | 67.09                              | 15.49 | 1.00    | 150.0    | $\pm 1.7\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 1.62    | 64.98                              | 14.34 |         | 150.0    |             |                                    |
|       |                             | Z | 1.81    | 66.73                              | 15.56 |         | 150.0    |             |                                    |
| 10388 | QPSK Waveform, 10 MHz       | X | 2.27    | 68.60                              | 16.12 | 0.00    | 150.0    | $\pm 1.0\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 2.13    | 66.82                              | 15.02 |         | 150.0    |             |                                    |
|       |                             | Z | 2.42    | 69.01                              | 16.29 |         | 150.0    |             |                                    |
| 10396 | 64-QAM Waveform, 100 kHz    | X | 2.79    | 70.26                              | 18.68 | 3.01    | 150.0    | $\pm 0.7\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 2.84    | 69.55                              | 18.23 |         | 150.0    |             |                                    |
|       |                             | Z | 2.93    | 69.98                              | 18.62 |         | 150.0    |             |                                    |
| 10399 | 64-QAM Waveform, 40 MHz     | X | 3.55    | 67.50                              | 15.98 | 0.00    | 150.0    | $\pm 0.8\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 3.48    | 66.73                              | 15.48 |         | 150.0    |             |                                    |
|       |                             | Z | 3.52    | 67.05                              | 15.80 |         | 150.0    |             |                                    |
| 10414 | WLAN CCDF, 64-QAM, 40 MHz   | X | 4.70    | 65.35                              | 15.36 | 0.00    | 150.0    | $\pm 1.8\%$ | $\pm 9.6\%$                        |
|       |                             | Y | 4.69    | 64.83                              | 15.04 |         | 150.0    |             |                                    |
|       |                             | Z | 4.88    | 65.50                              | 15.49 |         | 150.0    |             |                                    |

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 5 and 6).

<sup>B</sup> Linearization parameter uncertainty for maximum specified field strength.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## Parameters of Probe: EX3DV4 - SN:3902

### Sensor Model Parameters

|   | C1<br>fF | C2<br>fF | $\alpha$<br>$V^{-1}$ | T1<br>$msV^{-2}$ | T2<br>$msV^{-1}$ | T3<br>ms | T4<br>$V^{-2}$ | T5<br>$V^{-1}$ | T6   |
|---|----------|----------|----------------------|------------------|------------------|----------|----------------|----------------|------|
| x | 41.6     | 302.02   | 33.91                | 18.63            | 0.13             | 5.10     | 1.21           | 0.17           | 1.01 |
| y | 47.2     | 350.28   | 35.03                | 11.50            | 0.39             | 5.05     | 1.15           | 0.25           | 1.01 |
| z | 50.3     | 372.68   | 35.14                | 21.75            | 0.18             | 5.10     | 0.55           | 0.37           | 1.01 |

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle                               | 4.0°       |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 9 mm       |
| Tip Diameter                                  | 2.5 mm     |
| Probe Tip to Sensor X Calibration Point       | 1 mm       |
| Probe Tip to Sensor Y Calibration Point       | 1 mm       |
| Probe Tip to Sensor Z Calibration Point       | 1 mm       |
| Recommended Measurement Distance from Surface | 1.4 mm     |

**Note:** Measurement distance from surface can be increased to 3–4 mm for an *Area Scan* job.

## Parameters of Probe: EX3DV4 - SN:3902

### Calibration Parameter Determined in HSL

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 2450                 | 39.2                               | 1.80                            | 8.17    | 7.71    | 7.26    | 0.31               | 1.27                    | ±11.0%                   |
| 5250                 | 35.9                               | 4.71                            | 5.93    | 5.59    | 5.27    | 0.28               | 1.27                    | ±13.1%                   |
| 5600                 | 35.5                               | 5.07                            | 5.49    | 5.18    | 4.88    | 0.26               | 1.27                    | ±13.1%                   |
| 5750                 | 35.4                               | 5.22                            | 5.56    | 5.25    | 4.94    | 0.25               | 1.27                    | ±13.1%                   |
| 5850                 | 35.2                               | 5.32                            | 5.60    | 5.28    | 4.98    | 0.24               | 1.27                    | ±13.1%                   |

<sup>C</sup> Frequency validity above 300 MHz of  $\pm 100$  MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to  $\pm 50$  MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is  $\pm 10$ , 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4–9 MHz, and ConvF assessed at 13 MHz is 9–19 MHz. Above 5 GHz frequency validity can be extended to  $\pm 110$  MHz.

<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than  $\pm 5\%$  from the target values (typically better than  $\pm 3\%$ ) and are valid for TSL with deviations of up to  $\pm 10\%$  if SAR correction is applied.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than  $\pm 1\%$  for frequencies below 3 GHz and below  $\pm 2\%$  for frequencies between 3–6 GHz at any distance larger than half the probe tip diameter from the boundary.

<sup>H</sup> The stated uncertainty is the total calibration uncertainty ( $k = 2$ ) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.

Parameters of Probe: EX3DV4 - SN:3902

Calibration Parameter Determined in HSL

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 6500                 | 34.5                               | 6.07                            | 5.74    | 5.42    | 5.10    | 0.20               | 1.27                    | ±18.6%                   |

<sup>C</sup> Frequency validity at 6.5 GHz is -600/+700 MHz, and ±700 MHz at or above 7 GHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than ±10% from the target values (typically better than ±6%) and are valid for TSL with deviations of up to ±10%.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz; below ±2% for frequencies between 3–6 GHz; and below ±4% for frequencies between 6–10 GHz at any distance larger than half the probe tip diameter from the boundary.

<sup>H</sup> The stated uncertainty is the total calibration uncertainty (k = 2) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.



## Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide:R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

Receiving Pattern ( $\phi$ ),  $\vartheta = 0^\circ$




Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

Dynamic Range f(SAR<sub>HSL</sub>)  
(TEM cell, f<sub>eval</sub> = 1900MHz)




Uncertainty of Linearity Assessment:  $\pm 0.6\%$  (k=2)

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid





Accredited by the Swiss Accreditation Service (SAS)  
**The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates**

Accreditation No.: **SCS 0108**

Client

**UL**  
**Fremont, USA**

Certificate No.

**EX-7335\_Jan25**

## CALIBRATION CERTIFICATE

Object

**EX3DV4 - SN:7335**

Calibration procedure(s)

**QA CAL-01.v10, QA CAL-12.v10, QA CAL-14.v7, QA CAL-23.v6,  
QA CAL-25.v8  
Calibration procedure for dosimetric E-field probes**

Calibration date

**January 13, 2025**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature ( $22 \pm 3$ ) °C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards                    | ID         | Calibration Date (Certificate No.)     | Sched. Cal. |
|--------------------------------------|------------|----------------------------------------|-------------|
| Power Sensor R&S NRP-33T             | SN: 100967 | 28-Mar-24 (No. 217-04038)              | Mar-25      |
| Short [S6019I] + Attenuator [S6020I] | SN: L1119  | 26-Mar-24 (No. 217-04048)              | Mar-25      |
| OCP DAK-12                           | SN: 1016   | 24-Sept-24 (No. OCP-DAK12-1016_Sep24)  | Sep-25      |
| OCP DAK-3.5                          | SN: 1249   | 23-Sept-24 (No. OCP-DAK3.5-1249_Sep24) | Sep-25      |
| Reference Probe EX3DV4               | SN: 7349   | 10-Jan-25 (No. EX3-7349_Jan25)         | Jan-26      |
| DAE4                                 | SN: 1301   | 07-Nov-24 (No. DAE4-1301_Nov24)        | Nov-25      |

| Secondary Standards       | ID        | Check Date (in house)                            | Sched. Check |
|---------------------------|-----------|--------------------------------------------------|--------------|
| ACAP 2020 Calibration Box | SN: L1404 | 30-Sept-24 (No. Report_ACAP2020E-Cave_20240930s) | Sep-25       |

|               | Name               | Function              | Signature |
|---------------|--------------------|-----------------------|-----------|
| Calibrated by | Aidonia Georgiadou | Laboratory Technician |           |
| Approved by   | Sven Kühn          | Technical Manager     |           |

Issued: January 13, 2025

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

## Calibration Laboratory of

Schmid & Partner  
Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

**The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates**

Accreditation No.: **SCS 0108**

## Glossary

|                          |                                                                                                                                                      |
|--------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                             |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                            |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                           |
| DCP                      | diode compression point                                                                                                                              |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                         |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                        |
| Polarization $\varphi$   | $\varphi$ rotation around probe axis                                                                                                                 |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                               |

## Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices – Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

## Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

## Parameters of Probe: EX3DV4 - SN:7335

### Basic Calibration Parameters

|                                                           | Sensor X | Sensor Y | Sensor Z | Unc (k = 2) |
|-----------------------------------------------------------|----------|----------|----------|-------------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup> | 0.40     | 0.42     | 0.54     | ±10.1%      |
| DCP (mV) <sup>B</sup>                                     | 104.6    | 102.2    | 99.1     | ±4.7%       |

### Calibration Results for Modulation Response

| UID   | Communication System Name   |   | A<br>dB | B<br>dB $\sqrt{\mu\text{V}}$ | C     | D<br>dB | VR<br>mV | Max<br>dev. | Max<br>Unc <sup>E</sup><br>k = 2 |
|-------|-----------------------------|---|---------|------------------------------|-------|---------|----------|-------------|----------------------------------|
| 0     | CW                          | X | 0.00    | 0.00                         | 1.00  | 0.00    | 134.8    | ±0.9%       | ±4.7%                            |
|       |                             | Y | 0.00    | 0.00                         | 1.00  |         | 145.2    |             |                                  |
|       |                             | Z | 0.00    | 0.00                         | 1.00  |         | 132.0    |             |                                  |
| 10352 | Pulse Waveform (200Hz, 10%) | X | 6.79    | 75.38                        | 14.23 | 10.00   | 60.0     | ±2.8%       | ±9.6%                            |
|       |                             | Y | 8.73    | 79.99                        | 16.99 |         | 60.0     |             |                                  |
|       |                             | Z | 20.00   | 90.68                        | 20.49 |         | 60.0     |             |                                  |
| 10353 | Pulse Waveform (200Hz, 20%) | X | 20.00   | 86.85                        | 16.80 | 6.99    | 80.0     | ±1.5%       | ±9.6%                            |
|       |                             | Y | 20.00   | 90.32                        | 18.88 |         | 80.0     |             |                                  |
|       |                             | Z | 20.00   | 92.12                        | 20.24 |         | 80.0     |             |                                  |
| 10354 | Pulse Waveform (200Hz, 40%) | X | 20.00   | 92.35                        | 18.28 | 3.98    | 95.0     | ±1.2%       | ±9.6%                            |
|       |                             | Y | 20.00   | 93.76                        | 19.02 |         | 95.0     |             |                                  |
|       |                             | Z | 20.00   | 96.53                        | 21.20 |         | 95.0     |             |                                  |
| 10355 | Pulse Waveform (200Hz, 60%) | X | 20.00   | 108.49                       | 24.63 | 2.22    | 120.0    | ±1.3%       | ±9.6%                            |
|       |                             | Y | 20.00   | 100.26                       | 20.81 |         | 120.0    |             |                                  |
|       |                             | Z | 20.00   | 103.35                       | 23.23 |         | 120.0    |             |                                  |
| 10387 | QPSK Waveform, 1 MHz        | X | 2.10    | 74.38                        | 18.38 | 1.00    | 150.0    | ±2.3%       | ±9.6%                            |
|       |                             | Y | 1.63    | 65.65                        | 14.75 |         | 150.0    |             |                                  |
|       |                             | Z | 1.79    | 66.07                        | 15.28 |         | 150.0    |             |                                  |
| 10388 | QPSK Waveform, 10 MHz       | X | 2.20    | 70.02                        | 17.15 | 0.00    | 150.0    | ±1.1%       | ±9.6%                            |
|       |                             | Y | 2.14    | 67.28                        | 15.41 |         | 150.0    |             |                                  |
|       |                             | Z | 2.37    | 68.42                        | 15.98 |         | 150.0    |             |                                  |
| 10396 | 64-QAM Waveform, 100 kHz    | X | 1.92    | 65.71                        | 16.67 | 3.01    | 150.0    | ±1.0%       | ±9.6%                            |
|       |                             | Y | 2.79    | 70.22                        | 18.66 |         | 150.0    |             |                                  |
|       |                             | Z | 2.96    | 70.30                        | 18.75 |         | 150.0    |             |                                  |
| 10399 | 64-QAM Waveform, 40 MHz     | X | 3.43    | 67.83                        | 16.32 | 0.00    | 150.0    | ±0.8%       | ±9.6%                            |
|       |                             | Y | 3.48    | 66.90                        | 15.65 |         | 150.0    |             |                                  |
|       |                             | Z | 3.64    | 67.41                        | 15.96 |         | 150.0    |             |                                  |
| 10414 | WLAN CCDF, 64-QAM, 40 MHz   | X | 4.58    | 66.21                        | 15.89 | 0.00    | 150.0    | ±1.7%       | ±9.6%                            |
|       |                             | Y | 4.85    | 65.62                        | 15.48 |         | 150.0    |             |                                  |
|       |                             | Z | 4.85    | 65.18                        | 15.33 |         | 150.0    |             |                                  |

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Pages 5 and 6).

<sup>B</sup> Linearization parameter uncertainty for maximum specified field strength.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## Parameters of Probe: EX3DV4 - SN:7335

### Sensor Model Parameters

|   | C1<br>fF | C2<br>fF | $\alpha$<br>$V^{-1}$ | T1<br>$msV^{-2}$ | T2<br>$msV^{-1}$ | T3<br>ms | T4<br>$V^{-2}$ | T5<br>$V^{-1}$ | T6   |
|---|----------|----------|----------------------|------------------|------------------|----------|----------------|----------------|------|
| x | 24.4     | 175.23   | 33.34                | 11.37            | 0.00             | 5.00     | 0.38           | 0.10           | 1.00 |
| y | 43.7     | 321.67   | 34.60                | 7.12             | 0.45             | 4.99     | 1.70           | 0.09           | 1.01 |
| z | 53.4     | 399.50   | 35.71                | 17.88            | 0.00             | 5.07     | 1.03           | 0.28           | 1.01 |

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle                               | -0.8°      |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 9 mm       |
| Tip Diameter                                  | 2.5 mm     |
| Probe Tip to Sensor X Calibration Point       | 1 mm       |
| Probe Tip to Sensor Y Calibration Point       | 1 mm       |
| Probe Tip to Sensor Z Calibration Point       | 1 mm       |
| Recommended Measurement Distance from Surface | 1.4 mm     |

**Note:** Measurement distance from surface can be increased to 3–4 mm for an *Area Scan* job.

## Parameters of Probe: EX3DV4 - SN:7335

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 750                  | 41.9                               | 0.89                            | 9.91    | 10.10   | 9.53    | 0.37               | 1.27                    | ±11.0%                   |
| 900                  | 41.5                               | 0.97                            | 9.25    | 9.44    | 8.89    | 0.37               | 1.27                    | ±11.0%                   |
| 1450                 | 40.5                               | 1.20                            | 8.12    | 8.28    | 7.81    | 0.37               | 1.27                    | ±11.0%                   |
| 1640                 | 40.2                               | 1.31                            | 8.16    | 8.32    | 7.84    | 0.38               | 1.27                    | ±11.0%                   |
| 1750                 | 40.1                               | 1.37                            | 8.15    | 8.31    | 7.84    | 0.38               | 1.27                    | ±11.0%                   |
| 1900                 | 40.0                               | 1.40                            | 8.05    | 8.21    | 7.74    | 0.38               | 1.27                    | ±11.0%                   |
| 2100                 | 39.8                               | 1.49                            | 7.67    | 7.83    | 7.38    | 0.38               | 1.27                    | ±11.0%                   |
| 2300                 | 39.5                               | 1.67                            | 7.60    | 7.75    | 7.31    | 0.38               | 1.27                    | ±11.0%                   |
| 2450                 | 39.2                               | 1.80                            | 7.33    | 7.48    | 7.05    | 0.38               | 1.27                    | ±11.0%                   |
| 2600                 | 39.0                               | 1.96                            | 7.41    | 7.56    | 7.13    | 0.38               | 1.27                    | ±11.0%                   |
| 3300                 | 38.2                               | 2.71                            | 6.90    | 7.04    | 6.64    | 0.39               | 1.27                    | ±13.1%                   |
| 3500                 | 37.9                               | 2.91                            | 6.78    | 6.92    | 6.52    | 0.39               | 1.27                    | ±13.1%                   |
| 3700                 | 37.7                               | 3.12                            | 6.86    | 7.00    | 6.60    | 0.39               | 1.27                    | ±13.1%                   |
| 3900                 | 37.5                               | 3.32                            | 6.64    | 6.78    | 6.39    | 0.39               | 1.27                    | ±13.1%                   |
| 4100                 | 37.2                               | 3.53                            | 6.52    | 6.65    | 6.27    | 0.39               | 1.27                    | ±13.1%                   |
| 4200                 | 37.1                               | 3.63                            | 6.72    | 6.86    | 6.46    | 0.39               | 1.27                    | ±13.1%                   |
| 4400                 | 36.9                               | 3.84                            | 6.35    | 6.47    | 6.10    | 0.39               | 1.27                    | ±13.1%                   |
| 4600                 | 36.7                               | 4.04                            | 6.43    | 6.56    | 6.19    | 0.39               | 1.27                    | ±13.1%                   |
| 4800                 | 36.4                               | 4.25                            | 6.31    | 6.44    | 6.07    | 0.39               | 1.27                    | ±13.1%                   |
| 4950                 | 36.3                               | 4.40                            | 6.08    | 6.20    | 5.84    | 0.38               | 1.27                    | ±13.1%                   |
| 5250                 | 35.9                               | 4.71                            | 5.92    | 6.04    | 5.70    | 0.34               | 1.27                    | ±13.1%                   |
| 5600                 | 35.5                               | 5.07                            | 5.37    | 5.48    | 5.17    | 0.30               | 1.27                    | ±13.1%                   |
| 5750                 | 35.4                               | 5.22                            | 5.42    | 5.53    | 5.21    | 0.29               | 1.27                    | ±13.1%                   |
| 5850                 | 35.2                               | 5.32                            | 5.40    | 5.50    | 5.19    | 0.27               | 1.27                    | ±13.1%                   |

<sup>C</sup> Frequency validity above 300 MHz of ±100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ±50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ±10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4–9 MHz, and ConvF assessed at 13 MHz is 9–19 MHz. Above 5 GHz frequency validity can be extended to ±110 MHz.

<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than ±5% from the target values (typically better than ±3%) and are valid for TSL with deviations of up to ±10% if SAR correction is applied.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz and below ±2% for frequencies between 3–6 GHz at any distance larger than half the probe tip diameter from the boundary.

<sup>H</sup> The stated uncertainty is the total calibration uncertainty (k = 2) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.

## Parameters of Probe: EX3DV4 - SN:7335

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity <sup>F</sup> (S/m) | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc <sup>H</sup> (k = 2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|--------------------------|
| 6500                 | 34.5                               | 6.07                            | 5.71    | 5.82    | 5.49    | 0.20               | 1.27                    | ±18.6%                   |

<sup>C</sup> Frequency validity at 6.5 GHz is -600/+700 MHz, and ±700 MHz at or above 7 GHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


<sup>F</sup> The probes are calibrated using tissue simulating liquids (TSL) that deviate for  $\epsilon$  and  $\sigma$  by less than ±10% from the target values (typically better than ±6%) and are valid for TSL with deviations of up to ±10%.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ±1% for frequencies below 3 GHz; below ±2% for frequencies between 3–6 GHz; and below ±4% for frequencies between 6–10 GHz at any distance larger than half the probe tip diameter from the boundary.



<sup>H</sup> The stated uncertainty is the total calibration uncertainty (k = 2) of Norm-ConvF. This is equivalent to the uncertainty component with the symbol CF in Table 9 of IEC/IEEE 62209-1528:2020.

## Frequency Response of E-Field



(TEM-Cell:ifi110 EXX, Waveguide:R22)




Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

**Receiving Pattern ( $\phi$ ),  $\vartheta = 0^\circ$** 

Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )


**Dynamic Range  $f(\text{SAR}_{\text{head}})$** (TEM cell,  $f_{\text{eval}} = 1900\text{MHz}$ )Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ),  $f = 900\text{ MHz}$





Accredited by the Swiss Accreditation Service (SAS)  
**The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates**

Accreditation No.: **SCS 0108**

|        |                |                 |                        |
|--------|----------------|-----------------|------------------------|
| Client | <b>Apple</b>   | Certificate No. | <b>EUmm-9532_Feb25</b> |
|        | Cupertino, USA |                 |                        |

## CALIBRATION CERTIFICATE

|                          |                                                                                                                                        |
|--------------------------|----------------------------------------------------------------------------------------------------------------------------------------|
| Object                   | EUmmWV4 - SN:9532                                                                                                                      |
| Calibration procedure(s) | QA CAL-02.v9, QA CAL-25.v8, QA CAL-42.v3<br>Calibration procedure for E-field probes optimized for close near field evaluations in air |
| Calibration date         | February 17, 2025                                                                                                                      |

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3) °C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards       | ID         | Calibration Date (Certificate No.) | Sched. Cal. |
|-------------------------|------------|------------------------------------|-------------|
| Power sensor NRP33T     | SN: 100967 | 28-Mar-24 (No. 217-04038)          | Mar-25      |
| Power sensor NRP110T    | SN: 101244 | 04-Apr-24 (No. 0001A300740056)     | Apr-25      |
| Spectrum analyzer FSV40 | SN: 101832 | 29-Jan-25 (No. 4030A315009658)     | Jan-26      |
| Harmonic mixer FS-Z75   | SN: 101566 | 11-Apr-24 (No. 0001A300740054)     | Apr-25      |
| Harmonic mixer FS-Z110  | SN: 101633 | 05-Apr-24 (No. 0001A300740055)     | Apr-25      |
| Ref. Probe EUmmWV3      | SN: 9374   | 28-Aug-24 (No. EUmm-9374_Aug24)    | Aug-25      |
| DAE4ip                  | SN: 1662   | 05-Nov-24 (No. DAE4ip-1662_Nov24)  | Nov-25      |

| Secondary Standards  | ID         | Check Date (in house)             | Sched. Check           |
|----------------------|------------|-----------------------------------|------------------------|
| Generator APSIN26G   | SN: 2023   | 30-Nov-21 (in house check Jun-24) | In house check: Jun-25 |
| Power sensor NRP40T  | SN: 101439 | 08-Nov-21 (in house check Jun-24) | In house check: Jun-25 |
| Power sensor NRP110T | SN: 101226 | 15-Nov-21 (in house check Jun-24) | In house check: Jun-25 |

|                                                                                                                 |                |                       |                           |
|-----------------------------------------------------------------------------------------------------------------|----------------|-----------------------|---------------------------|
|                                                                                                                 | Name           | Function              | Signature                 |
| Calibrated by                                                                                                   | Joanna Lleshaj | Laboratory Technician |                           |
| Approved by                                                                                                     | Sven Kühn      | Technical Manager     |                           |
|                                                                                                                 |                |                       | Issued: February 18, 2025 |
| This calibration certificate shall not be reproduced except in full without written approval of the laboratory. |                |                       |                           |

## Calibration Laboratory of

Schmid & Partner  
Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

## Glossary

|                          |                                                                                                                                                      |
|--------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|
| NORM <sub>x,y</sub>      | sensitivity in free space                                                                                                                            |
| DCP                      | diode compression point                                                                                                                              |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                         |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                        |
| Polarization $\varphi$   | $\varphi$ rotation around probe axis                                                                                                                 |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                               |
| Sensor Angles            | sensor deviation from the probe axis, used to calculate the field orientation and polarization                                                       |
| $\vec{k}$                | is the wave propagation direction                                                                                                                    |

## Calibration is Performed According to the Following Standards:

- a) IEEE Std 1309-2005, "IEEE Standard for calibration of electromagnetic field sensors and probes, excluding antennas, from 9 kHz to 40 GHz", December 2005

## Methods Applied and Interpretation of Parameters:

- **NORM<sub>x,y</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). For frequencies  $> 6$  GHz, the far field in front of waveguide horn antennas is measured for a set of frequencies in various waveguide bands up to 110 GHz.
- **DCP<sub>x,y</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media.  
Note: As the field is measured with a diode detector sensor, it is warranted that the probe response is linear ( $E^2$ ) below the documented lowest calibrated value.
- **PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- The frequency sensor model parameters are determined prior to calibration based on a frequency sweep (sensor model involving resistors R, R<sub>p</sub>, inductance L and capacitors C, C<sub>p</sub>).
- **A<sub>x,y</sub>; B<sub>x,y</sub>; C<sub>x,y</sub>; D<sub>x,y</sub>; VR<sub>x,y</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- **Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- **Connector Angle**: The angle is assessed using the information gained by determining the **NORM<sub>x</sub>** (no uncertainty required).
- **Equivalent Sensor Angle**: The two probe sensors are mounted in the same plane at different angles. The angles are assessed using the information gained by determining the **NORM<sub>x</sub>** (no uncertainty required).
- **Spherical isotropy (3D deviation from isotropy)**: in a locally homogeneous field realized using an open waveguide / horn setup.

## Parameters of Probe: EUmmWV4 - SN:9532

### Basic Calibration Parameters

|                                              | Sensor X | Sensor Y | Unc (k = 2)  |
|----------------------------------------------|----------|----------|--------------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) | 0.01753  | 0.02053  | $\pm 10.1\%$ |
| DCP (mV) <sup>B</sup>                        | 106.0    | 105.0    | $\pm 4.7\%$  |
| Equivalent Sensor Angle                      | -59.2    | 35.8     |              |

### Calibration Results for Frequency Response (750 MHz – 110 GHz)

| Frequency<br>GHz | Target<br>E-Field<br>V/m | Deviation Sensor X<br>dB | Deviation Sensor Y<br>dB | Unc (k = 2)<br>dB |
|------------------|--------------------------|--------------------------|--------------------------|-------------------|
| 0.75             | 77.2                     | -0.34                    | -0.29                    | $\pm 0.43$        |
| 1.8              | 140.4                    | -0.02                    | -0.04                    | $\pm 0.43$        |
| 2.0              | 133.0                    | 0.13                     | 0.15                     | $\pm 0.43$        |
| 2.2              | 124.8                    | -0.07                    | -0.04                    | $\pm 0.43$        |
| 2.5              | 123.0                    | 0.09                     | 0.14                     | $\pm 0.43$        |
| 3.5              | 256.2                    | -0.07                    | -0.06                    | $\pm 0.43$        |
| 3.7              | 249.8                    | 0.09                     | 0.08                     | $\pm 0.43$        |
|                  |                          |                          |                          |                   |
| 6.6              | 63.4                     | -0.10                    | -0.27                    | $\pm 0.98$        |
| 8.0              | 58.5                     | -0.07                    | -0.16                    | $\pm 0.98$        |
| 10.0             | 57.9                     | -0.01                    | 0.02                     | $\pm 0.98$        |
| 15.0             | 45.6                     | 0.19                     | 0.21                     | $\pm 0.98$        |
|                  |                          |                          |                          |                   |
| 26.6             | 115.1                    | 0.17                     | 0.23                     | $\pm 0.98$        |
| 30.0             | 125.1                    | 0.01                     | 0.01                     | $\pm 0.98$        |
| 35.0             | 123.5                    | -0.15                    | -0.19                    | $\pm 0.98$        |
| 40.0             | 101.8                    | -0.23                    | -0.32                    | $\pm 0.98$        |
|                  |                          |                          |                          |                   |
| 50.0             | 60.8                     | 0.09                     | -0.02                    | $\pm 0.98$        |
| 55.0             | 73.7                     | -0.09                    | -0.05                    | $\pm 0.98$        |
| 60.0             | 76.4                     | 0.01                     | 0.03                     | $\pm 0.98$        |
| 65.0             | 72.0                     | 0.17                     | 0.13                     | $\pm 0.98$        |
| 70.0             | 68.5                     | 0.12                     | 0.07                     | $\pm 0.98$        |
| 75.0             | 67.9                     | -0.04                    | -0.08                    | $\pm 0.98$        |
|                  |                          |                          |                          |                   |
| 75.0             | 89.9                     | -0.07                    | -0.07                    | $\pm 0.98$        |
| 80.0             | 88.2                     | -0.14                    | -0.10                    | $\pm 0.98$        |
| 85.0             | 54.3                     | -0.03                    | -0.05                    | $\pm 0.98$        |
| 90.0             | 80.6                     | 0.02                     | 0.02                     | $\pm 0.98$        |
| 92.0             | 80.8                     | -0.01                    | 0.03                     | $\pm 0.98$        |
| 95.0             | 73.2                     | -0.05                    | -0.02                    | $\pm 0.98$        |
| 97.0             | 65.9                     | -0.03                    | -0.04                    | $\pm 0.98$        |
| 100.0            | 63.4                     | 0.03                     | 0.05                     | $\pm 0.98$        |
| 105.0            | 63.2                     | -0.10                    | -0.12                    | $\pm 0.98$        |
| 110.0            | 72.1                     | 0.07                     | 0.05                     | $\pm 0.98$        |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>B</sup> Linearization parameter uncertainty for maximum specified field strength.

## Parameters of Probe: EUmmWV4 - SN:9532

### Calibration Results for Modulation Response

| UID   | Communication System Name   |   | A<br>dB | B<br>dB $\sqrt{\mu V}$ | C     | D<br>dB | VR<br>mV | Max<br>dev. | Max<br>Unc <sup>E</sup><br>k = 2 |
|-------|-----------------------------|---|---------|------------------------|-------|---------|----------|-------------|----------------------------------|
| 0     | CW                          | X | 0.00    | 0.00                   | 1.00  | 0.00    | 123.9    | ±2.5%       | ±4.7%                            |
|       |                             | Y | 0.00    | 0.00                   | 1.00  |         | 69.0     |             |                                  |
| 10352 | Pulse Waveform (200Hz, 10%) | X | 2.34    | 60.00                  | 14.87 | 10.00   | 6.0      | ±1.0%       | ±9.6%                            |
|       |                             | Y | 2.06    | 60.00                  | 15.63 |         | 6.0      |             |                                  |
| 10353 | Pulse Waveform (200Hz, 20%) | X | 1.62    | 60.00                  | 13.75 | 6.99    | 12.0     | ±1.1%       | ±9.6%                            |
|       |                             | Y | 1.42    | 60.00                  | 14.64 |         | 12.0     |             |                                  |
| 10354 | Pulse Waveform (200Hz, 40%) | X | 0.99    | 60.27                  | 12.69 | 3.98    | 23.0     | ±1.5%       | ±9.6%                            |
|       |                             | Y | 0.89    | 60.00                  | 13.44 |         | 23.0     |             |                                  |
| 10355 | Pulse Waveform (200Hz, 60%) | X | 0.58    | 60.00                  | 12.08 | 2.22    | 27.0     | ±1.1%       | ±9.6%                            |
|       |                             | Y | 0.63    | 60.00                  | 12.37 |         | 27.0     |             |                                  |
| 10387 | QPSK Waveform, 1 MHz        | X | 1.15    | 60.00                  | 12.44 | 1.00    | 22.0     | ±1.2%       | ±9.6%                            |
|       |                             | Y | 1.27    | 60.00                  | 12.04 |         | 22.0     |             |                                  |
| 10388 | QPSK Waveform, 10 MHz       | X | 1.23    | 60.00                  | 12.30 | 0.00    | 22.0     | ±0.6%       | ±9.6%                            |
|       |                             | Y | 1.45    | 60.00                  | 11.85 |         | 22.0     |             |                                  |
| 10396 | 64-QAM Waveform, 100 kHz    | X | 3.26    | 66.13                  | 16.37 | 3.01    | 17.0     | ±0.6%       | ±9.6%                            |
|       |                             | Y | 2.13    | 60.00                  | 13.90 |         | 17.0     |             |                                  |
| 10399 | 64-QAM Waveform, 40 MHz     | X | 2.02    | 60.00                  | 12.67 | 0.00    | 19.0     | ±0.8%       | ±9.6%                            |
|       |                             | Y | 2.20    | 60.00                  | 12.43 |         | 19.0     |             |                                  |
| 10414 | WLAN CCDF, 64-QAM, 40 MHz   | X | 3.16    | 60.17                  | 13.12 | 0.00    | 12.0     | ±1.0%       | ±9.6%                            |
|       |                             | Y | 3.29    | 60.00                  | 12.87 |         | 12.0     |             |                                  |

Note: For details on UID parameters see Appendix

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## Parameters of Probe: EUmmWV4 - SN:9532

### Calibration Results for Linearity Response

| Frequency<br>GHz | Target E-Field<br>V/m | Deviation Sensor X<br>dB | Deviation Sensor Y<br>dB | Unc ( $k = 2$ )<br>dB |
|------------------|-----------------------|--------------------------|--------------------------|-----------------------|
| 0.9              | 50.0                  | -0.02                    | -0.13                    | $\pm 0.2$             |
| 0.9              | 100.0                 | -0.02                    | 0.01                     | $\pm 0.2$             |
| 0.9              | 500.0                 | -0.01                    | -0.01                    | $\pm 0.2$             |
| 0.9              | 1000.0                | 0.01                     | 0.01                     | $\pm 0.2$             |
| 0.9              | 1500.0                | -0.00                    | 0.01                     | $\pm 0.2$             |
| 0.9              | 2100.0                | -0.02                    | 0.00                     | $\pm 0.2$             |

### Sensor Frequency Model Parameters (750 MHz – 55 GHz)

|                             | Sensor X | Sensor Y |
|-----------------------------|----------|----------|
| R ( $\Omega$ )              | 61.95    | 93.28    |
| R <sub>p</sub> ( $\Omega$ ) | 83.25    | 116.19   |
| L (nH)                      | 0.05692  | 0.07457  |
| C (pF)                      | 0.2808   | 0.2696   |
| C <sub>p</sub> (pF)         | 0.1065   | 0.0780   |

### Sensor Frequency Model Parameters (55 GHz – 110 GHz)

|                             | Sensor X | Sensor Y |
|-----------------------------|----------|----------|
| R ( $\Omega$ )              | 18.10    | 24.81    |
| R <sub>p</sub> ( $\Omega$ ) | 102.72   | 127.19   |
| L (nH)                      | 0.05806  | 0.07297  |
| C (pF)                      | 0.0719   | 0.0599   |
| C <sub>p</sub> (pF)         | 0.0917   | 0.0704   |

### Sensor Model Parameters


|   | C1<br>fF | C2<br>fF | $\alpha$<br>$V^{-1}$ | T1<br>$ms V^{-2}$ | T2<br>$ms V^{-1}$ | T3<br>ms | T4<br>$V^{-2}$ | T5<br>$V^{-1}$ | T6   |
|---|----------|----------|----------------------|-------------------|-------------------|----------|----------------|----------------|------|
| x | 50.7     | 363.96   | 33.11                | 0.92              | 5.71              | 5.00     | 0.00           | 1.45           | 1.01 |
| y | 46.2     | 331.69   | 33.08                | 0.92              | 5.07              | 5.03     | 0.00           | 1.81           | 1.01 |

### Other Probe Parameters


|                                         |             |
|-----------------------------------------|-------------|
| Sensor Arrangement                      | Rectangular |
| Connector Angle                         | -72.3°      |
| Mechanical Surface Detection Mode       | enabled     |
| Optical Surface Detection Mode          | disabled    |
| Probe Overall Length                    | 320 mm      |
| Probe Body Diameter                     | 8 mm        |
| Tip Length                              | 23 mm       |
| Tip Diameter                            | 8.0 mm      |
| Probe Tip to Sensor X Calibration Point | 1.5 mm      |
| Probe Tip to Sensor Y Calibration Point | 1.5 mm      |

### Deviation from Isotropy in Air

30GHz: 3D isotropy, E-field parallel to probe axis



60GHz: 3D isotropy, E-field parallel to probe axis



Probe isotropy for  $E_{\text{tot}}$ : probe rotated  $\phi = 0^\circ$  to  $360^\circ$ , tilted from field propagation direction  $\vec{k}$   
 Parallel to the field propagation ( $\psi = 0^\circ - 90^\circ$ ) at 30 GHz: deviation within  $\pm 0.49$  dB  
 Parallel to the field propagation ( $\psi = 0^\circ - 90^\circ$ ) at 60 GHz: deviation within  $\pm 0.38$  dB