

RF System Check Test Report for CD1880:SN 1212

Hardware Setup

Probe | Calibration Date EF3DV3 - SN4041 | March 04, 2025
DAE | Calibration Date DAE4 Sn1784 | April 09, 2025
Software Version 1.2.6.2199

Communication Systems


Band Name CD1880
Channel | Frequency [MHz] 50 | 1880.0
Communication Systems' Name CW

Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	90.0	Step Y [mm]	5.0
Distance [mm]	15.0		

Results

E_{max} [V/m] **87.1**
Drift [dB] **-0.01**



RF System Check Test Report for CD1880:SN 1212

Hardware Setup

Probe | Calibration Date EF3DV3 - SN4041 | March 04, 2025
DAE | Calibration Date DAE4 Sn1784 | April 09, 2025
Software Version 1.2.6.2199

Communication Systems

Band Name CD1880
Channel | Frequency [MHz] 50 | 1880.0
Communication Systems' Name CW

Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	90.0	Step Y [mm]	5.0
Distance [mm]	15.0		

Results

E_{max} [V/m] **81.5**
Drift [dB] **-0.03**

