

RF System Check Test Report for CD1880:SN XXXX

Hardware Setup

Probe | Calibration DateEF3DV3 - SN4041 | March 04, 2025

DAE | Calibration DateDAE4ip Sn1621 | April 10, 2025

Software Version1.2.6.2199

Communication Systems

Band NameCD1880

Channel | Frequency [MHz]0 | 1730.0

Communication Systems' NameCW

Grid Settings

Extent X [mm]20.0Step X [mm]5.0

Extent Y [mm]90.0Step Y [mm]5.0

Distance [mm]15.0

Results

Emax [V/m]86.8

Drift [dB]0.06

