




Dynamic Range f(SAR_{brain}) (TEM-Cell:ifi110)



Conversion Factor Assessment




Head 900 MHz $\epsilon_r = 42 \pm 5\%$ $\sigma = 0.97 \pm 10\% \text{ mho/m}$

ConvF X	6.93 $\pm 7\%$ (k=2)	Boundary effect:
ConvF Y	6.93 $\pm 7\%$ (k=2)	Alpha 0.34
ConvF Z	6.93 $\pm 7\%$ (k=2)	Depth 2.74

Head 1800 MHz $\epsilon_r = 40 \pm 5\%$ $\sigma = 1.40 \pm 10\% \text{ mho/m}$

ConvF X	5.92 $\pm 7\%$ (k=2)	Boundary effect:
ConvF Y	5.92 $\pm 7\%$ (k=2)	Alpha 0.49
ConvF Z	5.92 $\pm 7\%$ (k=2)	Depth 2.13

Deviation from Isotropy in HSL
Error (θ, ϕ), $f = 900$ MHz



**Schmid & Partner
Engineering AG**

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

**Additional Conversion Factors
for Dosimetric E-Field Probe**

Type:

ET3DV6

Serial Number:

1577

Place of Assessment:

Zurich

Date of Assessment:


October 11, 2001

Probe Calibration Date:

April 20, 2001

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:



Dosimetric E-Field Probe ET3DV6 SN:1577

Conversion factor (\pm standard deviation)

450 MHz	ConvF	$8.0 \pm 8\%$	$\epsilon_r = 56.7 \pm 5\%$ $\sigma = 0.94 \pm 5\% \text{ mho/m}$ Muscle tissue
2450 MHz	ConvF	$4.4 \pm 10\%$	$\epsilon_r = 52.7 \pm 5\%$ $\sigma = 1.95 \pm 5\% \text{ mho/m}$ Muscle tissue
2450 MHz	ConvF	$4.9 \pm 10\%$	$\epsilon_r = 39.2 \pm 5\%$ $\sigma = 1.80 \pm 5\% \text{ mho/m}$ Head tissue

**Schmid & Partner
Engineering AG**

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

**Additional Conversion Factors
for Dosimetric E-Field Probe**

Type:

ET3DV6

Serial Number:

1577

Place of Assessment:

Zurich

Date of Assessment:


August 23, 2001

Probe Calibration Date:

April 20, 2001

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Approved by:



Dosimetric E-Field Probe ET3DV6 SN:1577

Conversion factor (\pm standard deviation)

900 MHz	ConvF	$6.6 \pm 8\%$	$\epsilon_r = 56.5 \pm 5\%$ $\sigma = 0.99 \pm 10\% \text{ mho/m}$ Muscle tissue
1800 MHz	ConvF	$5.4 \pm 8\%$	$\epsilon_r = 54.6 \pm 5\%$ $\sigma = 1.39 \pm 10\% \text{ mho/m}$ Muscle tissue