

Label specification

ascom



LAYOUT PRODUCT LABEL
TOTAL SIZE: 40*10 MM
Here shown NUCR-HU

CE

When used, the CE mark must have a height for a good visibility



yyyy-mm-dd

Insert date code at time of label printing.

S/N

Serial number, 8 characters.

UL

The symbol indicating separate collection for EEE consists of the crossed-out wheeled bin. The symbol must be printed visibly, legibly and indelibly.

yyyy-mm-dd

Insert date code at time of label printing.

R-NZ

Mark for NZL for radio products

Complies with IMDA Standards DA107674

Ascom (US) Inc.
300 Perimeter Park Dr.
Morrisville, NC 27560

Ascom (Sweden) AB
Grimboden 2
417 49 Göteborg
Sweden

SN: Mxxxxxx

UDI: (01)17350088670955(11)200819(21)Mxxxxxx

LAYOUT BOX LABEL
TOTAL SIZE: 89*51 MM, DRAWING M68480C
Here shown NUCR-HU

Change: Added HVIN number

Label pn: LBL0807

Label revision:A2

Product:NUCR-HU/XX

Description: Card Reader US

Latest updated: 2020-10-28

Signature: SEshch

Additional notes:

- All characters to be printed in approximate location shown, with approximate size and style. all dashed lines are for reference only and shall not be printed.
- Pre-production proofs must be submitted for approval prior to production run.

UDI:

Unique Device Identification (UDI) for Medical Devices.
Consists of three parts:
Fixed DI (Device Identifier) per product, prefixed by (01), what DI to use is supplied by Ascom
A variable PI (Product Identifier) consisting of production date prefixed by (11) (format YYMMDD)
and the unique device serial number prefixed by (21)
Example of a complete UDI: (01)1735008867075(11)150909(21)Mxxxxxx



For the complete UDI use the barcode in GS1 DataMatrix format > 5x5 mm
(According to the GS1 Data matrix standard the UDI is here without parentheses)

China RoHS 50 years



Symbol for manufacturer



This document contains confidential and proprietary information of Ascom and may not be copied, used or disclosed without the written permission of Ascom and must be returned upon request.