

## Section 15.231 and ANSI C63.4

This is a list of all test equipment used.

Test Equipment list for Honeywell OATS:

| Equipment             | Mfg                | Model  | Cal Date | Cal Due  |
|-----------------------|--------------------|--------|----------|----------|
| Spectrum Analyzer     | Rohde & Schwarz    | FSEA20 | 10/14/08 | 10/14/09 |
| Antenna ('Biconilog') | ETS (EMCO)Lindgren | 3149   | 03/09/09 | 03/09/10 |

PLEASE SEE PAGE 2-3 FOR TEST EQUIPMENT TRACEABILITY

If you need any additional information from Honeywell please contact:

Greg Barbato RF Engineer  
(Acting for Ken Eskildsen)  
Phone (Direct): (516) 577-5863  
Email: [greg.barbato@honeywell.com](mailto:greg.barbato@honeywell.com)



Cert I.D.: 72154

*Certificate of Calibration Conformance*

Page 1 of 5

The instrument identified below has been individually calibrated in compliance with the following standard(s):

SAE, ARP-958 - 2003, Electromagnetic Interference Measurement Antennas; Standard Calibration Method, Society of Automotive Engineers, Aerospace Recommended Practice. Fixed height, three antenna rotation, 1 meter separation. 3 meter separation performed per Annex C. Vertical calibration performed per above listed methodology.

Environment: Laboratory MTE is maintained in a temperature controlled environment with ambient conditions from 18 to 28 C, relative humidity less than 90%. The instrument under test has been calibrated on an open air test site (OATS) with environment temperature conditions ranging from 0 to 40 C which has no known influences on measurement quality.

|                                 |                                  |  |                         |
|---------------------------------|----------------------------------|--|-------------------------|
| <b>Manufacturer:</b>            | ETS-Lindgren                     | <b>Operating Range:</b>  | 80 MHz - 6 GHz          |
| <b>Model Number:</b>            | 3149.                            | <b>Instrument Type:</b>  | Biconilog (Type 5)      |
| <b>Serial Number/ ID:</b>       | 00029390                         | <b>Date Code:</b>  |                         |
| <b>Tracking Number:</b>         | S00015900                        | <b>Alternate ID:</b>   |                         |
| <b>Date Completed:</b>          | 09-Mar-09                        | <b>Customer:</b>   | HONEYWELL SECURITY (NY) |
| <b>Test Type:</b>               | 3 meter, Horizontal and Vertical |  |                         |
| <b>Calibration Uncertainty:</b> | 01m                              | 80 - 1000 MHz, +/-0.9 dB; 1000 - 2000 MHz, +/-0.8 dB; 2000 - 6000 MHz, +/-1.2 dB |                         |
| k=2, (95% Confidence Level)     | 03m                              | 80 - 1000 MHz, +/-0.9 dB; 1000 - 2000 MHz, +/-0.8 dB; 2000 - 6000 MHz, +/-1.3 dB |                         |
|                                 | 10m                              | 80 - 1000 MHz, +/-1.0 dB; 1000 - 2000 MHz, +/-1.4 dB; 2000 - 6000 MHz, +/-2.3 dB |                         |

**Test Remarks:** Replaced damaged bracket mount. Performed special calibration from 26 MHz to 6GHz. Provided data on disk.

Calibration Traceability: All Measuring and Test Equipment (MTE) identified below are traceable to the National Institute for Standards and Technology (NIST). Calibration Laboratory and Quality System controls are compliant with ISO/IEC 17025-2005.

**Standards and Equipment Used:**

**Make / Model / Name / S/N / Recall Date**

Anritsu MS4623A Network Analyzer 992201 05-Sep-09

**Condition of Instrument**

**Upon Receipt:**

Other

**On Release:**

In Tolerance to Internal Quality Standards

Calibration Completed by  
 Alan Schifferdecker, Calibration Technician

Attested and Issued on 09-Mar-09  
 Justin Farr, Calibration Supervisor

Issue Date: 10/14/2008



### General Calibration, Inc.

2 Mars Court, Boonton, New Jersey 07005  
Phone (973) 299-2950 Fax (973) 299-0595

Certificate #: 12962MR  
Purchase Order: 4965832  
Work Order #: MR219  
Customer #: 001464

GENERAL CALIBRATION, INC.  
2 MARS COURT  
MONTVILLE, NJ 07045

HONEYWELL SECURITY (1464)  
2 CORPORATE CENTER DRIVE  
MELVILLE, NY 11747

BarCode: 018675  
Manufacturer: R&S  
Description: SPECTRUM ANALYZER  
Current Location: ALARMNET  
Temp./RH: 22 C / 40 %  
Cal. Interval: 12 MONTHS  
Cal Date: 10/14/2008

Instrument I.D.: 10506  
Model Number: FSEA20  
Serial Number: DE23427  
Inspected By: MR1  
Job Title: METROLOGIST  
Calibration Result: PASS  
Cal. Due Date: 10/14/2009

Condition: Found In Tolerance and Left In Tolerance

|                     |     |                     |     |            |
|---------------------|-----|---------------------|-----|------------|
| GENERAL CALIBRATION | 434 | POWER SPLITTER      | N/A | 09/12/2009 |
| GENERAL CALIBRATION | 531 | MEASURING RECEIVER  | N/A | 09/22/2009 |
| GENERAL CALIBRATION | 636 | SYNTHESIZED SWEEPER | N/A | 09/03/2009 |
| GENERAL CALIBRATION | 666 | SENSOR MODULE       | N/A | 04/25/2009 |

The above instrument has been checked and calibrated against the above working standard(s) which are traceable to the NIST. The test limits stated in the report correspond to the published specifications of the equipment, at the points tested. Also, the collective uncertainties of measurement standards do not exceed 25% of the tolerance of the characteristics being calibrated, where possible. The metrology procedures utilized conform to and satisfy the requirements set forth in ANSI/NCSL Z540-1-1994, 10 CFR part 21, ISO 9001-2000, ISO 10012-2003, and MIL-STD 45662A.

Approved By *[Signature]*  
General Calibration, Inc. - Q. A. Manager