

Date: August 7, 2017

Subject: Ad Hoc Mode Function Declaration

FCC ID: CM6-WMIA198N

To whom it may concern,

We, Spacelabs Healthcare, Inc. declare that the device does not support any non-US channels, country code selection in all the operational mode(s), like Ad-Hoc...etc. for the following product. Also, the client software and associated drivers will not initiate any transmission on DFS frequencies without initiation by a master and the device without support DFS detection capability. This includes restriction on transmissions for beacons and support for ad-hoc peer-to-peer modes.

Sincerely,



Stephen J. Cantwell

Global R&D Product Safety Engineer

Spacelabs Healthcare

o (425) 363-5728

steve.cantwell@spacelabs.com