



Physio-Control, Inc. | Lifesaving starts here.™

June 6, 2018

ADDRESS

11811 Willows Road NE
Redmond, WA 98052

PHONE

GENERAL
425 867 4000

TOLL FREE
800 442 1142

www.physio-control.com

Federal Communications Commission
Authorization and Evaluation division
7435 Oakland Mills Road
Columbia, Maryland 21046

Subject: FCC Class II Permissive Change Application for LP20/20e
Defibrillator Monitor CodeManagement Module containing an 802.11 abgn radio;
Radio Description: Redpine RS9110-N-11-03 WLAN
FCC ID Numbers: CZO-CMMPHYSIO21, originally granted on 02/28/2013

To whom it may concern:

Physio-Control, now a part of Stryker, is herein submitting a Class II permissive change filing for the 802.11 abgn Module as listed in the Operational Description exhibit. Pursuant to FCC Part 15.247 & 15.407, Physio-Control Inc. is submitting a request to change DTS Grant (FCC Rule 15C) by removing Frequency Range 5745-5825 MHz. Additionally, we request those same 5745-5825 MHz frequencies be added to the NII Grant (FCC Rule 15E)

Sincerely,

Robert Zito

Sr. Manager, Compliance Engineering
Physio-Control, now part of Stryker
11811 Willows Rd NE, Redmond WA 98052
+1 (425) 867-4444
Bob.J.Zito@stryker.com
www.physio-control.com