

The Synergy Charger has been tested to, and has passed, the following emissions and immunity tests:

- IEC 60601-1-2 radiated and conducted emissions requirements for a Group 1 Class B device as stated in CISPR 11
- Harmonic distortion and voltage fluctuations affecting the power input source as stated in Table 2 of IEC 60601-1-2
- RF radiated immunity at a field level of 10 V/m between 80 MHz and 27 GHz as well as higher field levels from communications devices as stated in Table 9 of IEC 60601-1-2
- Immunity to power frequency magnetic fields at a field level of 30 A/m
- Immunity to ESD levels of +/- 8 kV conducted discharge and +/- 15 kV air discharge
- Immunity to Electrical fast transients on the power input at a level of +/- 2 kV at a 100 Hz repetition rate
- Immunity to Surges on the power input of +/- 1 kV line to line
- Immunity to conducted disturbances induced by RF fields on the power input as stated in Table 6 of IEC 60601-1-2
- Immunity to voltage dips and interruptions on the power input as stated in Table 6 of IEC 60601-1-2

#### WIRELESS NOTICES

FCC ID: EOA-24LIVIORCHG

IC: 6903A-24LIVIORCHG (Livio RIC R, Via RIC R, AGXsliv RIC R, Esentia RIC R, Livio AI RIC R, Via AI RIC R, AGXsliv AI RIC R, Esentia AI RIC R, Livio RIC Rechargeable, Via RIC Rechargeable, AGXsliv RIC Rechargeable, Esentia RIC Rechargeable, Livio AI RIC Rechargeable, Via AI RIC Rechargeable, AGXsliv AI RIC Rechargeable, Esentia AI RIC Rechargeable)

#### FCC NOTICE

This hearing aid and Synergy Charger comply with part 15 of the FCC rules and with IC: 6903A-24LIVIORCHG (Livio RIC R, Via RIC R, AGXsliv RIC R, Esentia RIC R, Livio AI RIC R, Via AI RIC R, AGXsliv AI RIC R, Esentia AI RIC R, Livio RIC Rechargeable, Via RIC Rechargeable, AGXsliv RIC Rechargeable, Esentia RIC Rechargeable, Livio AI RIC Rechargeable, Via AI RIC Rechargeable, AGXsliv AI RIC Rechargeable, Esentia AI RIC Rechargeable)

Note: The manufacturer is not responsible for any radio or TV interference caused by unauthorized modifications to this equipment. Such modifications could void the user's authority to operate the equipment.

Cet appareil est conforme à la partie 15 des règles de la FCC et avec les normes RSS de licence d'ISDE Canada. Le fonctionnement est soumis à deux conditions: (1) Cet appareil ne doit pas causer d'interférences nuisibles et (2) cet appareil doit accepter toute interférence reçue, y compris les interférences qui peuvent causer des fonctionnements du dispositif.

NOTE: Le manufacturier n'est pas responsable de l'interférence créée par la modification de cet équipement lors de l'écoute du téléviseur ou de la radio. De telles modifications pourraient entraîner la révocation de l'autorité de l'utilisateur à opérer cet équipement.

#### EU NOTICE

Hereby, Starkey Hearing Technologies declares that the RIC is in compliance with the essential requirements and other relevant provisions of Directive 2014/53/EU. A copy of the Declaration of Conformity can be obtained from the address below or from [docs.starkeyhearingtechnologies.com](http://docs.starkeyhearingtechnologies.com)

#### Starkey Hearing Technologies

6700 Washington Ave South  
Eden Prairie, MN 55344 USA



EC REP

Wm F. Austin House, Bramhall Technology Park  
Pepper Road, Hazel Grove, Stockport SK7 5BX  
United Kingdom

Class II Device



Waste from electronic equipment  
must be handled according to  
local regulations



Consult Operations Manual



#### INSTRUCTIONS FOR DISPOSAL OF OLD ELECTRONICS

Starkey Hearing Technologies encourages, the EU requires, and your local community laws may require, that your hearing aids and charger be disposed of via your local electronics recycling/disposal process.

The instructions below are provided for the benefit of disposal/recycling personnel. Please include this manual when disposing of your hearing aids and/or charger.

#### FOR DISPOSAL/RECYCLING PERSONNEL ONLY

These products contain Lithium Ion Polymer batteries. For instructions on removing the battery from the hearing aids, please visit [docs.starkeyhearingtechnologies.com](http://docs.starkeyhearingtechnologies.com). To remove the battery from the Synergy Charger:

- Remove the four skid pads on the bottom of the charger base to access fasteners.
- Remove fasteners using a Phillips head screwdriver
- Disassemble the charger base from the bezel to expose the battery cell
- Cut the THREE battery wire leads ONE AT A TIME close to the battery cell to avoid shorting
- Pry battery from base using a wide flat blade making sure not to puncture the battery cell