

Calibration Certificate

1800 MHz System Validation Dipole

Type

D1800V2

Serial Number:

207

Place of Calibration

Zurich

Date of Calibration

January 10, 2002

Calibration Interval

24 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by

N. Edosé Nwiana

Approved by:

Leonie Kahl

DASY

Dipole Validation Kit

Type: D1800V2

Serial: 207

Manufactured: July 1997

Calibrated: January 10, 2002

1. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with head simulating glycol solution of the following electrical parameters at 1800 MHz:

Relative Dielectricity	40.5	± 5%
Conductivity	1.35 mho/m	± 5%

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 5.31 at 1800 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 10mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW ± 3 %. The results are normalized to 1W input power.

2. SAR Measurement

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 1. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

averaged over cm ³ (1 g) of tissue	39.0 mW/g
averaged over 10 cm ³ (10 g) of tissue:	20.3 mW/g

Note: If the liquid parameters for validation are slightly different from the ones used for initial calibration, the SAR-values will be different as well. The estimated sensitivities of SAR-values and penetration depths to the liquid parameters are listed in the DASY Application Note 4: 'SAR Sensitivities'.

3. Dipole Impedance and Return Loss

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

Electrical delay:	1.232 ns	(one direction)
Transmission factor:	0.982	(voltage transmission, one direction)

The dipole was positioned at the flat phantom sections according to section 1 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 1800 MHz:	$\text{Re}\{Z\} = 46.0 \Omega$
----------------------------------	--

	$\text{Im}\{Z\} = -6.0 \Omega$
--	--

Return Loss at 1800 MHz	-22.5 dB
-------------------------	-----------------

4. Handling

Do not apply excessive force to the dipole arms, because they might bend. Bending of the dipole arms stresses the soldered connections near the feedpoint leading to a damage of the dipole.

5. Design

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

6. Power Test

After long term use with 40W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Validation Dipole D1800V2 SN:207, d = 10 mm

Frequency: 1800 MHz; Antenna Input Power: 250 [mW]


SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0

Probe: ET3DV6 - SN1507; ConvF(5.31,5.31,5.31) at 1800 MHz; IEEE1528 1800 MHz ; $\sigma = 1.35$ mho/m $\epsilon_r = 40.5$ $\rho = 1.00$ g/cm³


Cubes (2): Peak: 18.4 mW/g ± 0.06 dB, SAR (1g): 9.76 mW/g ± 0.04 dB, SAR (10g): 5.08 mW/g ± 0.02 dB, (Worst-case extrapolation)

Penetration depth: 8.2 (7.7, 9.3) [mm]

Powerdrift: -0.03 dB



SAR_{Tot} [mW/g]



10 Jan 2002 16:32:38

CH1 S11 1 U FS

1: 45.973 Ω -6.0176 Ω 14.694 pF


1 800.000 000 MHz

Del

Cor

Avg
16


↑



CH2 S11 LOG 5 dB/REF 0 dB 1:-22.454 dB 1 800.000 000 MHz

Cor

↑



START 1 500.000 000 MHz

STOP 2 100.000 000 MHz

Calibration Certificate

900 MHz System Validation Dipole

Type

D900V2

Serial Number:

056

Place of Calibration:

Zurich

Date of Calibration

January 29, 2002

Calibration Interval

24 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by

N. Kloss E. Meriana

Approved by

Alain Klatz

DASY

Dipole Validation Kit

Type: D900V2

Serial: 056

Manufactured: September 25, 1999

Calibrated: January 29, 2002

1. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with head simulating solution of the following electrical parameters at 900 MHz:

Relative Dielectricity	41.1	$\pm 5\%$
Conductivity	0.95 mho/m	$\pm 5\%$

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.48 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW $\pm 3\%$. The results are normalized to 1W input power.

2. SAR Measurement

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 1. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

averaged over 1 cm ³ (1 g) of tissue:	11.1 mW/g
averaged over 10 cm ³ (10 g) of tissue:	7.00 mW/g

Note: If the liquid parameters for validation are slightly different from the ones used for initial calibration, the SAR-values will be different as well.

3. Dipole Impedance and Return Loss

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

Electrical delay:	1.348 ns	(one direction)
Transmission factor:	0.986	(voltage transmission, one direction)

The dipole was positioned at the flat phantom sections according to section and the distance holder was in place during impedance measurements.

Feedpoint impedance at 900 MHz:	$\text{Re}\{Z\} = 50.2 \Omega$
---------------------------------	--------------------------------

	$\text{Im}\{Z\} = -0.7 \Omega$
--	--------------------------------

Return Loss at 900 MHz	-42.9 dB
------------------------	-----------------

4. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with body simulating solution of the following electrical parameters at 900 MHz:

Relative Dielectricity	54.8	$\pm 5\%$
Conductivity	1.03 mho/m	$\pm 5\%$

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.17 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW $\pm 3\%$. The results are normalized to 1W input power.

5. SAR Measurement

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 4. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

averaged over 1 cm³ (1 g) of tissue: **11.7 mW/g**

averaged over 10 cm³ (10 g) of tissue: **7.44 mW/g**

6. Dipole Impedance and Return Loss

The dipole was positioned at the flat phantom sections according to section 4 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 900 MHz: **Re{Z} = 45.7 Ω**

Im {Z} = -2.4 Ω

Return Loss at 900 MHz **-25.7 dB**

7. Handling

Do not apply excessive force to the dipole arms, because they might bend. Bending of the dipole arms stresses the soldered connections near the feedpoint leading to a damage of the dipole.

8. Design

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. •

9. Power Test

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Validation Dipole D900V2 SN:056, d = 15 mm

Frequency: 900 MHz; Antenna Input Power: 250 [mW]


SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0

Probe: ET3DV6 - SN1507; ConvF(6.48,6.48,6.48) at 900 MHz; IEEE1528 900 MHz; $\sigma = 0.95$ mho/m $\epsilon_r = 41.1$ $\rho = 1.00$ g/cm³

Cubes (2): Peak: 4.48 mW/g ± 0.01 dB, SAR (1g): 2.78 mW/g ± 0.02 dB, SAR (10g): 1.75 mW/g ± 0.02 dB, (Worst-case extrapolation)

Penetration depth: 11.5 (10.3, 13.0) [mm]

Powerdrift: -0.02 dB



CH1 S11 1 U FS


1: 50.168 Ω -728.52 m Ω 242.74 pF

900.000 000 MHz

De1

Cor


Avg
16



CH2 S11 LOG

Cor

↑



START 700.000 000 MHz

STOP 1 100.000 000 MH

Validation Dipole D900V2 SN:056, d = 15 mm

Frequency: 900 MHz; Antenna Input Power: 250 [mW]


SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0

Probe: ET3DV6 - SN1507; ConvF(6.17,6.17,6.17) at 900 MHz; Muscle 900 MHz; $\sigma = 1.03$ mho/m $\epsilon_r = 54.8$ $\rho = 1.00$ g/cm³

Cubes (2): Peak: 4.65 mW/g ± 0.01 dB, SAR (1g): 2.92 mW/g ± 0.00 dB, SAR (10g): 1.86 mW/g ± 0.01 dB, (Worst-case extrapolation)

Penetration depth: 12.0 (10.7, 13.7) [mm]

Powerdrift: -0.02 dB



CH1 S11 1 U FS

1: 45.686 Ω -2.4453 Ω 72.317 pF


900.000 000 MHz

De1

Cor

Avg
16


↑



CH2 S11 LOG 5 dB/REF 0 dB 1:-25.715 dB 900.000 000 MHz

Cor

Avg
16



CENTER 900.000 000 MHz

SPAN 400.000 000 MHz

Calibration Certificate

1800 MHz System Validation Dipole

Type

D1800V2

Serial Number:

256

Place of Calibration:

Zurich

Date of Calibration:

January 29, 2002

Calibration Interval:

24 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

M. Soloski Meriana

Approved by:

Blanc Katja

DASY

Dipole Validation Kit

Type: D1800V2

Serial: 256

Manufactured: December 23, 1999

Calibrated: January 29, 2002

1. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with head simulating glycol solution of the following electrical parameters at 1800 MHz:

Relative Dielectricity	39.6	$\pm 5\%$
Conductivity	1.37 mho/m	$\pm 5\%$

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 5.31 at 1800 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 10mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW $\pm 3\%$. The results are normalized to 1W input power.

2. SAR Measurement

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 1. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

averaged over 1 cm ³ (1 g) of tissue:	40.0 mW/g
averaged over 10 cm ³ (10 g) of tissue:	20.8 mW/g

Note: If the liquid parameters for validation are slightly different from the ones used for initial calibration, the SAR-values will be different as well. The estimated sensitivities of SAR-values and penetration depths to the liquid parameters are listed in the DASY Application Note 4: 'SAR Sensitivities'.

3. Dipole Impedance and Return Loss

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

Electrical delay:	1.180 ns	(one direction)
Transmission factor:	0.980	(voltage transmission, one direction)

The dipole was positioned at the flat phantom sections according to section 1 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 1800 MHz:	$\text{Re}\{Z\} = 47.3 \Omega$
	$\text{Im}\{Z\} = -6.6 \Omega$
Return Loss at 1800 MHz	-22.7 dB

4. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with body simulating glycol solution of the following electrical parameters at 1800 MHz:

Relative Dielectricity	53.5	$\pm 5\%$
Conductivity	1.45 mho/m	$\pm 5\%$

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 5.0 at 1800 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 10mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW $\pm 3\%$. The results are normalized to 1W input power.

5. SAR Measurement

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 4. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

averaged over 1 cm ³ (1 g) of tissue:	39.5 mW/g
averaged over 10 cm ³ (10 g) of tissue:	20.4 mW/g

6. Dipole Impedance and Return Loss

The dipole was positioned at the flat phantom sections according to section 4 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 1800 MHz:	Re{Z} = 43.2 Ω
	Im {Z} = -6.3 Ω
Return Loss at 1800 MHz	-20.1 dB

7. Handling

Do not apply excessive force to the dipole arms, because they might bend. Bending of the dipole arms stresses the soldered connections near the feedpoint leading to a damage of the dipole.

8. Design

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

9. Power Test

After long term use with 40W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Validation Dipole D1800V2 SN:256, d = 10 mm

Frequency: 1800 MHz; Antenna Input Power: 250 [mW]


SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0

Probe: ET3DV6 - SN1507; ConvF(5.31,5.31,5.31) at 1800 MHz; IEEE1528 1800 MHz; $\sigma = 1.37$ mho/m $\epsilon_r = 39.6$ $\rho = 1.00$ g/cm³

Cubes (2): Peak: 18.9 mW/g ± 0.04 dB, SAR (1g): 9.99 mW/g ± 0.01 dB, SAR (10g): 5.19 mW/g ± 0.01 dB, (Worst-case extrapolation)

Penetration depth: 8.2 (7.7, 9.2) [mm]

Powerdrift: -0.02 dB





CH1 S11 1 U FS

1: 47.275 Ω -6.5762 Ω 13.445 pF

1 800.000 000 MHz

De1

Avg
1K

Validation Dipole D1800V2 SN:256, d = 10 mm

Frequency: 1800 MHz; Antenna Input Power: 250 [mW]


SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0

Probe: ET3DV6 - SN1507; ConvF(5.00,5.00,5.00) at 1800 MHz; Muscle 1800 MHz; $\sigma = 1.45$ mho/m $\epsilon_r = 53.5$ $\rho = 1.00$ g/cm³


Cubes (2): Peak: 18.7 mW/g ± 0.02 dB, SAR (1g): 9.87 mW/g ± 0.01 dB, SAR (10g): 5.11 mW/g ± 0.01 dB, (Worst-case extrapolation)

Penetration depth: 8.4 (7.7, 9.8) [mm]

Powerdrift: -0.01 dB



SAR_{Tot} [mW/g]



CH1 S11 1 U FS


1: 43.176 Ω -6.2773 Ω 14.085 pF

29 Jan 2002 12:57:01

1 800.000 000 MHz

De1

Cor



CH2 S11 LOG 5 dB/REF 0 dB 1:-20.061 dB 1 800.000 000 MHz

