

Appendix 9

pages 1

VERIFY WITH DIPOLE VALIDATION KIT

Validation Kit 900 MHz, SN: 003, d = 15 mm, f = 900 MHz, Pin = 250 mW, Generic Flat Phantom, Measured 2000-2-29/NMP

$\sigma = 0.85$ [mho/m] $\epsilon_r = 43.3$ $\rho = 1.00$ [g/cm³]

Coarse Grid Dx = 20.0 Dy = 20.0 Dz = 5.0 [mm]

SAR [mW/g] Max 2.55

SAR (1g): 2.33 [mW/g] SAR (10g): 1.51 [mW/g]

