

MOTOROLA MOBILITY, EME Laboratories

Extension of Dipole Calibration Interval

Dipole SN: 863

Date of last external calibration: 3/17/2011

External calibration performed by: SPEAG

1st Extension

	Original Calibration Data: <3/17/11>	Internal Verification: <3/14/12>		
Return loss (dB): <i>Verify < -20dB</i>	-27.7	< -20dB		
Impedance, Real (Ω): <i>Verify within +/-5 Ω of original</i>	53.1	51.3		
Impedance, Imaginary (Ω): <i>Verify within +/-5 Ω of original</i>	2.9	-1.2		