Extension of Dipole Calibration Interval

)-1954, Rev.3

Dipole SN: 436tr

Date of last external calibration: 3/18/2011

External calibration performed by: SPEAG

		1st Extension	2nd Extension	
	Original	Internal	Internal	
	Calibration Data:	Verification:	Verification:	Accept / Reject
	3/18/2011	3/8/2012	1/4/2013	
Return loss (dB): Verify < -20dB & within 20% of original	-31.1	-36.7	-26.7	?
Impedance, Real (Ω): Verify within +/-5 Ω of original	51.4	50.4	51.9	?
Impedance, Imaginary (Ω): Verify within +/-5 Ω of original	-2.5	-1.3	-4.2	?

Conclusion:

Based on the requirements of KDB 865664 D01 SAR Measurement Requirements for 100MHz to 6GHz, it has been concluded that the dipole identified above has qualified for extension of its calibration interval for one additional year.

This review shall be repeated annually, but not to exceed a maximum 3 years from the most recent manufacturer's calibration.

First Extension Period:

Approved extension period: 1 Year

Previous due date: 3/18/2012 **NEW DUE DATE:** 3/18/2013

2 years from date of last external calibration

Authorized by: Marge Kaunas

Second Extension Period:

Approved extension period: 1 Year

Previous due date: 3/18/2013 **NEW DUE DATE:** 3/18/2014

Extension of Dipole Calibration Interval

FCD-1954, Rev.3

Dipole SN:	422tr
Date of last external calibration:	3/18/2011
External calibration performed by:	SPEAG

		1st Extension	2nd Extension	
	Original	Internal	Internal	
	Calibration Data:	Verification:	Verification:	Accept / Reject
	3/18/2011	3/8/2012	1/4/2013	
Return loss (dB): Verify < -20dB & within 20% of original	-26.9	-23	-27.4	Accept
Impedance, Real (Ω): Verify within +/-5 Ω of original	53.3	55	51.9	Accept
Impedance, Imaginary (Ω): Verify within +/-5 Ω of original	3.3	3.2	-1.6	Accept

Conclusion:

Based on the requirements of KDB 865664 D01 SAR Measurement Requirements for 100MHz to 6GHz, it has been concluded that the dipole identified above has qualified for extension of its calibration interval for one additional year.

This review shall be repeated annually, but not to exceed a maximum 3 years from the most recent manufacturer's calibration.

First Extension Period:

Approved extension period: 1 Year

Previous due date: 3/18/2012 **NEW DUE DATE: 3/18/2013**

2 years from date of last external calibration

Authorized by: Marge Kaunas

Second Extension Period:

Approved extension period: 1 Year

Previous due date: 3/18/2013 **NEW DUE DATE:** 3/18/2014

Extension of Dipole Calibration Interval

FCD-1954, Rev.3

Dipole SN: 2d190

Date of last external calibration: 1/5/2012

External calibration performed by: SPEAG

		1st Extension	2nd Extension	
	Original	Internal	Internal	
	Calibration Data:	Verification:	Verification:	Accept / Reject
	1/5/2012	1/8/2013	<date></date>	
Return loss (dB): Verify < -20dB & within 20% of original	-25.3	-26.4	1	Accept
Impedance, Real (Ω): Verify within +/-5 Ω of original	48.3	51.5	-	Accept
Impedance, Imaginary (Ω): Verify within +/-5 Ω of original	-5.1	-4.6	1	Accept

Conclusion:

Based on the requirements of KDB 865664 D01 SAR Measurement Requirements for 100MHz to 6GHz, it has been concluded that the dipole identified above has qualified for extension of its calibration interval for one additional year.

This review shall be repeated annually, but not to exceed a maximum 3 years from the most recent manufacturer's calibration.

First Extension Period:

Approved extension period: 1 Year

Previous due date: 1/5/2013 **NEW DUE DATE: 1/5/2014**

2 years from date of last external calibration

Authorized by: Marge Kaunas

Second Extension Period:

Approved extension period: 1 Year

Previous due date:

NEW DUE DATE:

Extension of Dipole Calibration Interval

FCD-1954, Rev.2

Dipole SN:	259tr
Date of last external calibration:	10/20/2011
External calibration performed by:	SPEAG

		1st Extension	2nd Extension	
	Original	Internal	Internal	
	Calibration Data:	Verification:	Verification:	Accept / Reject
	10/20/2011	10/16/2012	<date></date>	
Return loss (dB): Verify < -20dB & within 20% of original	-24.7	-26.6	-	Accept
Impedance, Real (Ω): Verify within +/-5 Ω of original	49.1	49.3	-	Accept
Impedance, Imaginary (Ω): Verify within +/-5 Ω of original	-5.7	-2.4	-	Accept

Conclusion:

Based on the requirements of KDB 50824 D02 Dipole SAR Validation Veritification v01, it has been concluded that the dipole identified above has qualified for extension of its calibration interval for one additional year. This review shall be repeated annually, but not to exceed a maximum 3 years from the most recent manufacturer's calibration.

First Extension Period:

Approved extension period: 1 Year

Previous due date: 10/20/2012 **NEW DUE DATE: 10/20/2013**

2 years from date of last external calibration

Authorized by: Marge Kaunas

Second Extension Period:

Approved extension period: 1 Year

Previous due date:

NEW DUE DATE:

Extension of Dipole Calibration Interval

FCD-1954, Rev.3

Dipole SN:	740
Date of last external calibration:	2/7/2012
External calibration performed by:	SPEAG

		1st Extension	2nd Extension	
	Original	Internal	Internal	
	Calibration Data:	Verification:	Verification:	Accept / Reject
	2/7/2012	1/8/2013	<date></date>	
Return loss (dB): Verify < -20dB & within 20% of original	-28.9	-27.3	-	Accept
Impedance, Real (Ω): Verify within +/-5 Ω of original	52.2	54.4	-	Accept
Impedance, Imaginary (Ω): Verify within +/-5 Ω of original	2.9	1.01	-	Accept

Conclusion:

Based on the requirements of KDB 865664 D01 SAR Measurement Requirements for 100MHz to 6GHz, it has been concluded that the dipole identified above has qualified for extension of its calibration interval for one additional year.

This review shall be repeated annually, but not to exceed a maximum 3 years from the most recent manufacturer's calibration.

First Extension Period:

Approved extension period: 1 Year

Previous due date: 2/7/2013 **NEW DUE DATE: 2/7/2014**

2 years from date of last external calibration

Authorized by: Marge Kaunas

Second Extension Period:

Approved extension period: 1 Year

Previous due date:

NEW DUE DATE: