

HAC_E_Dipole_835_160529

DUT: HAC-Dipole 835 MHz

Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration

- Probe: ER3DV6 - SN2358; ConvF(1, 1, 1); Calibrated: 2016/1/19;
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1399; Calibrated: 2015/11/23
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (7); SEMCAD X Version 14.6.10 (7164)

E Scan - measurement distance from the probe sensor center to CD835 = 10mm & 15mm/Hearing Aid Compatibility Test at 15mm distance (41x361x1): Interpolated grid:

dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 125.7 V/m; Power Drift = 0.02 dB

PMR not calibrated. PMF = 1.000 is applied.

E-field emissions = 111.9 V/m

Average value of Total=(111.9+110.7) / 2 = 111.3 V/m

PMF scaled E-field



Grid 1 M4 108.8 V/m	Grid 2 M4 111.9 V/m	Grid 3 M4 110.8 V/m
Grid 4 M4 62.61 V/m	Grid 5 M4 64.15 V/m	Grid 6 M4 63.46 V/m
Grid 7 M4 108.2 V/m	Grid 8 M4 110.7 V/m	Grid 9 M4 109.2 V/m

Cursor:

Total = 111.9 V/m

E Category: M4

Location: -1, -75, 9.7 mm



0 dB = 111.9 V/m = 40.98 dBV/m

HAC_E_Dipole_1880_160529

DUT: HAC Dipole 1880 MHz

Communication System: CW; Frequency: 1880 MHz; Duty Cycle: 1:1
 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³
 Ambient Temperature : 23.2 °C

DASY5 Configuration

- Probe: ER3DV6 - SN2358; ConvF(1, 1, 1); Calibrated: 2016/1/19;
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1399; Calibrated: 2015/11/23
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (7); SEMCAD X Version 14.6.10 (7164)

E Scan - measurement distance from the probe sensor center to CD1880 = 10mm & 15mm/Hearing Aid Compatibility Test at 15mm distance (41x181x1): Interpolated grid:
 dx=0.5000 mm, dy=0.5000 mm



Device Reference Point: 0, 0, -6.3 mm
 Reference Value = 138.8 V/m; Power Drift = 0.01 dB
 PMR not calibrated. PMF = 1.000 is applied.
 E-field emissions = 90.54 V/m
 Average value of Total=(86.98+90.54) / 2 = 88.76 V/m

PMF scaled E-field

Grid 1 M3 85.11 V/m	Grid 2 M3 86.98 V/m	Grid 3 M3 86.02 V/m
Grid 4 M3 70.00 V/m	Grid 5 M3 71.31 V/m	Grid 6 M3 70.51 V/m
Grid 7 M3 88.96 V/m	Grid 8 M3 90.54 V/m	Grid 9 M3 88.73 V/m

Cursor:

Total = 90.54 V/m
 E Category: M3
 Location: 0, 30.5, 9.7 mm



0 dB = 90.54 V/m = 39.14 dBV/m

HAC_E_Dipole_2600_160529

DUT: HAC Dipole 2600 MHz

Communication System: CW ; Frequency: 2600 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.2 °C

DASY5 Configuration

- Probe: ER3DV6 - SN2358; ConvF(1, 1, 1); Calibrated: 2016/1/19;
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn1399; Calibrated: 2015/11/23
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (7); SEMCAD X Version 14.6.10 (7164)

E Scan - measurement distance from the probe sensor center to CD2600 = 10mm & 15mm/Hearing Aid Compatibility Test at 15mm distance (41x181x1): Interpolated grid:

dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 69.34 V/m; Power Drift = -0.00 dB

PMR not calibrated. PMF = 1.000 is applied.

E-field emissions = 91.07 V/m

Average value of Total=(84.67+91.07) / 2 = 87.87 V/m

PMF scaled E-field



Grid 1 M3 83.48 V/m	Grid 2 M3 84.67 V/m	Grid 3 M3 83.61 V/m
Grid 4 M3 85.97 V/m	Grid 5 M3 87.78 V/m	Grid 6 M3 86.71 V/m
Grid 7 M3 89.36 V/m	Grid 8 M3 91.07 V/m	Grid 9 M3 89.37 V/m

Cursor:

Total = 91.07 V/m

E Category: M3

Location: 0, 20, 9.7 mm



0 dB = 91.07 V/m = 39.19 dBV/m