

Schmid & Partner Engineering AG

Zughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

Calibration Certificate

Dosimetric E-Field Probe

Type:

E33Dv6

Serial Number:

1371

Place of Calibration:

Zurich

Date of Calibration:

November 23, 2001

Calibration Interval:

12 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Whenever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:



Approved by:



**SENNER & BOELLER
Engineering AG**

Zughausstrasse 43, 8004 Zurich, Switzerland, Telephone +41 1 245 97 00, Fax +41 1 245 97 79

Probe ET3DV6

SN:1397

| | |
|----------------------|--------------------------|
| Manufactured: | October 24, 1999 |
| Repaired: | November 15, 2001 |
| Recalibrated: | November 23, 2001 |

Calibrated for System DASY3

ET3DV6 SN:1397

DASY3 - Parameters of Probe: ET3DV6 SN:1397

Sensitivity in Free Space

NormX 2.05 $\mu\text{V}/(\text{V}/\text{m})^2$
NormY 2.01 $\mu\text{V}/(\text{V}/\text{m})^2$
NormZ 2.03 $\mu\text{V}/(\text{V}/\text{m})^2$

Diode Compression

DCP X 97 mV
DCP Y 97 mV
DCP Z 97 mV


Sensitivity in Tissue Simulating Liquid

| | | | |
|---------|------------------|-----------------------------|---|
| Head | 450 MHz | $\epsilon_r = 43.5 \pm 5\%$ | $\sigma = 0.87 \pm 10\% \text{ mho}/\text{m}$ |
| ConvF X | 7.17 | extrapolated | Boundary effect: |
| ConvF Y | 7.17 | extrapolated | Alpha 0.40 |
| ConvF Z | 7.17 | extrapolated | Depth 2.15 |
| Head | 500 - 1000 MHz | $\epsilon_r = 39.0 - 43.5$ | $\sigma = 0.80 - 1.10 \text{ mho}/\text{m}$ |
| ConvF X | 6.61 $\pm 9.5\%$ | (k=2) | Boundary effect: |
| ConvF Y | 6.61 $\pm 9.5\%$ | (k=2) | Alpha 0.46 |
| ConvF Z | 6.61 $\pm 9.5\%$ | (k=2) | Depth 2.18 |
| Head | 1500 MHz | $\epsilon_r = 40.4 \pm 5\%$ | $\sigma = 1.23 \pm 10\% \text{ mho}/\text{m}$ |
| ConvF X | 5.85 | interpolated | Boundary effect: |
| ConvF Y | 5.85 | interpolated | Alpha 0.54 |
| ConvF Z | 5.85 | interpolated | Depth 2.21 |
| Head | 1700 - 1910 MHz | $\epsilon_r = 39.8 - 41.0$ | $\sigma = 1.20 - 1.55 \text{ mho}/\text{m}$ |
| ConvF X | 5.47 $\pm 9.5\%$ | (k=2) | Boundary effect: |
| ConvF Y | 5.47 $\pm 9.5\%$ | (k=2) | Alpha 0.57 |
| ConvF Z | 5.47 $\pm 9.5\%$ | (k=2) | Depth 2.23 |


Sensor Offset

Probe Tip to Sensor Center 2.7 mm
Optical Surface Detection 1.5 ± 0.2 mm


Receiving Pattern (ϕ), $\theta = 0^\circ$



ET3DV6 SN:1397




Isotropy Error (ϕ), $\theta = 0^\circ$





Frequency Response of E-Field


(TEM-Cell:if1110, Waveguide R22)



Dynamic Range f(SAR_{brain}) (Waveguide R22)



Conversion Factor Assessment



Head 800 - 1080 MHz $\epsilon_r = 39.0 - 43.6$ $\sigma = 0.80 - 1.10$ mho/m

ConvF X $6.61 \pm 9.5\%$ (k=2)

ConvF Y $6.61 \pm 9.6\%$ (k=2)

ConvF Z $6.61 \pm 9.5\%$ (k=2)

Boundary effect:

Alpha 0.45

Depth 2.15

Head 1700 - 1910 MHz $\epsilon_r = 38.5 - 41.0$ $\sigma = 1.20 - 1.55$ mho/m

ConvF X $6.47 \pm 9.5\%$ (k=2)

ConvF Y $6.47 \pm 9.6\%$ (k=2)

ConvF Z $6.47 \pm 9.5\%$ (k=2)

Boundary effect:

Alpha 0.57

Depth 2.23

Schmid & Partner Engineering AG

Zeughäuserstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 243 97 00, Fax +41 1 243 97 79

Additional Conversion Factors for Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1897

Place of Assessment:

Zurich

Date of Assessment:

November 25, 2001

Probe Calibration Date:

November 23, 2001

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:

Ilse Kötter

Dosimetric E-Field Probe ET3DV6 SN:1397Conversion factor (\pm standard deviation)

| | | | |
|----------|-------|--------------|---|
| 835 MHz | ConvF | 6.7 \pm 8% | $\epsilon_r = 41.5$ $\sigma = 0.98$ mho/m (head tissue) |
| 1950 MHz | ConvF | 5.2 \pm 8% | $\epsilon_r = 40.8$ $\sigma = 1.40$ mho/m (head tissue) |
| 835 MHz | ConvF | 6.5 \pm 8% | $\epsilon_r = 53.2$ $\sigma = 0.97$ mho/m (body tissue) |
| 900 MHz | ConvF | 6.4 \pm 8% | $\epsilon_r = 55.0$ $\sigma = 1.05$ mho/m (body tissue) |
| 1800 MHz | ConvF | 5.1 \pm 8% | $\epsilon_r = 53.3$ $\sigma = 1.52$ mho/m (body tissue) |
| 1950 MHz | ConvF | 4.9 \pm 8% | $\epsilon_r = 53.3$ $\sigma = 1.52$ mho/m (body tissue) |

Schmid & Partner Engineering AG

Zughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

Calibration Certificate

Dosimetric E-Field Probe

Type:

ET3D1V6

Serial Number:

1398

Place of Calibration:

Zurich

Date of Calibration:

August 31, 2001

Calibration Interval:

12 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

M. Tolosa

Approved by:

Alain Klotz

Engineering AG

Zeughausstrasse 42, 8004 Zurich, Switzerland, Telephone +41 1 245 97 00, Fax +41 1 245 97 79

Probe ET3DV6

SN:1398

| | |
|----------------------|-------------------------|
| Manufactured: | October 24, 1999 |
| Repaired: | August 24, 2001 |
| Calibrated: | August 31, 2001 |

Calibrated for System DASY3

ET3DV6 SN:1398

DASY3 - Parameters of Probe: ET3DV6 SN:1398

Sensitivity in Free Space

NormX 1.62 $\mu\text{V}/(\text{V}/\text{m})^2$
NormY 1.87 $\mu\text{V}/(\text{V}/\text{m})^2$
NormZ 1.71 $\mu\text{V}/(\text{V}/\text{m})^2$

Diode Compression

DCP X 97 mV
DCP Y 97 mV
DCP Z 97 mV


Sensitivity in Tissue Simulating Liquid

| | | | |
|---------|----------|-----------------------------|---|
| Head | 450 MHz | $\epsilon_r = 43.5 \pm 5\%$ | $\sigma = 0.87 \pm 10\% \text{ mho}/\text{m}$ |
| ConvF X | 6.95 | extrapolated | Boundary effect: |
| ConvF Y | 6.95 | extrapolated | Alpha 0.45 |
| ConvF Z | 6.95 | extrapolated | Depth 2.12 |
| Head | 900 MHz | $\epsilon_r = 42 \pm 6\%$ | $\sigma = 0.97 \pm 10\% \text{ mho}/\text{m}$ |
| ConvF X | 6.43 | $\pm 7\%$ (k=2) | Boundary effect: |
| ConvF Y | 6.43 | $\pm 7\%$ (k=2) | Alpha 0.49 |
| ConvF Z | 6.43 | $\pm 7\%$ (k=2) | Depth 2.18 |
| Head | 1800 MHz | $\epsilon_r = 40.4 \pm 8\%$ | $\sigma = 1.23 \pm 10\% \text{ mho}/\text{m}$ |
| ConvF X | 5.74 | interpolated | Boundary effect: |
| ConvF Y | 5.74 | interpolated | Alpha 0.54 |
| ConvF Z | 5.74 | interpolated | Depth 2.26 |
| Head | 1800 MHz | $\epsilon_r = 40 \pm 8\%$ | $\sigma = 1.49 \pm 10\% \text{ mho}/\text{m}$ |
| ConvF X | 5.39 | $\pm 7\%$ (k=2) | Boundary effect: |
| ConvF Y | 5.39 | $\pm 7\%$ (k=2) | Alpha 0.58 |
| ConvF Z | 5.39 | $\pm 7\%$ (k=2) | Depth 2.30 |


Sensor Offset

Probe Tip to Sensor Center 2.7 mm
Optical Surface Detection 1.2 ± 0.2 mm


Receiving Pattern (ϕ), $\theta = 0^\circ$



ET3DV8 SN:1388




Isotropy Error (ϕ), $\theta = 0^\circ$





Frequency Response of E-Field


(TEM-Cell: Ifi110, Waveguide R22)



Dynamic Range $f(\text{SAR}_{\text{brain}})$ (Waveguide R22)



Conversion Factor Assessment



| | | | | |
|------|---------|---------------------------|--------------------------------|------|
| Head | 900 MHz | $\epsilon_r = 42 \pm 5\%$ | $\sigma = 0.97 \pm 10\%$ mho/m | |
| | ConvF X | $6.43 \pm 7\%$ (k=2) | Boundary effect | |
| | ConvF Y | $6.43 \pm 7\%$ (k=2) | Alpha | 0.49 |
| | ConvF Z | $6.43 \pm 7\%$ (k=2) | Depth | 2.18 |

| | | | | |
|------|----------|---------------------------|--------------------------------|------|
| Head | 1800 MHz | $\epsilon_r = 40 \pm 5\%$ | $\sigma = 1.40 \pm 10\%$ mho/m | |
| | ConvF X | $6.39 \pm 7\%$ (k=2) | Boundary effect | |
| | ConvF Y | $6.39 \pm 7\%$ (k=2) | Alpha | 0.56 |
| | ConvF Z | $6.39 \pm 7\%$ (k=2) | Depth | 2.30 |

Schmid & Partner Engineering AG

Zughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

Additional Conversion Factors for Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1398

Place of Assessment:

Zurich

Date of Assessment:


September 7, 2001

Probe Calibration Date:

August 31, 2001

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:



Dosimetric E-Field Probe ET3DV6 SN:1398

Conversion factor (\pm standard deviation)

| | | | |
|----------|-------|--------------|---|
| 835 MHz | ConvF | 6.5 \pm 8% | $\epsilon_r = 41.5$ $\sigma = 0.90$ mho/m (head tissue) |
| 835 MHz | ConvF | 6.3 \pm 8% | $\epsilon_r = 53.2$ $\sigma = 0.97$ mho/m (body tissue) |
| 900 MHz | ConvF | 6.2 \pm 8% | $\epsilon_r = 55.0$ $\sigma = 1.05$ mho/m (body tissue) |
| 1950 MHz | ConvF | 5.1 \pm 8% | $\epsilon_r = 40.0$ $\sigma = 1.40$ mho/m (head tissue) |
| 1800 MHz | ConvF | 5.0 \pm 8% | $\epsilon_r = 53.3$ $\sigma = 1.52$ mho/m (body tissue) |
| 1950 MHz | ConvF | 4.7 \pm 8% | $\epsilon_r = 53.3$ $\sigma = 1.52$ mho/m (body tissue) |

Dosimetric E-Field Probe ET3DV6 SN:1398

Conversion factor (\pm standard deviation)

| | | | |
|----------|-------|---------------|---|
| 835 MHz | ConvF | $6.5 \pm 8\%$ | $\epsilon_r = 44.0$ $\sigma = 0.90 \text{ mho/m}$ (brain tissue) |
| 835 MHz | ConvF | $6.4 \pm 8\%$ | $\epsilon_r = 52.0$ $\sigma = 1.10 \text{ mho/m}$ (muscle tissue) |
| 900 MHz | ConvF | $6.4 \pm 8\%$ | $\epsilon_r = 42.5$ $\sigma = 0.86 \text{ mho/m}$ (brain tissue) |
| 925 MHz | ConvF | $6.3 \pm 8\%$ | $\epsilon_r = 44.0$ $\sigma = 0.93 \text{ mho/m}$ (brain tissue) |
| 925 MHz | ConvF | $6.3 \pm 8\%$ | $\epsilon_r = 52.0$ $\sigma = 1.20 \text{ mho/m}$ (muscle tissue) |
| 1800 MHz | ConvF | $5.4 \pm 8\%$ | $\epsilon_r = 40.3$ $\sigma = 1.35 \text{ mho/m}$ (brain tissue) |
| 1800 MHz | ConvF | $5.5 \pm 8\%$ | $\epsilon_r = 41.0$ $\sigma = 1.69 \text{ mho/m}$ (brain tissue) |
| 1900 MHz | ConvF | $5.2 \pm 8\%$ | $\epsilon_r = 39.9$ $\sigma = 1.42 \text{ mho/m}$ (brain tissue) |
| 1800 MHz | ConvF | $5.1 \pm 8\%$ | $\epsilon_r = 50.0$ $\sigma = 1.58 \text{ mho/m}$ (muscle tissue) |
| 1900 MHz | ConvF | $5.0 \pm 8\%$ | $\epsilon_r = 50.0$ $\sigma = 1.64 \text{ mho/m}$ (muscle tissue) |

