



August 19, 2010

Nemko USA Inc.  
11696 Sorrento Valley Rd  
San Diego, CA  
USA

FCC: IWC-DTS  
IC: 9113A-DTS

### **Modular Approval Requirements**

Limited Modular Approval is being requested for this device.  
The module meets all of the technical specifications applicable to the frequency band of operation.

The module has its own RF shielding.

*The module contains an internal shield. The shielding is achieved by the ground plane on the bottom layer of the module as well as the die and substrate ground. This shielding prevents RF coupling between the RF circuitry of the module and any wires or circuits in the device into which the module is installed. In addition the traces are kept very short to minimize undesirable radiation. See the “PCB Layout.pdf” document for ground plane details.*

All modulation and data input(s) are buffered.

*Data to the modulation circuit is buffered on the module via the WirelessUSB integrated circuit. Refer to the information regarding the on-board SERDES provided with the file “Technical Description.pdf”.*

The module has its own power supply regulation and local reference oscillator.

*The remote module contains its own power supply regulation and the rf reference oscillator is contained within the module (reference 12 MHz oscillator X1).*

*Power supply regulation is provided within the WirelessUSB integrated circuit that regulates the control voltage to the rf and data circuits. Details of clocking and power management are contained in the document “Technical Description.pdf”.*

The modular transmitter must comply with the antenna requirements of Section 15.203 and 15.204(c). The certification submission contains a detailed description of the configuration of all antennas that will be used with the module.

*The remote module uses a 4.1dBi Chip Antenna that will be internal to the end-products into which this module is installed.*

For Industry Canada, the module meets certification labeling requirements. Host devices that contain separately certified modules do not need to be re-certified, provided that they meet the following conditions:

**Telemetry • Fixed Cable • Biofeedback • Research • Clinical • EMG Software**



- The host device, as a stand alone unit without any separately certified modules, complies with all applicable Radio Standards Specifications.
- The host device and all the separately certified modules it contains jointly meet the safety requirements of RSS-102, if applicable.
- The host device complies with the certification labeling requirements of each of the modules it contains.

*The module is appropriately labeled (refer to the label and label location drawings contained within this application).*

For the FCC, the modular transmitter must be tested in a stand-alone configuration, i.e., the module must not be inside another device during testing. This is intended to demonstrate that the module is capable of complying with Part 15 emission limits regardless of the device into which it is eventually installed. Unless the transmitter module will be battery powered, it must comply with the AC line conducted requirements found in Section 15.207.

*Test data contained in this application is for the device tested as a stand-alone device. Radiated spurious emissions data demonstrating compliance with the requirements of Part 15 of the FCC rules for intentional radiators has been provided. Note that these tests were performed with no shield on the device to demonstrate compliance with the relevant limits.*

For the FCC, the modular transmitter must be labeled with its own FCC ID number, and, if the FCC ID is not visible when the module is installed inside another device, then the outside of the device into which the module is installed must also display a label referring to the enclosed module. This exterior label can use wording such as the following: "Contains Transmitter Module FCC ID: XYZMODEL1" or "Contains FCC ID: XYZMODEL1."

*The module is appropriately labeled (refer to the label and label location drawings contained within this application)*

The modular transmitter must comply with any applicable RF exposure requirements.

*The module meets the requirements for a portable device that may be used at separation distances of less than 2.5cm from the human body because its output power is below the threshold of 60/fGHz mW (25mW for a 2.4GHz device).*

Installation and operation of this device is limited to Noraxon U.S.A., Inc only.

Sincerely,

Frank Hosner  
Technical Director  
Noraxon U.S.A., Inc