

Federal Communication Commission
Equipment Authorization Division,
Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

15/09/2020

To Whom It May Concern

ID Label/ Label Location – Unitron HelloGO Products

Pursuant to paragraph §2.925 of 47 C.F.R. the hearing instruments Unitron HelloGO, which is subject of this filing, shall bear a permanently affixed, readily visible label listing the information as specified in §2.926 and §15.19(a) of the 47 C.F.R. Due to the specific use of the hearing instrument (as shown in figure 1) its external surface is in contact with the user's skin behind the ear.



Figure 1: Position of the Hearing Instrument on the User

Therefore, it is not appropriate to permanently affix a label on the outer surface of the device to prevent skin irritations. In this case the size of the FCC ID will be smaller then 2 points, which does not satisfy the requirements for "readily visible" label as defined in §2.925 (d)(2) and §2.925(g).

Due to the small size of the devices and pursuant to paragraph § 15.19(a)(5) of 47 C.F.R., the FCC ID and the statement specified in §15.19(a)(3) are placed in the user manual as shown below and affixed on the device packaging.

	USA - FCC ID:	Canada - IC:
HelloGO	KWC-HGO	2262A-HGO



Kevin Dafoe
Director, Systems & Supply Chain and Regulatory Affairs

SONOVA
HEAR THE WORLD