

## ID Label / Label Location Phonak Roger 21

Pursuant to §2.925 of 47 C.F.R. the model Roger 21, which is the object of this filing, shall bear a permanently affixed, readily visible label listing the information as specified in §2.926 and §15.19(a) of 47 C.F.R. Due to the specific use of this product, its external surface is in a permanent contact with the skin behind the ear. Therefore, it is not appropriate to permanently affix a label on the outer surface of the device which is in contact with the skin in order to prevent any possible skin irritations. The only possible place where the FCC ID and the serial number of the device can be put is the short side of the battery compartment of the attached hearing aid, which is with a surface smaller than 15mm x 4mm.. However, in this case the size of the FCC ID will be smaller than 2 points, which does not satisfy the requirements for "readily visible" label as defined in §2.925(d)(2) and §2.925(g).

Therefore, due to the very small size of the device and pursuant to paragraph §15.19(a)(5) of 47 C.F.R., the FCC ID and the statement specified in §15.19(a)(3) are placed in the Declaration of conformity as shown in the example on Figure 1.

### Roger 21

(02)

1.



2.



3.



4.





**Declaration of Conformity**

Roger receivers are products of Phonak AG, which have been tested and confirmed to be in compliance with the essential requirements and other relevant provisions of the Medical Devices Directive 93/42/EEC and Directive 1999/5/EC. To obtain a copy of the Declaration of conformity, please contact the manufacturer or the local Phonak representative whose address can be taken from the list on [www.phonak.com](http://www.phonak.com) (worldwide locations).

This device complies with Part 15 of the FCC Rules and with RSS210 of Industry Canada. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation. The device has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. The device generates uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If the device does cause harmful interference to radio or television reception, which can be determined by turning the device off and on, the user is encouraged to try to correct the interference by one of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the device and receiver.
- Connect the device into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

**FCC Caution: Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the device.**

Figure 1. Example of labeling information in the Declaration of Conformity of Phonak Roger 21