APPENDIX A: SAR DISTRIBUTION COMPARISON FOR ACCURACY VERIFICATION

Date/Time: 3/16/2010 1:41:57 AM

Test Laboratory: RIM TESTING SERVICES
File Name:
DipoleValidation_835MHz_Amb_Tem_22.7_Liq_Tem_22.0_C_03_16_10.da4

DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN:446
Program Name: System Performance Check at 835 MHz

Communication System: CW; Frequency: 835 MHz;Duty Cycle: 1:1
Medium parameters used: f = 835 MHz; σ = 0.856 mho/m; ε_r = 43.1; ρ = 1000 kg/m^3
Phantom section: Flat Section

DASY4 Configuration:
- Probe: ET3DV6 - SN1644; ConvF(6.08, 6.08, 6.08); Calibrated: 11/11/2009
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn473; Calibrated: 1/4/2010
- Phantom: SAM 1; Type: SAM 4.0; Serial: 1076
- Measurement SW: DASY4, V4.7 Build 80; Postprocessing SW: SEMCAD, V1.8 Build 186

d=15mm, Pin=1000mW/Zoom Scan (5x5x7) 2 (5x5x7)/Cube 0: Measurement grid: dx=7.5mm, dy=7.5mm, dz=5mm
Reference Value = 111.2 V/m; Power Drift = -0.034 dB
Peak SAR (extrapolated) = 13.0 W/kg
SAR(1 g) = 9.05 mW/g; SAR(10 g) = 5.97 mW/g
Maximum value of SAR (measured) = 9.77 mW/g

d=15mm, Pin=1000mW/Area Scan (31x121x1): Measurement grid: dx=15mm, dy=15mm
Maximum value of SAR (interpolated) = 9.74 mW/g



0 dB = 9.74mW/g

Date/Time: 3/16/2010 7:11:04 PM

Test Laboratory: RIM TESTING SERVICES
File Name:
DipoleValidation_1900MHz_Amb_Tem_22.6_Liq_Tem_21.2_C_03_16_10.da4

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN:545
Program Name: System Performance Check at 1900 MHz

Communication System: CW; Frequency: 1900 MHz;Duty Cycle: 1:1
Medium parameters used: f = 1900 MHz; σ = 1.43 mho/m; ε_r = 40.2; ρ = 1000 kg/m^3
Phantom section: Flat Section

DASY4 Configuration:
- Probe: ET3DV6 - SN1644; ConvF(5.17, 5.17, 5.17); Calibrated: 11/11/2009
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn473; Calibrated: 1/4/2010
- Phantom: SAM 1; Type: SAM 4.0; Serial: 1076
- Measurement SW: DASY4, V4.7 Build 80; Postprocessing SW: SEMCAD, V1.8 Build 186

d=10mm, Pin=1000mW/Zoom Scan (5x5x7) 2 (5x5x7)/Cube 0: Measurement grid: dx=7.5mm, dy=7.5mm, dz=5mm
Reference Value = 176.1 V/m; Power Drift = 0.026 dB
Peak SAR (extrapolated) = 64.9 W/kg
SAR(1 g) = 36.9 mW/g; SAR(10 g) = 19.2 mW/g
Maximum value of SAR (measured) = 41.6 mW/g

d=10mm, Pin=1000mW/Area Scan (31x61x1): Measurement grid: dx=15mm, dy=15mm
Maximum value of SAR (interpolated) = 42.2 mW/g



0 dB = 42.2mW/g

Date/Time: 3/15/2010 11:55:17 AM

File Name: DipoleValidation_2450MHz_Amb_Tem_23.0_Liq_Tem_21.2C.da4

DUT: Dipole 2450 MHz; Type: D2450V2; - SN:747
Program Name: System Performance Check at 2450 MHz

Communication System: CW; Frequency: 2450 MHz;Duty Cycle: 1:1
Medium parameters used: f = 2450 MHz; σ = 1.88 mho/m; ε_r = 37.9; ρ = 1000 kg/m^3
Phantom section: Flat Section

DASY4 Configuration:
- Probe: ET3DV6 - SN1644; ConvF(4.5, 4.5, 4.5); Calibrated: 11/11/2009
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn473; Calibrated: 1/4/2010
- Phantom: SAM 1; Type: SAM 4.0; Serial: 1076
- Measurement SW: DASY4, V4.7 Build 80; Postprocessing SW: SEMCAD, V1.8 Build 186

d=10mm, Pin=1000mW/Zoom Scan (5x5x7) 2 (5x5x7)/Cube 0: Measurement grid: dx=7.5mm, dy=7.5mm, dz=5mm
Reference Value = 185.2 V/m; Power Drift = 0.023 dB
Peak SAR (extrapolated) = 130.5 W/kg
SAR(1 g) = 57.2 mW/g; SAR(10 g) = 26.1 mW/g
Maximum value of SAR (measured) = 63.7 mW/g

d=10mm, Pin=1000mW/Area Scan (31x41x1): Measurement grid: dx=15mm, dy=15mm
Maximum value of SAR (interpolated) = 64.2 mW/g



0 dB = 64.2mW/g