

Medtronic

RADIO REGULATORY COMPLIANCE INFORMATION

FCC Compliance Information:

The following provision applies to the low frequency communications system in the device: This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation. The user is cautioned that changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

Cobalt™ and Crome™ Models
FCC ID: LF5BLEIMPLANT4

The device uses Bluetooth¹ wireless technology utilizing a 2 Mhz bandwidth and capable of transmitting up to 6.5 feet (2 m) in an environment that is conducive for wireless transmission and reception.

¹ The Bluetooth® word mark is a registered trademark of Bluetooth SIG, Inc. and any use of this mark by Medtronic is under license.

Caution: Federal Law (USA) restricts this device to sale by or on the order of a physician.

Medtronic, Inc.

710 Medtronic Parkway
Minneapolis, MN 55432
USA

www.medtronic.com

+1 763 514 4000

Medtronic USA, Inc.

Toll-free in the USA (24-hour technical
consultation for physicians and medical
professionals)

Bradycardia: +1 800 505 4636

Tachycardia: +1 800 723 4636

Technical manuals:

www.medtronic.com/manuals



* M 9 9 5 4 2 8 A 0 0 1 *

© 2010 Medtronic, Inc.

M995428A001 Rev.A

2019-04-30