


Conversion Factor Assessment



Body **835 MHz** $\epsilon_r = 55.2 \pm 5\%$ $\sigma = 0.97 \pm 5\% \text{ mho/m}$

Valid for f=793-877 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C


ConvF X	6.1 $\pm 9.5\%$ (k=2)	Boundary effect:
ConvF Y	6.1 $\pm 9.5\%$ (k=2)	Alpha 0.24
ConvF Z	6.1 $\pm 9.5\%$ (k=2)	Depth 2.00

Body **1900 MHz** $\epsilon_r = 53.3 \pm 5\%$ $\sigma = 1.52 \pm 5\% \text{ mho/m}$

Valid for f=1805-1995 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

ConvF X	4.6 $\pm 9.5\%$ (k=2)	Boundary effect:
ConvF Y	4.6 $\pm 9.5\%$ (k=2)	Alpha 0.24
ConvF Z	4.6 $\pm 9.5\%$ (k=2)	Depth 2.64

Conversion Factor Assessment



Body 900 MHz $\epsilon_r = 55.0 \pm 5\%$ $\sigma = 1.05 \pm 5\%$ mho/m

Valid for f=855-945 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C


ConvF X	6.1 $\pm$ 9.5% (k=2)	Boundary effect:
ConvF Y	6.1 $\pm$ 9.5% (k=2)	Alpha 0.27
ConvF Z	6.1 $\pm$ 9.5% (k=2)	Depth 1.82

Body 1800 MHz $\epsilon_r = 53.3 \pm 5\%$ $\sigma = 1.52 \pm 5\%$ mho/m

Valid for f=1710-1890 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

ConvF X	4.7 $\pm$ 9.5% (k=2)	Boundary effect:
ConvF Y	4.7 $\pm$ 9.5% (k=2)	Alpha 0.23
ConvF Z	4.7 $\pm$ 9.5% (k=2)	Depth 2.99

Conversion Factor Assessment



Head **2450 MHz** $\epsilon_r = 39.2 \pm 5\%$ $\sigma = 1.80 \pm 5\%$ mho/m

Valid for f=2400-2500 MHz with Head Tissue Simulating Liquid according to EN 60381, P1528-200X

ConvF X	4.5 $\pm 9.5\%$ (k=2)	Boundary effect:
ConvF Y	4.5 $\pm 9.5\%$ (k=2)	Alpha 0.40
ConvF Z	4.5 $\pm 9.5\%$ (k=2)	Depth 1.62

Body **2450 MHz** $\epsilon_r = 52.7 \pm 5\%$ $\sigma = 1.95 \pm 5\%$ mho/m

Valid for f=2400-2500 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

ConvF X	4.2 $\pm 9.5\%$ (k=2)	Boundary effect:
ConvF Y	4.2 $\pm 9.5\%$ (k=2)	Alpha 0.32
ConvF Z	4.2 $\pm 9.5\%$ (k=2)	Depth 1.98

Zeughausstrasse 43, 8004 Zurich, Switzerland
Phone +41 1 245 9700, Fax +41 1 245 9779
info@speag.com, http://www.speag.com

Additional Conversion Factors

for Dosimetric E-Field Probe

Type:

ES3DV2

Serial Number:

3019

Place of Assessment:

Zurich

Date of Assessment:

October 13, 2003

Probe Calibration Date:

October 9, 2003

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:



ES3DV2-SN:3019

October 13, 2003

Zeughausstrasse 43, 8004 Zurich, Switzerland
Phone +41 1 245 9700, Fax +41 1 245 9779
info@speag.com, http://www.speag.com

Dosimetric E-Field Probe ES3DV2 SN:3019

Conversion factor ($\pm$ standard deviation)



150 MHz	ConvF	8.7 $\pm$ 8 %	$\epsilon_r = 52.3 \pm 5\%$ $\sigma = 0.76 \pm 5\%$ mho/m (head tissue)
150 MHz	ConvF	8.3 $\pm$ 8 %	$\epsilon_r = 61.9 \pm 5\%$ $\sigma = 0.80 \pm 5\%$ mho/m (body tissue)
450 MHz	ConvF	7.4 $\pm$ 8 %	$\epsilon_r = 43.5 \pm 5\%$ $\sigma = 0.87 \pm 5\%$ mho/m (head tissue)
450 MHz	ConvF	7.3 $\pm$ 8 %	$\epsilon_r = 56.7 \pm 5\%$ $\sigma = 0.94 \pm 5\%$ mho/m (body tissue)

ES3DV2-SN:3019

October 13, 2003



Freq. (MHz)	Amplitude (dB)	Phase (deg)	Rel. Perm	Condy (S/m)	Freq. (MHz)	Rel. Perm.	Condy (S/m)
5725	-53.5	152	51	6.829	400	45.3	0.87
5800	-54	138	50.81	6.891	835	41.5	0.9
5825	-54.75	134	50.65	6.996	900	41.5	0.97
					1450	40.5	1.2
					1800	40	1.4
					2000	40	1.4
					2450	39.2	1.8
					3000	38.5	2.4
					5800	48.2	6
					400	49.83	0.96 10
					835	45.65	0.99 High %
					900	45.65	1.07
					1450	44.55	1.32
					1800	44.00	1.54
					2000	44.00	1.54
					2450	43.12	1.98
					3000	42.35	2.64
					5800	53.02	6.60
					400	40.77	0.78 10
					835	37.35	0.81 Low %
					900	37.35	0.87
					1450	36.45	1.08
					1800	36.00	1.26
					2000	36.00	1.26
					2450	35.28	1.62
					3000	34.65	2.16
					5800	43.38	5.4

0



Freq. (MHz)	Amplitude (dB)	Phase (deg)	Rel. Perm	Condy (S/m)	Freq. (MHz)	Rel. Perm.	Condy (S/m)
5725	-51.65	-20.5	38.7	5.849	400	45.3	0.87
5800	-52.5	-35	38.68	5.959	835	41.5	0.9
5825	-53	-35.6	38.39	6.008	900	41.5	0.97
					1450	40.5	1.2
					1800	40	1.4
					2000	40	1.4
					2450	39.2	1.8
					3000	38.5	2.4
					5800	35.33	5.27
					400	49.83	0.96 10
					835	45.65	0.99 High %
					900	45.65	1.07
					1450	44.55	1.32
					1800	44.00	1.54
					2000	44.00	1.54
					2450	43.12	1.98
					3000	42.35	2.64
					5800	38.86	5.80
					400	40.77	0.78 10
					835	37.35	0.81 Low %
					900	37.35	0.87
					1450	36.45	1.08
					1800	36.00	1.26
					2000	36.00	1.26
					2450	35.28	1.62
					3000	34.65	2.16
					5800	31.797	4.743

5 0



Body 2450 MHz Liquid Measurement, 2004-02-10

frequency	e'	e''	2450 MHz Body Liquid
2400000000.0000	52.6735		13.9488
2402000000.0000	52.5854		14.0026
2404000000.0000	52.7475		13.9783
2406000000.0000	52.8836		14.0266
2408000000.0000	52.9645		14.0183
2410000000.0000	52.8734		14.0039
2412000000.0000	52.7795		14.0342
2414000000.0000	52.7748		14.0026
2416000000.0000	52.8348		14.0219
2418000000.0000	52.8524		14.0258
2420000000.0000	52.8201		14.0268
2422000000.0000	52.7181		14.0445
2424000000.0000	52.5875		14.0311
2426000000.0000	52.4679		14.0684
2428000000.0000	52.4572		14.1060
2430000000.0000	52.4977		14.1055
2432000000.0000	52.4841		14.0912
2434000000.0000	52.5003		14.0501
2436000000.0000	52.6003		14.1069
2438000000.0000	52.6164		14.0984
2440000000.0000	52.6844		14.0881
2442000000.0000	52.7092		14.1037
2444000000.0000	52.6703		14.1109
2446000000.0000	52.5996		14.0723
2448000000.0000	52.5549		14.1160
2450000000.0000	52.5417		14.1503
2452000000.0000	52.5663		14.1605
2454000000.0000	52.6005		14.1704
2456000000.0000	52.6193		14.1507
2458000000.0000	52.5963		14.1875
2460000000.0000	52.5909		14.2037
2462000000.0000	52.6125		14.2219
2464000000.0000	52.6138		14.2225
2466000000.0000	52.6254		14.2317
2468000000.0000	52.6151		14.2436
2470000000.0000	52.5709		14.2483
2472000000.0000	52.5904		14.2328
2474000000.0000	52.6791		14.2458
2476000000.0000	52.7674		14.2655
2478000000.0000	52.8528		14.2699
2480000000.0000	52.7638		14.2545
2482000000.0000	52.7294		14.2813
2484000000.0000	52.6734		14.3042
2486000000.0000	52.6331		14.2941
2488000000.0000	52.6112		14.2960
2490000000.0000	52.6158		14.2949
2492000000.0000	52.6292		14.3181
2494000000.0000	52.6620		14.3215
2496000000.0000	52.7338		14.3311
2498000000.0000	52.7413		14.3290
2500000000.0000	52.7454		14.3182

2/10/2004

1.9286

$$\sigma = \omega \epsilon_0 \epsilon'' = 2 \pi f \epsilon_0 \epsilon'' = 1.9286$$

where $f = 2450 \times 10^6$

$$\epsilon_0 = 8.854 \times 10^{-12}$$

$$\epsilon'' = 14.1503$$

Head 2450 MHz Liquid Measurement, 2004-02-10

NonG
2/10/2004

frequency	e'	e''	2450 MHz head Liquid
2300000000.0000	39.5703		13.1568
2304166666.6667	39.5415		13.1731
2308333333.3333	39.4667		13.1753
2312500000.0000	39.4315		13.2026
2316666666.6667	39.4022		13.2276
2320833333.3333	39.3488		13.2427
2325000000.0000	39.3614		13.2583
2329166666.6667	39.3418		13.2726
2333333333.3333	39.3485		13.3395
2337500000.0000	39.3775		13.3799
2341666666.6667	39.3961		13.4183
2345833333.3333	39.3991		13.4666
2350000000.0000	39.4172		13.4830
2354166666.6667	39.4612		13.5303
2358333333.3333	39.4690		13.5563
2362500000.0000	39.4958		13.5823
2366666666.6667	39.4769		13.6115
2370833333.3333	39.4918		13.6205
2375000000.0000	39.5275		13.6372
2379166666.6667	39.5328		13.6540
2383333333.3333	39.5053		13.6462
2387500000.0000	39.5291		13.6646
2391666666.6667	39.5017		13.6824
2395833333.3333	39.4912		13.6617
2400000000.0000	39.4149		13.6865
2404166666.6667	39.3328		13.6458
2408333333.3333	39.2857		13.6250
2412500000.0000	39.2926		13.6664
2416666666.6667	39.1856		13.5818
2420833333.3333	39.1275		13.6054
2425000000.0000	39.1302		13.6842
2429166666.6667	39.0423		13.6252
2433333333.3333	38.9632		13.5945
2437500000.0000	39.0356		13.7377
2441666666.6667	39.0281		13.7639
2445833333.3333	39.0185		13.7910
2450000000.0000	39.0062		13.8263
2454166666.6667	39.0048		13.8700
2458333333.3333	38.9243		13.8131
2462500000.0000	38.9485		13.8584
2466666666.6667	39.0303		13.9177
2470833333.3333	39.0456		13.9368
2475000000.0000	39.1292		14.0190
2479166666.6667	39.0972		14.0417
2483333333.3333	39.0786		14.1126
2487500000.0000	39.1383		14.1530
2491666666.6667	39.2013		14.1851
2495833333.3333	39.1794		14.2151
2500000000.0000	39.1380		14.2148

1.8845

$$\sigma = \omega \epsilon_0 \epsilon'' = 2 \pi f \epsilon_0 \epsilon'' = 1.8845$$

where $f = 2450 \times 10^6$
 $\epsilon_0 = 8.854 \times 10^{-12}$
 $\epsilon'' = 13.8263$

3 - EUT SUMMARY

Applicant:	AMBIT Microsystems Corporation
Product Description:	Wireless MinPCI Card
Model Name:	AIR-MPI350
FCC ID:	MCLAIRMPI350DE
Serial Number:	000000001SC/00D059C99003
Transmitter Frequency:	2412-2462 MHz
Maximum Output Power:	20.67 dBm (117 mW)
RF Exposure environment:	General Population/Uncontrolled
Power Supply:	DELL AC Adapter, M/N: AA22850
Applicable Standard	FCC CFR 47, Part 15 Subpart C
Application Type:	Certification

¹Specific Absorption Rate (SAR) is a measure of the rate of energy absorption due to exposure to an RF transmitting source (wireless portable device).

²IEEE/ANSI Std. C95.1-1992 limits are used to determine compliance with FCC ET Docket 93-62.

Note: The test data was good for test sample only. It may have deviation for other test samples.

** The test data gathered are from production sample, serial number:MPI350-001, provided by the manuafactuer.*

4 - SYSTEM TEST CONFIGURATION

4.1 Justification

The system was configured for testing in a typical fashion (as normally used by a typical user).

4.2 EUT Exercise Software and Procedure

The EUT exercising program used during SAR testing was designed to exercise the various system components in a manner similar to a typical use. The software, PRISM utilities, contained on the hard drive, is auto starting on power-up. Once loaded, the program sequentially exercises each system component.

The testing procedure is as follows:

1. Click PRISM test utilities on Window
2. Select wireless LAN Adapter under adapters list
3. Select low, mid and high channels under Radio Channels
4. Select Tx Rate of 11MB
5. Click on “continuous Tx” bottom

4.3 Special Accessories

All interface cables used for compliance testing are shielded as normally supplied by INMAC, Monster Cable and their respective support equipment manufacturer. The EUT is featured shielded metal connectors.


4.4 Equipment Modifications

No modification(s) were made to ensure that the EUT complies with the applicable limits.

5 - CONDUCTED OUTPUT POWER MEASUREMENT

5.1 Measurement Procedure

1. Place the EUT on a bench and set it in transmitting mode.
2. Remove the antenna from the EUT and then connect a low loss RF cable from the antenna port to a spectrum analyzer.
3. Add a correction factor to the display.



5.2 Test Results

HFT06

Antenna	Frequency (MHz)	Peak Output Power (dBm)	Correction Factor (dBm)	Corrected Factor (dBm)	Output Power (W)	Standard (W)	Result
Main	2462	12.50	8.0	20.50	0.112	≤ 1 W	Compliant
	2437	12.67	8.0	20.67	0.117	≤ 1 W	Compliant
	2412	12.50	8.0	20.50	0.112	≤ 1 W	Compliant
Auxiliary	2463	12.50	8.0	20.50	0.112	≤ 1 W	Compliant
	2437	12.67	8.0	20.67	0.117	≤ 1 W	Compliant
	2412	12.33	8.0	20.33	0.108	≤ 1 W	Compliant

CAQ-S




Antenna	Frequency (MHz)	Peak Output Power (dBm)	Correction Factor (dBm)	Corrected Factor (dBm)	Output Power (W)	Standard (W)	Result
Main	2437	12.67	8.0	20.67	0.117	≤ 1 W	Compliant
Auxiliary	2437	12.50	8.0	20.50	0.112	≤ 1 W	Compliant




$$\text{*Correction Factor} = 10 \log \frac{6\text{dB Bandwidth}}{\text{RBW}} = 10 \log \frac{13}{2} \approx 8.0$$

5.3 Measurement Plots



Please refer to the plots hereinafter.

HFT06





CAQ-S



6 - DOSIMETRIC ASSESSMENT SETUP


These measurements were performed with the automated near-field scanning system DASY3 from Schmid & Partner Engineering AG (SPEAG). The system is based on a high precision robot (working range greater than 0.9m) which positions the probes with a positional repeatability of better than $\pm 0.02\text{mm}$. Special E- and H-field probes have been developed for measurements close to material discontinuity, the sensors of which are directly loaded with a Schottky diode and connected via highly resistive lines to the data acquisition unit. The system is described in detail in [3].

The SAR measurements were conducted with the dosimetric probe ET3DV6 SN: 1604 (manufactured by SPEAG), designed in the classical triangular configuration [3] and optimized for dosimetric evaluation. The probe has been calibrated according to the procedure described in [7] with accuracy of better than $\pm 10\%$. The spherical isotropy was evaluated with the procedure described in [8] and found to be better than $\pm 0.25\text{dB}$.

The phantom used was the "Generic Twin Phantom" described in [4]. The ear was simulated as a spacer of 4 mm thickness between the earpiece of the phone and the tissue simulating liquid. The Tissue simulation liquid used for each test is in according with the FCC OET65 supplement C as listed below.

Ingredients (% by weight)	Frequency (MHz)									
	450		835		915		1900		2450	
Tissue Type	Head	Body	Head	Body	Head	Body	Head	Body	Head	Body
Water	38.56	51.16	41.45	52.4	41.05	56.0	54.9	40.4	62.7	73.2
Salt (NaCl)	3.95	1.49	1.45	1.4	1.35	0.76	0.18	0.5	0.5	0.04
Sugar	56.32	46.78	56.0	45.0	56.5	41.76	0.0	58.0	0.0	0.0
HEC	0.98	0.52	1.0	1.0	1.0	1.21	0.0	1.0	0.0	0.0
Bactericide	0.19	0.05	0.1	0.1	0.1	0.27	0.0	0.1	0.0	0.0
Triton x-100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	36.8	0.0
DGBE	0.0	0.0	0.0	0.0	0.0	0.0	44.92	0.0	0.0	26.7
Dielectric Constant	43.42	58.0	42.54	56.1	42.0	56.8	39.9	54.0	39.2	52.7
Conductivity (s/m)	0.85	0.83	0.91	0.95	1.0	1.07	1.42	1.45	1.80	1.91

6.1 Measurement System Diagram




The DAS3 system for performing compliance tests consist of the following items:

1. A standard high precision 6-axis robot (Stäubli RX family) with controller and software.
2. An arm extension for accommodating the data acquisition electronics (DAE).
3. A dosimetric probe, i.e., an isotropic E-field probe optimized and calibrated for usage in tissue simulating liquid. The probe is equipped with an optical surface detector system.
4. A data acquisition electronic (DAE), which performs the signal amplification, signal multiplexing, AD-conversion, offset measurements, mechanical surface detection, collision detection, etc. The unit is battery powered with standard or rechargeable batteries. The signal is optically transmitted to the EOC.
5. A unit to operate the optical surface detector, which is connected to the EOC. The Electro-optical coupler (EOC) performs the conversion from the optical into a digital electric signal of the DAE. The EOC is connected to the PC plug-in card. The functions of the PC plug-in card based on a DSP is to perform the time critical task such as signal filtering, surveillance of the robot operation fast movement interrupts.
6. A computer operating Windows 95 or larger
7. DAS3 software
8. Remote control with teaches pendant and additional circuitry for robot safety such as warning lamps, etc.
9. The generic twin phantom enabling testing left-hand and right-hand usage.
10. The device holder for handheld EUT.
11. Tissue simulating liquid mixed according to the given recipes (see Application Note).
12. System validation dipoles to validate the proper functioning of the system.


6.2 System Components

ET3DV6 Probe Specification

Construction Symmetrical design with triangular core
Built-in optical fiber for surface detection System
Built-in shielding against static charges
Calibration In air from 10 MHz to 2.5 GHz
In brain and muscle simulating tissue at
Frequencies of 450 MHz, 900 MHz and
1.8 GHz (accuracy $\pm 8\%$)
Frequency 10 MHz to > 6 GHz; Linearity: ± 0.2 dB
(30 MHz to 3 GHz)
Directivity ± 0.2 dB in brain tissue (rotation around
probe axis)
 ± 0.4 dB in brain tissue (rotation normal probe axis)
Dynamic 5 mW/g to > 100 mW/g;
Range Linearity: ± 0.2 dB
Surface ± 0.2 mm repeatability in air and clear liquids
Detection over diffuse reflecting surfaces.
Dimensions Overall length: 330 mm
Tip length: 16 mm
Body diameter: 12 mm
Tip diameter: 6.8 mm
Distance from probe tip to dipole centers: 2.7 mm
Application General dosimetric up to 3 GHz
Compliance tests of mobile phones
Fast automatic scanning in arbitrary phantoms



Photograph of the probe



Inside view of
ET3DV6 E-field Probe

The SAR measurements were conducted with the dosimetric probe ET3DV6 designed in the classical triangular configuration and optimized for dosimetric evaluation. The probe is constructed using the thick film technique; with printed resistive lines on ceramic substrates. The probe is equipped with an optical multi-fiber line ending at the front of the probe tip. It is connected to the EOC box on the robot arm and provides an automatic detection of the phantom surface. Half of the fibers are connected to a pulsed infrared transmitter, the other half to a synchronized receiver. As the probe approaches the surface, the reflection from the surface produces a coupling from the transmitting to the receiving fibers. This reflection increases first during the approach, reaches maximum and then decreases. If the probe is flatly touching the surface, the coupling is zero. The distance of the coupling maximum to the surface is independent of the surface reflectivity and largely independent of the surface to probe angle. The DASY3 software reads the reflection during a software approach and looks for the maximum using a 2nd order fitting. The approach is stopped when reaching the maximum.

E-Field Probe Calibration Process

Each probe is calibrated according to a dosimetric assessment procedure described in [6] with accuracy better than +/- 10%. The spherical isotropy was evaluated with the procedure described in [7] and found to be better than +/-0.25dB. The sensitivity parameters (NormX, NormY, NormZ), the diode compression parameter (DCP) and the conversion factor (ConvF) of the probe are tested.

The free space E-field from amplified probe outputs is determined in a test chamber. This is performed in a TEM cell for frequencies below 1 GHz, and in a waveguide above 1 GHz for free space. For the free space calibration, the probe is placed in the volumetric center of the cavity and at the proper orientation with the field. The probe is then rotated 360 degrees.

E-field temperature correlation calibration is performed in a flat phantom filled with the appropriate simulated brain tissue. The measured free space E-field in the medium correlates to temperature rise in dielectric medium. For temperature correlation calibration a RF transparent thermistor-based temperature probe is used in conjunction with the E-field probe.

Data Evaluation

The DASY3 software automatically executes the following procedures to calculate the field units from the microvolt readings at the probe connector. The parameters used in the evaluation are stored in the configuration modules of the software:

Probe Parameter:	-Sensitivity	Norm _i , a _{i0} , a _{i1} , a _{i2}
	-Conversion Factor	ConvFi
	-Diode compression point	Dcp _i
Device parameter:	-Frequency	f
	-Crest Factor	cf
Media parameter:	-Conductivity	σ
	-Density	ρ

These parameters must be set correctly in the software. They can either be found in the component documents or be imported into the software from the configuration files issued for the DASY3 components. In the direct measuring mode of the multi-meter option, the parameters of the actual system setup are used. In the scan visualization and export modes, the parameters stored in the corresponding document files are used.

The first step of the evaluation is a linearization of the filtered input signal to account for the compression characteristics of the detector diode. The compensation depends on the input signal, the diode type and the DC-transmission factor from the diode to the evaluation electronics. If the exciting field is pulsed, the crest factor of the signal must be known to correctly compensate for peak power. The formula for each channel can be given as:

$$V_i = U_i + (U_i)^2 \text{ cf} / \text{dcp}_i$$

With V_i = compensated signal of channel i (i=x, y, z)
 U_i = input signal of channel i (i=x, y, z)
 cf = crest factor of exciting field (DASY parameter)
 dcp_i = diode compression point (DASY parameter)

From the compensated input signals the primary field data for each channel can be evaluated:

$$\begin{aligned} \text{E-field probes:} \quad E_i &= \sqrt{\frac{V_i}{\text{Norm}_i \cdot \text{ConvF}}} \\ \text{H-field probes:} \quad H_i &= \sqrt{V_i} \cdot \frac{a_{i0} + a_{i1}f + a_{i2}f^2}{f} \end{aligned}$$

With V_i = compensated signal of channel i (i=x, y, z)
 Norm_i = sensor sensitivity of channel i (i=x, y, z)
 $\mu\text{V}/(\text{V/m})^2$ for E-field probes
 ConvF = sensitivity enhancement in solution
 a_{ij} = sensor sensitivity factors for H-field probes
 f = carrier frequency [GHz]
 E_i = electric field strength of channel i in V/m
 H_i = diode compression point (DASY parameter)

The RSS value of the field components gives the total field strength (Hermitian magnitude):

$$E_{\text{tot}} = \text{Square Root} [(E_x)^2 + (E_y)^2 + (E_z)^2]$$

The primary field data are used to calculate the derived field units.

$$\text{SAR} = (E_{\text{tot}})^2 \cdot \sigma / (\rho \cdot 1000)$$

With SAR = local specific absorption rate in mW/g
 E_{tot} = total field strength in V/m
 σ = conductivity in [mho/m] or [Siemens/m]
 ρ = equivalent tissue density in g/cm^3

Note that the density is normally set to 1 (or 1.06), to account for actual brain density rather than the density of the simulation liquid.

The power flow density is calculated assuming the excitation field as a free space field.

$$P_{\text{pwe}} = (E_{\text{tot}})^2 / 3770 \text{ or } P_{\text{pwe}} = (H_{\text{tot}})^2 \cdot 37.7$$

With P_{pwe} = equivalent power density of a plane wave in mW/cm³
 E_{tot} = total electric field strength in V/m
 H_{tot} = total magnetic field strength in V/m


Generic Twin Phantom

The Generic Twin Phantom is constructed of a fiberglass shell integrated in a wooden table. The shape of the shell is based on data from an anatomical study designed to determine the maximum exposure in at least 90% of all users [9][10]. It enables the dosimetric evaluation of left and right hand phone usage as well as body mounted usage at the flat phantom region. A cover prevents the evaporation of the liquid. Reference markings on the Phantom allows the complete setup of all predefined phantom positions and measurement grids by manually teaching three points in the robot.

Shell Thickness 2 ± 0.1 mm

Filling Volume Approx. 20 liters

Dimensions 810 x 1000 x 500 mm (H x L x W)




Generic Twin Phantom

Device Holder

In combination with the Generic Twin Phantom V3.0, the Mounting Device enables the rotation of the mounted transmitter in spherical coordinates whereby the rotation points is the ear opening. The devices can be easily, accurately, and repeatedly positioned according to the FCC and CENELEC specifications. The device holder can be locked at different phantom locations (left head, right head, flat phantom).

* Note: A simulating human hand is not used due to the complex anatomical and geometrical structure of the hand that may produced infinite number of configurations [10]. To produce the worst-case condition (the hand absorbs antenna output power), the hand is omitted during the tests.



Device Holder