

Measurement Report for Device

Device Under Test Properties

Model, Manufacturer	Dimensions [mm]	Software Version	DUT Type
Device,	100.0 x 100.0 x 172.0	3.2.0.1840	5G Verification Source

Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Frequency [MHz]	Conversion Factor
5G	FRONT, 10.00	10000.0	1.0

Hardware Setup

Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave - 1109	Air -	EUmmWV4 - SN9461_F1-55GHz, 2024-10-16	DAE4 Sn1805, 2024-05-22

Scans Setup

Scan Type	5G Scan
Grid Extents [mm]	120.0 x 120.0
Grid Steps [lambda]	0.25 x 0.25
Sensor Surface [mm]	10.0

Measurement Results

Date	2024-12-26
Avg. Area [cm ²]	4.00
psPDn+ [W/m ²]	59.7
psPDtot+ [W/m ²]	59.9
H _{max} [A/m]	0.415
E _{max} [V/m]	161
max _(Stot) [W/m ²]	65.9
Power Drift [dB]	-0.03
IPDn	107

