

1900MHz Brain Dipole Validation

Generic Twin Phantom; Flat Section; Probe: ET3DV5 - SN1370 -- Probe Cal Date 02/00


Med. Parameters 1900 MHz Brain: $\sigma = 1.82$ mho/m $\epsilon_r = 40.4$ $\rho = 1.00$ g/cm³; Antenna Position -- Out; Crest Factor 1.0

SAR (1g): 11.0 mW/g, SAR (10g): 5.50 mW/g

1900MHz Brain Dipole Validation (D1900V2 S/N: 502)

Frequency: 1900 MHz; Antenna Input Power: 250 [mW]

PCTEST Brain Tissue Simulating Liquid



SAR_{Tot} [mW/g]

