

La Crosse Technology Ltd.

Date: 2019.7.31

To: Federal Communications Commission
7435 Oakland Mills Road
Columbia, MD

FCC ID: OMOC80765

To Whom It May Concern:

This letter is to ascertain that La Crosse Technology Ltd. Product PROJECTION ALARM CLOCK OMOC8076, has been the units used for conducting FCC compliance testing, and it meets KDB 680106 D01, Clause 5(b) all 6 conditions as stated below hence PAG is not required.

1	Power transfer frequency is less than 1 MHz
2	Output power from each primary coil is less than or equal to 15 watts
3	The transfer system includes only single primary and secondary coils. This includes charging systems that may have multiple primary coils and clients that are able to detect and allow coupling only between individual pairs of coils
4	Client device is placed directly in contact with the transmitter.
5	Mobile exposure conditions only (portable exposure conditions are not covered by this exclusion).
6	The aggregate H-field strengths at 15 cm surrounding the device and 20 cm above the top surface from all simultaneous transmitting coils are demonstrated to be less than 50% of the MPE limit.

If you have any question or concerns, pls. contact us.

Sincerely,

Client's signature



Name: Allan McCormick

Address: 2809 Losey Blvd. South La Crosse Wisconsin 54601 United States