

Supplement to the CARESCAPE ONE User Manual


2108306-001 paper

1st edition

2018-01-16

© 2018 General Electric Company. All rights reserved.

All specifications are subject to change without notice.



GE Medical Systems *Information Technologies*, Inc.

8200 West Tower Avenue

Milwaukee, WI 53223 USA

Tel: + 1 414 355 5000

1 800 558 5120 (US only)

www.gehealthcare.com

GE Medical Systems *Information Technologies*, Inc.

Munzingerstasse 5

79111 Freiburg

Germany

Tel: + 49 761 45 43 - 0

Intended use of this supplement

This document is intended to provide supplemental information about wireless LAN as specified by the Federal Communications Commission (FCC) and Innovation, Science and Economic Development (ISED) Canada, formerly Industry Canada (IC).

NOTE

Although the CARESCAPE ONE contains wireless LAN components, there is no wireless LAN functionality in the current version.

FCC information

Contains FCC ID: OU5-CSONE01

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

CAUTION

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment. This transmitter must not be co-located or operated in conjunction with any other antenna or transmitter.

CAUTION

To comply with the FCC RF exposure requirements, the device with wireless LAN must be operated with a separation distance of 20 cm or more from a person's body.

IC information

Contains IC: 4048B-CSONE01

This device complies with Industry Canada's license-exempt RSSs. Operation is subject to the following two conditions:

1. This device may not cause interference; and
2. This device must accept any interference, including interference that may cause undesired operation of the device.

Le présent appareil est conforme aux CNR d'Industrie Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes :

1. l'appareil ne doit pas produire de brouillage;
2. l'utilisateur de l'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

5150-5350 MHz band is restricted to indoor operation only.

La bande 5150-5350 MHz est restreinte à une utilisation à l'intérieur seulement.

High-power radars are allocated as primary users (i.e. priority users) for the bands 5250-5350 MHz and 5650-5860 MHz and that these radars could cause interference and/or damage to LE-LAN devices.

Les radars de haute puissance sont désignés utilisateurs principaux (c.-à-d., qu'ils ont la priorité) pour les bandes 5250-5350 MHz et 5650-5860 MHz, et ces radars pourraient causer du brouillage et/ou des dommages aux dispositifs LAN-EL.

This equipment complies with IC radiation exposure limits set forth for an uncontrolled environment and meets RSS-102 of the IC radio frequency (RF) Exposure rules. This equipment should be installed and operated keeping the radiator at least 20 cm or more away from a person's body.

Cet équipement est conforme aux limites d'exposition aux rayonnements énoncées pour un environnement non contrôlé et respecte les règles d'exposition aux fréquences radioélectriques (RF) CNR-102 de l'IC. Cet équipement doit être installé et utilisé en gardant une distance de 20 cm ou plus entre le radiateur et le corps humain.