

FCC - TEST REPORT

Report Number	: 68.930.19.0006.01	Date of Issue: <u>June 27, 2019</u>
Model	: TMB-1872-B	
Product Type	: Blood Pressure Monitor	
Applicant	: Guangdong Transtek Medical Electronics Co., Ltd	
Address	: Zone A, No.105, Dongli Road Torch Development District	
	: Zhongshan, Guangdong China 528437	
Factory	: Guangdong Transtek Medical Electronics Co., Ltd	
Address	: Zone B, No.105, Dongli Road Torch Development District	
	: Zhongshan, Guangdong China 528437	

Test Result	: n Positive o Negative
-------------	-------------------------------------

Total pages including Appendices	: 34
----------------------------------	-------------

TÜV SÜD Certification and Testing (China) Co., Ltd. Shenzhen Branch is a subcontractor to TÜV SÜD Product Service GmbH according to the principles outlined in ISO 17025.

TÜV SÜD Certification and Testing (China) Co., Ltd. Shenzhen Branch reports apply only to the specific samples tested under stated test conditions. Construction of the actual test samples has been documented. It is the manufacturer's responsibility to assure that additional production units of this model are manufactured with identical electrical and mechanical components. The manufacturer/importer is responsible to the Competent Authorities in Europe for any modifications made to the production units which result in non-compliance to the relevant regulations. TÜV SÜD Certification and Testing (China) Co., Ltd. Shenzhen Branch shall have no liability for any deductions, inferences or generalizations drawn by the client or others from TÜV SÜD Certification and Testing (China) Co., Ltd. Shenzhen Branch issued reports.

This report is the confidential property of the client. As a mutual protection to our clients, the public and ourselves, extracts from the test report shall not be reproduced except in full without our written approval.

1 Table of Contents

1	Table of Contents.....	2
2	Details about the Test Laboratory	3
3	Description of the Equipment Under Test	4
4	Summary of Test Standards	5
5	Summary of Test Results	6
6	General Remarks.....	7
7	Test Setups	8
8	Systems test configuration.....	9
9	Technical Requirement.....	10
9.1	Conducted Emission	10
9.2	Conducted peak output power.....	13
9.3	Power spectral density	16
9.4	6 dB Bandwidth and 99% Occupied Bandwidth.....	19
9.5	Spurious RF conducted emissions.....	23
9.6	Band edge	27
9.7	Spurious radiated emissions for transmitter	29
10	Test Equipment List.....	33
11	System Measurement Uncertainty	34



2 Details about the Test Laboratory

Details about the Test Laboratory

Test Site 1

Company name: TÜV SÜD Certification and Testing (China) Co., Ltd. Shenzhen Branch
Building 12 & 13, Zhiheng Wisdomland Business Park, Nantou Checkpoint
Road 2, Nanshan District
Shenzhen 518052
P.R. China

Telephone: 86 755 8828 6998

Fax: 86 755 8288 5299

FCC Registration No.: 514049

3 Description of the Equipment Under Test

Product:	Blood Pressure Monitor
Model no.:	TMB-1872-B
FCC ID:	OU9TMB1872BS
Options and accessories:	USB Cable, Adapter
Rating:	6VDC, (Supplied by 4×1.5V AAA batteries) 5VDC, 1000mA (Charged by AC/AD adapter)
Adapter	Model: BLJ06L050100U-U Input: 100-240V~50/60Hz 0.2A max Output: 5V/1000mA
RF Transmission Frequency:	2402MHz-2480MHz
No. of Operated Channel:	40
Modulation:	GFSK
Antenna Type:	Integrated Antenna
Antenna Gain:	0dBi
Description of the EUT:	The Equipment Under Test (EUT) is Blood Pressure Monitor operated at 2.4GHz.

4 Summary of Test Standards

Test Standards	
FCC Part 15 Subpart C 10-1-2017 Edition	PART 15 - RADIO FREQUENCY DEVICES Subpart C - Intentional Radiators

All the test methods were according to KDB558074 D01 v05r02 DTS Measurement Guidance and ANSI C63.10 (2013).

5 Summary of Test Results

Technical Requirements						
FCC Part 15 Subpart C						
Test Condition		Pages	Test Site	Test Result		
				Pass	Fail	N/A
§15.207	Conducted emission AC power port	10	Site 1	☒	☐	☐
§15.247 (b) (1)	Conducted peak output power	13	Site 1	☒	☐	☐
§15.247(a)(1)	20dB bandwidth	---	---	☐	☐	☒
§15.247(a)(1)	Carrier frequency separation	---	---	☐	☐	☒
§15.247(a)(1)(iii)	Number of hopping frequencies	---	---	☐	☐	☒
§15.247(a)(1)(iii)	Dwell Time	---	---	☐	☐	☒
§15.247(a)(2)	6dB bandwidth and 99% Occupied Bandwidth	16	Site 1	☒	☐	☐
§15.247(e)	Power spectral density	19	Site 1	☒	☐	☐
§15.247(d)	Spurious RF conducted emissions	23	Site 1	☒	☐	☐
§15.247(d)	Band edge	27	Site 1	☒	☐	☐
§15.247(d) & §15.209 & §15.205	Spurious radiated emissions for transmitter	29	Site 1	☒	☐	☐
§15.203	Antenna requirement	See note 2		☒	☐	☐

Note 1: N/A=Not Applicable.

Note 2: The EUT uses a Integrated antenna, which gain is 0dBi. In accordance to §15.203, it is considered sufficiently to comply with the provisions of this section.



6 General Remarks

Remarks

This submittal(s) (test report) is intended for FCC ID: OU9TMB1872BS complies with Section 15.205, 15.207 15.209, 15.247 of the FCC Part 15, Subpart C.

SUMMARY:

All tests according to the regulations cited on page 5 were

n - Performed

o - **Not** Performed

The Equipment under Test

n - **Fulfills** the general approval requirements.

o - **Does not** fulfill the general approval requirements.

Sample Received Date: February 21, 2019

Testing Start Date: February 21, 2019

Testing End Date: April 22, 2019

- TÜV SÜD Certification and Testing (China) Co., Ltd. Shenzhen Branch -

Reviewed by:

Prepared by:

Tested by:


Laurent Yuan
EMC Project Manager

Mark Chen
EMC Project Engineer


Carry Cai
EMC Test Engineer

7 Test Setups


Below 1GHz




Above 1GHz



Conducted RF test setups



AC Power Line Conducted Emission test setups



8 Systems test configuration

Auxiliary Equipment Used during Test:

DESCRIPTION	MANUFACTURER	MODEL NO.(SHIELD)	S/N(LENGTH)
Notebook	Lenovo	X220	---
---	---	---	---

Test software: EMI_Test_Tool_v1.3.exe test tool, which used to control the EUT in continues transmitting mode.

The system was configured to channel 0, 19, and 39 for the test.

9 Technical Requirement

9.1 Conducted Emission

Test Method


1. The EUT was placed on a table, which is 0.8m above ground plane
2. The power line of the EUT is connected to the AC mains through a Artificial Mains Network (A.M.N.).
3. Maximum procedure was performed to ensure EUT compliance
4. A EMI test receiver is used to test the emissions from both sides of AC line

Limit

Frequency MHz	QP Limit dB μ V	AV Limit dB μ V
0.150-0.500	66-56*	56-46*
0.500-5	56	46
5-30	60	50

*Decreases with the logarithm of the frequency.

Product Type : Blood Pressure Monitor
 M/N : TMB-1872-B
 Operating Condition : AC Powered(Adapter)+ Measuring+ BT Link
 Test Specification : Line
 Comment : 5VDC(Supplies by AC/AD Adapter)



Critical_Freqs

Frequency (MHz)	MaxPeak* (dBμV)	Average (dBμV)	Limit (dBμV)	Margin (dB)	Line	Corr.** (dB/m)
0.170000	39.58	---	64.96	25.39	L1	10.2
0.334000	33.15	---	59.35	26.20	L1	10.2
0.434000	31.29	---	57.18	25.89	L1	10.3
1.274000	21.27	---	56.00	34.73	L1	10.3
2.690000	20.58	---	56.00	35.42	L1	10.3
15.458000	28.96	---	60.00	31.04	L1	10.8


Remark

*Level=Reading Level + Correction Factor

**Correction Factor=Cable Loss + LISN Factor

(The Reading Level is recorded by software which is not shown in the sheet)

Product Type : Blood Pressure Monitor
 M/N : TMB-1872-B
 Operating Condition : AC Powered(Adapter)+ Measuring+ BT Link
 Test Specification : Neutral
 Comment : 5VDC(Supplies by AC/AD Adapter)



Critical_Freqs

Frequency (MHz)	MaxPeak* (dBμV)	Average (dBμV)	Limit (dBμV)	Margin (dB)	Line	Corr.** (dB/m)
0.162000	39.91	---	65.36	25.45	N	10.2
0.318000	33.70	---	59.76	26.06	N	10.3
0.438000	31.25	---	57.10	25.85	N	10.3
0.794000	25.95	---	56.00	30.05	N	10.3
2.014000	27.02	---	56.00	28.98	N	10.3
3.934000	25.63	---	56.00	30.37	N	10.4

Remark

*Level=Reading Level + Correction Factor

**Correction Factor=Cable Loss + LISN Factor

(The Reading Level is recorded by software which is not shown in the sheet)

9.2 Conducted peak output power

Test Method

1. Use the following spectrum analyzer settings:
RBW > the 6dB bandwidth of the emission being measured, VBW $\geq$ 3RBW, Span $\geq$ 3RBW
Sweep = auto, Detector function = peak, Trace = max hold.
2. Add a correction factor to the display.
3. Allow the trace to stabilize. Use the marker-to-peak function to set the marker to the peak of the emission. The indicated level is the peak output power.

Limits

According to §15.247 (b) (1), conducted peak output power limit as below:


Frequency Range MHz	Limit W	Limit dBm
2400-2483.5	≤ 1	≤ 30

Test result as below table

Frequency MHz	Conducted Peak Output Power dBm	Result
Low channel 2402MHz	4.70	Pass
Middle channel 2440MHz	4.29	Pass
High channel 2480MHz	3.79	Pass




Low channel 2402MHz



Date: 14 APR 2019 16:40:15


Middle channel 2440MHz



Date: 14 APR 2019 16:42:05



High channel 2480MHz



Date: 14 APR 2019 16:43:59

9.3 Power spectral density

Test Method

This procedure shall be used if maximum peak conducted output power was used to demonstrate compliance:

1. Set analyzer center frequency to DTS channel center frequency. RBW=10kHz, VBW $\geq$ 3RBW, Span=1.5 times DTS bandwidth, Detector=Peak, Sweep=auto, Trace= max hold.
2. Allow trace to fully stabilize, use the peak marker function to determine the maximum amplitude level within the RBW.
3. Repeat above procedures until other frequencies measured were completed.

Limit

Limit [dBm/3kHz]


≤ 8

Test result

Frequency MHz	Power spectral density dBm/10kHz	Result
Top channel 2402MHz	-8.01	Pass
Middle channel 2440MHz	-8.51	Pass
Bottom channel 2480MHz	-8.92	Pass




Low channel 2402MHz



Date: 14 APR 2019 16:40:21


Middle channel 2440MHz



Date: 14 APR 2019 16:42:11



High channel 2480MHz



Date: 14 APR 2019 16:44:05

9.4 6 dB Bandwidth and 99% Occupied Bandwidth

Test Method

1. Use the following spectrum analyzer settings:
RBW=100K, VBW $\geq$ 3RBW, Sweep = auto, Detector function = peak, Trace = max hold
2. Use the automatic bandwidth measurement capability of an instrument, may be employed using the X dB bandwidth mode with X set to 6 dB, care shall be taken so that the bandwidth measurement is not influenced by any intermediate power nulls in the fundamental emission that might be $\geq$ 6 dB.
3. Allow the trace to stabilize, record the X dB Bandwidth value.

Limit

Limit [kHz]

≥ 500


Test result

Frequency MHz	6dB bandwidth kHz	99 bandwidth kHz	Result
Bottom channel 2402MHz	1340	2398	Pass
Middle channel 2440MHz	1520	2422	Pass
Top channel 2480MHz	1416	2338	Pass




6 dB Bandwidth

Low channel 2402MHz




Date: 14 APR 2019 16:40:08




Date: 14 APR 2019 16:39:57



Middle channel 2440MHz




Date: 14 APR 2019 16:41:58




Date: 14 APR 2019 16:41:47



High channel 2480MHz



Date: 14 APR 2019 16:43:52



Date: 14 APR 2019 16:43:40

9.5 Spurious RF conducted emissions

Test Method

1. Establish a reference level by using the following procedure:
 - a. Set RBW=100 kHz. VBW $\geq$ 3RBW. Detector =peak, Sweep time = auto couple, Trace mode = max hold.
 - b. Allow trace to fully stabilize, use the peak marker function to determine the maximum PSD level.
2. Use the maximum PSD level to establish the reference level.
 - a. Set the center frequency and span to encompass frequency range to be measured.
 - b. Use the peak marker function to determine the maximum amplitude level. Ensure that the amplitude of all unwanted emissions outside of the authorized frequency band (excluding restricted frequency bands) are attenuated by at least the minimum requirements, report the three highest emissions relative to the limit.
3. Repeat above procedures until other frequencies measured were completed.



Limit

Frequency Range MHz	Limit (dBc)
30-25000	-20




Spurious RF conducted emissions

2402MHz






2440MHz




Date: 14 APR 2019 16:42:26




Date: 14 APR 2019 16:42:38



2480MHz



Date: 14 APR 2019 16:44:30



Date: 14 APR 2019 16:44:41

9.6 Band edge

Test Method

- 1 Use the following spectrum analyzer settings:
Span = wide enough to capture the peak level of the in-band emission and all spurious
RBW = 100 kHz, VBW $\geq$ RBW, Sweep = auto, Detector function = peak, Trace = max hold.
- 2 Allow the trace to stabilize, use the peak and delta measurement to record the result.
- 3 The level displayed must comply with the limit specified in this Section.


Limit

Frequency Range MHz	Limit (dBc)
30-25000	-20




Band edge testing

2402MHz



Date:14 APR 2019 16:40:31

2480MHz



Date:14 APR 2019 16:44:15

9.7 Spurious radiated emissions for transmitter

Test Method

- 1: The EUT was placed on a turn table which is 1.5m above ground plane for above 1GHz and 0.8m above ground for below 1GHz at 3 meter chamber room for test. The table was rotated 360 degrees to determine the position of the highest radiation.
- 2: The EUT was set 3 meters away from the interference – receiving antenna, which was mounted on the top of a variable – height antenna tower.
- 3: The height of antenna is varied from one meter to four meters above the ground to determine the maximum value of the field strength. Both horizontal and vertical polarizations of the antenna are set to make the measurement.
- 4: For each suspected emission, the EUT was arranged to its worst case and then the antenna was tuned to heights from 1 meter to 4 meters and the rotatable table was turned from 0 degrees to 360 degrees to find the maximum reading.
- 5: Use the following spectrum analyzer settings According to C63.10:

For Above 1GHz

Span = wide enough to capture the peak level of the in-band emission and all spurious
RBW = 1MHz, VBW ≥ RBW for peak measurement and VBW = 10Hz for average measurement, Sweep = auto, Detector function = peak, Trace = max hold.

For Below 1GHz

Use the following spectrum analyzer settings:

Span = wide enough to capture the peak level of the in-band emission and all spurious
RBW = 100 KHz, VBW ≥ RBW for peak measurement, Sweep = auto, Detector function = peak, Trace = max hold.

Note:

- 1: The resolution bandwidth and video bandwidth of test receiver/spectrum analyzer is 120 KHz for Quasi-peak detection (QP) at frequency below 1GHz.
- 2: The resolution bandwidth of test receiver/spectrum analyzer is 1MHz and the video bandwidth is 3MHz for peak detection (PK) at frequency above 1GHz.
- 3: The resolution bandwidth of test receiver/spectrum analyzer is 1MHz and the video bandwidth is 3MHz for RMS Average ((duty cycle < 98%) for Average detection (AV) at frequency above 1GHz, then the measurement results was added to a correction factor (20log(1/duty cycle)).
- 4: The resolution bandwidth of test receiver/spectrum analyzer is 1MHz and the video bandwidth is 10Hz (duty cycle > 98%) for Average detection (AV) at frequency above 1GHz.

Limit

The radio emission outside the operating frequency band shall be at least 20 dB below that in the 100 kHz bandwidth within the band that contains the highest level of the desired power. Radiated emissions which fall in the restricted bands, as defined in section 15.205, must comply with the radiated emission limits specified in section 15.209.

Frequency MHz	Field Strength uV/m	Field Strength dBµV/m	Detector
30-88	100	40	QP
88-216	150	43.5	QP
216-960	200	46	QP
960-1000	500	54	QP
Above 1000	500	54	AV
Above 1000	5000	74	PK

Spurious radiated emissions for transmitter

Transmitting spurious emission test result as below:

Low channel 2402MHz Test Result

Frequency Band	Frequency MHz	Emission Level dBuV/m	Polarization	Limit dBuV/m	Detector	Margin dBuV/m	Correct factor (dB/m)	Result
30-1000MHz	886.13*	27.04	H	46	QP	18.96	-15.8	Pass
	877.19*	31.07	V	46	QP	14.93	-15.9	Pass
1000-25000MHz	4804*	43.87	H	74	PK	30.13	2.7	Pass
	--	--	H	54	AV	--	--	Pass
	4804*	41.80	V	74	PK	32.20	2.7	Pass
	--	--	V	54	AV	--	--	Pass

Middle channel 2440MHz Test Result

Frequency Band	Frequency MHz	Emission Level dBuV/m	Polarization	Limit dBuV/m	Detector	Margin dBuV/m	Correct factor (dB/m)	Result
30-1000MHz	--	--	H	43.5	QP	--	--	Pass
	--	--	H	46	QP	--	--	Pass
1000-25000MHz	4880*	47.90	H	74	PK	26.10	5.2	Pass
	--	--	H	54	AV	--	--	Pass
	4880*	48.25	V	74	PK	25.75	5.2	Pass
	--	--	V	54	AV	--	--	Pass

High channel 2480MHz Test Result

Frequency Band	Frequency MHz	Emission Level dBuV/m	Polarization	Limit dBμV/m	Detector	Margin dBuV/m	Correct factor (dB/m)	Result
30-1000MHz	--	--	H	43.5	QP	--	--	Pass
	--	--	H	46	QP	--	--	Pass
1000-25000MHz	4960*	44.02	H	74	PK	29.98	3.3	Pass
	--	--	H	54	AV	--	--	Pass
	4960*	43.32	V	74	PK	30.68	3.3	Pass
	--	--	V	54	AV	--	--	Pass

Remark:

- (1) "*" means the emission(s) appear within the restrict bands shall follow the requirement of section 15.205.
- (2) Data of measurement within this frequency range shown "--" in the table above means the reading of emissions are the noise floor or attenuated more than 10dB below the permissible limits or the field strength is too small to be measured.
- (3) Level=Reading Level + Correction Factor
 Above 1GHz: Corrector factor = Antenna Factor + Cable Loss- Amplifier Gain
 Below 1GHz: Corrector factor = Antenna Factor + Cable Loss
 (The Reading Level is recorded by software which is not shown in the sheet)

10 Test Equipment List

List of Test Instruments

Radiated Spurious Emission Test

Description	Manufacturer	Model no.	Serial no.	Cal. due date
Signal Analyzer	Rohde & Schwarz	FSV40	101031	2019-7-6
Trilog Super Broadband Test Antenna	Schwarzbeck	VULB 9163	708	2019-7-13
Horn Antenna	Rohde & Schwarz	HF907	102295	2019-7-13
Wideband Horn Antenna	Q-PAR	QWH-SL-18-40-K-SG	12827	2019-7-12
Pre-amplifier	Rohde & Schwarz	SCU 18	102230	2019-7-6
Pre-amplifier	Rohde & Schwarz	SCU 40A	100432	2019-7-6
Fully Anechoic Chamber	TDK	8X4X4	--	2020-7-7
Test software	Rohde & Schwarz	EMC32	Version 9.15.00	N/A

RF Conducted Test

Description	Manufacturer	Model no.	Serial no.	cal. due date
Signal Analyzer	Rohde & Schwarz	FSV40	101030	2019-7-6

11 System Measurement Uncertainty

For a 95% confidence level, the measurement expanded uncertainties for defined systems, in accordance with the recommendations of ISO 17025 were:

System Measurement Uncertainty	
Test Items	Extended Uncertainty
Uncertainty for Radiated Spurious Emission 25MHz-3000MHz	Horizontal: 4.80dB; Vertical: 4.87dB;
Uncertainty for Radiated Spurious Emission 3000MHz-18000MHz	Horizontal: 4.59dB; Vertical: 4.58dB;
Uncertainty for Radiated Spurious Emission 18000MHz-40000MHz	Horizontal: 5.05dB; Vertical: 5.04dB;
Uncertainty for Conducted RF test with TS 8997	RF Power Conducted: 1.16dB Frequency test involved: 0.6×10^{-7} or 1%