

 Celltech <small>Testing and Engineering Services Lab</small>	<u>Date(s) of Evaluation</u> March 28 - April 03, 2007	<u>Test Report Serial No.</u> 032807OWD-T826-S90F	<u>Report Revision No.</u> Revision 1.1	 ILAC-MRA  ACREDITED
	<u>Test Report Issue Date</u> April 25, 2007	<u>Description of Test(s)</u> Specific Absorption Rate	<u>RF Exposure Category</u> Occupational (Controlled)	

APPENDIX E - SYSTEM VALIDATION

Company:	M/A-COM, Inc.	Model:	P5300	FCC ID:	OWDTR-0047-E	IC ID:	3636B-0047	
DUT Type:	Portable Analog/Digital 900 MHz PTT Radio Transceiver				Freq.:	896-902 MHz / 935-941 MHz		
2007 Celltech Labs Inc.	This document is not to be reproduced in whole or in part without the prior written permission of Celltech Labs Inc.				Page 107 of 109			

 Celltech Testing and Engineering Services Ltd.	Date of Evaluation:	March 28, 2007	Document Serial No.:	SV900B-032807-R1.0		
	Evaluation Type:	System Validation	Validation Dipole:	900 MHz	Fluid Type:	Brain

900 MHz SYSTEM VALIDATION

Type:

900 MHz Validation Dipole

Asset Number:

00020

Serial Number:

054

Place of Validation:

Celltech Labs Inc.

Date of Validation:

March 28, 2007

Performed by:

Sean Johnston

Approved by:


Spencer Watson

1. Validation Dipole Construction & Electrical Characteristics

The validation dipole was constructed in accordance with the IEEE Standard "Annex G (informative) Reference dipoles for use in system validation". The electrical properties were measured using an HP 8753ET Network Analyzer. The network analyzer was calibrated to the validation dipole N-type connector feed point using an HP85032E Type N calibration kit. The dipole was placed parallel to a planar phantom at a separation distance of 15.0mm from the simulating fluid using a loss-less dielectric spacer. The measured input impedance is:



Feed point impedance at 900 MHz $\text{Re}\{Z\} = 58.063\Omega$
 $\text{Im}\{Z\} = 10.963\Omega$

Return Loss at 900 MHz -20.596dB



 Celltech <small>Testing and Engineering Services Labs</small>	Date of Evaluation:	March 28, 2007	Document Serial No.:	SV900B-032807-R1.0		
	Evaluation Type:	System Validation	Validation Dipole:	900 MHz	Fluid Type:	Brain

2. Validation Dipole VSWR Data



3. Validation Dipole Dimensions

Frequency (MHz)	L (mm)	h (mm)	d (mm)
300	420.0	250.0	6.2
450	288.0	167.0	6.2
835	161.0	89.8	3.6
900	149.0	83.3	3.6
1450	89.1	51.7	3.6
1800	72.0	41.7	3.6
1900	68.0	39.5	3.6
2000	64.5	37.5	3.6
2450	51.8	30.6	3.6
3000	41.5	25.0	3.6

4. Validation Phantom

The validation phantom is the SAM (Specific Anthropomorphic Mannequin) phantom manufactured by Schmid & Partner Engineering AG. The SAM phantom is a Fiberglass shell integrated in a wooden table. The shape of the shell corresponds to the phantom defined by SCC34-SC2. It enables the dosimetric evaluation of left and right hand phone usage as well as body mounted usage at the flat phantom region. A cover prevents evaporation of the liquid. Reference markings on the phantom allow the complete setup of all predefined phantom positions and measurement grids by manually teaching three points in the robot.

Shell Thickness: 2.0 ± 0.1 mm
Filling Volume: Approx. 25 liters
Dimensions: 50 cm (W) x 100 cm (L)

Celltech Testing and Engineering Services Labs	Date of Evaluation:	March 28, 2007	Document Serial No.:	SV900B-032807-R1.0		
	Evaluation Type:	System Validation	Validation Dipole:	900 MHz	Fluid Type:	Brain

5. 900 MHz System Validation Setup




 Celltech Testing and Engineering Services Labs	Date of Evaluation:	March 28, 2007	Document Serial No.:	SV900B-032807-R1.0		
	Evaluation Type:	System Validation	Validation Dipole:	900 MHz	Fluid Type:	Brain

6. 900 MHz Validation Dipole Setup



7. SAR Measurement

Measurements were made using a dosimetric E-field probe ET3DV6 (S/N: 1387, Conversion Factor 6.0). The SAR measurement was performed with the E-field probe in mechanical detection mode only. The setup and determination of the forward power into the dipole was performed using the following procedures.



First the power meter PM1 (including attenuator Att1) is connected to the cable to measure the forward power at the location of the dipole connector (X). The signal generator is adjusted for the desired forward power at the dipole connector (taking into account the attenuation of Att1) as read by power meter PM2. After connecting the cable to the dipole, the signal generator is readjusted for the same reading at power meter PM2. If the signal generator does not allow adjustment in 0.01dB steps, the remaining difference at PM2 must be taken into consideration. PM3 records the reflected power from the dipole to ensure that the value is not changed from the previous value. The reflected power should be 20dB below the forward power.

 Celltech <small>Testing and Engineering Services Labs</small>	Date of Evaluation:	March 28, 2007	Document Serial No.:	SV900B-032807-R1.0		
	Evaluation Type:	System Validation	Validation Dipole:	900 MHz	Fluid Type:	Brain

8. Measurement Conditions

The phantom was filled with 900 MHz Brain tissue simulant.

Relative Permittivity: 39.6 (-4.5% from target)
 Conductivity: 0.94 mho/m (-3.0% from target)
 Fluid Temperature: 22.0°C
 Fluid Depth: ≥ 15.0 cm

Environmental Conditions:

Ambient Temperature: 22.5°C
 Barometric Pressure: 101.9 kPa
 Humidity: 33%

The 900 MHz Brain tissue simulant consisted of the following ingredients:

Ingredient	Percentage by weight
Water	40.71%
Sugar	56.63%
Salt	1.48%
Dowicil 75	0.19%
HEC	0.99%
Target Dielectric Parameters at 22 °C	$\epsilon_r = 41.5 (+/- 5\%)$ $\sigma = 0.97 \text{ S/m} (+/- 5\%)$

9. System Validation SAR Results

SAR @ 0.25W Input averaged over 1g (W/kg)				SAR @ 1W Input averaged over 1g (W/kg)			
IEEE Target		Measured	Deviation	IEEE Target		Measured	Deviation
2.70	+/- 10%	2.65	-1.8%	10.8	+/- 10%	10.6	-1.8%
SAR @ 0.25W Input averaged over 10g (W/kg)				SAR @ 1W Input averaged over 10g (W/kg)			
IEEE Target		Measured	Deviation	IEEE Target		Measured	Deviation
1.73	+/- 10%	1.66	-4.0%	6.92	+/- 10%	6.64	-4.0%
The results have been normalized to 1W (forward power) into the dipole.							

System Validation - 900 MHz Dipole - March 28, 2007

DUT: Dipole 900 MHz; Asset: 00020; Serial: 054

Ambient Temp: 22.5°C; Fluid Temp: 22.0°C; Barometric Pressure: 101.9 kPa; Humidity: 33%

Communication System: CW

Forward Conducted Power: 250 mW

Frequency: 900 MHz; Duty Cycle: 1:1

Medium: HSL900 Medium parameters used: $f = 900$ MHz; $\sigma = 0.94$ mho/m; $\epsilon_r = 39.6$; $\rho = 1000$ kg/m³

- Probe: ET3DV6 - SN1387; ConvF(6, 6, 6); Calibrated: 16/03/2007
- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)
- Electronics: DAE4 Sn353; Calibrated: 21/06/2006
- Phantom: SAM 4.0; Type: Fiberglas; Serial: 1033
- Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

900 MHz System Validation/Area Scan (6x10x1):

Measurement grid: dx=10mm, dy=10mm

900 MHz System Validation/Zoom Scan 2 (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm


Reference Value = 54.5 V/m; Power Drift = 0.030 dB

Peak SAR (extrapolated) = 4.48 W/kg

SAR(1 g) = 2.65 mW/g; SAR(10 g) = 1.66 mW/g

Maximum value of SAR (measured) = 2.84 mW/g





10. Measured Fluid Dielectric Parameters

900 MHz System Validation (Brain)

Celltech Labs Inc.

Test Result for UIM Dielectric Parameter

Wed 28/Mar/2007

Frequency (GHz)

FCC_eHFCC OET 65 Supplement C (June 2001) Limits for Head Epsilon

FCC_sHFCC OET 65 Supplement C (June 2001) Limits for Head Sigma

Test_e Epsilon of UIM

Test_s Sigma of UIM

Freq	FCC_eH	FCC_sH	Test_e	Test_s
0.8000	41.68	0.90	40.80	0.84
0.8100	41.63	0.90	40.73	0.85
0.8200	41.58	0.90	40.58	0.87
0.8300	41.53	0.90	40.54	0.87
0.8400	41.50	0.91	40.37	0.88
0.8500	41.50	0.92	40.20	0.89
0.8600	41.50	0.93	40.05	0.90
0.8700	41.50	0.94	40.03	0.91
0.8800	41.50	0.95	39.76	0.92
0.8900	41.50	0.96	39.76	0.93
0.9000	41.50	0.97	39.56	0.94
0.9100	41.50	0.98	39.45	0.95
0.9200	41.49	0.98	39.44	0.96
0.9300	41.47	0.99	39.34	0.97
0.9400	41.45	0.99	39.30	0.98
0.9500	41.43	0.99	39.02	0.98
0.9600	41.42	1.00	39.12	1.00
0.9700	41.40	1.00	38.87	1.01
0.9800	41.38	1.01	38.79	1.01
0.9900	41.36	1.01	38.79	1.03
1.0000	41.34	1.01	38.56	1.03