

Advanced
Compliance Laboratory

6 Randolph Way
Hillsborough, NJ 08844
Tel: (908) 927 9288
Fax: (908) 927 0728

ELECTROMAGNETIC EMISSION COMPLIANCE REPORT

of

HEART RATE MONITOR
MODEL: DS-14 HRM/FM
FCC ID: OWFHRM-FM

February 21, 2005

This report concerns (check one): Original grant ☒ Class II change ☐
Equipment type: Low Power Intentional Radiator

Deferred grant requested per 47 CF 0.457(d)(1)(ii)? yes ☐ no ☒
If yes, defer until: _____ (date)
Company agrees to notify the Commission by _____ (date)
of the intended date of announcement of the product so that the grant can be
issued on that date.

Transition Rules Request per 15.37? yes ☐ no ☒
If no, assumed Part 15, Subpart B for unintentional radiators - the new 47 CFR
[10-1-90 Edition] provision.

Report prepared for: KNOW-HOW DEVELOPMENT LIMITED
Report prepared by: Advanced Compliance Lab
Report number: 0048-050110-01



The test result in this report IS supported and covered by the NVLAP accreditation

Table of Contents

Report Cover Page	1
Table of Contents	2
Figures	3
 1. GENERAL INFORMATION	 4
1.1 Verification of Compliance.....	4
1.2 Equipment Modifications	5
1.3 Product Information	6
1.4 Test Methodology	6
1.5 Test Facility.....	6
1.6 Test Equipment.....	6
1.7 Statement of the Document Use	7
 2. PRODUCT LABELING	 8
 3. SYSTEM TEST CONFIGURATION.....	 9
3.1 Justification.....	9
3.2 Special Accessories.....	9
3.3 Configuration of Tested System	9
 4. SYSTEM SCHEMATICS.....	 12
 5. RADIATED EMISSION DATA	 13
5.1 Field Strength Calculation.....	13
5.2 Test Methods and Conditions	13
5.3 Test Data	13
 6. PHOTOS OF TESTED EUT	 15

Figures

Figure 2.1 FCC ID Label	8
Figure 2.2 Location of Label on Back of the EUT	8
Figure 3.1 Radiated Test Setup, Position 1	10
Figure 3.2 Radiated Test Setup, Position 2	10
Figure 3.3 Radiated Test Setup, Position 3	11
Figure 4.1 EUT Schematics.....	12
Figure 6.1 Front View	16
Figure 6.2 Rear View.....	17
Figure 6.2 Inside View, Cover Opened.....	18
Figure 6.3 Component Side.....	19
Figure 6.4 Foil Side	20

1. GENERAL INFORMATION

1.1 Verification of Compliance

EUT: HEART RATE MONITOR

Model: DS-14 HRM/FM

Applicant: KNOW-HOW DEVELOPMENT LIMITED.
Block D, 4/F., Luk Hop Ind. Bldg.,
8 Luk Hop Street, San Po Kong,
Kowloon, Hong Kong

Test Type: FCC Part 15C CERTIFICATION

Result: PASS

Tested by: ADVANCED COMPLIANCE LABORATORY


Test Date: February 7, 2005

Report Number: 0048-050110-01

The above equipment was tested by Compliance Laboratory, Advanced Technologies, Inc. for compliance with the requirement set forth in the FCC rules and regulations Part 15 subpart C. This said equipment in the configuration described in the report, shows the maximum emission levels emanating from equipment are within the compliance requirements.

The estimated uncertainty of the test result is given as following. The method of uncertainty calculation is provided in Advanced Compliance Lab. Doc. No. 0048-01-01.

	Prob. Dist.	Uncertainty(dB)	Uncertainty(dB)	Uncertainty(dB)
		30-1000MHz	1-6.5GHz	Conducted
Combined Std. Uncertainty u_c	norm.	± 2.36	± 2.99	± 1.83



Wei Li
Lab Manager
Advanced Compliance Lab

Date: February 21, 2005

1.2 Equipment Modifications

N/A

1.3 Product Information

System Configuration

ITEM	DESCRIPTION	FCC ID	CABLE
Product	HEART RATE MONITOR DS-14 HRM-FM ⁽¹⁾	OWFHRM-FM	
Housing	PLASTICS		
Power Supply	12V DC Battery		
Operation/OSC Freq.	75KHz		
Receiver	DS-14 HRM-FM (RX)		

(1) EUT submitted for grant.

1.4 Test Methodology

Radiated tests were performed according to the procedures in ANSI C63.4-2001 at an antenna to EUT distance of 3 meters.

1.5 Test Facility

The open area test site and conducted measurement facility used to collect the radiated and conducted data are located at Hillsborough, New Jersey. This site has been accepted by FCC to perform measurements under Part 15 or 18 in a letter dated May 19, 1997 (Refer to: 31040/PRV 1300F2). The NVLAP Lab code for accreditation of FCC EMC Test Method is: 200101-0.

1.6 Test Equipment

Manufacture	Model	Serial No.	Description	Last Cal dd/mm/yy	Cal Due dd/mm/yy
Hewlett-Packard	HP8546A	3448A00290	EMI Receiver	18/01/05	18/01/06
EMCO	3104C	9307-4396	20-300MHz Biconical Antenna	12/02/04	12/02/05
EMCO	3146	9008-2860	200-1000MHz Log-Periodic Antenna	09/02/04	09/02/05
Fischer Custom	LISN-2	900-4-0008	Line Impedance Stabilization Networks	23/08/04	23/08/05
Fischer Custom	LISN-2	900-4-0009	Line Impedance Stabilization Networks	23/08/04	23/08/05
EMCO	6502	2665	10KHz-30MHz Active Loop Antenna	27/02/04	27/02/05
EMCO	3115	4945	Double Ridge Guide Horn Antenna	11/08/04	11/08/05

All Test Equipment Used are Calibrated Traceable to NIST Standards.

1.7 Statement for the Document Use

This report shall not be reproduced except in full, without the written approval of the laboratory. And this report must not be used by the client to claim product endorsement by NVLAP or any agency of the U.S. Government.

2. PRODUCT LABELING

FCC ID: OWFHRM-FM

This device complies with part 15 of the FCC Rules. Operating is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

Figure 2.1 FCC ID Label



Figure 2.2 Location of the Label

3. SYSTEM TEST CONFIGURATION

3.1 Justification

The system was configured for testing in a typical fashion (as a customer would normally use it). The EUT was set for transmitting signal at frequency 75KHz continuously for testing purpose.

3.2 Special Accessories

N/A

3.3 Configuration of Tested System

Figure 3.1 to Figure 3.3 illustrate this system, which is tested standing along.



Figure 3.1 Radiated Test Setup, position 1



Figure 3.2 Radiated Test Setup, position 2



Figure 3.3 Radiated Test Setup, position 3

4. SYSTEM SCHEMATICS




Figure 4.1 System Schematics

5. RADIATED EMISSION DATA

5.1 Field Strength Calculation

The corrected field strength is automatically calculated by EMI Receiver using following:

$$FS = RA + AF + CF + AG$$

where FS: Corrected Field Strength in dBμV/m

RA: Amplitude of EMI Receiver before correction in dBμV

AF: Antenna Factor in dB/m

CF: Cable Attenuation Factor in dB

AG: Built-in Preamplifier Gain in dB (Stored in receiver as part of the calibration data)

5.2 Test Methods and Conditions

The initial step in collecting radiated data is a EMI Receiver scan of the measurement range below 30MHz using peak detector and 9KHz IF bandwidth / 30KHz video bandwidth. For the range 30MHz - 1GHz, 120KHz IF bandwidth / 120KHz video bandwidth are used. Both bandwidths are 1MHz for above 1GHz measurement. Up to 10th harmonics were investigated.

5.3 Test Data

The following data lists the significant emission frequencies, polarity and position, peak reading of the EMI Receiver, the FCC limit, and the difference between the peak reading and the limit. Explanation of the correction and calculation are given in section 5.1.

Test Personnel:



Typed/Printed Name: Edward Lee

Date: February 21, 2005

Radiated Test Data (CH-75KHz)

Frequency (MHz)	Polarity [H, V] Position	Height (m)	Azimuth (Degree)	Peak(2,3) Reading (dBμV/m)	300m Limit(1) (dBμV/m)	Difference from limit (dB)
75.9	X	1.1	280	27.8	30.1	-2.3
152.0	X	1.1	280	17.2	24.1	-6.9
227.1	X	1.1	280	15.4	20.4	-5
303.5	X	1.1	280	12.1	18.0	-5.9
75.9	Y	1.1	100	26.9	30.1	-3.2
152.0	Y	1.1	100	16.9	24.1	-7.2
227.0	Y	1.1	100	16.0	20.4	-4.4
303.0	Y	1.1	100	13.2	18.0	-4.8
75.9	Z	1.1	180	24.7	30.1	-5.4
152.1	Z	1.1	180	15.2	24.1	-8.9
303.4	Z	1.1	190	11.0	20.4	-9.4

- (1) The limit for emissions is defined in Sec. 15.209: 2400 μ V/F(kHz)@300m (average detector).
- (2) If each peak reading is less than the FCC average limit, it'll be not necessary to show the calculated average reading based on the pulse train characteristics.
- (3) Radiated tests were performed according to the procedures in ANSI C63.4-2001 at an antenna to EUT distance of 3 meters. An inverse proportional extrapolation was performed to compare the signal level to 30-meter limit.

6. PHOTOS OF TESTED EUT

The following photos show the inside details of the EUT.