

Test Calibration certificate

Appendix B.1 Calibration certificate for Probe(S/N: 7413)

**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **SGS Korea (Dymstec)**

Certificate No: **EX3-7413_Sep18**

CALIBRATION CERTIFICATE

Object **EX3DV4 - SN:7413**

Calibration procedure(s) **QA CAL-01.v9, QA CAL-14.v4, QA CAL-23.v5, QA CAL-25.v6**
Calibration procedure for dosimetric E-field probes

Calibration date: **September 25, 2018**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&E critical for calibration)

Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-18 (No. 217-02672/02673)	Apr-19
Power sensor NRP-Z91	SN: 103244	04-Apr-18 (No. 217-02672)	Apr-19
Power sensor NRP-Z91	SN: 103245	04-Apr-18 (No. 217-02673)	Apr-19
Reference 20 dB Attenuator	SN: S5277 (20x)	04-Apr-18 (No. 217-02682)	Apr-19
Reference Probe ES3DV2	SN: 3013	30-Dec-17 (No. ES3-3013, Dec17)	Dec-18
DAE4	SN: 660	21-Dec-17 (No. DAE4-660, Dec17)	Dec-18
Secondary Standards	ID	Check Date (in house)	Scheduled Check
Power meter E4419B	SN: GB41293874	06-Apr-18 (in house check Jun-18)	In house check: Jun-20
Power sensor E4412A	SN: MY41498087	06-Apr-18 (in house check Jun-18)	In house check: Jun-20
Power sensor E4412A	SN: 000110210	06-Apr-18 (in house check Jun-18)	In house check: Jun-20
RF generator HP 8648C	SN: US3642U01700	04-Aug-99 (in house check Jun-18)	In house check: Jun-20
Network Analyzer E6358A	SN: US41080477	31-Mar-14 (in house check Oct-17)	In house check: Oct-18

Calibrated by:	Name Claudio Leubler	Function Laboratory Technician	Signature 
Approved by:	Name Katja Pokovic	Function Technical Manager	Signature 
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			

Issued: September 27, 2018

Certificate No: EX3-7413_Sep18

Page 1 of 39

Report File No : F690501/RF-SAR002704

Date of Issue : 2019-01-08

(All SGS services are rendered in accordance with the applicable SGS conditions of service available on request and accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>.)

RTT5041-76(2015.10.01) (2)

A4 (210mm x 297mm)

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Glossary:

TSL	tissue simulating liquid
NORM _{x,y,z}	sensitivity in free space
ConvF	sensitivity in TSL / NORM _{x,y,z}
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization ϕ	ϕ rotation around probe axis
Polarization θ	θ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\theta = 0$ is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- NORM_{x,y,z}:** Assessed for E-field polarization $\theta = 0$ ($f \leq 900$ MHz in TEM-cell; $f > 1800$ MHz: R22 waveguide). NORM_{x,y,z} are only intermediate values, i.e., the uncertainties of NORM_{x,y,z} does not affect the E²-field uncertainty inside TSL (see below ConvF).
- NORM(f)_{x,y,z} = NORM_{x,y,z} * frequency_response** (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCPx,y,z:** DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR:** PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A_{x,y,z}; B_{x,y,z}; C_{x,y,z}; D_{x,y,z}; VR_{x,y,z}:** A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters:** Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f \leq 800$ MHz) and inside waveguide using analytical field distributions based on power measurements for $f > 800$ MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{x,y,z} * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy):** in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset:** The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle:** The angle is assessed using the information gained by determining the NORM_x (no uncertainty required).

EX3DV4 – SN:7413

September 25, 2018

Probe EX3DV4

SN:7413

Manufactured: March 10, 2016
Repaired: September 14, 2018
Calibrated: September 25, 2018

Calibrated for DASY/EASY Systems
(Note: non-compatible with DASY2 system!)

EX3DV4- SN:7413

September 25, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7413

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm ($\mu\text{V}/(\text{V}/\text{m})^2$) ^A	0.60	0.67	0.42	$\pm 10.1 \%$
DCP (mV) ^B	101.9	98.7	99.3	

Modulation Calibration Parameters

UID	Communication System Name		A dB	B dB $\sqrt{\mu\text{V}}$	C	D dB	VR mV	Unc ^E (k=2)
0	CW	X	0.0	0.0	1.0	0.00	156.7	$\pm 2.5 \%$
		Y	0.0	0.0	1.0		169.0	
		Z	0.0	0.0	1.0		167.8	

Note: For details on UID parameters see Appendix.

Sensor Model Parameters

	C1 fF	C2 fF	α V^{-1}	T1 $\text{ms} \cdot \text{V}^{-2}$	T2 $\text{ms} \cdot \text{V}^{-1}$	T3 ms	T4 V^{-2}	T5 V^{-1}	T6
X	48.17	356.6	35.13	18.43	0.254	5.100	0.843	0.368	1.007
Y	40.77	309.9	36.58	13.30	0.336	5.100	0.635	0.423	1.009
Z	43.41	323.1	35.34	12.13	0.366	5.077	1.618	0.248	1.008

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of Norm X,Y,Z do not affect the E^2 -field uncertainty inside TSL (see Pages 5 and 6).

^B Numerical linearization parameter: uncertainty not required.

^E Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

EX3DV4- SN:7413

September 25, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7413

Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) ^C	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^G (mm)	Unc (k=2)
835	41.5	0.90	9.54	9.54	9.54	0.41	0.86	± 12.0 %
900	41.5	0.97	9.39	9.39	9.39	0.48	0.83	± 12.0 %
1750	40.1	1.37	8.18	8.18	8.18	0.36	0.80	± 12.0 %
1900	40.0	1.40	7.92	7.92	7.92	0.29	0.80	± 12.0 %
2300	39.5	1.67	7.73	7.73	7.73	0.28	0.87	± 12.0 %
2450	39.2	1.80	7.34	7.34	7.34	0.29	0.90	± 12.0 %
2600	39.0	1.96	7.11	7.11	7.11	0.33	0.92	± 12.0 %
5200	36.0	4.66	5.30	5.30	5.30	0.50	1.80	± 13.1 %
5300	35.9	4.76	5.15	5.15	5.15	0.50	1.80	± 13.1 %
5600	35.5	5.07	4.57	4.57	4.57	0.50	1.80	± 13.1 %
5800	35.3	5.27	4.71	4.71	4.71	0.50	1.80	± 13.1 %

^C Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

^F At frequencies below 3 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ϵ and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

EX3DV4- SN:7413

September 25, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7413

Calibration Parameter Determined in Body Tissue Simulating Media

f (MHz) ^C	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^G (mm)	Unc (k=2)
2300	52.9	1.81	7.45	7.45	7.45	0.37	0.85	± 12.0 %
2450	52.7	1.95	7.46	7.46	7.46	0.32	0.96	± 12.0 %
5200	49.0	5.30	4.86	4.86	4.86	0.50	1.90	± 13.1 %
5300	48.9	5.42	4.71	4.71	4.71	0.50	1.90	± 13.1 %
5600	48.5	5.77	4.26	4.26	4.26	0.50	1.90	± 13.1 %
5800	48.2	6.00	4.37	4.37	4.37	0.50	1.90	± 13.1 %

^C Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.


^F At frequencies below 3 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ϵ and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

EX3DV4- SN:7413

September 25, 2018

Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)



Uncertainty of Frequency Response of E-field: $\pm 6.3\%$ (k=2)

EX3DV4- SN:7413



September 25, 2018

Receiving Pattern (ϕ), $\theta = 0^\circ$

f=600 MHz,TEM



f=1800 MHz,R22





Uncertainty of Axial Isotropy Assessment: $\pm 0.5\%$ (k=2)

EX3DV4- SN:7413

September 25, 2018

Dynamic Range $f(\text{SAR}_{\text{head}})$ (TEM cell, $f_{\text{eval}} = 1900 \text{ MHz}$)




Uncertainty of Linearity Assessment: $\pm 0.6\%$ ($k=2$)

EX3DV4- SN:7413


September 25, 2018

Conversion Factor Assessment



Deviation from Isotropy in Liquid

Error (ϕ , θ), $f = 900$ MHz



EX3DV4- SN:7413

September 25, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7413

Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (°)	-17.1
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm

EX3DV4- SN:7413

September 25, 2018

Appendix: Modulation Calibration Parameters

UID	Communication System Name		A dB	B dB μ V	C	D dB	VR mV	Max Unc ^ε (k=2)
0	CW	X	0.00	0.00	1.00	0.00	156.7	± 2.5 %
		Y	0.00	0.00	1.00		169.0	
		Z	0.00	0.00	1.00		167.8	
10010- CAA	SAR Validation (Square, 100ms, 10ms)	X	5.25	75.30	14.26	10.00	20.0	± 9.6 %
		Y	2.70	68.26	11.38		20.0	
		Z	2.95	69.13	11.83		20.0	
10011- CAB	UMTS-FDD (WCDMA)	X	1.09	68.55	15.97	0.00	150.0	± 9.6 %
		Y	0.84	64.38	12.97		150.0	
		Z	0.93	65.80	14.13		150.0	
10012- CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)	X	1.21	64.59	15.79	0.41	150.0	± 9.6 %
		Y	1.09	62.81	14.24		150.0	
		Z	1.13	63.34	14.75		150.0	
10013- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps)	X	4.91	66.88	17.26	1.46	150.0	± 9.6 %
		Y	4.76	66.57	16.97		150.0	
		Z	4.80	66.64	17.00		150.0	
10021- DAC	GSM-FDD (TDMA, GMSK)	X	100.00	116.61	28.42	9.39	50.0	± 9.6 %
		Y	100.00	115.65	27.89		50.0	
		Z	100.00	115.19	27.71		50.0	
10023- DAC	GPRS-FDD (TDMA, GMSK, TN 0)	X	100.00	116.16	28.25	9.57	50.0	± 9.6 %
		Y	100.00	115.03	27.64		50.0	
		Z	100.00	114.62	27.49		50.0	
10024- DAC	GPRS-FDD (TDMA, GMSK, TN 0-1)	X	100.00	116.43	27.52	6.56	60.0	± 9.6 %
		Y	100.00	115.44	26.78		60.0	
		Z	100.00	115.69	26.88		60.0	
10025- DAC	EDGE-FDD (TDMA, 8PSK, TN 0)	X	13.48	111.72	46.07	12.57	50.0	± 9.6 %
		Y	5.15	78.28	31.20		50.0	
		Z	5.02	77.07	30.27		50.0	
10026- DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1)	X	20.33	114.60	41.34	9.56	60.0	± 9.6 %
		Y	9.12	93.97	34.18		60.0	
		Z	8.83	92.66	33.46		60.0	
10027- DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2)	X	100.00	118.14	27.61	4.80	80.0	± 9.6 %
		Y	100.00	116.29	26.38		80.0	
		Z	100.00	117.54	26.92		80.0	
10028- DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)	X	100.00	121.19	28.27	3.55	100.0	± 9.6 %
		Y	100.00	117.36	26.14		100.0	
		Z	100.00	120.24	27.36		100.0	
10029- DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2)	X	8.92	92.61	32.33	7.80	80.0	± 9.6 %
		Y	5.55	81.63	27.96		80.0	
		Z	5.46	81.00	27.54		80.0	
10030- CAA	IEEE 802.15.1 Bluetooth (GFSK, DH1)	X	100.00	115.39	26.67	5.30	70.0	± 9.6 %
		Y	100.00	113.46	25.44		70.0	
		Z	100.00	114.31	25.81		70.0	
10031- CAA	IEEE 802.15.1 Bluetooth (GFSK, DH3)	X	100.00	123.56	27.83	1.88	100.0	± 9.6 %
		Y	100.00	108.89	21.26		100.0	
		Z	100.00	117.22	24.69		100.0	

EX3DV4- SN:7413

September 25, 2018

10032-CAA	IEEE 802.15.1 Bluetooth (GFSK, DH5)	X	100.00	134.27	31.06	1.17	100.0	± 9.6 %
		Y	100.00	104.40	18.62		100.0	
		Z	100.00	121.08	25.26		100.0	
10033-CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH1)	X	100.00	131.06	35.71	5.30	70.0	± 9.6 %
		Y	44.11	115.67	31.08		70.0	
		Z	34.54	112.94	30.73		70.0	
10034-CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH3)	X	14.55	100.35	26.07	1.88	100.0	± 9.6 %
		Y	2.90	76.31	17.14		100.0	
		Z	3.46	79.49	18.88		100.0	
10035-CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH5)	X	4.69	85.05	21.15	1.17	100.0	± 9.6 %
		Y	1.67	70.13	14.26		100.0	
		Z	2.00	73.02	16.07		100.0	
10036-CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH1)	X	100.00	131.52	35.92	5.30	70.0	± 9.6 %
		Y	100.00	128.74	34.26		70.0	
		Z	80.34	126.67	34.16		70.0	
10037-CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH3)	X	12.31	98.03	25.40	1.88	100.0	± 9.6 %
		Y	2.63	75.19	16.71		100.0	
		Z	3.12	78.20	18.39		100.0	
10038-CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH5)	X	4.81	85.77	21.53	1.17	100.0	± 9.6 %
		Y	1.68	70.42	14.50		100.0	
		Z	2.01	73.36	16.33		100.0	
10039-CAB	CDMA2000 (1xRTT, RC1)	X	2.11	74.20	16.71	0.00	150.0	± 9.6 %
		Y	1.07	65.38	11.41		150.0	
		Z	1.41	68.81	13.70		150.0	
10042-CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Halfrate)	X	100.00	112.51	25.91	7.78	50.0	± 9.6 %
		Y	100.00	111.05	25.04		50.0	
		Z	100.00	111.39	25.21		50.0	
10044-CAA	IS-91/EIA/TIA-553 FDD (FDMA, FM)	X	0.00	101.43	1.78	0.00	150.0	± 9.6 %
		Y	0.00	126.53	5.73		150.0	
		Z	0.00	104.01	9.90		150.0	
10048-CAA	DECT (TDD, TDMA/FDM, GFSK, Full Slot, 24)	X	100.00	116.96	29.64	13.80	25.0	± 9.6 %
		Y	100.00	112.78	28.01		25.0	
		Z	100.00	111.93	27.73		25.0	
10049-CAA	DECT (TDD, TDMA/FDM, GFSK, Double Slot, 12)	X	100.00	114.83	27.85	10.79	40.0	± 9.6 %
		Y	100.00	113.29	27.16		40.0	
		Z	100.00	112.94	27.06		40.0	
10056-CAA	UMTS-TDD (TD-SCDMA, 1.28 Mcps)	X	100.00	126.97	34.91	9.03	50.0	± 9.6 %
		Y	100.00	124.79	33.66		50.0	
		Z	100.00	124.90	33.76		50.0	
10058-DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2-3)	X	6.06	83.60	27.93	6.55	100.0	± 9.6 %
		Y	4.30	76.23	24.76		100.0	
		Z	4.25	75.85	24.48		100.0	
10059-CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps)	X	1.28	66.20	16.70	0.61	110.0	± 9.6 %
		Y	1.13	63.85	14.87		110.0	
		Z	1.16	64.38	15.37		110.0	
10060-CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps)	X	100.00	140.76	37.06	1.30	110.0	± 9.6 %
		Y	6.64	96.84	25.27		110.0	
		Z	10.42	105.12	28.14		110.0	

EX3DV4- SN:7413

September 25, 2018

10061-CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)	X	7.93	98.79	28.94	2.04	110.0	± 9.6 %
		Y	2.68	79.66	21.94		110.0	
		Z	2.71	80.09	22.29		110.0	
10062-CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps)	X	4.70	66.82	16.63	0.49	100.0	± 9.6 %
		Y	4.53	66.43	16.27		100.0	
		Z	4.59	66.58	16.38		100.0	
10063-CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps)	X	4.72	66.94	16.75	0.72	100.0	± 9.6 %
		Y	4.55	66.54	16.39		100.0	
		Z	4.61	66.69	16.48		100.0	
10064-CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps)	X	5.01	67.21	16.99	0.86	100.0	± 9.6 %
		Y	4.82	66.80	16.64		100.0	
		Z	4.88	66.94	16.72		100.0	
10065-CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps)	X	4.88	67.14	17.12	1.21	100.0	± 9.6 %
		Y	4.70	66.72	16.76		100.0	
		Z	4.76	66.83	16.82		100.0	
10066-CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps)	X	4.91	67.18	17.31	1.46	100.0	± 9.6 %
		Y	4.72	66.75	16.95		100.0	
		Z	4.78	66.87	17.00		100.0	
10067-CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps)	X	5.20	67.35	17.77	2.04	100.0	± 9.6 %
		Y	5.02	67.05	17.48		100.0	
		Z	5.07	67.08	17.48		100.0	
10068-CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps)	X	5.26	67.45	18.03	2.55	100.0	± 9.6 %
		Y	5.07	67.05	17.70		100.0	
		Z	5.11	67.10	17.69		100.0	
10069-CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps)	X	5.33	67.43	18.22	2.67	100.0	± 9.6 %
		Y	5.14	67.08	17.91		100.0	
		Z	5.19	67.11	17.89		100.0	
10071-CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 9 Mbps)	X	5.01	67.00	17.60	1.99	100.0	± 9.6 %
		Y	4.86	66.69	17.31		100.0	
		Z	4.90	66.74	17.31		100.0	
10072-CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 12 Mbps)	X	5.00	67.38	17.86	2.30	100.0	± 9.6 %
		Y	4.83	67.00	17.54		100.0	
		Z	4.87	67.04	17.54		100.0	
10073-CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 18 Mbps)	X	5.06	67.57	18.23	2.83	100.0	± 9.6 %
		Y	4.90	67.21	17.92		100.0	
		Z	4.93	67.22	17.88		100.0	
10074-CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 24 Mbps)	X	5.05	67.48	18.40	3.30	100.0	± 9.6 %
		Y	4.90	67.15	18.10		100.0	
		Z	4.92	67.12	18.04		100.0	
10075-CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 36 Mbps)	X	5.09	67.64	18.76	3.82	90.0	± 9.6 %
		Y	4.93	67.24	18.41		90.0	
		Z	4.95	67.20	18.35		90.0	
10076-CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 48 Mbps)	X	5.09	67.40	18.87	4.15	90.0	± 9.6 %
		Y	4.96	67.08	18.58		90.0	
		Z	4.97	67.01	18.48		90.0	
10077-CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 54 Mbps)	X	5.11	67.46	18.97	4.30	90.0	± 9.6 %
		Y	4.99	67.17	18.68		90.0	
		Z	4.99	67.09	18.59		90.0	

EX3DV4- SN:7413

September 25, 2018

10081-CAB	CDMA2000 (1xRTT, RC3)	X	0.91	67.28	13.32	0.00	150.0	± 9.6 %
		Y	0.57	61.99	9.06		150.0	
		Z	0.70	63.97	10.89		150.0	
10082-CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Fullrate)	X	0.81	60.00	4.80	4.77	80.0	± 9.6 %
		Y	0.70	60.00	4.41		80.0	
		Z	0.69	60.00	4.46		80.0	
10090-DAC	GPRS-FDD (TDMA, GMSK, TN 0-4)	X	100.00	116.46	27.56	6.56	60.0	± 9.6 %
		Y	100.00	115.52	26.83		60.0	
		Z	100.00	115.73	26.92		60.0	
10097-CAB	UMTS-FDD (HSDPA)	X	1.87	68.26	16.03	0.00	150.0	± 9.6 %
		Y	1.61	65.91	14.19		150.0	
		Z	1.73	66.91	15.00		150.0	
10098-CAB	UMTS-FDD (HSUPA, Subtest 2)	X	1.84	68.23	16.01	0.00	150.0	± 9.6 %
		Y	1.58	65.84	14.15		150.0	
		Z	1.69	66.85	14.96		150.0	
10099-DAC	EDGE-FDD (TDMA, 8PSK, TN 0-4)	X	20.65	114.96	41.45	9.56	60.0	± 9.6 %
		Y	9.20	94.17	34.25		60.0	
		Z	8.91	92.85	33.52		60.0	
10100-CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	X	3.21	70.90	16.99	0.00	150.0	± 9.6 %
		Y	2.78	68.46	15.60		150.0	
		Z	2.95	69.46	16.18		150.0	
10101-CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	X	3.27	67.78	16.08	0.00	150.0	± 9.6 %
		Y	3.03	66.54	15.24		150.0	
		Z	3.13	67.08	15.59		150.0	
10102-CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	X	3.37	67.71	16.15	0.00	150.0	± 9.6 %
		Y	3.14	66.58	15.37		150.0	
		Z	3.24	67.09	15.70		150.0	
10103-CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	X	7.90	79.98	22.39	3.98	65.0	± 9.6 %
		Y	6.07	75.90	20.77		65.0	
		Z	6.05	75.73	20.64		65.0	
10104-CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	X	7.08	76.27	21.73	3.98	65.0	± 9.6 %
		Y	6.01	73.62	20.55		65.0	
		Z	6.02	73.55	20.46		65.0	
10105-CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	X	6.81	75.42	21.68	3.98	65.0	± 9.6 %
		Y	5.55	71.90	20.07		65.0	
		Z	5.98	73.28	20.65		65.0	
10108-CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	X	2.80	70.10	16.81	0.00	150.0	± 9.6 %
		Y	2.40	67.73	15.38		150.0	
		Z	2.56	68.70	15.98		150.0	
10109-CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	X	2.92	67.65	16.00	0.00	150.0	± 9.6 %
		Y	2.67	66.30	15.02		150.0	
		Z	2.78	66.89	15.44		150.0	
10110-CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	X	2.28	69.24	16.46	0.00	150.0	± 9.6 %
		Y	1.91	66.74	14.77		150.0	
		Z	2.06	67.75	15.48		150.0	
10111-CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	X	2.65	68.54	16.33	0.00	150.0	± 9.6 %
		Y	2.34	66.83	14.98		150.0	
		Z	2.48	67.65	15.62		150.0	

EX3DV4- SN:7413

September 25, 2018

10112-CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	X	3.05	67.61	16.04	0.00	150.0	± 9.6 %
		Y	2.80	66.39	15.12		150.0	
		Z	2.91	66.94	15.52		150.0	
10113-CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	X	2.80	68.64	16.44	0.00	150.0	± 9.6 %
		Y	2.49	67.06	15.17		150.0	
		Z	2.64	67.85	15.78		150.0	
10114-CAC	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	X	5.12	67.27	16.47	0.00	150.0	± 9.6 %
		Y	4.97	66.81	16.15		150.0	
		Z	5.04	67.05	16.28		150.0	
10115-CAC	IEEE 802.11n (HT Greenfield, 81 Mbps, 16-QAM)	X	5.41	67.39	16.54	0.00	150.0	± 9.6 %
		Y	5.22	66.90	16.21		150.0	
		Z	5.30	67.12	16.33		150.0	
10116-CAC	IEEE 802.11n (HT Greenfield, 135 Mbps, 64-QAM)	X	5.22	67.46	16.49	0.00	150.0	± 9.6 %
		Y	5.05	66.99	16.17		150.0	
		Z	5.12	67.22	16.29		150.0	
10117-CAC	IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK)	X	5.09	67.14	16.42	0.00	150.0	± 9.6 %
		Y	4.95	66.72	16.12		150.0	
		Z	5.01	66.92	16.23		150.0	
10118-CAC	IEEE 802.11n (HT Mixed, 81 Mbps, 16-QAM)	X	5.49	67.59	16.65	0.00	150.0	± 9.6 %
		Y	5.30	67.10	16.32		150.0	
		Z	5.37	67.31	16.42		150.0	
10119-CAC	IEEE 802.11n (HT Mixed, 135 Mbps, 64-QAM)	X	5.20	67.40	16.47	0.00	150.0	± 9.6 %
		Y	5.04	66.96	16.17		150.0	
		Z	5.10	67.17	16.28		150.0	
10140-CAE	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	X	3.40	67.72	16.07	0.00	150.0	± 9.6 %
		Y	3.16	66.59	15.29		150.0	
		Z	3.27	67.10	15.62		150.0	
10141-CAE	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	X	3.53	67.80	16.23	0.00	150.0	± 9.6 %
		Y	3.29	66.75	15.49		150.0	
		Z	3.40	67.23	15.81		150.0	
10142-CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	X	2.06	69.35	16.19	0.00	150.0	± 9.6 %
		Y	1.65	66.31	14.02		150.0	
		Z	1.82	67.58	14.97		150.0	
10143-CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	X	2.54	69.44	16.13	0.00	150.0	± 9.6 %
		Y	2.10	66.83	14.10		150.0	
		Z	2.30	68.12	15.09		150.0	
10144-CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	X	2.29	67.05	14.48	0.00	150.0	± 9.6 %
		Y	1.92	64.88	12.62		150.0	
		Z	2.08	65.86	13.46		150.0	
10145-CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	X	1.29	65.99	12.36	0.00	150.0	± 9.6 %
		Y	0.85	61.51	8.56		150.0	
		Z	1.02	63.27	10.20		150.0	
10146-CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	X	2.10	67.27	12.22	0.00	150.0	± 9.6 %
		Y	1.36	62.63	8.83		150.0	
		Z	1.71	64.75	10.26		150.0	
10147-CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	X	2.58	69.80	13.51	0.00	150.0	± 9.6 %
		Y	1.47	63.38	9.34		150.0	
		Z	1.95	66.21	11.10		150.0	

EX3DV4- SN:7413

September 25, 2018

10149-CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	X	2.93	67.71	16.04	0.00	150.0	± 9.6 %
		Y	2.68	66.36	15.06		150.0	
		Z	2.79	66.96	15.49		150.0	
10150-CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	X	3.05	67.67	16.08	0.00	150.0	± 9.6 %
		Y	2.81	66.44	15.16		150.0	
		Z	2.91	67.00	15.57		150.0	
10151-CAG	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	X	8.73	83.44	23.83	3.98	65.0	± 9.6 %
		Y	6.65	79.25	22.19		65.0	
		Z	6.61	78.98	22.03		65.0	
10152-CAG	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	X	6.70	76.61	21.61	3.98	65.0	± 9.6 %
		Y	5.55	73.67	20.22		65.0	
		Z	5.56	73.55	20.15		65.0	
10153-CAG	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	X	7.10	77.55	22.36	3.98	65.0	± 9.6 %
		Y	5.94	74.73	21.05		65.0	
		Z	5.94	74.58	20.96		65.0	
10154-CAG	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	X	2.33	69.66	16.72	0.00	150.0	± 9.6 %
		Y	1.94	67.04	14.97		150.0	
		Z	2.10	68.12	15.72		150.0	
10155-CAG	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	X	2.65	68.55	16.34	0.00	150.0	± 9.6 %
		Y	2.34	66.85	15.00		150.0	
		Z	2.49	67.67	15.64		150.0	
10156-CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	X	1.92	69.56	16.05	0.00	150.0	± 9.6 %
		Y	1.47	65.89	13.42		150.0	
		Z	1.65	67.43	14.59		150.0	
10157-CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	X	2.14	67.76	14.59	0.00	150.0	± 9.6 %
		Y	1.70	64.86	12.24		150.0	
		Z	1.89	66.17	13.32		150.0	
10158-CAG	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	X	2.81	68.70	16.48	0.00	150.0	± 9.6 %
		Y	2.50	67.12	15.22		150.0	
		Z	2.64	67.92	15.83		150.0	
10159-CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	X	2.26	68.24	14.88	0.00	150.0	± 9.6 %
		Y	1.78	65.16	12.45		150.0	
		Z	1.99	66.59	13.58		150.0	
10160-CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	X	2.78	69.00	16.51	0.00	150.0	± 9.6 %
		Y	2.48	67.31	15.34		150.0	
		Z	2.60	67.99	15.83		150.0	
10161-CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	X	2.95	67.62	16.02	0.00	150.0	± 9.6 %
		Y	2.69	66.35	15.03		150.0	
		Z	2.81	66.94	15.47		150.0	
10162-CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	X	3.06	67.75	16.12	0.00	150.0	± 9.6 %
		Y	2.80	66.56	15.18		150.0	
		Z	2.92	67.12	15.60		150.0	
10166-CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	X	3.62	69.91	19.31	3.01	150.0	± 9.6 %
		Y	3.39	69.13	18.87		150.0	
		Z	3.59	70.09	19.30		150.0	
10167-CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	X	4.54	73.21	19.92	3.01	150.0	± 9.6 %
		Y	4.12	72.00	19.29		150.0	
		Z	4.64	73.93	20.08		150.0	

EX3DV4- SN:7413

September 25, 2018

10168-CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	X	5.07	75.55	21.26	3.01	150.0	± 9.6 %
		Y	4.61	74.46	20.73		150.0	
		Z	5.32	76.88	21.69		150.0	
10169-CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	X	3.04	69.71	19.26	3.01	150.0	± 9.6 %
		Y	2.81	68.10	18.39		150.0	
		Z	3.06	69.94	19.24		150.0	
10170-CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	X	4.37	76.54	21.90	3.01	150.0	± 9.6 %
		Y	3.75	73.68	20.61		150.0	
		Z	4.78	78.31	22.45		150.0	
10171-AAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	X	3.54	72.09	19.05	3.01	150.0	± 9.6 %
		Y	3.11	69.76	17.90		150.0	
		Z	3.66	72.72	19.11		150.0	
10172-CAG	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	X	23.39	113.60	36.11	6.02	65.0	± 9.6 %
		Y	6.34	87.79	28.08		65.0	
		Z	7.14	90.12	28.69		65.0	
10173-CAG	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	X	94.69	132.80	38.51	6.02	65.0	± 9.6 %
		Y	17.96	103.89	31.21		65.0	
		Z	39.25	117.30	34.48		65.0	
10174-CAG	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	X	52.23	119.67	34.57	6.02	65.0	± 9.6 %
		Y	11.25	94.23	27.69		65.0	
		Z	31.03	111.09	32.12		65.0	
10175-CAG	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	X	3.01	69.40	19.02	3.01	150.0	± 9.6 %
		Y	2.78	67.82	18.15		150.0	
		Z	3.02	69.59	18.98		150.0	
10176-CAG	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	X	4.38	76.56	21.91	3.01	150.0	± 9.6 %
		Y	3.76	73.70	20.62		150.0	
		Z	4.79	78.34	22.47		150.0	
10177-CAI	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	X	3.03	69.55	19.11	3.01	150.0	± 9.6 %
		Y	2.80	67.95	18.23		150.0	
		Z	3.05	69.75	19.07		150.0	
10178-CAG	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM)	X	4.33	76.33	21.79	3.01	150.0	± 9.6 %
		Y	3.73	73.53	20.53		150.0	
		Z	4.73	78.07	22.33		150.0	
10179-CAG	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	X	3.92	74.20	20.35	3.01	150.0	± 9.6 %
		Y	3.40	71.60	19.12		150.0	
		Z	4.15	75.30	20.61		150.0	
10180-CAG	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM)	X	3.53	72.02	19.00	3.01	150.0	± 9.6 %
		Y	3.10	69.71	17.86		150.0	
		Z	3.65	72.64	19.06		150.0	
10181-CAE	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	X	3.03	69.53	19.11	3.01	150.0	± 9.6 %
		Y	2.79	67.93	18.23		150.0	
		Z	3.04	69.73	19.06		150.0	
10182-CAE	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	X	4.32	76.31	21.78	3.01	150.0	± 9.6 %
		Y	3.73	73.51	20.52		150.0	
		Z	4.72	78.04	22.32		150.0	
10183-AAD	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	X	3.52	71.99	18.99	3.01	150.0	± 9.6 %
		Y	3.10	69.69	17.85		150.0	
		Z	3.64	72.62	19.05		150.0	

EX3DV4- SN:7413

September 25, 2018

10184-CAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	X	3.04	69.58	19.13	3.01	150.0	± 9.6 %
		Y	2.80	67.97	18.25		150.0	
		Z	3.05	69.78	19.09		150.0	
10185-CAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	X	4.35	76.38	21.82	3.01	150.0	± 9.6 %
		Y	3.74	73.58	20.55		150.0	
		Z	4.75	78.14	22.37		150.0	
10186-AAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	X	3.54	72.07	19.03	3.01	150.0	± 9.6 %
		Y	3.11	69.75	17.88		150.0	
		Z	3.66	72.69	19.09		150.0	
10187-CAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	X	3.05	69.63	19.19	3.01	150.0	± 9.6 %
		Y	2.81	68.03	18.32		150.0	
		Z	3.06	69.85	19.16		150.0	
10188-CAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	X	4.50	77.09	22.20	3.01	150.0	± 9.6 %
		Y	3.85	74.16	20.90		150.0	
		Z	4.95	79.02	22.82		150.0	
10189-AAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	X	3.62	72.52	19.31	3.01	150.0	± 9.6 %
		Y	3.17	70.13	18.14		150.0	
		Z	3.76	73.23	19.41		150.0	
10193-CAC	IEEE 802.11n (HT Greenfield, 6.5 Mbps, BPSK)	X	4.53	66.71	16.19	0.00	150.0	± 9.6 %
		Y	4.36	66.27	15.79		150.0	
		Z	4.43	66.50	15.96		150.0	
10194-CAC	IEEE 802.11n (HT Greenfield, 39 Mbps, 16-QAM)	X	4.70	67.02	16.31	0.00	150.0	± 9.6 %
		Y	4.51	66.55	15.93		150.0	
		Z	4.59	66.79	16.09		150.0	
10195-CAC	IEEE 802.11n (HT Greenfield, 65 Mbps, 64-QAM)	X	4.74	67.05	16.33	0.00	150.0	± 9.6 %
		Y	4.55	66.58	15.95		150.0	
		Z	4.63	66.82	16.11		150.0	
10196-CAC	IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)	X	4.53	66.77	16.21	0.00	150.0	± 9.6 %
		Y	4.35	66.30	15.79		150.0	
		Z	4.43	66.54	15.97		150.0	
10197-CAC	IEEE 802.11n (HT Mixed, 39 Mbps, 16-QAM)	X	4.71	67.04	16.33	0.00	150.0	± 9.6 %
		Y	4.52	66.56	15.94		150.0	
		Z	4.60	66.80	16.10		150.0	
10198-CAC	IEEE 802.11n (HT Mixed, 65 Mbps, 64-QAM)	X	4.74	67.07	16.34	0.00	150.0	± 9.6 %
		Y	4.55	66.59	15.96		150.0	
		Z	4.63	66.83	16.11		150.0	
10219-CAC	IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)	X	4.48	66.78	16.17	0.00	150.0	± 9.6 %
		Y	4.30	66.31	15.75		150.0	
		Z	4.38	66.55	15.93		150.0	
10220-CAC	IEEE 802.11n (HT Mixed, 43.3 Mbps, 16-QAM)	X	4.70	67.01	16.31	0.00	150.0	± 9.6 %
		Y	4.51	66.53	15.93		150.0	
		Z	4.60	66.77	16.08		150.0	
10221-CAC	IEEE 802.11n (HT Mixed, 72.2 Mbps, 64-QAM)	X	4.75	66.99	16.33	0.00	150.0	± 9.6 %
		Y	4.56	66.53	15.95		150.0	
		Z	4.64	66.77	16.10		150.0	
10222-CAC	IEEE 802.11n (HT Mixed, 15 Mbps, BPSK)	X	5.07	67.15	16.42	0.00	150.0	± 9.6 %
		Y	4.92	66.70	16.10		150.0	
		Z	4.98	66.92	16.22		150.0	

EX3DV4- SN:7413

September 25, 2018

10223-CAC	IEEE 802.11n (HT Mixed, 90 Mbps, 16-QAM)	X	5.36	67.33	16.53	0.00	150.0	± 9.6 %
		Y	5.21	66.95	16.25		150.0	
		Z	5.28	67.16	16.37		150.0	
10224-CAC	IEEE 802.11n (HT Mixed, 150 Mbps, 64-QAM)	X	5.11	67.26	16.40	0.00	150.0	± 9.6 %
		Y	4.95	66.80	16.08		150.0	
		Z	5.02	67.03	16.20		150.0	
10225-CAB	UMTS-FDD (HSPA+)	X	2.82	66.33	15.45	0.00	150.0	± 9.6 %
		Y	2.60	65.32	14.44		150.0	
		Z	2.70	65.81	14.90		150.0	
10226-CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	X	100.00	134.07	38.92	6.02	65.0	± 9.6 %
		Y	19.74	105.81	31.86		65.0	
		Z	46.04	120.45	35.40		65.0	
10227-CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	X	84.99	128.22	36.71	6.02	65.0	± 9.6 %
		Y	19.82	104.08	30.65		65.0	
		Z	42.57	116.55	33.59		65.0	
10228-CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	X	27.92	117.59	37.30	6.02	65.0	± 9.6 %
		Y	8.73	94.73	30.56		65.0	
		Z	11.03	99.17	31.75		65.0	
10229-CAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	X	95.41	132.92	38.55	6.02	65.0	± 9.6 %
		Y	18.10	104.01	31.25		65.0	
		Z	39.70	117.48	34.53		65.0	
10230-CAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	X	72.64	125.24	35.90	6.02	65.0	± 9.6 %
		Y	17.99	102.25	30.04		65.0	
		Z	36.59	113.74	32.78		65.0	
10231-CAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	X	25.59	115.63	36.67	6.02	65.0	± 9.6 %
		Y	8.31	93.62	30.11		65.0	
		Z	10.38	97.82	31.23		65.0	
10232-CAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM)	X	95.46	132.94	38.55	6.02	65.0	± 9.6 %
		Y	18.07	103.99	31.24		65.0	
		Z	39.64	117.47	34.53		65.0	
10233-CAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM)	X	72.47	125.22	35.90	6.02	65.0	± 9.6 %
		Y	17.93	102.20	30.03		65.0	
		Z	36.43	113.68	32.76		65.0	
10234-CAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	X	23.75	113.83	36.03	6.02	65.0	± 9.6 %
		Y	8.00	92.69	29.67		65.0	
		Z	9.89	96.67	30.73		65.0	
10235-CAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	X	96.22	133.11	38.59	6.02	65.0	± 9.6 %
		Y	18.12	104.05	31.27		65.0	
		Z	39.80	117.56	34.56		65.0	
10236-CAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	X	74.63	125.69	36.00	6.02	65.0	± 9.6 %
		Y	18.24	102.46	30.10		65.0	
		Z	37.31	114.04	32.85		65.0	
10237-CAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	X	25.84	115.88	36.74	6.02	65.0	± 9.6 %
		Y	8.32	93.69	30.14		65.0	
		Z	10.40	97.91	31.26		65.0	
10238-CAF	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	X	95.47	132.96	38.56	6.02	65.0	± 9.6 %
		Y	18.03	103.97	31.24		65.0	
		Z	39.55	117.45	34.52		65.0	

EX3DV4- SN:7413

September 25, 2018

10239-CAF	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	X	72.28	125.20	35.90	6.02	65.0	± 9.6 %
		Y	17.85	102.15	30.02		65.0	
		Z	36.26	113.62	32.75		65.0	
10240-CAF	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	X	25.72	115.79	36.72	6.02	65.0	± 9.6 %
		Y	8.30	93.65	30.12		65.0	
		Z	10.36	97.86	31.25		65.0	
10241-CAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	X	9.65	85.53	27.47	6.98	65.0	± 9.6 %
		Y	8.16	82.80	26.38		65.0	
		Z	8.52	83.49	26.51		65.0	
10242-CAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	X	9.07	84.20	26.88	6.98	65.0	± 9.6 %
		Y	7.14	79.98	25.15		65.0	
		Z	8.30	83.00	26.25		65.0	
10243-CAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	X	6.96	79.80	26.08	6.98	65.0	± 9.6 %
		Y	5.70	76.03	24.38		65.0	
		Z	5.63	75.51	24.02		65.0	
10244-CAC	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	X	8.41	81.52	20.94	3.98	65.0	± 9.6 %
		Y	5.74	75.85	18.12		65.0	
		Z	6.26	77.17	18.80		65.0	
10245-CAC	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	X	7.98	80.42	20.47	3.98	65.0	± 9.6 %
		Y	5.48	74.87	17.66		65.0	
		Z	5.97	76.17	18.34		65.0	
10246-CAC	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK)	X	10.90	89.52	24.06	3.98	65.0	± 9.6 %
		Y	5.41	78.63	19.50		65.0	
		Z	5.72	79.67	20.17		65.0	
10247-CAF	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	X	6.46	78.33	20.71	3.98	65.0	± 9.6 %
		Y	4.73	73.54	18.13		65.0	
		Z	4.85	73.96	18.53		65.0	
10248-CAF	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	X	6.29	77.34	20.29	3.98	65.0	± 9.6 %
		Y	4.65	72.78	17.77		65.0	
		Z	4.78	73.21	18.17		65.0	
10249-CAF	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	X	12.77	92.98	26.19	3.98	65.0	± 9.6 %
		Y	7.09	83.49	22.48		65.0	
		Z	7.11	83.55	22.67		65.0	
10250-CAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	X	7.20	80.17	23.07	3.98	65.0	± 9.6 %
		Y	5.69	76.44	21.30		65.0	
		Z	5.69	76.32	21.30		65.0	
10251-CAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	X	6.65	77.37	21.57	3.98	65.0	± 9.6 %
		Y	5.35	74.06	19.88		65.0	
		Z	5.38	74.00	19.92		65.0	
10252-CAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	X	10.47	89.10	25.88	3.98	65.0	± 9.6 %
		Y	7.05	82.76	23.44		65.0	
		Z	6.95	82.34	23.30		65.0	
10253-CAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	X	6.50	75.90	21.30	3.98	65.0	± 9.6 %
		Y	5.45	73.20	19.95		65.0	
		Z	5.45	73.05	19.89		65.0	
10254-CAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	X	6.88	76.80	21.99	3.98	65.0	± 9.6 %
		Y	5.80	74.15	20.68		65.0	
		Z	5.80	73.99	20.62		65.0	

EX3DV4- SN:7413

September 25, 2018

10255-CAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	X	8.11	82.37	23.67	3.98	65.0	± 9.6 %
		Y	6.29	78.44	22.05		65.0	
		Z	6.24	78.11	21.89		65.0	
10256-CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	X	6.32	76.57	17.95	3.98	65.0	± 9.6 %
		Y	3.97	70.17	14.49		65.0	
		Z	4.39	71.61	15.33		65.0	
10257-CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	X	5.90	75.19	17.29	3.98	65.0	± 9.6 %
		Y	3.77	69.15	13.91		65.0	
		Z	4.17	70.53	14.74		65.0	
10258-CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	X	7.33	82.37	20.68	3.98	65.0	± 9.6 %
		Y	3.56	71.85	15.67		65.0	
		Z	3.95	73.52	16.73		65.0	
10259-CAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	X	6.76	79.02	21.56	3.98	65.0	± 9.6 %
		Y	5.13	74.76	19.33		65.0	
		Z	5.20	74.92	19.56		65.0	
10260-CAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	X	6.70	78.49	21.35	3.98	65.0	± 9.6 %
		Y	5.13	74.38	19.16		65.0	
		Z	5.20	74.56	19.41		65.0	
10261-CAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	X	10.53	89.56	25.50	3.98	65.0	± 9.6 %
		Y	6.62	82.08	22.47		65.0	
		Z	6.57	81.89	22.52		65.0	
10262-CAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	X	7.18	80.11	23.03	3.98	65.0	± 9.6 %
		Y	5.68	76.37	21.24		65.0	
		Z	5.68	76.26	21.25		65.0	
10263-CAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	X	6.63	77.34	21.56	3.98	65.0	± 9.6 %
		Y	5.34	74.03	19.87		65.0	
		Z	5.37	73.98	19.91		65.0	
10264-CAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	X	10.33	88.82	25.76	3.98	65.0	± 9.6 %
		Y	6.97	82.52	23.32		65.0	
		Z	6.87	82.10	23.19		65.0	
10265-CAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	X	6.70	76.61	21.61	3.98	65.0	± 9.6 %
		Y	5.55	73.67	20.22		65.0	
		Z	5.56	73.56	20.16		65.0	
10266-CAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	X	7.10	77.53	22.35	3.98	65.0	± 9.6 %
		Y	5.93	74.71	21.04		65.0	
		Z	5.93	74.57	20.95		65.0	
10267-CAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	X	8.71	83.38	23.80	3.98	65.0	± 9.6 %
		Y	6.63	79.20	22.16		65.0	
		Z	6.60	78.93	22.01		65.0	
10268-CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	X	7.17	75.90	21.68	3.98	65.0	± 9.6 %
		Y	6.16	73.50	20.59		65.0	
		Z	6.17	73.40	20.49		65.0	
10269-CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	X	7.09	75.35	21.50	3.98	65.0	± 9.6 %
		Y	6.14	73.08	20.45		65.0	
		Z	6.15	72.97	20.35		65.0	
10270-CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	X	7.67	78.81	22.16	3.98	65.0	± 9.6 %
		Y	6.35	75.98	21.01		65.0	
		Z	6.35	75.83	20.88		65.0	

Certificate No: EX3-7413_Sep18

Page 22 of 39

Report File No : F690501/RF-SAR002704

Date of Issue : 2019-01-08

(All SGS services are rendered in accordance with the applicable SGS conditions of service available on request and accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>.)

RTT5041-76(2015.10.01) (2)

A4 (210mm x 297mm)

EX3DV4- SN:7413

September 25, 2018

10274-CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.10)	X	2.61	66.78	15.41	0.00	150.0	± 9.6 %
		Y	2.40	65.62	14.31		150.0	
		Z	2.49	66.16	14.81		150.0	
10275-CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.4)	X	1.67	68.69	16.01	0.00	150.0	± 9.6 %
		Y	1.37	65.61	13.82		150.0	
		Z	1.48	66.80	14.73		150.0	
10277-CAA	PHS (QPSK)	X	2.17	62.10	7.54	9.03	50.0	± 9.6 %
		Y	1.88	60.98	6.50		50.0	
		Z	1.95	61.23	6.78		50.0	
10278-CAA	PHS (QPSK, BW 884MHz, Rolloff 0.5)	X	12.02	86.99	21.36	9.03	50.0	± 9.6 %
		Y	4.64	72.74	15.30		50.0	
		Z	5.23	74.54	16.27		50.0	
10279-CAA	PHS (QPSK, BW 884MHz, Rolloff 0.38)	X	12.30	87.29	21.53	9.03	50.0	± 9.6 %
		Y	4.79	73.08	15.51		50.0	
		Z	5.39	74.89	16.48		50.0	
10290-AAB	CDMA2000, RC1, SO55, Full Rate	X	1.58	70.12	14.69	0.00	150.0	± 9.6 %
		Y	0.93	63.75	10.30		150.0	
		Z	1.16	66.22	12.20		150.0	
10291-AAB	CDMA2000, RC3, SO55, Full Rate	X	0.89	66.99	13.16	0.00	150.0	± 9.6 %
		Y	0.56	61.88	8.98		150.0	
		Z	0.69	63.81	10.79		150.0	
10292-AAB	CDMA2000, RC3, SO32, Full Rate	X	1.29	73.01	16.30	0.00	150.0	± 9.6 %
		Y	0.62	63.59	10.25		150.0	
		Z	0.83	66.86	12.73		150.0	
10293-AAB	CDMA2000, RC3, SO3, Full Rate	X	2.44	82.42	20.40	0.00	150.0	± 9.6 %
		Y	0.79	66.28	12.06		150.0	
		Z	1.21	71.93	15.49		150.0	
10295-AAB	CDMA2000, RC1, SO3, 1/8th Rate 25 fr.	X	15.41	96.52	28.44	9.03	50.0	± 9.6 %
		Y	17.39	96.44	27.32		50.0	
		Z	12.32	90.98	25.84		50.0	
10297-AAD	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	X	2.81	70.20	16.88	0.00	150.0	± 9.6 %
		Y	2.41	67.82	15.44		150.0	
		Z	2.57	68.80	16.04		150.0	
10298-AAD	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, QPSK)	X	1.65	68.60	14.62	0.00	150.0	± 9.6 %
		Y	1.13	63.90	11.12		150.0	
		Z	1.33	65.83	12.67		150.0	
10299-AAD	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	X	2.89	70.98	14.91	0.00	150.0	± 9.6 %
		Y	1.96	66.13	11.78		150.0	
		Z	2.54	69.05	13.41		150.0	
10300-AAD	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	X	2.03	65.64	11.70	0.00	150.0	± 9.6 %
		Y	1.55	62.92	9.44		150.0	
		Z	1.83	64.42	10.48		150.0	
10301-AAA	IEEE 802.16e WiMAX (29:18, 5ms, 10MHz, QPSK, PUSC)	X	4.89	66.04	17.74	4.17	50.0	± 9.6 %
		Y	4.72	65.90	17.43		50.0	
		Z	4.71	65.67	17.40		50.0	
10302-AAA	IEEE 802.16e WiMAX (29:18, 5ms, 10MHz, QPSK, PUSC, 3 CTRL symbols)	X	5.35	66.65	18.48	4.96	50.0	± 9.6 %
		Y	5.11	66.10	17.92		50.0	
		Z	5.13	66.03	17.97		50.0	

EX3DV4- SN:7413

September 25, 2018

10303-AAA	IEEE 802.16e WiMAX (31:15, 5ms, 10MHz, 64QAM, PUSC)	X	5.11	66.33	18.34	4.96	50.0	± 9.6 %
		Y	4.87	65.75	17.73		50.0	
		Z	4.89	65.66	17.79		50.0	
10304-AAA	IEEE 802.16e WiMAX (29:18, 5ms, 10MHz, 64QAM, PUSC)	X	4.91	66.14	17.78	4.17	50.0	± 9.6 %
		Y	4.67	65.58	17.19		50.0	
		Z	4.70	65.53	17.28		50.0	
10305-AAA	IEEE 802.16e WiMAX (31:15, 10ms, 10MHz, 64QAM, PUSC, 15 symbols)	X	4.59	68.53	20.21	6.02	35.0	± 9.6 %
		Y	4.40	68.07	19.35		35.0	
		Z	4.34	67.55	19.31		35.0	
10306-AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, 64QAM, PUSC, 18 symbols)	X	4.85	67.27	19.64	6.02	35.0	± 9.6 %
		Y	4.68	66.96	19.01		35.0	
		Z	4.65	66.58	18.96		35.0	
10307-AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, QPSK, PUSC, 18 symbols)	X	4.76	67.50	19.64	6.02	35.0	± 9.6 %
		Y	4.57	67.06	18.94		35.0	
		Z	4.54	66.70	18.90		35.0	
10308-AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, 16QAM, PUSC)	X	4.75	67.74	19.80	6.02	35.0	± 9.6 %
		Y	4.56	67.30	19.10		35.0	
		Z	4.53	66.91	19.05		35.0	
10309-AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, 16QAM, AMC 2x3, 18 symbols)	X	4.92	67.53	19.81	6.02	35.0	± 9.6 %
		Y	4.72	67.12	19.14		35.0	
		Z	4.70	66.76	19.09		35.0	
10310-AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, QPSK, AMC 2x3, 18 symbols)	X	4.81	67.35	19.63	6.02	35.0	± 9.6 %
		Y	4.63	67.03	18.99		35.0	
		Z	4.60	66.64	18.94		35.0	
10311-AAD	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	X	3.18	69.46	16.50	0.00	150.0	± 9.6 %
		Y	2.76	67.20	15.20		150.0	
		Z	2.93	68.15	15.75		150.0	
10313-AAA	iDEN 1:3	X	10.55	87.80	21.52	6.99	70.0	± 9.6 %
		Y	4.36	76.88	17.80		70.0	
		Z	4.32	76.98	17.91		70.0	
10314-AAA	iDEN 1:6	X	22.03	106.52	30.50	10.00	30.0	± 9.6 %
		Y	6.79	87.25	24.62		30.0	
		Z	7.08	87.93	24.86		30.0	
10315-AAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 96pc duty cycle)	X	1.10	64.34	15.61	0.17	150.0	± 9.6 %
		Y	1.00	62.56	13.99		150.0	
		Z	1.04	63.18	14.59		150.0	
10316-AAB	IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 96pc duty cycle)	X	4.59	66.81	16.38	0.17	150.0	± 9.6 %
		Y	4.42	66.36	15.99		150.0	
		Z	4.49	66.55	16.12		150.0	
10317-AAC	IEEE 802.11a WiFi 5 GHz (OFDM, 6 Mbps, 96pc duty cycle)	X	4.59	66.81	16.38	0.17	150.0	± 9.6 %
		Y	4.42	66.36	15.99		150.0	
		Z	4.49	66.55	16.12		150.0	
10400-AAD	IEEE 802.11ac WiFi (20MHz, 64-QAM, 99pc duty cycle)	X	4.69	67.08	16.32	0.00	150.0	± 9.6 %
		Y	4.49	66.58	15.92		150.0	
		Z	4.57	66.82	16.07		150.0	
10401-AAD	IEEE 802.11ac WiFi (40MHz, 64-QAM, 99pc duty cycle)	X	5.38	67.22	16.45	0.00	150.0	± 9.6 %
		Y	5.22	66.80	16.15		150.0	
		Z	5.28	66.98	16.25		150.0	

EX3DV4- SN:7413

September 25, 2018

10402-AAD	IEEE 802.11ac WiFi (80MHz, 64-QAM, 99pc duty cycle)	X	5.63	67.54	16.46	0.00	150.0	± 9.6 %
		Y	5.48	67.08	16.16		150.0	
		Z	5.54	67.30	16.27		150.0	
10403-AAB	CDMA2000 (1xEV-DO, Rev. 0)	X	1.58	70.12	14.69	0.00	115.0	± 9.6 %
		Y	0.93	63.75	10.30		115.0	
		Z	1.16	66.22	12.20		115.0	
10404-AAB	CDMA2000 (1xEV-DO, Rev. A)	X	1.58	70.12	14.69	0.00	115.0	± 9.6 %
		Y	0.93	63.75	10.30		115.0	
		Z	1.16	66.22	12.20		115.0	
10406-AAB	CDMA2000, RC3, SO32, SCH0, Full Rate	X	100.00	121.45	30.20	0.00	100.0	± 9.6 %
		Y	39.42	106.90	26.01		100.0	
		Z	100.00	116.01	27.58		100.0	
10410-AAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK, UL Subframe=2,3,4,7,8,9, Subframe Conf=4)	X	100.00	124.72	31.61	3.23	80.0	± 9.6 %
		Y	100.00	126.56	32.14		80.0	
		Z	100.00	124.31	31.11		80.0	
10415-AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 99pc duty cycle)	X	1.01	63.32	14.91	0.00	150.0	± 9.6 %
		Y	0.93	61.82	13.40		150.0	
		Z	0.98	62.45	14.03		150.0	
10416-AAA	IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 99pc duty cycle)	X	4.53	66.74	16.26	0.00	150.0	± 9.6 %
		Y	4.36	66.30	15.87		150.0	
		Z	4.43	66.53	16.03		150.0	
10417-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 99pc duty cycle)	X	4.53	66.74	16.26	0.00	150.0	± 9.6 %
		Y	4.36	66.30	15.87		150.0	
		Z	4.43	66.53	16.03		150.0	
10418-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps, 99pc duty cycle, Long preamble)	X	4.52	66.91	16.28	0.00	150.0	± 9.6 %
		Y	4.35	66.46	15.90		150.0	
		Z	4.42	66.69	16.06		150.0	
10419-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps, 99pc duty cycle, Short preamble)	X	4.54	66.86	16.28	0.00	150.0	± 9.6 %
		Y	4.37	66.41	15.89		150.0	
		Z	4.44	66.64	16.05		150.0	
10422-AAB	IEEE 802.11n (HT Greenfield, 7.2 Mbps, BPSK)	X	4.66	66.85	16.29	0.00	150.0	± 9.6 %
		Y	4.48	66.41	15.92		150.0	
		Z	4.56	66.64	16.07		150.0	
10423-AAB	IEEE 802.11n (HT Greenfield, 43.3 Mbps, 16-QAM)	X	4.82	67.16	16.40	0.00	150.0	± 9.6 %
		Y	4.62	66.69	16.02		150.0	
		Z	4.71	66.93	16.18		150.0	
10424-AAB	IEEE 802.11n (HT Greenfield, 72.2 Mbps, 64-QAM)	X	4.74	67.12	16.38	0.00	150.0	± 9.6 %
		Y	4.54	66.64	15.99		150.0	
		Z	4.63	66.88	16.15		150.0	
10425-AAB	IEEE 802.11n (HT Greenfield, 15 Mbps, BPSK)	X	5.33	67.37	16.53	0.00	150.0	± 9.6 %
		Y	5.17	66.95	16.23		150.0	
		Z	5.23	67.13	16.33		150.0	
10426-AAB	IEEE 802.11n (HT Greenfield, 90 Mbps, 16-QAM)	X	5.34	67.40	16.53	0.00	150.0	± 9.6 %
		Y	5.19	67.03	16.27		150.0	
		Z	5.25	67.20	16.36		150.0	

EX3DV4- SN:7413

September 25, 2018

10427-AAB	IEEE 802.11n (HT Greenfield, 150 Mbps, 64-QAM)	X	5.35	67.38	16.52	0.00	150.0	± 9.6 %
		Y	5.17	66.92	16.20		150.0	
		Z	5.25	67.15	16.33		150.0	
10430-AAD	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1)	X	4.25	70.86	18.18	0.00	150.0	± 9.6 %
		Y	3.95	70.17	17.39		150.0	
		Z	4.15	70.83	17.92		150.0	
10431-AAD	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1)	X	4.21	67.33	16.27	0.00	150.0	± 9.6 %
		Y	3.97	66.72	15.70		150.0	
		Z	4.08	67.02	15.95		150.0	
10432-AAC	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1)	X	4.51	67.18	16.33	0.00	150.0	± 9.6 %
		Y	4.30	66.66	15.89		150.0	
		Z	4.39	66.92	16.07		150.0	
10433-AAC	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1)	X	4.76	67.15	16.40	0.00	150.0	± 9.6 %
		Y	4.56	66.67	16.01		150.0	
		Z	4.65	66.91	16.17		150.0	
10434-AAA	W-CDMA (BS Test Model 1, 64 DPCH)	X	4.36	71.77	18.17	0.00	150.0	± 9.6 %
		Y	3.97	70.72	17.12		150.0	
		Z	4.23	71.64	17.82		150.0	
10435-AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	100.00	124.50	31.51	3.23	80.0	± 9.6 %
		Y	100.00	126.33	32.03		80.0	
		Z	100.00	124.07	30.99		80.0	
10447-AAD	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	X	3.52	67.40	15.63	0.00	150.0	± 9.6 %
		Y	3.20	66.35	14.59		150.0	
		Z	3.34	66.87	15.07		150.0	
10448-AAD	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1, Clipping 44%)	X	4.05	67.11	16.14	0.00	150.0	± 9.6 %
		Y	3.83	66.50	15.56		150.0	
		Z	3.93	66.81	15.81		150.0	
10449-AAC	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1, Clipping 44%)	X	4.32	67.01	16.24	0.00	150.0	± 9.6 %
		Y	4.13	66.47	15.77		150.0	
		Z	4.22	66.74	15.97		150.0	
10450-AAC	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	X	4.52	66.93	16.26	0.00	150.0	± 9.6 %
		Y	4.34	66.43	15.85		150.0	
		Z	4.42	66.68	16.02		150.0	
10451-AAA	W-CDMA (BS Test Model 1, 64 DPCH, Clipping 44%)	X	3.41	67.59	15.25	0.00	150.0	± 9.6 %
		Y	3.02	66.19	13.94		150.0	
		Z	3.20	66.88	14.55		150.0	
10456-AAB	IEEE 802.11ac WiFi (160MHz, 64-QAM, 99pc duty cycle)	X	6.19	67.91	16.67	0.00	150.0	± 9.6 %
		Y	6.11	67.67	16.50		150.0	
		Z	6.13	67.75	16.53		150.0	
10457-AAA	UMTS-FDD (DC-HSDPA)	X	3.78	65.39	15.97	0.00	150.0	± 9.6 %
		Y	3.68	65.00	15.57		150.0	
		Z	3.73	65.20	15.73		150.0	
10458-AAA	CDMA2000 (1xEV-DO, Rev. B, 2 carriers)	X	4.01	71.10	17.60	0.00	150.0	± 9.6 %
		Y	3.55	69.58	16.15		150.0	
		Z	3.83	70.70	17.02		150.0	
10459-AAA	CDMA2000 (1xEV-DO, Rev. B, 3 carriers)	X	5.05	68.33	18.07	0.00	150.0	± 9.6 %
		Y	4.82	68.20	17.59		150.0	
		Z	4.98	68.57	17.96		150.0	

EX3DV4- SN:7413

September 25, 2018

10460-AAA	UMTS-FDD (WCDMA, AMR)	X	0.96	69.56	16.94	0.00	150.0	± 9.6 %
		Y	0.71	64.53	13.34		150.0	
		Z	0.80	66.21	14.73		150.0	
10461-AAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	100.00	131.08	34.55	3.29	80.0	± 9.6 %
		Y	100.00	131.23	34.36		80.0	
		Z	100.00	130.22	33.85		80.0	
10462-AAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	100.00	108.36	23.96	3.23	80.0	± 9.6 %
		Y	5.36	78.49	16.13		80.0	
		Z	5.52	77.71	15.34		80.0	
10463-AAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	13.34	85.09	17.45	3.23	80.0	± 9.6 %
		Y	1.49	65.04	10.66		80.0	
		Z	1.36	63.93	9.77		80.0	
10464-AAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	100.00	128.51	33.18	3.23	80.0	± 9.6 %
		Y	100.00	128.38	32.86		80.0	
		Z	100.00	127.25	32.30		80.0	
10465-AAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	100.00	107.59	23.60	3.23	80.0	± 9.6 %
		Y	3.21	73.22	14.34		80.0	
		Z	3.12	72.21	13.49		80.0	
10466-AAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	5.81	77.13	15.10	3.23	80.0	± 9.6 %
		Y	1.31	63.78	10.06		80.0	
		Z	1.21	62.87	9.25		80.0	
10467-AAE	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	100.00	128.82	33.32	3.23	80.0	± 9.6 %
		Y	100.00	128.73	33.01		80.0	
		Z	100.00	127.60	32.46		80.0	
10468-AAE	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	100.00	107.83	23.71	3.23	80.0	± 9.6 %
		Y	3.64	74.56	14.81		80.0	
		Z	3.56	73.52	13.96		80.0	
10469-AAE	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	5.95	77.36	15.17	3.23	80.0	± 9.6 %
		Y	1.31	63.82	10.08		80.0	
		Z	1.21	62.90	9.26		80.0	
10470-AAE	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	100.00	128.87	33.33	3.23	80.0	± 9.6 %
		Y	100.00	128.77	33.02		80.0	
		Z	100.00	127.64	32.47		80.0	
10471-AAE	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	100.00	107.74	23.66	3.23	80.0	± 9.6 %
		Y	3.59	74.40	14.75		80.0	
		Z	3.49	73.31	13.87		80.0	
10472-AAE	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	5.83	77.15	15.09	3.23	80.0	± 9.6 %
		Y	1.30	63.76	10.03		80.0	
		Z	1.20	62.83	9.22		80.0	
10473-AAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	100.00	128.83	33.32	3.23	80.0	± 9.6 %
		Y	100.00	128.73	33.00		80.0	
		Z	100.00	127.60	32.45		80.0	
10474-AAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	100.00	107.75	23.66	3.23	80.0	± 9.6 %
		Y	3.55	74.29	14.71		80.0	
		Z	3.44	73.19	13.83		80.0	
10475-AAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	5.73	77.01	15.05	3.23	80.0	± 9.6 %
		Y	1.30	63.72	10.02		80.0	
		Z	1.20	62.81	9.21		80.0	

EX3DV4- SN:7413

September 25, 2018

10477-AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	100.00	107.51	23.55	3.23	80.0	± 9.6 %
		Y	3.22	73.26	14.33		80.0	
		Z	3.09	72.15	13.45		80.0	
10478-AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	5.57	76.72	14.94	3.23	80.0	± 9.6 %
		Y	1.29	63.66	9.98		80.0	
		Z	1.19	62.74	9.17		80.0	
10479-AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	20.22	102.43	28.33	3.23	80.0	± 9.6 %
		Y	14.12	96.98	26.33		80.0	
		Z	13.97	96.66	26.28		80.0	
10480-AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	25.60	98.12	24.83	3.23	80.0	± 9.6 %
		Y	11.88	87.54	21.18		80.0	
		Z	15.83	91.00	22.19		80.0	
10481-AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	16.81	91.34	22.50	3.23	80.0	± 9.6 %
		Y	7.47	80.76	18.62		80.0	
		Z	9.52	83.56	19.52		80.0	
10482-AAB	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	5.43	80.93	20.45	2.23	80.0	± 9.6 %
		Y	2.31	69.11	14.99		80.0	
		Z	2.69	71.25	16.26		80.0	
10483-AAB	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	7.46	81.02	19.81	2.23	80.0	± 9.6 %
		Y	3.79	71.88	15.63		80.0	
		Z	4.60	74.47	16.88		80.0	
10484-AAB	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	6.47	78.84	19.06	2.23	80.0	± 9.6 %
		Y	3.45	70.43	15.04		80.0	
		Z	4.13	72.83	16.26		80.0	
10485-AAE	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	5.13	80.55	21.37	2.23	80.0	± 9.6 %
		Y	2.88	72.04	17.45		80.0	
		Z	3.10	73.10	18.13		80.0	
10486-AAE	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	4.08	73.33	18.03	2.23	80.0	± 9.6 %
		Y	2.73	67.84	14.95		80.0	
		Z	2.96	68.95	15.73		80.0	
10487-AAE	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.00	72.63	17.73	2.23	80.0	± 9.6 %
		Y	2.73	67.43	14.74		80.0	
		Z	2.95	68.51	15.51		80.0	
10488-AAE	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	4.62	77.03	20.75	2.23	80.0	± 9.6 %
		Y	3.20	71.56	18.23		80.0	
		Z	3.33	72.09	18.53		80.0	
10489-AAE	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	3.95	71.37	18.47	2.23	80.0	± 9.6 %
		Y	3.20	68.42	16.79		80.0	
		Z	3.29	68.76	17.05		80.0	
10490-AAE	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.01	71.04	18.33	2.23	80.0	± 9.6 %
		Y	3.29	68.28	16.74		80.0	
		Z	3.38	68.60	16.99		80.0	
10491-AAE	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	4.50	74.08	19.69	2.23	80.0	± 9.6 %
		Y	3.47	70.22	17.86		80.0	
		Z	3.57	70.62	18.06		80.0	
10492-AAE	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	4.16	69.97	18.15	2.23	80.0	± 9.6 %
		Y	3.56	67.77	16.91		80.0	
		Z	3.63	68.00	17.06		80.0	

EX3DV4- SN:7413

September 25, 2018

10493-AAE	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.22	69.76	18.07	2.23	80.0	± 9.6 %
		Y	3.62	67.66	16.86		80.0	
		Z	3.69	67.87	17.01		80.0	
10494-AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	5.10	76.29	20.39	2.23	80.0	± 9.6 %
		Y	3.72	71.55	18.30		80.0	
		Z	3.86	72.09	18.54		80.0	
10495-AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	4.22	70.45	18.38	2.23	80.0	± 9.6 %
		Y	3.58	68.06	17.10		80.0	
		Z	3.65	68.32	17.25		80.0	
10496-AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.27	70.03	18.23	2.23	80.0	± 9.6 %
		Y	3.66	67.84	17.04		80.0	
		Z	3.73	68.08	17.18		80.0	
10497-AAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	3.91	75.68	17.46	2.23	80.0	± 9.6 %
		Y	1.45	63.46	11.16		80.0	
		Z	1.79	65.97	12.81		80.0	
10498-AAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	2.10	64.95	11.88	2.23	80.0	± 9.6 %
		Y	1.26	60.00	8.20		80.0	
		Z	1.34	60.44	8.90		80.0	
10499-AAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	1.97	63.98	11.26	2.23	80.0	± 9.6 %
		Y	1.28	60.00	8.05		80.0	
		Z	1.31	60.00	8.51		80.0	
10500-AAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	4.70	78.39	20.88	2.23	80.0	± 9.6 %
		Y	2.99	71.70	17.72		80.0	
		Z	3.15	72.42	18.20		80.0	
10501-AAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	4.02	72.49	18.17	2.23	80.0	± 9.6 %
		Y	2.97	68.33	15.77		80.0	
		Z	3.13	69.03	16.30		80.0	
10502-AAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.05	72.22	18.00	2.23	80.0	± 9.6 %
		Y	3.02	68.16	15.62		80.0	
		Z	3.18	68.87	16.16		80.0	
10503-AAE	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	4.55	76.79	20.65	2.23	80.0	± 9.6 %
		Y	3.16	71.37	18.14		80.0	
		Z	3.29	71.89	18.44		80.0	
10504-AAE	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	3.93	71.27	18.41	2.23	80.0	± 9.6 %
		Y	3.18	68.32	16.73		80.0	
		Z	3.27	68.66	16.99		80.0	
10505-AAE	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	3.99	70.94	18.28	2.23	80.0	± 9.6 %
		Y	3.27	68.19	16.68		80.0	
		Z	3.36	68.51	16.93		80.0	
10506-AAE	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	5.05	76.11	20.31	2.23	80.0	± 9.6 %
		Y	3.69	71.42	18.22		80.0	
		Z	3.83	71.94	18.46		80.0	
10507-AAE	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	4.20	70.39	18.35	2.23	80.0	± 9.6 %
		Y	3.57	68.00	17.06		80.0	
		Z	3.64	68.26	17.21		80.0	

EX3DV4- SN:7413

September 25, 2018

10508-AAE	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.25	69.96	18.19	2.23	80.0	± 9.6 %
		Y	3.65	67.78	17.00		80.0	
		Z	3.72	68.01	17.13		80.0	
10509-AAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	5.11	73.77	19.36	2.23	80.0	± 9.6 %
		Y	4.07	70.36	17.80		80.0	
		Z	4.19	70.79	17.98		80.0	
10510-AAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	4.64	69.77	18.15	2.23	80.0	± 9.6 %
		Y	4.05	67.77	17.11		80.0	
		Z	4.12	68.01	17.21		80.0	
10511-AAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.67	69.41	18.03	2.23	80.0	± 9.6 %
		Y	4.12	67.57	17.06		80.0	
		Z	4.18	67.78	17.15		80.0	
10512-AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	5.63	76.15	20.14	2.23	80.0	± 9.6 %
		Y	4.20	71.68	18.21		80.0	
		Z	4.37	72.28	18.46		80.0	
10513-AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	4.55	70.17	18.32	2.23	80.0	± 9.6 %
		Y	3.93	67.94	17.19		80.0	
		Z	4.00	68.22	17.31		80.0	
10514-AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.53	69.60	18.13	2.23	80.0	± 9.6 %
		Y	3.97	67.59	17.08		80.0	
		Z	4.03	67.83	17.19		80.0	
10515-AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 99pc duty cycle)	X	0.98	63.53	14.99	0.00	150.0	± 9.6 %
		Y	0.89	61.91	13.38		150.0	
		Z	0.94	62.58	14.05		150.0	
10516-AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 99pc duty cycle)	X	0.69	73.37	18.90	0.00	150.0	± 9.6 %
		Y	0.42	64.50	12.96		150.0	
		Z	0.49	66.76	14.93		150.0	
10517-AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 99pc duty cycle)	X	0.84	65.76	15.82	0.00	150.0	± 9.6 %
		Y	0.71	62.86	13.34		150.0	
		Z	0.77	63.92	14.32		150.0	
10518-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 99pc duty cycle)	X	4.52	66.82	16.24	0.00	150.0	± 9.6 %
		Y	4.35	66.37	15.84		150.0	
		Z	4.42	66.60	16.01		150.0	
10519-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 99pc duty cycle)	X	4.70	67.05	16.35	0.00	150.0	± 9.6 %
		Y	4.50	66.57	15.95		150.0	
		Z	4.59	66.81	16.12		150.0	
10520-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps, 99pc duty cycle)	X	4.56	67.01	16.28	0.00	150.0	± 9.6 %
		Y	4.36	66.49	15.85		150.0	
		Z	4.45	66.75	16.03		150.0	
10521-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps, 99pc duty cycle)	X	4.49	67.01	16.27	0.00	150.0	± 9.6 %
		Y	4.29	66.47	15.83		150.0	
		Z	4.38	66.73	16.01		150.0	
10522-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 99pc duty cycle)	X	4.55	67.11	16.35	0.00	150.0	± 9.6 %
		Y	4.35	66.60	15.94		150.0	
		Z	4.44	66.86	16.11		150.0	

Certificate No: EX3-7413_Sep18

Page 30 of 39

EX3DV4- SN:7413

September 25, 2018

10523-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 99pc duty cycle)	X	4.43	66.98	16.21	0.00	150.0	± 9.6 %
		Y	4.25	66.51	15.80		150.0	
		Z	4.33	66.75	15.97		150.0	
10524-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 99pc duty cycle)	X	4.49	67.02	16.32	0.00	150.0	± 9.6 %
		Y	4.29	66.52	15.91		150.0	
		Z	4.38	66.77	16.08		150.0	
10525-AAB	IEEE 802.11ac WiFi (20MHz, MCS0, 99pc duty cycle)	X	4.48	66.08	15.92	0.00	150.0	± 9.6 %
		Y	4.31	65.60	15.52		150.0	
		Z	4.39	65.85	15.68		150.0	
10526-AAB	IEEE 802.11ac WiFi (20MHz, MCS1, 99pc duty cycle)	X	4.65	66.44	16.06	0.00	150.0	± 9.6 %
		Y	4.44	65.91	15.64		150.0	
		Z	4.53	66.18	15.82		150.0	
10527-AAB	IEEE 802.11ac WiFi (20MHz, MCS2, 99pc duty cycle)	X	4.57	66.41	16.00	0.00	150.0	± 9.6 %
		Y	4.37	65.86	15.57		150.0	
		Z	4.46	66.13	15.75		150.0	
10528-AAB	IEEE 802.11ac WiFi (20MHz, MCS3, 99pc duty cycle)	X	4.59	66.42	16.03	0.00	150.0	± 9.6 %
		Y	4.38	65.88	15.61		150.0	
		Z	4.47	66.15	15.78		150.0	
10529-AAB	IEEE 802.11ac WiFi (20MHz, MCS4, 99pc duty cycle)	X	4.59	66.42	16.03	0.00	150.0	± 9.6 %
		Y	4.38	65.88	15.61		150.0	
		Z	4.47	66.15	15.78		150.0	
10531-AAB	IEEE 802.11ac WiFi (20MHz, MCS6, 99pc duty cycle)	X	4.58	66.52	16.04	0.00	150.0	± 9.6 %
		Y	4.35	65.92	15.59		150.0	
		Z	4.45	66.21	15.78		150.0	
10532-AAB	IEEE 802.11ac WiFi (20MHz, MCS7, 99pc duty cycle)	X	4.44	66.38	15.98	0.00	150.0	± 9.6 %
		Y	4.23	65.77	15.52		150.0	
		Z	4.32	66.07	15.71		150.0	
10533-AAB	IEEE 802.11ac WiFi (20MHz, MCS8, 99pc duty cycle)	X	4.60	66.48	16.03	0.00	150.0	± 9.6 %
		Y	4.39	65.94	15.60		150.0	
		Z	4.48	66.21	15.78		150.0	
10534-AAB	IEEE 802.11ac WiFi (40MHz, MCS0, 99pc duty cycle)	X	5.12	66.49	16.07	0.00	150.0	± 9.6 %
		Y	4.95	66.01	15.74		150.0	
		Z	5.02	66.25	15.87		150.0	
10535-AAB	IEEE 802.11ac WiFi (40MHz, MCS1, 99pc duty cycle)	X	5.18	66.66	16.15	0.00	150.0	± 9.6 %
		Y	5.00	66.17	15.81		150.0	
		Z	5.08	66.42	15.95		150.0	
10536-AAB	IEEE 802.11ac WiFi (40MHz, MCS2, 99pc duty cycle)	X	5.05	66.63	16.11	0.00	150.0	± 9.6 %
		Y	4.88	66.13	15.77		150.0	
		Z	4.95	66.37	15.90		150.0	
10537-AAB	IEEE 802.11ac WiFi (40MHz, MCS3, 99pc duty cycle)	X	5.11	66.59	16.09	0.00	150.0	± 9.6 %
		Y	4.94	66.10	15.76		150.0	
		Z	5.01	66.34	15.89		150.0	
10538-AAB	IEEE 802.11ac WiFi (40MHz, MCS4, 99pc duty cycle)	X	5.20	66.60	16.14	0.00	150.0	± 9.6 %
		Y	5.01	66.11	15.80		150.0	
		Z	5.09	66.34	15.93		150.0	
10540-AAB	IEEE 802.11ac WiFi (40MHz, MCS6, 99pc duty cycle)	X	5.13	66.61	16.16	0.00	150.0	± 9.6 %
		Y	4.94	66.08	15.80		150.0	
		Z	5.02	66.33	15.94		150.0	

EX3DV4- SN:7413

September 25, 2018

10541-AAB	IEEE 802.11ac WiFi (40MHz, MCS7, 99pc duty cycle)	X	5.10	66.49	16.09	0.00	150.0	± 9.6 %
		Y	4.93	65.98	15.74		150.0	
		Z	5.00	66.24	15.88		150.0	
10542-AAB	IEEE 802.11ac WiFi (40MHz, MCS8, 99pc duty cycle)	X	5.26	66.56	16.14	0.00	150.0	± 9.6 %
		Y	5.08	66.09	15.82		150.0	
		Z	5.16	66.32	15.94		150.0	
10543-AAB	IEEE 802.11ac WiFi (40MHz, MCS9, 99pc duty cycle)	X	5.33	66.58	16.17	0.00	150.0	± 9.6 %
		Y	5.15	66.12	15.86		150.0	
		Z	5.22	66.34	15.98		150.0	
10544-AAB	IEEE 802.11ac WiFi (80MHz, MCS0, 99pc duty cycle)	X	5.43	66.61	16.06	0.00	150.0	± 9.6 %
		Y	5.28	66.14	15.76		150.0	
		Z	5.35	66.37	15.88		150.0	
10545-AAB	IEEE 802.11ac WiFi (80MHz, MCS1, 99pc duty cycle)	X	5.61	66.99	16.20	0.00	150.0	± 9.6 %
		Y	5.46	66.57	15.93		150.0	
		Z	5.52	66.76	16.02		150.0	
10546-AAB	IEEE 802.11ac WiFi (80MHz, MCS2, 99pc duty cycle)	X	5.49	66.80	16.13	0.00	150.0	± 9.6 %
		Y	5.32	66.27	15.79		150.0	
		Z	5.39	66.53	15.92		150.0	
10547-AAB	IEEE 802.11ac WiFi (80MHz, MCS3, 99pc duty cycle)	X	5.56	66.84	16.14	0.00	150.0	± 9.6 %
		Y	5.40	66.35	15.83		150.0	
		Z	5.46	66.59	15.94		150.0	
10548-AAB	IEEE 802.11ac WiFi (80MHz, MCS4, 99pc duty cycle)	X	5.76	67.65	16.52	0.00	150.0	± 9.6 %
		Y	5.58	67.09	16.17		150.0	
		Z	5.64	67.31	16.28		150.0	
10550-AAB	IEEE 802.11ac WiFi (80MHz, MCS6, 99pc duty cycle)	X	5.51	66.82	16.15	0.00	150.0	± 9.6 %
		Y	5.37	66.40	15.87		150.0	
		Z	5.42	66.59	15.96		150.0	
10551-AAB	IEEE 802.11ac WiFi (80MHz, MCS7, 99pc duty cycle)	X	5.52	66.87	16.13	0.00	150.0	± 9.6 %
		Y	5.34	66.30	15.78		150.0	
		Z	5.42	66.58	15.92		150.0	
10552-AAB	IEEE 802.11ac WiFi (80MHz, MCS8, 99pc duty cycle)	X	5.44	66.68	16.05	0.00	150.0	± 9.6 %
		Y	5.29	66.21	15.74		150.0	
		Z	5.36	66.45	15.86		150.0	
10553-AAB	IEEE 802.11ac WiFi (80MHz, MCS9, 99pc duty cycle)	X	5.52	66.71	16.09	0.00	150.0	± 9.6 %
		Y	5.36	66.21	15.77		150.0	
		Z	5.43	66.46	15.90		150.0	
10554-AAC	IEEE 802.11ac WiFi (160MHz, MCS0, 99pc duty cycle)	X	5.83	66.96	16.14	0.00	150.0	± 9.6 %
		Y	5.70	66.50	15.86		150.0	
		Z	5.76	66.73	15.97		150.0	
10555-AAC	IEEE 802.11ac WiFi (160MHz, MCS1, 99pc duty cycle)	X	5.95	67.24	16.26	0.00	150.0	± 9.6 %
		Y	5.81	66.76	15.97		150.0	
		Z	5.87	66.99	16.08		150.0	
10556-AAC	IEEE 802.11ac WiFi (160MHz, MCS2, 99pc duty cycle)	X	5.97	67.29	16.28	0.00	150.0	± 9.6 %
		Y	5.84	66.84	16.00		150.0	
		Z	5.89	67.05	16.10		150.0	
10557-AAC	IEEE 802.11ac WiFi (160MHz, MCS3, 99pc duty cycle)	X	5.94	67.20	16.26	0.00	150.0	± 9.6 %
		Y	5.79	66.71	15.96		150.0	
		Z	5.85	66.95	16.07		150.0	

EX3DV4- SN:7413

September 25, 2018

10558-AAC	IEEE 802.11ac WiFi (160MHz, MCS4, 99pc duty cycle)	X	5.99	67.35	16.35	0.00	150.0	± 9.6 %
		Y	5.83	66.84	16.04		150.0	
		Z	5.89	67.09	16.16		150.0	
10560-AAC	IEEE 802.11ac WiFi (160MHz, MCS6, 99pc duty cycle)	X	5.99	67.22	16.32	0.00	150.0	± 9.6 %
		Y	5.83	66.72	16.02		150.0	
		Z	5.89	66.96	16.13		150.0	
10561-AAC	IEEE 802.11ac WiFi (160MHz, MCS7, 99pc duty cycle)	X	5.91	67.18	16.34	0.00	150.0	± 9.6 %
		Y	5.76	66.71	16.04		150.0	
		Z	5.82	66.93	16.15		150.0	
10562-AAC	IEEE 802.11ac WiFi (160MHz, MCS8, 99pc duty cycle)	X	6.02	67.53	16.51	0.00	150.0	± 9.6 %
		Y	5.84	66.94	16.16		150.0	
		Z	5.91	67.21	16.29		150.0	
10563-AAC	IEEE 802.11ac WiFi (160MHz, MCS9, 99pc duty cycle)	X	6.20	67.70	16.55	0.00	150.0	± 9.6 %
		Y	5.92	66.86	16.08		150.0	
		Z	6.00	67.12	16.21		150.0	
10564-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 9 Mbps, 99pc duty cycle)	X	4.85	66.90	16.41	0.46	150.0	± 9.6 %
		Y	4.68	66.49	16.04		150.0	
		Z	4.75	66.68	16.18		150.0	
10565-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 12 Mbps, 99pc duty cycle)	X	5.07	67.33	16.71	0.46	150.0	± 9.6 %
		Y	4.88	66.90	16.36		150.0	
		Z	4.96	67.11	16.49		150.0	
10566-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 18 Mbps, 99pc duty cycle)	X	4.91	67.19	16.54	0.46	150.0	± 9.6 %
		Y	4.71	66.72	16.16		150.0	
		Z	4.80	66.94	16.30		150.0	
10567-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 24 Mbps, 99pc duty cycle)	X	4.94	67.57	16.88	0.46	150.0	± 9.6 %
		Y	4.74	67.09	16.51		150.0	
		Z	4.83	67.34	16.66		150.0	
10568-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 36 Mbps, 99pc duty cycle)	X	4.82	66.99	16.33	0.46	150.0	± 9.6 %
		Y	4.63	66.51	15.94		150.0	
		Z	4.71	66.71	16.06		150.0	
10569-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 48 Mbps, 99pc duty cycle)	X	4.90	67.67	16.95	0.46	150.0	± 9.6 %
		Y	4.72	67.25	16.61		150.0	
		Z	4.80	67.48	16.76		150.0	
10570-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 54 Mbps, 99pc duty cycle)	X	4.93	67.51	16.88	0.46	150.0	± 9.6 %
		Y	4.73	67.07	16.52		150.0	
		Z	4.82	67.30	16.67		150.0	
10571-AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 90pc duty cycle)	X	1.21	65.23	16.13	0.46	130.0	± 9.6 %
		Y	1.07	63.17	14.42		130.0	
		Z	1.11	63.71	14.95		130.0	
10572-AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 90pc duty cycle)	X	1.22	65.86	16.52	0.46	130.0	± 9.6 %
		Y	1.08	63.61	14.71		130.0	
		Z	1.12	64.20	15.26		130.0	
10573-AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 90pc duty cycle)	X	5.69	103.43	29.16	0.46	130.0	± 9.6 %
		Y	0.94	72.79	17.37		130.0	
		Z	1.18	76.82	19.70		130.0	
10574-AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 90pc duty cycle)	X	1.41	72.65	19.91	0.46	130.0	± 9.6 %
		Y	1.07	67.50	16.70		130.0	
		Z	1.16	68.81	17.68		130.0	

EX3DV4- SN:7413

September 25, 2018

10575-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps, 90pc duty cycle)	X	4.64	66.72	16.48	0.46	130.0	± 9.6 %
		Y	4.47	66.30	16.11		130.0	
		Z	4.54	66.48	16.23		130.0	
10576-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 9 Mbps, 90pc duty cycle)	X	4.67	66.89	16.55	0.46	130.0	± 9.6 %
		Y	4.50	66.48	16.18		130.0	
		Z	4.56	66.65	16.30		130.0	
10577-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 12 Mbps, 90pc duty cycle)	X	4.86	67.17	16.71	0.46	130.0	± 9.6 %
		Y	4.67	66.73	16.34		130.0	
		Z	4.75	66.92	16.46		130.0	
10578-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 18 Mbps, 90pc duty cycle)	X	4.76	67.32	16.81	0.46	130.0	± 9.6 %
		Y	4.57	66.85	16.43		130.0	
		Z	4.65	67.06	16.56		130.0	
10579-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 24 Mbps, 90pc duty cycle)	X	4.53	66.64	16.15	0.46	130.0	± 9.6 %
		Y	4.33	66.11	15.72		130.0	
		Z	4.41	66.30	15.84		130.0	
10580-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 36 Mbps, 90pc duty cycle)	X	4.58	66.69	16.18	0.46	130.0	± 9.6 %
		Y	4.38	66.19	15.76		130.0	
		Z	4.45	66.37	15.88		130.0	
10581-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 48 Mbps, 90pc duty cycle)	X	4.66	67.38	16.76	0.46	130.0	± 9.6 %
		Y	4.47	66.90	16.38		130.0	
		Z	4.55	67.10	16.51		130.0	
10582-AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 54 Mbps, 90pc duty cycle)	X	4.47	66.41	15.95	0.46	130.0	± 9.6 %
		Y	4.27	65.89	15.51		130.0	
		Z	4.34	66.06	15.62		130.0	
10583-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 90pc duty cycle)	X	4.64	66.72	16.48	0.46	130.0	± 9.6 %
		Y	4.47	66.30	16.11		130.0	
		Z	4.54	66.48	16.23		130.0	
10584-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 90pc duty cycle)	X	4.67	66.89	16.55	0.46	130.0	± 9.6 %
		Y	4.50	66.48	16.18		130.0	
		Z	4.56	66.65	16.30		130.0	
10585-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 90pc duty cycle)	X	4.86	67.17	16.71	0.46	130.0	± 9.6 %
		Y	4.67	66.73	16.34		130.0	
		Z	4.75	66.92	16.46		130.0	
10586-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps, 90pc duty cycle)	X	4.76	67.32	16.81	0.46	130.0	± 9.6 %
		Y	4.57	66.85	16.43		130.0	
		Z	4.65	67.06	16.56		130.0	
10587-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps, 90pc duty cycle)	X	4.53	66.64	16.15	0.46	130.0	± 9.6 %
		Y	4.33	66.11	15.72		130.0	
		Z	4.41	66.30	15.84		130.0	
10588-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 90pc duty cycle)	X	4.58	66.69	16.18	0.46	130.0	± 9.6 %
		Y	4.38	66.19	15.76		130.0	
		Z	4.45	66.37	15.88		130.0	
10589-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 90pc duty cycle)	X	4.66	67.38	16.76	0.46	130.0	± 9.6 %
		Y	4.47	66.90	16.38		130.0	
		Z	4.55	67.10	16.51		130.0	
10590-AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 90pc duty cycle)	X	4.47	66.41	15.95	0.46	130.0	± 9.6 %
		Y	4.27	65.89	15.51		130.0	
		Z	4.34	66.06	15.62		130.0	

EX3DV4- SN:7413

September 25, 2018

10591-AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS0, 90pc duty cycle)	X	4.79	66.77	16.58	0.46	130.0	± 9.6 %
		Y	4.63	66.39	16.23		130.0	
		Z	4.69	66.55	16.34		130.0	
10592-AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS1, 90pc duty cycle)	X	4.94	67.11	16.71	0.46	130.0	± 9.6 %
		Y	4.76	66.69	16.36		130.0	
		Z	4.83	66.87	16.47		130.0	
10593-AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS2, 90pc duty cycle)	X	4.86	67.02	16.59	0.46	130.0	± 9.6 %
		Y	4.67	66.57	16.22		130.0	
		Z	4.75	66.76	16.34		130.0	
10594-AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS3, 90pc duty cycle)	X	4.92	67.18	16.74	0.46	130.0	± 9.6 %
		Y	4.73	66.75	16.38		130.0	
		Z	4.80	66.93	16.50		130.0	
10595-AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS4, 90pc duty cycle)	X	4.88	67.14	16.64	0.46	130.0	± 9.6 %
		Y	4.69	66.71	16.29		130.0	
		Z	4.77	66.89	16.40		130.0	
10596-AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS5, 90pc duty cycle)	X	4.82	67.14	16.65	0.46	130.0	± 9.6 %
		Y	4.63	66.69	16.28		130.0	
		Z	4.70	66.87	16.39		130.0	
10597-AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS6, 90pc duty cycle)	X	4.77	67.05	16.53	0.46	130.0	± 9.6 %
		Y	4.58	66.56	16.14		130.0	
		Z	4.65	66.76	16.26		130.0	
10598-AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS7, 90pc duty cycle)	X	4.75	67.27	16.78	0.46	130.0	± 9.6 %
		Y	4.56	66.77	16.39		130.0	
		Z	4.64	66.98	16.52		130.0	
10599-AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS0, 90pc duty cycle)	X	5.45	67.29	16.76	0.46	130.0	± 9.6 %
		Y	5.30	66.91	16.49		130.0	
		Z	5.35	67.03	16.55		130.0	
10600-AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS1, 90pc duty cycle)	X	5.56	67.63	16.91	0.46	130.0	± 9.6 %
		Y	5.42	67.30	16.67		130.0	
		Z	5.46	67.38	16.69		130.0	
10601-AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS2, 90pc duty cycle)	X	5.47	67.42	16.82	0.46	130.0	± 9.6 %
		Y	5.31	67.06	16.56		130.0	
		Z	5.36	67.18	16.61		130.0	
10602-AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS3, 90pc duty cycle)	X	5.57	67.47	16.77	0.46	130.0	± 9.6 %
		Y	5.45	67.25	16.58		130.0	
		Z	5.49	67.33	16.61		130.0	
10603-AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS4, 90pc duty cycle)	X	5.64	67.74	17.03	0.46	130.0	± 9.6 %
		Y	5.52	67.53	16.85		130.0	
		Z	5.54	67.56	16.85		130.0	
10604-AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS5, 90pc duty cycle)	X	5.47	67.28	16.79	0.46	130.0	± 9.6 %
		Y	5.40	67.19	16.66		130.0	
		Z	5.42	67.21	16.67		130.0	
10605-AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS6, 90pc duty cycle)	X	5.56	67.55	16.93	0.46	130.0	± 9.6 %
		Y	5.42	67.23	16.68		130.0	
		Z	5.46	67.33	16.72		130.0	
10606-AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS7, 90pc duty cycle)	X	5.32	66.95	16.49	0.46	130.0	± 9.6 %
		Y	5.17	66.56	16.20		130.0	
		Z	5.21	66.66	16.24		130.0	

Certificate No: EX3-7413_Sep18

Page 35 of 39

Report File No : F690501/RF-SAR002704

Date of Issue : 2019-01-08

(All SGS services are rendered in accordance with the applicable SGS conditions of service available on request and accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>.)

RTT5041-76(2015.10.01) (2)

A4 (210mm x 297mm)

EX3DV4- SN:7413

September 25, 2018

10607-AAB	IEEE 802.11ac WiFi (20MHz, MCS0, 90pc duty cycle)	X	4.63	66.11	16.21	0.46	130.0	± 9.6 %
		Y	4.46	65.69	15.85		130.0	
		Z	4.53	65.87	15.97		130.0	
10608-AAB	IEEE 802.11ac WiFi (20MHz, MCS1, 90pc duty cycle)	X	4.82	66.51	16.38	0.46	130.0	± 9.6 %
		Y	4.62	66.04	16.00		130.0	
		Z	4.70	66.24	16.12		130.0	
10609-AAB	IEEE 802.11ac WiFi (20MHz, MCS2, 90pc duty cycle)	X	4.71	66.37	16.22	0.46	130.0	± 9.6 %
		Y	4.51	65.87	15.82		130.0	
		Z	4.59	66.08	15.95		130.0	
10610-AAB	IEEE 802.11ac WiFi (20MHz, MCS3, 90pc duty cycle)	X	4.76	66.52	16.38	0.46	130.0	± 9.6 %
		Y	4.56	66.03	15.99		130.0	
		Z	4.64	66.24	16.12		130.0	
10611-AAB	IEEE 802.11ac WiFi (20MHz, MCS4, 90pc duty cycle)	X	4.67	66.33	16.23	0.46	130.0	± 9.6 %
		Y	4.47	65.84	15.84		130.0	
		Z	4.55	66.04	15.96		130.0	
10612-AAB	IEEE 802.11ac WiFi (20MHz, MCS5, 90pc duty cycle)	X	4.68	66.50	16.28	0.46	130.0	± 9.6 %
		Y	4.47	65.97	15.87		130.0	
		Z	4.55	66.18	16.00		130.0	
10613-AAB	IEEE 802.11ac WiFi (20MHz, MCS6, 90pc duty cycle)	X	4.68	66.37	16.16	0.46	130.0	± 9.6 %
		Y	4.47	65.81	15.74		130.0	
		Z	4.55	66.04	15.87		130.0	
10614-AAB	IEEE 802.11ac WiFi (20MHz, MCS7, 90pc duty cycle)	X	4.63	66.55	16.38	0.46	130.0	± 9.6 %
		Y	4.42	66.00	15.96		130.0	
		Z	4.51	66.24	16.11		130.0	
10615-AAB	IEEE 802.11ac WiFi (20MHz, MCS8, 90pc duty cycle)	X	4.67	66.18	16.02	0.46	130.0	± 9.6 %
		Y	4.47	65.69	15.62		130.0	
		Z	4.55	65.88	15.74		130.0	
10616-AAB	IEEE 802.11ac WiFi (40MHz, MCS0, 90pc duty cycle)	X	5.27	66.55	16.38	0.46	130.0	± 9.6 %
		Y	5.11	66.11	16.07		130.0	
		Z	5.17	66.30	16.17		130.0	
10617-AAB	IEEE 802.11ac WiFi (40MHz, MCS1, 90pc duty cycle)	X	5.34	66.72	16.44	0.46	130.0	± 9.6 %
		Y	5.17	66.30	16.15		130.0	
		Z	5.24	66.47	16.23		130.0	
10618-AAB	IEEE 802.11ac WiFi (40MHz, MCS2, 90pc duty cycle)	X	5.23	66.74	16.46	0.46	130.0	± 9.6 %
		Y	5.07	66.32	16.17		130.0	
		Z	5.13	66.49	16.25		130.0	
10619-AAB	IEEE 802.11ac WiFi (40MHz, MCS3, 90pc duty cycle)	X	5.24	66.54	16.30	0.46	130.0	± 9.6 %
		Y	5.08	66.11	16.00		130.0	
		Z	5.14	66.28	16.08		130.0	
10620-AAB	IEEE 802.11ac WiFi (40MHz, MCS4, 90pc duty cycle)	X	5.33	66.58	16.37	0.46	130.0	± 9.6 %
		Y	5.16	66.14	16.07		130.0	
		Z	5.22	66.31	16.15		130.0	
10621-AAB	IEEE 802.11ac WiFi (40MHz, MCS5, 90pc duty cycle)	X	5.33	66.70	16.54	0.46	130.0	± 9.6 %
		Y	5.17	66.27	16.25		130.0	
		Z	5.24	66.47	16.34		130.0	
10622-AAB	IEEE 802.11ac WiFi (40MHz, MCS6, 90pc duty cycle)	X	5.34	66.86	16.61	0.46	130.0	± 9.6 %
		Y	5.16	66.37	16.29		130.0	
		Z	5.24	66.62	16.41		130.0	

EX3DV4- SN:7413

September 25, 2018

10623-AAB	IEEE 802.11ac WiFi (40MHz, MCS7, 90pc duty cycle)	X	5.22	66.41	16.27	0.46	130.0	± 9.6 %
		Y	5.05	65.93	15.94		130.0	
		Z	5.12	66.14	16.04		130.0	
10624-AAB	IEEE 802.11ac WiFi (40MHz, MCS8, 90pc duty cycle)	X	5.41	66.59	16.42	0.46	130.0	± 9.6 %
		Y	5.25	66.17	16.13		130.0	
		Z	5.31	66.35	16.21		130.0	
10625-AAB	IEEE 802.11ac WiFi (40MHz, MCS9, 90pc duty cycle)	X	5.75	67.47	16.91	0.46	130.0	± 9.6 %
		Y	5.43	66.62	16.41		130.0	
		Z	5.56	66.99	16.59		130.0	
10626-AAB	IEEE 802.11ac WiFi (80MHz, MCS0, 90pc duty cycle)	X	5.57	66.61	16.33	0.46	130.0	± 9.6 %
		Y	5.44	66.19	16.06		130.0	
		Z	5.49	66.38	16.14		130.0	
10627-AAB	IEEE 802.11ac WiFi (80MHz, MCS1, 90pc duty cycle)	X	5.79	67.13	16.55	0.46	130.0	± 9.6 %
		Y	5.67	66.78	16.33		130.0	
		Z	5.70	66.90	16.37		130.0	
10628-AAB	IEEE 802.11ac WiFi (80MHz, MCS2, 90pc duty cycle)	X	5.60	66.69	16.28	0.46	130.0	± 9.6 %
		Y	5.44	66.20	15.96		130.0	
		Z	5.50	66.40	16.05		130.0	
10629-AAB	IEEE 802.11ac WiFi (80MHz, MCS3, 90pc duty cycle)	X	5.67	66.74	16.29	0.46	130.0	± 9.6 %
		Y	5.53	66.32	16.02		130.0	
		Z	5.57	66.48	16.08		130.0	
10630-AAB	IEEE 802.11ac WiFi (80MHz, MCS4, 90pc duty cycle)	X	6.04	68.05	16.95	0.46	130.0	± 9.6 %
		Y	5.84	67.46	16.59		130.0	
		Z	5.88	67.61	16.65		130.0	
10631-AAB	IEEE 802.11ac WiFi (80MHz, MCS5, 90pc duty cycle)	X	5.98	67.95	17.08	0.46	130.0	± 9.6 %
		Y	5.77	67.36	16.73		130.0	
		Z	5.85	67.60	16.84		130.0	
10632-AAB	IEEE 802.11ac WiFi (80MHz, MCS6, 90pc duty cycle)	X	5.76	67.19	16.72	0.46	130.0	± 9.6 %
		Y	5.65	66.88	16.51		130.0	
		Z	5.68	67.00	16.55		130.0	
10633-AAB	IEEE 802.11ac WiFi (80MHz, MCS7, 90pc duty cycle)	X	5.67	66.87	16.39	0.46	130.0	± 9.6 %
		Y	5.51	66.40	16.10		130.0	
		Z	5.57	66.61	16.18		130.0	
10634-AAB	IEEE 802.11ac WiFi (80MHz, MCS8, 90pc duty cycle)	X	5.65	66.89	16.46	0.46	130.0	± 9.6 %
		Y	5.49	66.42	16.16		130.0	
		Z	5.55	66.63	16.25		130.0	
10635-AAB	IEEE 802.11ac WiFi (80MHz, MCS9, 90pc duty cycle)	X	5.53	66.25	15.89	0.46	130.0	± 9.6 %
		Y	5.36	65.75	15.56		130.0	
		Z	5.42	65.94	15.64		130.0	
10636-AAC	IEEE 802.11ac WiFi (160MHz, MCS0, 90pc duty cycle)	X	5.98	66.96	16.41	0.46	130.0	± 9.6 %
		Y	5.87	66.57	16.16		130.0	
		Z	5.91	66.74	16.23		130.0	
10637-AAC	IEEE 802.11ac WiFi (160MHz, MCS1, 90pc duty cycle)	X	6.13	67.32	16.57	0.46	130.0	± 9.6 %
		Y	6.00	66.92	16.32		130.0	
		Z	6.04	67.09	16.39		130.0	
10638-AAC	IEEE 802.11ac WiFi (160MHz, MCS2, 90pc duty cycle)	X	6.13	67.31	16.55	0.46	130.0	± 9.6 %
		Y	6.01	66.91	16.30		130.0	
		Z	6.05	67.08	16.36		130.0	

Certificate No: EX3-7413_Sep18

Page 37 of 39

Report File No : F690501/RF-SAR002704

Date of Issue : 2019-01-08

(All SGS services are rendered in accordance with the applicable SGS conditions of service available on request and accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>.)

RTT5041-76(2015.10.01) (2)

A4 (210mm x 297mm)

EX3DV4- SN:7413

September 25, 2018

10639-AAC	IEEE 802.11ac WiFi (160MHz, MCS3, 90pc duty cycle)	X	6.11	67.26	16.57	0.46	130.0	± 9.6 %
		Y	5.97	66.82	16.30		130.0	
		Z	6.02	67.01	16.37		130.0	
10640-AAC	IEEE 802.11ac WiFi (160MHz, MCS4, 90pc duty cycle)	X	6.11	67.28	16.52	0.46	130.0	± 9.6 %
		Y	5.97	66.81	16.23		130.0	
		Z	6.02	67.00	16.31		130.0	
10641-AAC	IEEE 802.11ac WiFi (160MHz, MCS5, 90pc duty cycle)	X	6.16	67.18	16.49	0.46	130.0	± 9.6 %
		Y	6.04	66.81	16.26		130.0	
		Z	6.08	66.95	16.30		130.0	
10642-AAC	IEEE 802.11ac WiFi (160MHz, MCS6, 90pc duty cycle)	X	6.20	67.42	16.77	0.46	130.0	± 9.6 %
		Y	6.06	67.00	16.51		130.0	
		Z	6.11	67.19	16.59		130.0	
10643-AAC	IEEE 802.11ac WiFi (160MHz, MCS7, 90pc duty cycle)	X	6.04	67.12	16.52	0.46	130.0	± 9.6 %
		Y	5.91	66.72	16.27		130.0	
		Z	5.95	66.87	16.33		130.0	
10644-AAC	IEEE 802.11ac WiFi (160MHz, MCS8, 90pc duty cycle)	X	6.19	67.58	16.78	0.46	130.0	± 9.6 %
		Y	6.00	66.99	16.42		130.0	
		Z	6.06	67.23	16.52		130.0	
10645-AAC	IEEE 802.11ac WiFi (160MHz, MCS9, 90pc duty cycle)	X	6.45	67.98	16.93	0.46	130.0	± 9.6 %
		Y	6.13	67.06	16.43		130.0	
		Z	6.19	67.25	16.50		130.0	
10646-AAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Subframe=2,7)	X	76.90	147.03	48.69	9.30	60.0	± 9.6 %
		Y	15.83	110.01	38.63		60.0	
		Z	17.24	111.04	38.54		60.0	
10647-AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Subframe=2,7)	X	56.16	140.36	47.20	9.30	60.0	± 9.6 %
		Y	13.25	106.35	37.62		60.0	
		Z	14.31	107.27	37.51		60.0	
10648-AAA	CDMA2000 (1x Advanced)	X	0.70	64.06	11.11	0.00	150.0	± 9.6 %
		Y	0.49	60.68	7.77		150.0	
		Z	0.58	62.06	9.29		150.0	
10652-AAD	LTE-TDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	X	3.82	67.93	17.16	2.23	80.0	± 9.6 %
		Y	3.39	66.35	16.07		80.0	
		Z	3.46	66.60	16.29		80.0	
10653-AAD	LTE-TDD (OFDMA, 10 MHz, E-TM 3.1, Clipping 44%)	X	4.28	66.97	17.14	2.23	80.0	± 9.6 %
		Y	3.94	65.81	16.37		80.0	
		Z	4.00	65.97	16.49		80.0	
10654-AAD	LTE-TDD (OFDMA, 15 MHz, E-TM 3.1, Clipping 44%)	X	4.25	66.57	17.11	2.23	80.0	± 9.6 %
		Y	3.95	65.46	16.40		80.0	
		Z	3.99	65.62	16.50		80.0	
10655-AAE	LTE-TDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	X	4.31	66.55	17.14	2.23	80.0	± 9.6 %
		Y	4.02	65.42	16.44		80.0	
		Z	4.05	65.58	16.53		80.0	
10658-AAA	Pulse Waveform (200Hz, 10%)	X	100.00	113.12	27.05	10.00	50.0	± 9.6 %
		Y	100.00	111.85	26.41		50.0	
		Z	100.00	111.59	26.33		50.0	
10659-AAA	Pulse Waveform (200Hz, 20%)	X	100.00	111.97	25.70	6.99	60.0	± 9.6 %
		Y	100.00	110.52	24.79		60.0	
		Z	100.00	110.73	24.88		60.0	

EX3DV4- SN:7413

September 25, 2018

10660-AAA	Pulse Waveform (200Hz, 40%)	X	100.00	113.56	25.24	3.98	80.0	± 9.6 %
		Y	100.00	109.61	23.11		80.0	
		Z	100.00	111.53	23.93		80.0	
10661-AAA	Pulse Waveform (200Hz, 60%)	X	100.00	118.18	26.00	2.22	100.0	± 9.6 %
		Y	100.00	107.17	20.93		100.0	
		Z	100.00	113.13	23.41		100.0	
10662-AAA	Pulse Waveform (200Hz, 80%)	X	100.00	129.82	28.88	0.97	120.0	± 9.6 %
		Y	100.00	93.72	14.22		120.0	
		Z	100.00	111.54	21.23		120.0	
10670-AAA	Bluetooth Low Energy	X	100.00	119.81	27.11	2.19	100.0	± 9.6 %
		Y	100.00	112.39	23.42		100.0	
		Z	100.00	117.76	25.68		100.0	

^ε Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Appendix B.2 Calibration certificate for DAE(S/N: 1507)

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **SGS Korea (Dymstec)**

Certificate No: **DAE4-1507_Sep18**

CALIBRATION CERTIFICATE

Object **DAE4 - SD 000 D04 BM - SN: 1507**

Calibration procedure(s) **QA CAL-06.v29**
Calibration procedure for the data acquisition electronics (DAE)

Calibration date: **September 19, 2018**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Keithley Multimeter Type 2001	SN: 0810278	03-Sep-18 (No.23488)	Sep-19
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Auto DAE Calibration Unit	SE UWS 053 AA 1001	04-Jan-18 (in house check)	In house check: Jan-19
Calibrator Box V2.1	SE UMS 006 AA 1002	04-Jan-18 (in house check)	In house check: Jan-19

Calibrated by:	Name Dominique Steffen	Function Laboratory Technician	Signature 
Approved by:	Sven Kühn	Deputy Manager	

Issued: September 19, 2018

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: DAE4-1507_Sep18

Page 1 of 5

Calibration Laboratory of
Schmid & Partner
Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Glossary

DAE data acquisition electronics
Connector angle information used in DASY system to align probe sensor X to the robot coordinate system.

Methods Applied and Interpretation of Parameters

- **DC Voltage Measurement:** Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- **Connector angle:** The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.
 - **DC Voltage Measurement Linearity:** Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
 - **Common mode sensitivity:** Influence of a positive or negative common mode voltage on the differential measurement.
 - **Channel separation:** Influence of a voltage on the neighbor channels not subject to an input voltage.
 - **AD Converter Values with inputs shorted:** Values on the internal AD converter corresponding to zero input voltage
 - **Input Offset Measurement:** Output voltage and statistical results over a large number of zero voltage measurements.
 - **Input Offset Current:** Typical value for information; Maximum channel input offset current, not considering the input resistance.
 - **Input resistance:** Typical value for information; DAE input resistance at the connector, during internal auto-zeroing and during measurement.
 - **Low Battery Alarm Voltage:** Typical value for information. Below this voltage, a battery alarm signal is generated.
 - **Power consumption:** Typical value for information. Supply currents in various operating modes.

DC Voltage Measurement

A/D - Converter Resolution nominal

High Range: 1LSB = 6.1 μ V, full range = -100...+300 mV

Low Range: 1LSB = 61nV, full range = -1.....+3mV

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Calibration Factors	X	Y	Z
High Range	404.395 $\pm$ 0.02% (k=2)	404.370 $\pm$ 0.02% (k=2)	404.111 $\pm$ 0.02% (k=2)
Low Range	3.98054 $\pm$ 1.50% (k=2)	3.99165 $\pm$ 1.50% (k=2)	3.98650 $\pm$ 1.50% (k=2)

Connector Angle

Connector Angle to be used in DASY system	105.0 $^{\circ}$ $\pm$ 1 $^{\circ}$
---	-------------------------------------

Appendix (Additional assessments outside the scope of SCS0108)

1. DC Voltage Linearity

High Range	Reading (μV)	Difference (μV)	Error (%)
Channel X + Input	199997.76	-0.95	-0.00
Channel X + Input	20002.80	1.19	0.01
Channel X - Input	-20000.04	1.75	-0.01
Channel Y + Input	199998.72	0.28	0.00
Channel Y + Input	20000.45	-1.16	-0.01
Channel Y - Input	-20002.80	-1.06	0.01
Channel Z + Input	199999.52	1.05	0.00
Channel Z + Input	20000.12	-1.31	-0.01
Channel Z - Input	-20001.31	0.57	-0.00

Low Range	Reading (μV)	Difference (μV)	Error (%)
Channel X + Input	2000.46	-0.41	-0.02
Channel X + Input	200.88	-0.49	-0.24
Channel X - Input	-199.18	-0.55	0.28
Channel Y + Input	2000.53	-0.26	-0.01
Channel Y + Input	200.21	-1.09	-0.54
Channel Y - Input	-199.26	-0.69	0.35
Channel Z + Input	2000.53	-0.34	-0.02
Channel Z + Input	200.31	-0.85	-0.42
Channel Z - Input	-199.31	-0.60	0.30

2. Common mode sensitivity

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Common mode Input Voltage (mV)	High Range Average Reading (μV)	Low Range Average Reading (μV)
Channel X	200	0.55	-1.37
	- 200	1.69	-0.29
Channel Y	200	8.36	8.15
	- 200	-8.96	-9.13
Channel Z	200	-19.28	-19.87
	- 200	18.54	18.77

3. Channel separation

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Input Voltage (mV)	Channel X (μV)	Channel Y (μV)	Channel Z (μV)
Channel X	200	-	-0.66	-3.73
Channel Y	200	7.44	-	1.06
Channel Z	200	10.02	5.64	-

4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	16195	13616
Channel Y	15639	15888
Channel Z	15484	14107

5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Input 10M Ω

	Average (μ V)	min. Offset (μ V)	max. Offset (μ V)	Std. Deviation (μ V)
Channel X	-0.18	-1.19	1.41	0.43
Channel Y	-0.51	-1.87	0.91	0.54
Channel Z	-0.07	-1.06	1.80	0.59

6. Input Offset Current

Nominal Input circuitry offset current on all channels: <25fA

7. Input Resistance (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)
Supply (+ Vcc)	+7.9
Supply (- Vcc)	-7.6

9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+6	+14
Supply (- Vcc)	-0.01	-8	-9

Appendix B.3 Calibration certificate for Dipole

**Calibration Laboratory of
Schmid & Partner
Engineering AG**
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **SGS Korea (Dymstec)**

Certificate No: **D2450V2-734_May18**

CALIBRATION CERTIFICATE

Object **D2450V2 - SN:734**

Calibration procedure(s) **QA CAL-05.v10**
Calibration procedure for dipole validation kits above 700 MHz



Calibration date: **May 30, 2018**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-18 (No. 217-02672/02673)	Apr-19
Power sensor NRP-Z91	SN: 103244	04-Apr-18 (No. 217-02672)	Apr-19
Power sensor NRP-Z91	SN: 103245	04-Apr-18 (No. 217-02673)	Apr-19
Reference 20 dB Attenuator	SN: 5058 (20k)	04-Apr-18 (No. 217-02682)	Apr-19
Type-N mismatch combination	SN: 5047.2 / 06327	04-Apr-18 (No. 217-02683)	Apr-19
Reference Probe EX3DV4	SN: 7349	30-Dec-17 (No. EX3-7349_Dec17)	Dec-18
DAE4	SN: 601	26-Oct-17 (No. DAE4-601_Oct17)	Oct-18
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power meter EPM-442A	SN: GB37480704	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
Power sensor HP 8481A	SN: US37292783	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
Power sensor HP 8481A	SN: MY41092317	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
RF generator R&S SMT-06	SN: 100972	15-Jun-15 (in house check Oct-16)	In house check: Oct-18
Network Analyzer HP 8753E	SN: US37390585	18-Oct-01 (in house check Oct-17)	In house check: Oct-18

Calibrated by: **Manu Seitz** Function: **Laboratory Technician**

Approved by: **Katja Pokovic** Technical Manager

Signature



Issued: May 31, 2018

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D2450V2-734_May18

Page 1 of 8

Calibration Laboratory of
Schmid & Partner
Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Glossary:

TSL tissue simulating liquid
 ConvF sensitivity in TSL / NORM x,y,z
 N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.1
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz $\pm$ 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	38.2 $\pm$ 6 %	1.85 mho/m $\pm$ 6 %
Head TSL temperature change during test	< 0.5 °C	----	----

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.1 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	51.4 W/kg $\pm$ 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	6.08 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	24.0 W/kg $\pm$ 16.5 % (k=2)

Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52.7	1.95 mho/m
Measured Body TSL parameters	(22.0 $\pm$ 0.2) °C	52.3 $\pm$ 6 %	1.99 mho/m $\pm$ 6 %
Body TSL temperature change during test	< 0.5 °C	----	----

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	12.5 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	49.4 W/kg $\pm$ 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	250 mW input power	5.86 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	23.3 W/kg $\pm$ 16.5 % (k=2)

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	54.8 Ω + 9.2 j Ω
Return Loss	- 20.2 dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	50.1 Ω + 8.6 j Ω
Return Loss	- 21.3 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.152 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	May 07, 2003

DASY5 Validation Report for Head TSL

Date: 30.05.2018

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:734

Communication System: UID 0 - CW; Frequency: 2450 MHz

Medium parameters used: $f = 2450$ MHz; $\sigma = 1.85$ S/m; $\epsilon_r = 38.2$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN7349; ConvF(7.88, 7.88, 7.88) @ 2450 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 115.5 V/m; Power Drift = -0.04 dB


Peak SAR (extrapolated) = 25.8 W/kg

SAR(1 g) = 13.1 W/kg; SAR(10 g) = 6.08 W/kg

Maximum value of SAR (measured) = 21.4 W/kg



Impedance Measurement Plot for Head TSL



DASY5 Validation Report for Body TSL

Date: 29.05.2018

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:734

Communication System: UID 0 - CW; Frequency: 2450 MHz

Medium parameters used: $f = 2450$ MHz; $\sigma = 1.99$ S/m; $\epsilon_r = 52.3$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN7349; ConvF(8.01, 8.01, 8.01) @ 2450 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (back); Type: QD 000 P50 AA; Serial: 1002
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 107.3 V/m; Power Drift = -0.03 dB


Peak SAR (extrapolated) = 24.6 W/kg

SAR(1 g) = 12.5 W/kg; SAR(10 g) = 5.86 W/kg

Maximum value of SAR (measured) = 20.5 W/kg



Impedance Measurement Plot for Body TSL



Calibration Laboratory of
Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **SGS Korea (Dymstec)**

Certificate No: **D5GHzV2-1130_May18**

CALIBRATION CERTIFICATE

Object **D5GHzV2 - SN:1130**

Calibration procedure(s) **QA CAL-22.v3**
Calibration procedure for dipole validation kits between 3-6 GHz

Calibration date: **May 25, 2018**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature $(22 \pm 3)^{\circ}\text{C}$ and humidity $< 70\%$.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-18 (No. 217-02672/02673)	Apr-19
Power sensor NRP-Z91	SN: 103244	04-Apr-18 (No. 217-02672)	Apr-19
Power sensor NRP-Z91	SN: 103245	04-Apr-18 (No. 217-02673)	Apr-19
Reference 20 dB Attenuator	SN: 5058 (20k)	04-Apr-18 (No. 217-02682)	Apr-19
Type-N mismatch combination	SN: 5047.2 / 06327	04-Apr-18 (No. 217-02683)	Apr-19
Reference Probe EX3DV4	SN: 3503	30-Dec-17 (No. EX3-3503_Dec17)	Dec-18
DAE4	SN: 601	26-Oct-17 (No. DAE4-601_Oct17)	Oct-18

Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power meter EPM-442A	SN: GB37480704	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
Power sensor HP 8481A	SN: US37292783	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
Power sensor HP 8481A	SN: MY41092317	07-Oct-15 (in house check Oct-16)	In house check: Oct-18
RF generator R&S SMT-06	SN: 100972	15-Jun-15 (in house check Oct-16)	In house check: Oct-18
Network Analyzer HP 8753E	SN: US37390585	18-Oct-01 (in house check Oct-17)	In house check: Oct-18

Calibrated by:	Name	Function	Signature
	Manu Seitz	Laboratory Technician	
Approved by:	Name	Function	Signature
	Katja Pokovic	Technical Manager	

Issued: May 25, 2018

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D5GHzV2-1130_May18

Page 1 of 14

Report File No : F690501/RF-SAR002704

Date of Issue : 2019-01-08

(All SGS services are rendered in accordance with the applicable SGS conditions of service available on request and accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>.)

RTT5041-76(2015.10.01) (2)

A4 (210mm x 297mm)

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.1
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
Frequency	5200 MHz $\pm$ 1 MHz 5300 MHz $\pm$ 1 MHz 5600 MHz $\pm$ 1 MHz 5800 MHz $\pm$ 1 MHz	

Head TSL parameters at 5200 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	36.0	4.66 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	36.2 $\pm$ 6 %	4.52 mho/m $\pm$ 6 %
Head TSL temperature change during test	< 0.5 °C	----	----

SAR result with Head TSL at 5200 MHz

SAR averaged over 1 cm³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.75 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	77.5 W/kg $\pm$ 19.9 % (k=2)

SAR averaged over 10 cm³ (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.23 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.3 W/kg $\pm$ 19.5 % (k=2)

Head TSL parameters at 5300 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.76 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	36.1 $\pm$ 6 %	4.63 mho/m $\pm$ 6 %
Head TSL temperature change during test	< 0.5 °C	----	----

SAR result with Head TSL at 5300 MHz

SAR averaged over 1 cm³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.31 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	83.1 W / kg $\pm$ 19.9 % (k=2)

SAR averaged over 10 cm³ (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.38 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	23.8 W/kg $\pm$ 19.5 % (k=2)

Head TSL parameters at 5600 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.7 ± 6 %	4.94 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C	----	----

SAR result with Head TSL at 5600 MHz

SAR averaged over 1 cm³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.53 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	85.3 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm³ (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.45 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	24.5 W/kg ± 19.5 % (k=2)

Head TSL parameters at 5800 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.3	5.27 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.4 ± 6 %	5.15 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C	----	----

SAR result with Head TSL at 5800 MHz

SAR averaged over 1 cm³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.30 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	83.0 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm³ (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.36 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	23.6 W/kg ± 19.5 % (k=2)

Body TSL parameters at 5200 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	49.0	5.30 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	47.0 ± 6 %	5.41 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C	----	----

SAR result with Body TSL at 5200 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.45 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	73.9 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	100 mW input power	2.09 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	20.7 W/kg ± 19.5 % (k=2)

Body TSL parameters at 5300 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.9	5.42 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	46.8 ± 6 %	5.54 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C	----	----

SAR result with Body TSL at 5300 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.66 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	76.0 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	100 mW input power	2.15 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	21.3 W/kg ± 19.5 % (k=2)

Body TSL parameters at 5600 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.5	5.77 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	46.3 ± 6 %	5.95 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C	----	----

SAR result with Body TSL at 5600 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	8.08 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	80.1 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	100 mW input power	2.26 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	22.4 W/kg ± 19.5 % (k=2)

Body TSL parameters at 5800 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.2	6.00 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	45.9 ± 6 %	6.22 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C	----	----

SAR result with Body TSL at 5800 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.63 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	75.7 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	100 mW input power	2.11 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	20.9 W/kg ± 19.5 % (k=2)

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL at 5200 MHz

Impedance, transformed to feed point	53.4 Ω - 9.7 j Ω
Return Loss	- 20.0 dB

Antenna Parameters with Head TSL at 5300 MHz

Impedance, transformed to feed point	52.1 Ω - 4.3 j Ω
Return Loss	- 26.6 dB

Antenna Parameters with Head TSL at 5600 MHz

Impedance, transformed to feed point	58.5 Ω - 2.9 j Ω
Return Loss	- 21.6 dB

Antenna Parameters with Head TSL at 5800 MHz

Impedance, transformed to feed point	56.8 Ω - 0.8 j Ω
Return Loss	- 23.9 dB

Antenna Parameters with Body TSL at 5200 MHz

Impedance, transformed to feed point	51.6 Ω - 8.6 j Ω
Return Loss	- 21.3 dB

Antenna Parameters with Body TSL at 5300 MHz

Impedance, transformed to feed point	51.0 Ω - 4.0 j Ω
Return Loss	- 27.9 dB

Antenna Parameters with Body TSL at 5600 MHz

Impedance, transformed to feed point	57.9 Ω - 3.2 j Ω
Return Loss	- 22.0 dB

Antenna Parameters with Body TSL at 5800 MHz

Impedance, transformed to feed point	55.9 Ω - 2.5 j Ω
Return Loss	- 24.3 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.206 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	September 08, 2011

DASY5 Validation Report for Head TSL

Date: 25.05.2018

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole D5GHzV2; Type: D5GHzV2; Serial: D5GHzV2 - SN:1130

Communication System: UID 0 - CW; Frequency: 5200 MHz, Frequency: 5300 MHz, Frequency: 5600 MHz, Frequency: 5800 MHz

Medium parameters used: $f = 5200$ MHz; $\sigma = 4.52$ S/m; $\epsilon_r = 36.2$; $\rho = 1000$ kg/m³,

Medium parameters used: $f = 5300$ MHz; $\sigma = 4.63$ S/m; $\epsilon_r = 36.1$; $\rho = 1000$ kg/m³,

Medium parameters used: $f = 5600$ MHz; $\sigma = 4.94$ S/m; $\epsilon_r = 35.7$; $\rho = 1000$ kg/m³,

Medium parameters used: $f = 5800$ MHz; $\sigma = 5.15$ S/m; $\epsilon_r = 35.4$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3503; ConvF(5.75, 5.75, 5.75) @ 5200 MHz, ConvF(5.5, 5.5, 5.5) @ 5300 MHz, ConvF(5.05, 5.05, 5.05) @ 5600 MHz, ConvF(4.96, 4.96, 4.96) @ 5800 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 73.29 V/m; Power Drift = -0.04 dB

Peak SAR (extrapolated) = 27.9 W/kg

SAR(1 g) = 7.75 W/kg; SAR(10 g) = 2.23 W/kg

Maximum value of SAR (measured) = 17.6 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 74.92 V/m; Power Drift = -0.09 dB

Peak SAR (extrapolated) = 30.2 W/kg

SAR(1 g) = 8.31 W/kg; SAR(10 g) = 2.38 W/kg

Maximum value of SAR (measured) = 18.9 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm


Reference Value = 75.83 V/m; Power Drift = -0.08 dB

Peak SAR (extrapolated) = 32.1 W/kg


SAR(1 g) = 8.53 W/kg; SAR(10 g) = 2.45 W/kg

Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5800 MHz/Zoom Scan,
dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
 Reference Value = 73.55 V/m; Power Drift = -0.09 dB
 Peak SAR (extrapolated) = 32.8 W/kg
SAR(1 g) = 8.3 W/kg; SAR(10 g) = 2.36 W/kg
 Maximum value of SAR (measured) = 19.6 W/kg



Impedance Measurement Plot for Head TSL



DASY5 Validation Report for Body TSL

Date: 24.05.2018

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole D5GHzV2; Type: D5GHzV2; Serial: D5GHzV2 - SN:1130

Communication System: UID 0 - CW; Frequency: 5200 MHz, Frequency: 5300 MHz, Frequency: 5600 MHz, Frequency: 5800 MHz

Medium parameters used: $f = 5200$ MHz; $\sigma = 5.41$ S/m; $\epsilon_r = 47$; $\rho = 1000$ kg/m³,

Medium parameters used: $f = 5300$ MHz; $\sigma = 5.54$ S/m; $\epsilon_r = 46.8$; $\rho = 1000$ kg/m³,

Medium parameters used: $f = 5600$ MHz; $\sigma = 5.95$ S/m; $\epsilon_r = 46.3$; $\rho = 1000$ kg/m³,

Medium parameters used: $f = 5800$ MHz; $\sigma = 6.22$ S/m; $\epsilon_r = 45.9$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3503; ConvF(5.35, 5.35, 5.35) @ 5200 MHz, ConvF(5.15, 5.15, 5.15) @ 5300 MHz, ConvF(4.65, 4.65, 4.65) @ 5600 MHz, ConvF(4.53, 4.53, 4.53) @ 5800 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (back); Type: QD 000 P50 AA; Serial: 1002
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 68.61 V/m; Power Drift = -0.08 dB

Peak SAR (extrapolated) = 27.6 W/kg

SAR(1 g) = 7.45 W/kg; SAR(10 g) = 2.09 W/kg

Maximum value of SAR (measured) = 17.5 W/kg

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 68.51 V/m; Power Drift = -0.08 dB

Peak SAR (extrapolated) = 29.9 W/kg

SAR(1 g) = 7.66 W/kg; SAR(10 g) = 2.15 W/kg

Maximum value of SAR (measured) = 17.9 W/kg

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm


Reference Value = 68.63 V/m; Power Drift = -0.08 dB

Peak SAR (extrapolated) = 33.8 W/kg


SAR(1 g) = 8.08 W/kg; SAR(10 g) = 2.26 W/kg

Maximum value of SAR (measured) = 19.5 W/kg

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5800 MHz/Zoom Scan,
dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
 Reference Value = 66.70 V/m; Power Drift = -0.08 dB
 Peak SAR (extrapolated) = 33.0 W/kg
SAR(1 g) = 7.63 W/kg; SAR(10 g) = 2.11 W/kg
 Maximum value of SAR (measured) = 18.6 W/kg



Impedance Measurement Plot for Body TSL



-THE END-