

Annex F. Calibration Certificates

ID	Device	Type/Model	Serial Number	Manufacturer	Calibration Certificate
088-000	E-field Probe	EUmmWV2	9366	SPEAG	
003-007	Dosimetric E-field Probe	EX3DV4	7465	SPEAG	
096-000	6500MHz System Validation Dipole	D6.5GHzV2	1013	SPEAG	

Dipole calibration

According to the KDB 865664 D01, a dipole must be calibrated using a fully validated SAR system according to the tissue dielectric parameters and SAR probe calibration frequency required for device testing. However, instead of the typical annual calibration recommended by measurement standards, longer calibration intervals of up to three years may be considered when it is demonstrated that the SAR target, impedance and return loss of a dipole have remain stable according to the following requirements.

1. When the most recent return-loss result, measured at least annually, deviates by less than 20% from the previous measurement (i.e. value in dB $\times$ 0.2) and not meeting the required 20 dB minimum return-loss requirement.
2. When the most recent measurement of the real and imaginary parts of the impedance, measured at least annually, deviates by less than 5 Ω from the previous measurement

Dipole ID #096-000			
Dipole 6500MHz Head TSL			
	Return Loss [dB]	Impedance [Ω]	Date
Initial Calibration	-24.5	49.4 - 5.9 j	2020-08-25
Last calibration	-27.2	49.8 - 4.4 j	2021-09-10