

Arima Communication, Model No: Arima M2001C

Date of Test: February 4, 2001

Muscle	
Ingredient	Frequency (900 MHz)
Water	54.05 %
Sugar	45.75 %
Salt	0.1 %
Preservative	0.1 %

The dielectric parameters were verified prior to assessment using the HP 85070A dielectric probe kit and the HP 8753C network Analyzer. The dielectric parameters were:

Frequency (MHz)	[*]	^{*(mho/m)}	^{**(kg/m³)}
835	56.5 ± 5%	0.97 ± 10%	1000

^{*} worst case uncertainty of the HP 85070A dielectric probe kit

^{**} worst case assumption

Muscle	
Ingredient	Frequency (1900 MHz)
Water	54.5 %
Cellulose	0.1 %
Salt	0 %
Preservative	0.1 %
Sugar	45.3 %

The dielectric parameters were verified prior to assessment using the HP 85070A dielectric probe kit and the HP 8753C network Analyzer. The dielectric parameters were:

Frequency (MHz)	[*]	^{*(mho/m)}	^{**(kg/m³)}
1900	52.2 ± 5%	1.65 ± 10%	1000

^{*} worst case uncertainty of the HP 85070A dielectric probe kit

^{**} worst case assumption

Note: The amount of each ingredient specified in the tables are not the exact amounts of the final test solution. The final test solution was adjusted by adding small amounts of either water, sugar, and/or salt to calibrate the solution to meet the proper dielectric parameters.

3.3 E-Field Probe Calibration

Probes were calibrated by the manufacturer in the TEM cell ifi 110. To ensure consistency, a strict protocol was followed. The conversion factor (ConF) between this calibration and the measurement in the tissue simulation solution was performed by comparison with temperature measurement and computer simulations. Probe calibration factors are included in Appendix C.

3.4 Measurement Uncertainty

The uncertainty budget has been determined for the DASY3 measurement system according to the NIS81 [5] and the NIST 1297 [6] documents and is given in the following table. The extended uncertainty (K=2) was assessed to be 23.5 %

UNCERTAINTY BUDGET				
Uncertainty Description	Error	Distrib.	Weight	Std.Dev.
Probe Uncertainty				
Axial isotropy	±0.2 dB	U-shape	0.5	±2.4 %
Spherical isotropy	±0.4 dB	U-shape	0.5	±4.8 %
Isotropy from gradient	±0.5 dB	U-shape	0	
Spatial resolution	±0.5 %	Normal	1	±0.5 %
Linearity error	±0.2 dB	Rectang.	1	±2.7 %
Calibration error	±3.3 %	Normal	1	±3.3 %
SAR Evaluation Uncertainty				
Data acquisition error	±1 %	Rectang.	1	±0.6 %
ELF and RF disturbances	±0.25 %	Normal	1	±0.25 %
Conductivity assessment	±10 %	Rectang.	1	±5.8 %
Spatial Peak SAR Evaluation Uncertainty				
Extrapol boundary effect	±3 %	Normal	1	±3 %
Probe positioning error	±0.1 mm	Normal	1	±1 %
Integrat. and cube orient	±3 %	Normal	1	±3 %
Cube shape inaccuracies	±2 %	Rectang.	1	±1.2 %
Device positioning	±6 %	Normal	1	±6 %
Combined Uncertainties				±11.7 %

3.5 Measurement Tractability

All measurements described in this report are traceable to National Institute of Standards and Technology (NIST) standards or appropriate national standards.

Arima Communication, Model No: Arima M2001C

Date of Test: February 4, 2001

4.0 WARNING LABEL INFORMATION - USA

See User Manual.

5.0 REFERENCES

- [1] ANSI, *ANSI/IEEE C95.1-1991: IEEE Standard for Safety Levels with Respect to Human Exposure to Radio Frequency Electromagnetic Fields, 3kHz to 300 GHz*, The Institute of electrical and Electronics Engineers, Inc., New York, NY 10017, 1992
- [2] Federal Communications Commission, "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields", OET Bulletin 65, FCC, Washington, D.C. 20554, 1997
- [3] Thomas Schmid, Oliver Egger, and Niels Kuster, "Automated E-field scanning system for dosimetric assessments", *IEEE Transaction on Microwave Theory and Techniques*, vol. 44, pp. 105-113, Jan. 1996.
- [4] Niels Kuster, Ralph Kastle, and Thomas Schmid, "Dosimetric evaluation of mobile communications equipment with known precision", *IEICE Transactions on Communications*, vol. E80-B, no. 5, pp.645-652, May 1997.
- [5] NIS81, NAMAS, "The treatment of uncertainty in EMC measurement", Tech. Rep., NAMAS Executive, National Physical Laboratory, Teddington, Middlesex, England, 1994.
- [6] Barry N. Taylor and Chris E. Kuyatt, "Guidelines for evaluating and expressing the uncertainty of NIST measurement results", Tech. Rep., National Institute of Standards and Technology, 1994.

Arima Communication, Model No: Arima M2001C

Date of Test: February 4, 2001

APPENDIX A - SAR Evaluation Data

Please note that the graphical visualization of the phone position onto the SAR distribution gives only limited information on the current distribution of the device, since the curvature of the head results in graphical distortion. Full information can only be obtained either by H-field scans in free space or SAR evaluation with a flat phantom.

Powerdrift is the measurement of power drift of the device over one complete SAR scan.

ARIMA, M2001C



Generic Twin Phantom; Left Hand X Section; Position: (80°, 65°); Frequency: 1850 MHz

Probe: ET3DV5 - SN1333; ConvF(5,03,5,03,5,03); Crest factor: 8.0; Brain1900 MHz; $\sigma = 1.77 \text{ mho/m}$ $\epsilon_r = 42.8$ $\rho = 1.00 \text{ g/cm}^3$

Cube 5x5x7; SAR (1g): 1.12 mW/g, SAR (10g): 0.641 mW/g, (Worst-case extrapolation)

Coarse: Dx = 20.0, Dy = 20.0, Dz = 10.0

Poweredrift: 0.02 dB

SAR_{Tot} [mW/g]

ARIMA, M2001C





Generic Twin Phantom; Left Hand X Section; Position: (80°, 65°); Frequency: 1850 MHz

Probe: ET3DV5 - SN1333; ConvF(5,03,5,03,5,03); Crest factor: 8.0; Brain1900 MHz: $\sigma = 1.77 \text{ mho/m}$ $\epsilon_r = 42.8$ $\rho = 1.00 \text{ g/cm}^3$

Cube 5x5x7; SAR (1g): 1.32 mW/g, SAR (10g): 0.745 mW/g, (Worst-case extrapolation)

Coarse: Dx = 20.0, Dy = 20.0, Dz = 10.0

Powerdrift: 0.06 dB

SAR_{Tot} [mW/g]

ARIMA, M2001C


Generic Twin Phantom; Left Hand X Section; Position: (80°, 65°); Frequency: 1880 MHz

Probe: ET3DV5 - SN1333; ConvF(5.03,5.03,5.03); Crest factor: 8.0; Brain1900 MHz; $\sigma = 1.77$ mho/n; $\epsilon_r = 42.8$; $\rho = 1.00$ g/cm³

Cube 5x5x7: SAR (1g): 1.18 mW/g, SA

Coarse: Dx = 20.0, Dy = 20.0, Dz = 10.0

Powerdrift: 0.03 dB



ARIMA, M2001C #4


Generic Twin Phantom; Left Hand X Section; Position: (80°, 65°); Frequency: 1910 MHz

Probe: ET3DV5 - SN1333; ConvF(5,03,5,03); Crest factor: 8.0; Brain1900 MHz: $\sigma = 1.77 \text{ mho/m}$ $\epsilon_r = 42.8$ $\rho = 1.00 \text{ g/cm}^3$

Cube 5x5x7: SAR (1g): 0.915 mW/g, SAR (10g): 0.519 mW/g, (Worst-case extrapolation)

Coarse: Dx = 20.0, Dy = 20.0, Dz = 10.0

Powerdrift: -0.26 dB



SAR_{Tot} [mW/g]

	9.04E-1
	8.14E-1
	7.23E-1
	6.33E-1
	5.42E-1
	4.52E-1
	3.62E-1
	2.71E-1
	1.81E-1
	9.04E-2

ARIMA, M2001C




Generic Twin Phantom; Right Hand Section; Position: (80°, 65°); Frequency: 1850 MHz

Probe: FT3DV5 - SN1333; ConvF(5,03,5,03,5,03); Crest factor: 8.0; Brain1900 MHz, $\sigma = 1.77 \text{ mho/m}$ $\epsilon_r = 42.8$ $\rho = 1.00 \text{ g/cm}^3$

Cube 5x5x7; SAR (1g): 1.25 mW/g, SAR (10g): 0.696 mW/g, (Worst-case extrapolation)

Coarse: Dx = 20.0, Dy = 20.0, Dz = 10.0

Powerdrift: 0.17 dB

SAR_{Tot} [mW/g]

	1.20E+0
	1.08E+0
	9.56E-1
	8.37E-1
	7.17E-1
	5.98E-1
	4.78E-1
	3.58E-1
	2.39E-1
	1.20E-1

Arima Communication, Model No: Arima M2001C

Date of Test: February 4, 2001

APPENDIX B - E-Field Probe Calibration Data

See attached.

**Schmid & Partner
Engineering AG**

Staffelstrasse 8, 8045 Zurich, Switzerland, Telefon +41 1 280 08 60, Fax +41 1 280 08 64

Calibration Certificate

Dosimetric E-Field Probe

Type:

ET3DV5

Serial Number:

1333

Place of Calibration:

Zurich

Date of Calibration:

April 10, 2000

Calibration Interval:

12 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

Ulrich Käfer

Approved by:

C. E. J.

**Schmid & Partner
Engineering AG**

Staffelstrasse 8, 8045 Zurich, Switzerland, Telefon +41 1 280 08 60, Fax +41 1 280 08 64

Probe ET3DV5

SN:1333

Manufactured:	December 20, 1997
Last calibration:	March 18, 1999
Recalibrated:	April 10, 2000

Calibrated for System DASY3

DASY3 - Parameters of Probe: ET3DV5 SN:1333

Sensitivity in Free Space

NormX	2.39 $\mu\text{V}/(\text{V}/\text{m})^2$
NormY	2.36 $\mu\text{V}/(\text{V}/\text{m})^2$
NormZ	2.34 $\mu\text{V}/(\text{V}/\text{m})^2$

Diode Compression

DCP X	100 mV
DCP Y	100 mV
DCP Z	100 mV

Sensitivity in Tissue Simulating Liquid


Brain	450 MHz	$\epsilon_r = 48 \pm 5\%$	$\sigma = 0.50 \pm 10\% \text{ mho/m}$
	ConvF X	6.03 extrapolated	Boundary effect:
	ConvF Y	6.03 extrapolated	Alpha 0.13
	ConvF Z	6.03 extrapolated	Depth 3.57
Brain	900 MHz	$\epsilon_r = 42.5 \pm 5\%$	$\sigma = 0.86 \pm 10\% \text{ mho/m}$
	ConvF X	5.70 $\pm 7\%$ (k=2)	Boundary effect:
	ConvF Y	5.70 $\pm 7\%$ (k=2)	Alpha 0.34
	ConvF Z	5.70 $\pm 7\%$ (k=2)	Depth 3.00
Brain	1500 MHz	$\epsilon_r = 41 \pm 5\%$	$\sigma = 1.32 \pm 10\% \text{ mho/m}$
	ConvF X	5.25 interpolated	Boundary effect:
	ConvF Y	5.25 interpolated	Alpha 0.61
	ConvF Z	5.25 interpolated	Depth 2.23
Brain	1800 MHz	$\epsilon_r = 41 \pm 5\%$	$\sigma = 1.69 \pm 10\% \text{ mho/m}$
	ConvF X	5.03 $\pm 7\%$ (k=2)	Boundary effect:
	ConvF Y	5.03 $\pm 7\%$ (k=2)	Alpha 0.74
	ConvF Z	5.03 $\pm 7\%$ (k=2)	Depth 1.85

Sensor Offset


Probe Tip to Sensor Center	2.7	mm
Optical Surface Detection	1.9 \pm 0.2	mm

Receiving Pattern (ϕ), $\theta = 0^\circ$


$f = 30$ MHz, TEM cell ifi110




$f = 100$ MHz, TEM cell ifi110




$f = 300$ MHz, TEM cell ifi110




$f = 900$ MHz, TEM cell ifi110





Isotropy Error (ϕ), $\theta = 0^\circ$



6.0 DOCUMENT HISTORY

Revision/ Job Number	Writer Initials	Date	Change
1.0 /J20039587	SS	February 16, 2001	Original document