

Patient Monitoring
Quality & Regulatory Group
3000 Minuteman Road
Mailstop 0480
Andover, MA 01810
Tel: (978) 659-4348
Fax: (978) 685-5624

Date: 21 March 2012

Subject: Ad-Hoc Mode Function Declaration

FCC ID: PQC-MX40SH1B4

To whom it may concern:

We, Philips Medical Systems, declare that the device (FCC ID: PQC-MX40SH1B4)

1. Works as a client device.
2. Ad-hoc will not be supported.
3. Is not able to operate in non-US frequencies, country code selection will be disabled.
4. Cannot operate as a client device on frequencies within the band 5600 – 5650 MHz in the US.
5. In non associated mode, that software and associated drivers will not initiate any transmission on DFS frequencies. This includes transmissions for beacon ad-hoc peer to peer modes. (KDB848637)


Please contact the signatory below or the authorized agent for this filing.

Regards,

Applicant: Philips Medical Systems, Inc.

Address: 3000 Minuteman Drive, Andover, MA 01810

Signature:



Name and Job title: Claire Arakaki, Regulatory Affairs Specialist

Email: Claire.arakaki@philips.com

Telephone: 978.659.4348