

Philips Medizin Systeme Böblingen GmbH
Hewlett-Packard-Str. 2
71034 Böblingen
Germany

Declaration about Hardware – CTG Base Station/ Model - 866074

To Whom It May Concern:
Product/Model: CTG Base Station/ 866074
FCC ID: PQC-OBBSBV1

We hereby declare that Product/ Model: CTG Base Station/ 866074 (FCC ID: PQC-OBBSBV1) uses the same Hardware and Software (as far as the SRR part is concerned) as Product/ Model: 2.4 GHz Transceiver/ SRR Module BV3 (FCC ID: PQC-SRRBV3). No RF related modifications (redesign, component replacements, pcb redesign) have been made.
Based on this fact conducted signal measurements in FCC part 15.247 reports (report No. 1-5420/12-01-11-A, dated 2013-10-17) issued for Product/Model: 2.4 GHz Transceiver/ SRR Module BV3 also apply to Product/Model: CTG Base Station/ 866074.

We hereby also declare that Product/ Model: CTG Base Station/ 866074 (FCC ID: PQC-OBBSBV1) uses the same Hardware and Software (as far as the 608-614MHz part is concerned) as Product/ Model: CTG Base Station/ 866074 (FCC ID: PQC-OBRBSBV1). No RF related modifications (redesign, component replacements, pcb redesign) have been made.
Based on this fact conducted signal measurements in FCC part 95H (report No. 1-4695/12-01-03-D, dated 2013-11-08) issued for Product/Model: CTG Base Station/ 866074 also apply to Product/Model: CTG Base Station/ 866074.

If you have any questions feel free to contact us at the address shown.

Sincerely



Hansjörg Geywitz (Signature)
Hewlett-Packard-Str. 2
71034 Böblingen, Germany
Tel: 07031-4630
Email: hansjoerg.geywitz@philips.com