



Philips Medizin Systeme Böblingen GmbH

Postfach 1471, 71004 Böblingen, Germany

Federal Communication Commission
Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Aug 23, 2010

TO WHOM IT MAY CONCERN

Modular Approval Request

The **PQC-SRRBV2 Single Short Range Radio Module** consists of a transmitter and a receiver working in the 2.4 GHz band

- 15.212(a)(1)(i) It does not have its own RF shielding
- 15.212(a)(1)(ii) It has digital input buffers
- 15.212(a)(1)(iii) It has its own power supply regulation
- 15.212(a)(1)(iv) The module has been tested with one onboard antenna. It has no antenna plug.
- 15.212(a)(1)(v) The module has been tested in a stand-alone configuration.
- 15.212(a)(1)(vi) The module will be labelled with its own FCC ID number
- 15.212(a)(1)(vii) The module complies with all applicable FCC rules. Instructions for maintaining compliance are given in the documents accompanying the host device, in which the module will be incorporated
- 15.212(a)(1)(viii) The module complies with all applicable SAR and RF exposure requirements.

See also 15.212 (b)

If you have any questions, please feel free to contact us at the address shown below

Sincerely,

Name: Markus Stacha

Company: Philips Medizin Systeme Boeblingen GmbH

Address: Hewlett-Packard-Str. 2

71034 Boeblingen / Germany

Phone: +49(0)7031 463-2840

Fax: +49(0)7031 463-2442

Email: markus.stacha@philips.com



Philips Medizin Systeme Böblingen GmbH
Hewlett-Packard-Str. 2, 71034 Böblingen, Germany

Telefon +49 (0) 7031 463-0

Telefax +49 (0) 7031 463-2202

Geschäftsführer: Michael Dreher (Sprecher), Dr. Werner Haas

Registergericht Böblingen Reg.-Nr. HRB 245187, Sitz der Gesellschaft: Böblingen